

Data Path : Z:\HPCHEM1\BNA G\Data\BG031218\
 Data File : BG033084.D
 Acq On : 12 Mar 2018 10:45
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 12 16:23:53 2018
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG022118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 09 15:10:14 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	67	0.00
2	1,4-Dioxane	0.543	0.506	6.8	63	0.00
3	Pyridine	1.495	1.446	3.3	62	0.00
4	n-Nitrosodimethylamine	0.600	0.676	-12.7	74	0.00
5 S	2-Fluorophenol	1.250	1.223	2.2	64	0.00
6	Aniline	2.359	2.153	8.7	61	0.00
7 S	Phenol-d6	1.742	1.684	3.3	64	0.00
8	2-Chlorophenol	1.368	1.305	4.6	63	0.00
9	Benzaldehyde	1.004	0.814	18.9	56	0.00
10 C	Phenol	1.757	1.693	3.6	65	0.00
11	bis(2-Chloroethyl)ether	1.436	1.284	10.6	60	0.00
12	1,3-Dichlorobenzene	1.603	1.461	8.9	61	0.00
13 C	1,4-Dichlorobenzene	1.613	1.476	8.5	61	0.00
14	1,2-Dichlorobenzene	1.546	1.410	8.8	62	0.00
15	Benzyl Alcohol	1.281	1.225	4.4	64	0.00
16	2,2'-oxybis(1-Chloropropane	2.469	2.569	-4.1	70	0.00
17	2-Methylphenol	1.295	1.214	6.3	62	0.00
18	Hexachloroethane	0.549	0.525	4.4	64	0.00
19 P	n-Nitroso-di-n-propylamine	1.230	1.103	10.3	61	0.00
20	3+4-Methylphenols	1.801	1.718	4.6	64	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	67	0.00
22	Acetophenone	0.505	0.476	5.7	63	0.00
23 S	Nitrobenzene-d5	0.242	0.335	-38.4#	86	0.00
24	Nitrobenzene	0.289	0.352	-21.8	79	0.00
25	Isophorone	0.702	0.620	11.7	60	0.00
26 C	2-Nitrophenol	0.103	0.151	-46.6#	104	0.00
27	2,4-Dimethylphenol	0.305	0.271	11.1	61	0.00
28	bis(2-Chloroethoxy)methane	0.409	0.361	11.7	59	0.00
29 C	2,4-Dichlorophenol	0.277	0.265	4.3	64	0.00
30	1,2,4-Trichlorobenzene	0.324	0.296	8.6	61	0.00
31	Naphthalene	1.045	0.970	7.2	63	0.00
32	Benzoic acid	0.174	0.166	4.6	68	-0.02
33	4-Chloroaniline	0.467	0.412	11.8	60	0.00
34 C	Hexachlorobutadiene	0.182	0.172	5.5	63	0.00
35	Caprolactam	0.111	0.106	4.5	62	0.00
36 C	4-Chloro-3-methylphenol	0.322	0.328	-1.9	67	0.00
37	2-Methylnaphthalene	0.756	0.683	9.7	61	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	69	0.00
39	1,2,4,5-Tetrachlorobenzene	0.594	0.563	5.2	66	0.00
40 P	Hexachlorocyclopentadiene	0.303	0.286	5.6	64	0.00
41 S	2,4,6-Tribromophenol	0.210	0.232	-10.5	74	0.00
42 C	2,4,6-Trichlorophenol	0.374	0.366	2.1	67	0.00
43	2,4,5-Trichlorophenol	0.433	0.421	2.8	66	0.00
44 S	2-Fluorobiphenyl	1.387	1.286	7.3	64	0.00

Data Path : Z:\HPCHEM1\BNA G\Data\BG031218\
 Data File : BG033084.D
 Acq On : 12 Mar 2018 10:45
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 12 16:23:53 2018
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG022118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 09 15:10:14 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.748	1.590	9.0	63	0.00
46	2-Chloronaphthalene	1.306	1.189	9.0	63	0.00
47	2-Nitroaniline	0.282	0.427	-51.4#	105	0.00
48	Acenaphthylene	2.137	1.985	7.1	64	0.00
49	Dimethylphthalate	1.698	1.522	10.4	62	0.00
50	2,6-Dinitrotoluene	0.244	0.333	-36.5#	93	0.00
51 C	Acenaphthene	1.331	1.212	8.9	63	0.00
52	3-Nitroaniline	0.308	0.389	-26.3#	86	0.00
53 P	2,4-Dinitrophenol	0.070	0.061	12.9	62	0.00
54	Dibenzofuran	1.895	1.754	7.4	64	0.00
55 P	4-Nitrophenol	0.233	0.196	15.9	57	0.00
56	2,4-Dinitrotoluene	0.293	0.463	-58.0#	113	0.00
57	Fluorene	1.564	1.449	7.4	64	0.00
58	2,3,4,6-Tetrachlorophenol	0.336	0.349	-3.9	70	0.00
59	Diethylphthalate	1.791	1.631	8.9	63	0.00
60	4-Chlorophenyl-phenylether	0.765	0.699	8.6	64	0.00
61	4-Nitroaniline	0.336	0.364	-8.3	72	0.00
62	Azobenzene	1.602	1.503	6.2	65	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	72	0.00
64	4,6-Dinitro-2-methylphenol	0.054	0.077	-42.6#	108	0.00
65 c	n-Nitrosodiphenylamine	0.651	0.578	11.2	64	0.00
66	4-Bromophenyl-phenylether	0.211	0.193	8.5	67	0.00
67	Hexachlorobenzene	0.229	0.209	8.7	67	0.00
68	Atrazine	0.214	0.189	11.7	63	0.00
69 C	Pentachlorophenol	0.144	0.124	13.9	62	0.00
70	Phenanthrene	1.116	1.023	8.3	66	0.00
71	Anthracene	1.120	1.030	8.0	66	0.00
72	Carbazole	1.104	0.994	10.0	65	0.00
73	Di-n-butylphthalate	1.355	1.272	6.1	66	0.00
74 C	Fluoranthene	1.233	1.176	4.6	69	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	79	0.00
76	Benzidine	0.682	0.336	50.7#	40#	0.00
77	Pyrene	1.410	1.225	13.1	69	0.00
78 S	Terphenyl-d14	0.890	0.802	9.9	72	0.00
79	Butylbenzylphthalate	0.625	0.596	4.6	72	0.00
80	Benzo(a)anthracene	1.283	1.165	9.2	72	0.00
81	3,3'-Dichlorobenzidine	0.475	0.392	17.5	64	0.00
82	Chrysene	1.225	1.112	9.2	72	0.00
83	Bis(2-ethylhexyl)phthalate	0.926	0.878	5.2	71	0.00
84 c	Di-n-octyl phthalate	1.411	1.411	0.0	73	0.00
85	Indeno(1,2,3-cd)pyrene	1.429	1.201	16.0	65	-0.02
86 I	Perylene-d12	1.000	1.000	0.0	80	0.00
87	Benzo(b)fluoranthene	1.183	1.067	9.8	71	0.00

Data Path : Z:\HPCHEM1\BNA G\Data\BG031218\
Data File : BG033084.D
Acq On : 12 Mar 2018 10:45
Operator : SJ/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Mar 12 16:23:53 2018
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG022118.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 09 15:10:14 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.175	1.079	8.2	74	0.00
89 C	Benzo(a)pyrene	1.125	1.029	8.5	73	0.00
90	Dibenzo(a,h)anthracene	1.132	0.927	18.1	64	-0.02
91	Benzo(a,h,i)perylene	1.116	0.955	14.4	68	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 1