

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG031219\
 Data File : BG039870.D
 Acq On : 12 Mar 2019 12:17
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 13 05:46:41 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG030419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 13 05:34:05 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	-0.02
2	1,4-Dioxane	0.541	0.539	0.4	96	-0.01
3	Pyridine	1.449	1.555	-7.3	92	-0.01
4	n-Nitrosodimethylamine	0.687	0.760	-10.6	99	-0.02
5 S	2-Fluorophenol	1.172	1.228	-4.8	95	-0.01
6	Aniline	2.290	2.325	-1.5	93	-0.02
7 S	Phenol-d6	1.748	1.862	-6.5	96	-0.02
8	2-Chlorophenol	1.422	1.437	-1.1	92	-0.01
9	Benzaldehyde	1.128	1.092	3.2	88	-0.01
10 C	Phenol	1.894	2.019	-6.6	95	-0.01
11	bis(2-Chloroethyl)ether	1.476	1.496	-1.4	94	-0.02
12	1,3-Dichlorobenzene	1.609	1.637	-1.7	93	-0.01
13 C	1,4-Dichlorobenzene	1.638	1.664	-1.6	94	-0.01
14	1,2-Dichlorobenzene	1.583	1.612	-1.8	94	-0.01
15	Benzyl Alcohol	1.387	1.479	-6.6	94	-0.01
16	2,2'-oxybis(1-Chloropropane	2.953	2.984	-1.0	93	-0.02
17	2-Methylphenol	1.207	1.270	-5.2	94	-0.02
18	Hexachloroethane	0.588	0.592	-0.7	95	-0.02
19 P	n-Nitroso-di-n-propylamine	1.258	1.256	0.2	89	-0.02
20	3+4-Methylphenols	1.677	1.780	-6.1	95	-0.02
21 I	Naphthalene-d8	1.000	1.000	0.0	94	-0.02
22	Acetophenone	0.537	0.525	2.2	90	-0.02
23 S	Nitrobenzene-d5	0.370	0.379	-2.4	94	-0.02
24	Nitrobenzene	0.388	0.396	-2.1	94	-0.01
25	Isophorone	0.775	0.770	0.6	91	-0.01
26 C	2-Nitrophenol	0.187	0.201	-7.5	96	-0.02
27	2,4-Dimethylphenol	0.291	0.293	-0.7	91	-0.02
28	bis(2-Chloroethoxy)methane	0.471	0.451	4.2	88	-0.01
29 C	2,4-Dichlorophenol	0.328	0.342	-4.3	94	-0.01
30	1,2,4-Trichlorobenzene	0.369	0.361	2.2	92	-0.01
31	Naphthalene	1.059	1.048	1.0	92	-0.01
32	Benzoic acid	0.181	0.165	8.8	84	0.00
33	4-Chloroaniline	0.449	0.458	-2.0	92	-0.01
34 C	Hexachlorobutadiene	0.252	0.240	4.8	89	-0.02
35	Caprolactam	0.125	0.129	-3.2	93	-0.02
36 C	4-Chloro-3-methylphenol	0.376	0.392	-4.3	93	-0.01
37	2-Methylnaphthalene	0.792	0.792	0.0	93	-0.01
38	1-Methylnaphthalene	0.769	0.770	-0.1	92	-0.01
39 I	Acenaphthene-d10	1.000	1.000	0.0	94	-0.01
40	1,2,4,5-Tetrachlorobenzene	0.678	0.673	0.7	93	-0.01
41 P	Hexachlorocyclopentadiene	0.363	0.321	11.6	81	-0.02
42 S	2,4,6-Tribromophenol	0.233	0.242	-3.9	92	-0.01
43 C	2,4,6-Trichlorophenol	0.397	0.421	-6.0	96	-0.01
44	2,4,5-Trichlorophenol	0.447	0.466	-4.3	92	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG031219\
 Data File : BG039870.D
 Acq On : 12 Mar 2019 12:17
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 13 05:46:41 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG030419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 13 05:34:05 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.405	1.358	3.3	91	-0.02
46	1,1'-Biphenyl	1.645	1.633	0.7	93	-0.01
47	2-Chloronaphthalene	1.238	1.222	1.3	93	-0.01
48	2-Nitroaniline	0.398	0.432	-8.5	97	0.00
49	Acenaphthylene	2.031	2.037	-0.3	92	-0.02
50	Dimethylphthalate	1.672	1.648	1.4	91	-0.01
51	2,6-Dinitrotoluene	0.355	0.371	-4.5	92	-0.01
52 C	Acenaphthene	1.223	1.223	0.0	93	-0.01
53	3-Nitroaniline	0.356	0.380	-6.7	95	-0.01
54 P	2,4-Dinitrophenol	0.197	0.219	-11.2	99	0.00
55	Dibenzofuran	2.006	2.016	-0.5	94	-0.01
56 P	4-Nitrophenol	0.319	0.304	4.7	85	-0.01
57	2,4-Dinitrotoluene	0.498	0.535	-7.4	95	-0.01
58	Fluorene	1.577	1.581	-0.3	94	-0.01
59	2,3,4,6-Tetrachlorophenol	0.437	0.465	-6.4	95	0.00
60	Diethylphthalate	1.656	1.635	1.3	90	-0.01
61	4-Chlorophenyl-phenylether	0.842	0.832	1.2	92	-0.01
62	4-Nitroaniline	0.386	0.407	-5.4	92	-0.01
63	Azobenzene	1.501	1.518	-1.1	94	-0.01
64 I	Phenanthrene-d10	1.000	1.000	0.0	95	-0.01
65	4,6-Dinitro-2-methylphenol	0.118	0.114	3.4	90	-0.01
66 c	n-Nitrosodiphenylamine	0.579	0.580	-0.2	93	-0.01
67	4-Bromophenyl-phenylether	0.224	0.217	3.1	92	-0.01
68	Hexachlorobenzene	0.232	0.233	-0.4	95	-0.01
69	Atrazine	0.228	0.228	0.0	93	0.00
70 C	Pentachlorophenol	0.147	0.135	8.2	85	0.00
71	Phenanthrene	1.102	1.072	2.7	92	0.00
72	Anthracene	1.074	1.064	0.9	94	-0.01
73	Carbazole	0.968	0.956	1.2	94	-0.01
74	Di-n-butylphthalate	1.139	1.140	-0.1	94	-0.01
75 C	Fluoranthene	1.302	1.293	0.7	95	-0.01
76 I	Chrysene-d12	1.000	1.000	0.0	100	-0.02
77	Benzidine	0.635	0.695	-9.4	98	0.00
78	Pyrene	1.300	1.269	2.4	95	-0.01
79 S	Terphenyl-d14	0.908	0.883	2.8	97	-0.01
80	Butylbenzylphthalate	0.504	0.521	-3.4	99	-0.01
81	Benzo(a)anthracene	1.253	1.242	0.9	98	-0.01
82	3,3'-Dichlorobenzidine	0.458	0.451	1.5	94	-0.01
83	Chrysene	1.215	1.198	1.4	99	-0.01
84	Bis(2-ethylhexyl)phthalate	0.708	0.727	-2.7	98	-0.02
85 c	Di-n-octyl phthalate	1.172	1.228	-4.8	98	-0.02
86	Indeno(1,2,3-cd)pyrene	1.379	1.361	1.3	96	-0.03
87 I	Perylene-d12	1.000	1.000	0.0	98	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG031219\
Data File : BG039870.D
Acq On : 12 Mar 2019 12:17
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Mar 13 05:46:41 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG030419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 13 05:34:05 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.193	1.176	1.4	97	-0.02
89	Benzo(k)fluoranthene	1.110	1.078	2.9	97	-0.02
90 C	Benzo(a)pyrene	1.102	1.102	0.0	98	-0.02
91	Dibenzo(a,h)anthracene	1.080	1.005	6.9	92	-0.03
92	Benzo(q,h,i)perylene	1.085	1.020	6.0	92	-0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0