

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG031424\
 Data File : BG060635.D
 Acq On : 14 Mar 2024 12:28
 Operator : MA/JU
 Sample : PB159530TB
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PB159530TB

Quant Time: Mar 14 14:06:09 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG031324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 14 00:45:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.325	152	92511	20.000	ng	0.00
21) Naphthalene-d8	11.169	136	421195	20.000	ng	0.00
39) Acenaphthene-d10	14.947	164	296205	20.000	ng	0.00
64) Phenanthrene-d10	17.697	188	661407	20.000	ng	0.00
76) Chrysene-d12	22.015	240	577301	20.000	ng	-0.01
86) Perylene-d12	25.570	264	650037	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.811	112	845653	143.697	ng	0.00
7) Phenol-d6	7.438	99	1185631	138.885	ng	0.00
23) Nitrobenzene-d5	9.524	82	710236	87.758	ng	0.00
42) 2,4,6-Tribromophenol	16.433	330	452326	131.781	ng	0.00
45) 2-Fluorobiphenyl	13.572	172	1822966	83.210	ng	0.00
79) Terphenyl-d14	20.276	244	2725525	85.337	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG031424\
Data File : BG060635.D
Acq On : 14 Mar 2024 12:28
Operator : MA/JU
Sample : PB159530TB
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB159530TB

Quant Time: Mar 14 14:06:09 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG031324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 14 00:45:22 2024
Response via : Initial Calibration

