

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG031521\
 Data File : BG048377.D
 Acq On : 15 Mar 2021 9:43
 Operator : CG/JU
 Sample : PB135005BL
 Misc : SFAM-MDL-W-BL02
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SBLK005

Quant Time: Mar 15 11:29:33 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SFAM-EPA-BG031121.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Mar 15 10:45:12 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.11	152	54114	20.00	ng/ul	0.00
20) Naphthalene-d8	10.94	136	213856	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.76	164	164465	20.00	ng/ul	0.00
64) Phenanthrene-d10	17.51	188	432050	20.00	ng/ul	0.00
79) Chrysene-d12	21.80	240	503557	20.00	ng/ul	0.00
88) Perylene-d12	25.15	264	471536	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.51	96	5565	4.46	ng/uL	0.00
4) Pyridine-d5	3.93	84	92221	24.12	ng/ul	0.00
7) Phenol-d5	7.25	99	111927	23.81	ng/ul	0.00
9) Bis-(2-Chloroethyl)ether-d	7.44	67	72258	24.89	ng/ul	0.00
11) 2-Chlorophenol-d4	7.64	132	81628	24.98	ng/ul	0.00
15) 4-Methylphenol-d8	8.81	113	92636	24.02	ng/ul	0.00
21) Nitrobenzene-d5	9.28	128	39776	26.86	ng/ul	-0.01
24) 2-Nitrophenol-d4	10.01	143	42032	25.93	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.55	165	92724	23.28	ng/ul	0.00
31) 4-Chloroaniline-d4	11.06	131	116796	23.28	ng/ul	0.00
46) Dimethylphthalate-d6	14.16	166	348432	25.75	ng/ul	0.00
49) Acenaphthylene-d8	14.45	160	400209	26.94	ng/ul	0.00
54) 4-Nitrophenol-d4	14.92	143	50473	25.05	ng/ul	0.00
60) Fluorene-d10	15.75	176	300009	25.12	ng/ul	0.00
65) 4,6-Dinitro-2-methylphenol	15.85	200	48277	21.87	ng/ul	0.00
73) Anthracene-d10	17.61	188	525879	26.64	ng/ul	0.00
81) Pyrene-d10	19.89	212	652457	25.34	ng/ul	0.00
92) Benzo(a)pyrene-d12	24.92	264	686019	26.84	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG031521\
Data File : BG048377.D
Acq On : 15 Mar 2021 9:43
Operator : CG/JU
Sample : PB135005BL
Misc : SFAM-MDL-W-BL02
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SBLK005

Quant Time: Mar 15 11:29:33 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SFAM-EPA-BG031121.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Mar 15 10:45:12 2021
Response via : Initial Calibration

