

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG031521\
 Data File : BG048379.D
 Acq On : 15 Mar 2021 11:20
 Operator : CG/JU
 Sample : PB135014BL
 Misc : SFAM-MDL-W-BL02
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SBLK014

Quant Time: Mar 15 13:01:54 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SFAM-EPA-BG031121.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Mar 15 10:45:12 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.13	152	52813	20.00	ng/ul	0.00
20) Naphthalene-d8	10.95	136	206274	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.76	164	161891	20.00	ng/ul	0.00
64) Phenanthrene-d10	17.51	188	426760	20.00	ng/ul	0.00
79) Chrysene-d12	21.80	240	482302	20.00	ng/ul	0.00
88) Perylene-d12	25.16	264	463741	20.00	ng/ul	0.01

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.51	96	5294	4.34	ng/uL	0.00
4) Pyridine-d5	3.93	84	84663	22.69	ng/ul	0.00
7) Phenol-d5	7.26	99	113497	24.74	ng/ul	0.00
9) Bis-(2-Chloroethyl)ether-d	7.44	67	75176	26.53	ng/ul	0.00
11) 2-Chlorophenol-d4	7.64	132	84137	26.39	ng/ul	0.00
15) 4-Methylphenol-d8	8.81	113	90426	24.02	ng/ul	0.00
21) Nitrobenzene-d5	9.28	128	38728	27.12	ng/ul	0.00
24) 2-Nitrophenol-d4	10.01	143	43100	27.56	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.55	165	93711	24.39	ng/ul	0.00
31) 4-Chloroaniline-d4	11.07	131	114649	23.69	ng/ul	0.00
46) Dimethylphthalate-d6	14.16	166	341894	25.67	ng/ul	0.00
49) Acenaphthylene-d8	14.45	160	390006	26.67	ng/ul	0.00
54) 4-Nitrophenol-d4	14.92	143	49029	24.72	ng/ul	0.00
60) Fluorene-d10	15.75	176	295477	25.14	ng/ul	0.00
65) 4,6-Dinitro-2-methylphenol	15.85	200	49515	22.70	ng/ul	0.00
73) Anthracene-d10	17.61	188	510935	26.20	ng/ul	0.00
81) Pyrene-d10	19.89	212	638855	25.90	ng/ul	0.00
92) Benzo(a)pyrene-d12	24.92	264	667421	26.55	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG031521\
Data File : BG048379.D
Acq On : 15 Mar 2021 11:20
Operator : CG/JU
Sample : PB135014BL
Misc : SFAM-MDL-W-BL02
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SBLK014

Quant Time: Mar 15 13:01:54 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SFAM-EPA-BG031121.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Mar 15 10:45:12 2021
Response via : Initial Calibration

