

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG031524\
 Data File : BG060660.D
 Acq On : 15 Mar 2024 11:44
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 15 13:18:36 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG031324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 14 00:45:22 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	97	0.00
2	1,4-Dioxane	40.000	37.593	6.0	93	0.00
3	Pyridine	40.000	39.711	0.7	95	0.00
4	n-Nitrosodimethylamine	40.000	40.238	-0.6	97	0.00
5 S	2-Fluorophenol	80.000	80.299	-0.4	95	0.00
6	Aniline	40.000	39.302	1.7	92	0.00
7 S	Phenol-d6	80.000	80.358	-0.4	94	0.00
8	2-Chlorophenol	40.000	40.329	-0.8	95	0.00
9	Benzaldehyde	40.000	37.870	5.3	87	0.00
10 C	Phenol	40.000	39.694	0.8	92	0.00
11	bis(2-Chloroethyl)ether	40.000	39.330	1.7	93	0.00
12	1,3-Dichlorobenzene	40.000	38.312	4.2	93	0.00
13 C	1,4-Dichlorobenzene	40.000	38.351	4.1	91	0.00
14	1,2-Dichlorobenzene	40.000	37.769	5.6	91	0.00
15	Benzyl Alcohol	40.000	39.147	2.1	89	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.922	5.2	90	-0.01
17	2-Methylphenol	40.000	39.509	1.2	91	0.00
18	Hexachloroethane	40.000	39.423	1.4	94	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.918	2.7	90	0.00
20	3+4-Methylphenols	40.000	39.829	0.4	91	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	97	0.00
22	Acetophenone	40.000	37.297	6.8	89	0.00
23 S	Nitrobenzene-d5	80.000	78.585	1.8	92	0.00
24	Nitrobenzene	40.000	38.795	3.0	90	0.00
25	Isophorone	40.000	37.519	6.2	89	0.00
26 C	2-Nitrophenol	40.000	40.366	-0.9	99	0.00
27	2,4-Dimethylphenol	40.000	37.496	6.3	92	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.505	3.7	93	0.00
29 C	2,4-Dichlorophenol	40.000	39.387	1.5	92	0.00
30	1,2,4-Trichlorobenzene	40.000	37.717	5.7	93	0.00
31	Naphthalene	40.000	37.401	6.5	91	0.00
32	Benzoic acid	40.000	39.661	0.8	96	-0.01
33	4-Chloroaniline	40.000	38.043	4.9	91	0.00
34 C	Hexachlorobutadiene	40.000	37.488	6.3	91	0.00
35	Caprolactam	40.000	36.082	9.8	85	0.00
36 C	4-Chloro-3-methylphenol	40.000	37.767	5.6	89	0.00
37	2-Methylnaphthalene	40.000	37.597	6.0	91	0.00
38	1-Methylnaphthalene	40.000	37.507	6.2	90	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	94	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.883	5.3	89	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.675	3.3	87	0.00
42 S	2,4,6-Tribromophenol	80.000	83.575	-4.5	93	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.868	0.3	91	0.00
44	2,4,5-Trichlorophenol	40.000	38.981	2.5	88	0.00
45 S	2-Fluorobiphenyl	80.000	76.369	4.5	90	0.00
46	1,1'-Biphenyl	40.000	38.218	4.5	90	0.00
47	2-Chloronaphthalene	40.000	38.008	5.0	88	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG031524\
 Data File : BG060660.D
 Acq On : 15 Mar 2024 11:44
 Operator : MA/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 15 13:18:36 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG031324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 14 00:45:22 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.801	-2.0	88	0.00
49	Acenaphthylene	40.000	38.358	4.1	90	0.00
50	Dimethylphthalate	40.000	38.053	4.9	88	0.00
51	2,6-Dinitrotoluene	40.000	41.361	-3.4	91	0.00
52 C	Acenaphthene	40.000	37.976	5.1	90	0.00
53	3-Nitroaniline	40.000	39.482	1.3	89	0.00
54 P	2,4-Dinitrophenol	40.000	36.287	9.3	87	0.00
55	Dibenzofuran	40.000	37.286	6.8	88	0.00
56 P	4-Nitrophenol	40.000	38.175	4.6	87	0.00
57	2,4-Dinitrotoluene	40.000	42.171	-5.4	92	0.00
58	Fluorene	40.000	37.504	6.2	88	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.562	1.1	89	0.00
60	Diethylphthalate	40.000	37.421	6.4	87	-0.01
61	4-Chlorophenyl-phenylether	40.000	37.521	6.2	89	0.00
62	4-Nitroaniline	40.000	39.313	1.7	87	0.00
63	Azobenzene	40.000	37.325	6.7	87	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	89	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.013	-0.0	90	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.297	1.8	87	0.00
67	4-Bromophenyl-phenylether	40.000	41.836	-4.6	94	0.00
68	Hexachlorobenzene	40.000	38.485	3.8	86	0.00
69	Atrazine	40.000	39.731	0.7	86	0.00
70 C	Pentachlorophenol	40.000	39.560	1.1	83	0.00
71	Phenanthrene	40.000	38.116	4.7	85	0.00
72	Anthracene	40.000	38.698	3.3	86	0.00
73	Carbazole	40.000	38.403	4.0	85	0.00
74	Di-n-butylphthalate	40.000	39.904	0.2	85	0.00
75 C	Fluoranthene	40.000	38.561	3.6	84	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	90	-0.01
77	Benzydine	40.000	44.328	-10.8	89	0.00
78	Pyrene	40.000	38.162	4.6	85	0.00
79 S	Terphenyl-d14	80.000	76.265	4.7	87	0.00
80	Butylbenzylphthalate	40.000	41.767	-4.4	89	-0.01
81	Benzo(a)anthracene	40.000	38.371	4.1	85	-0.01
82	3,3'-Dichlorobenzidine	40.000	40.152	-0.4	85	0.00
83	Chrysene	40.000	38.241	4.4	85	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.044	-2.6	87	-0.01
85 c	Di-n-octyl phthalate	40.000	41.978	-4.9	88	-0.02
86 I	Perylene-d12	20.000	20.000	0.0	89	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.013	-0.0	87	-0.03
88	Benzo(b)fluoranthene	40.000	38.918	2.7	84	-0.01
89	Benzo(k)fluoranthene	40.000	39.497	1.3	87	-0.01
90 C	Benzo(a)pyrene	40.000	39.238	1.9	86	-0.01
91	Dibenzo(a,h)anthracene	40.000	39.878	0.3	87	-0.03
92	Benzo(g,h,i)perylene	40.000	39.820	0.4	86	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG031524\
Data File : BG060660.D
Acq On : 15 Mar 2024 11:44
Operator : MA/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Mar 15 13:18:36 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG031324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 14 00:45:22 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0