


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG031616\
Data File : BG021424.D
Acq On : 16 Mar 2016 10:08
Operator : UM/SJ
Sample : DFTPP08
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP08

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG031616.M
Title : SVOA CALIBRATION
Last Update : Wed Mar 16 16:06:35 2016



AutoFind: Scans 2554, 2555, 2556; Background Corrected with Scan 2549

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.3	8470	PASS
68	69	0.00	2	1.6	150	PASS
69	198	0.00	100	43.5	9619	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	49.7	10990	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	22125	PASS
199	198	5	9	7.1	1568	PASS
275	198	10	60	29.1	6429	PASS
365	198	1	100	4.6	1028	PASS
441	442	0.01	100	16.2	2329	PASS
442	198	50	100	65.1	14414	PASS
443	442	15	24	18.4	2645	PASS

m/z

Abundance

Instrument :

BNA_G

ClientSampleId :

DFTPP08

m/z	Abundance
39.00	705.0
50.00	2790.0
51.00	10300.0
52.00	481.0
55.90	374.0
57.00	641.0
63.00	525.0
64.90	255.0
68.00	230.0
68.90	10662.0
72.90	183.0
74.00	1269.0
75.00	1908.0
76.00	556.0
77.00	11120.0
78.00	872.0
79.00	954.0
79.90	677.0
80.90	1127.0
81.90	256.0
82.90	161.0
84.90	153.0
86.00	233.0
86.80	154.0
91.00	207.0
92.00	238.0
92.90	1513.0
97.90	1178.0
98.90	877.0
100.90	515.0
102.90	200.0
103.90	299.0
104.90	258.0
106.90	3780.0
107.90	545.0
108.90	176.0
109.90	6318.0
110.90	1021.0
115.80	309.0
116.90	3470.0
117.90	228.0
121.90	274.0
122.80	443.0
123.80	199.0
124.90	222.0
126.90	12241.0
127.90	959.0
128.90	5527.0
129.90	472.0
133.90	221.0
134.80	453.0
136.90	245.0
140.90	668.0
141.90	239.0
146.90	372.0
147.90	889.0
148.80	254.0
152.90	232.0
153.90	204.0
155.00	445.0
155.90	651.0
159.90	231.0
160.90	310.0
164.90	316.0
165.90	241.0
166.90	1448.0
167.90	561.0
172.90	226.0
173.90	270.0
174.90	633.0
176.90	287.0
178.90	1041.0
179.90	739.0
181.00	273.0
184.90	542.0
185.90	3852.0
187.00	1031.0
188.80	268.0
191.90	242.0
193.00	514.0
195.80	504.0
197.90	22544.0
198.90	1821.0
203.00	240.0
203.90	1097.0
204.90	1951.0
205.90	6997.0
206.90	1099.0
207.90	388.0
210.80	198.0
215.80	199.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

216.90	1880.0
217.90	202.0
220.90	809.0
222.90	428.0
223.90	3709.0
224.90	1051.0
226.90	1936.0
227.90	309.0
228.90	257.0
230.90	190.0
233.90	173.0
236.90	174.0
241.90	191.0
242.90	254.0
243.90	2871.0
245.00	400.0
245.90	919.0
246.80	154.0
254.90	15683.0
255.90	1947.0
256.90	204.0
257.90	1040.0
264.90	421.0
272.90	539.0
273.90	1431.0
274.90	6180.0
275.90	751.0
276.90	646.0
295.90	2253.0
296.90	352.0
302.80	171.0
314.90	284.0
323.00	535.0
333.80	345.0
351.80	196.0
364.90	805.0
371.90	237.0
422.90	724.0
440.90	1731.0
441.90	11343.0
442.90	2252.0
443.90	275.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

m/z	Abundance
39.00	664.0
49.00	157.0
50.00	2732.0
51.00	9647.0
52.00	619.0
56.00	254.0
57.00	732.0
61.90	181.0
62.90	476.0
65.00	194.0
67.90	222.0
68.90	11179.0
73.00	163.0
74.00	1394.0
75.00	1768.0
76.00	580.0
77.00	11966.0
78.00	875.0
78.90	838.0
79.90	709.0
80.90	898.0
81.80	225.0
82.90	205.0
84.90	225.0
85.90	387.0
90.90	192.0
92.90	1596.0
97.90	1383.0
99.00	988.0
100.90	554.0
103.90	403.0
104.90	291.0
106.90	3991.0
107.90	579.0
109.90	6771.0
110.90	1216.0
115.90	227.0
116.90	4001.0
117.90	314.0
121.90	364.0
122.90	576.0
123.80	255.0
124.90	243.0
126.90	12322.0
127.90	952.0
128.90	5812.0
129.90	485.0
133.90	205.0
134.90	654.0
135.90	225.0
136.90	229.0
140.90	756.0
141.90	276.0
142.90	172.0
145.90	187.0
147.00	499.0
147.90	1036.0
152.90	265.0
153.80	157.0
154.90	395.0
155.90	738.0
157.00	177.0
159.80	221.0
160.90	266.0
164.90	311.0
165.80	302.0
166.90	1636.0
167.90	661.0
171.90	165.0
173.90	312.0
174.90	531.0
176.90	438.0
178.90	1058.0
179.90	821.0
181.00	311.0
184.90	587.0
185.90	3899.0
186.90	1628.0
187.90	152.0
188.80	381.0
190.90	181.0
191.90	481.0
192.90	487.0
195.90	748.0
197.90	26480.0
198.80	1790.0
199.90	157.0
202.90	281.0
203.90	1063.0
204.90	2060.0
205.90	8253.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

206.90	1118.0
207.90	451.0
208.80	185.0
210.90	347.0
214.80	209.0
215.90	216.0
216.90	2462.0
217.90	327.0
220.90	1047.0
221.70	331.0
222.90	378.0
223.90	4652.0
224.90	1051.0
226.00	186.0
226.90	2314.0
227.90	319.0
228.80	512.0
234.80	214.0
236.90	178.0
240.80	163.0
241.80	315.0
243.00	230.0
244.00	3365.0
244.90	491.0
245.90	848.0
246.80	181.0
254.90	18352.0
255.90	2568.0
256.90	286.0
257.90	1422.0
258.90	217.0
264.90	621.0
272.90	646.0
273.90	1466.0
274.90	7740.0
275.90	1040.0
276.90	616.0
277.80	174.0
284.90	174.0
292.80	186.0
295.90	2911.0
296.80	275.0
302.90	326.0
314.90	280.0
315.90	171.0
322.90	692.0
326.90	199.0
333.80	448.0
334.90	154.0
345.90	221.0
351.90	248.0
352.90	276.0
354.00	153.0
364.90	1230.0
365.80	169.0
371.90	308.0
373.00	169.0
402.90	153.0
420.90	210.0
422.90	904.0
423.90	188.0
440.90	2878.0
441.90	16936.0
442.90	3268.0
443.80	255.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

m/z	Abundance
39.00	351.0
50.00	1463.0
51.00	5464.0
52.00	369.0
55.90	161.0
56.90	355.0
63.00	186.0
68.90	7017.0
73.00	198.0
73.90	665.0
74.90	1222.0
76.00	353.0
77.00	7469.0
78.00	482.0
78.90	652.0
80.00	441.0
80.90	665.0
81.90	200.0
85.90	164.0
92.90	1117.0
97.90	867.0
98.90	604.0
100.90	263.0
103.00	199.0
103.80	225.0
105.00	187.0
107.00	2561.0
108.00	387.0
109.90	4400.0
110.90	526.0
115.90	184.0
116.90	2471.0
118.00	199.0
121.90	182.0
122.90	233.0
123.90	153.0
126.90	8408.0
127.90	643.0
128.90	3635.0
129.80	231.0
134.90	301.0
136.90	167.0
140.90	483.0
142.00	166.0
146.90	418.0
147.90	577.0
148.90	169.0
152.90	278.0
153.90	183.0
155.00	197.0
156.00	331.0
159.90	165.0
160.80	233.0
164.80	233.0
165.90	199.0
166.90	898.0
167.90	604.0
168.90	177.0
173.80	237.0
174.90	440.0
176.80	196.0
178.80	740.0
179.90	629.0
180.80	188.0
184.90	572.0
185.90	2884.0
186.90	838.0
188.80	196.0
191.80	200.0
192.90	309.0
195.90	694.0
197.90	17352.0
198.80	1095.0
202.80	205.0
203.90	983.0
204.90	1185.0
205.90	5068.0
206.90	1009.0
207.90	274.0
210.80	247.0
215.80	197.0
216.80	1715.0
220.80	716.0
221.80	243.0
222.90	468.0
223.90	3099.0
224.90	839.0
226.90	1544.0
227.80	244.0
228.90	378.0
233.90	183.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

242.90	238.0
243.90	2440.0
244.90	350.0
245.90	606.0
254.90	12875.0
255.90	1782.0
256.80	198.0
257.90	994.0
264.90	411.0
272.90	465.0
273.90	892.0
274.90	5369.0
275.90	779.0
276.90	573.0
295.90	1864.0
296.90	363.0
302.90	225.0
314.70	292.0
322.90	638.0
333.90	359.0
364.80	1051.0
371.90	297.0
402.90	166.0
422.90	769.0
424.00	174.0
440.90	2380.0
441.90	14964.0
442.90	2415.0
444.00	239.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08