

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG031620\
 Data File : BG044819.D
 Acq On : 16 Mar 2020 18:58
 Operator : CG/JU
 Sample : L1912-25
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 Z1-5

Manual Integrations
 APPROVED

mohammad
 3/18/2020 8:44:19 AM

Quant Time: Mar 17 04:35:01 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG031020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 16 17:37:15 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.05	152	100305	20.00	ng	0.00
21) Naphthalene-d8	10.85	136	398028	20.00	ng	0.00
39) Acenaphthene-d10	14.68	164	273975	20.00	ng	0.00
64) Phenanthrene-d10	17.42	188	611187	20.00	ng	0.00
76) Chrysene-d12	21.70	240	524323	20.00	ng	0.00
87) Perylene-d12	24.91	264	598460	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.62	112	586239	90.98	ng	0.00
7) Phenol-d6	7.21	99	837152	89.42	ng	0.00
23) Nitrobenzene-d5	9.20	82	543971	59.97	ng	0.00
42) 2,4,6-Tribromophenol	16.17	330	344888	96.14	ng	0.00
45) 2-Fluorobiphenyl	13.30	172	1077035	56.30	ng	0.00
79) Terphenyl-d14	20.04	244	1569974	56.01	ng	0.00
Target Compounds						
50) Dimethylphthalate	14.13	163	140788	6.452	ng	Qvalue # 97
75) Fluoranthene	19.47	202	186802	4.804	ng	95
78) Pyrene	19.83	202	176749	4.595	ng	98
81) Benzo(a)anthracene	21.68	228	134533	3.563	ng	97
83) Chrysene	21.74	228	118834	3.295	ng	99
88) Benzo(b)fluoranthene	23.89	252	160423	4.060	ng	98
90) Benzo(a)pyrene	24.76	252	104890m	2.837	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG031620\
Data File : BG044819.D
Acq On : 16 Mar 2020 18:58
Operator : CG/JU
Sample : L1912-25
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
Z1-5

Manual Integrations
APPROVED

mohammad
3/18/2020 8:44:19 AM

Quant Time: Mar 17 04:35:01 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG031020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 16 17:37:15 2020
Response via : Initial Calibration

