

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG031720\
 Data File : BG044836.D
 Acq On : 17 Mar 2020 17:12
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 17 18:14:41 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG031020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 16 17:37:15 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	78	0.00
2	1,4-Dioxane	0.619	0.509	17.8	64	0.00
3	Pyridine	1.832	1.530	16.5	64	0.00
4	n-Nitrosodimethylamine	0.695	0.581	16.4	64	0.00
5 S	2-Fluorophenol	1.285	1.172	8.8	72	0.00
6	Aniline	2.323	2.122	8.7	72	0.00
7 S	Phenol-d6	1.867	1.727	7.5	74	0.00
8	2-Chlorophenol	1.283	1.223	4.7	77	0.00
9	Benzaldehyde	1.089	0.978	10.2	68	0.00
10 C	Phenol	1.848	1.686	8.8	73	0.00
11	bis(2-Chloroethyl)ether	1.475	1.302	11.7	71	0.00
12	1,3-Dichlorobenzene	1.580	1.447	8.4	73	0.00
13 C	1,4-Dichlorobenzene	1.562	1.416	9.3	72	0.00
14	1,2-Dichlorobenzene	1.481	1.362	8.0	74	0.00
15	Benzyl Alcohol	1.498	1.403	6.3	74	0.00
16	2,2'-oxybis(1-Chloropropane	2.603	2.096	19.5	63	0.00
17	2-Methylphenol	1.252	1.175	6.2	74	0.00
18	Hexachloroethane	0.615	0.572	7.0	74	0.00
19 P	n-Nitroso-di-n-propylamine	1.318	1.186	10.0	70	0.00
20	3+4-Methylphenols	1.726	1.635	5.3	73	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	78	0.00
22	Acetophenone	0.567	0.513	9.5	74	0.00
23 S	Nitrobenzene-d5	0.456	0.419	8.1	75	0.00
24	Nitrobenzene	0.473	0.422	10.8	73	0.00
25	Isophorone	0.890	0.804	9.7	74	0.00
26 C	2-Nitrophenol	0.154	0.178	-15.6	94	0.00
27	2,4-Dimethylphenol	0.290	0.269	7.2	77	0.00
28	bis(2-Chloroethoxy)methane	0.531	0.473	10.9	72	0.00
29 C	2,4-Dichlorophenol	0.338	0.328	3.0	78	0.00
30	1,2,4-Trichlorobenzene	0.405	0.377	6.9	78	0.00
31	Naphthalene	1.108	1.006	9.2	74	0.00
32	Benzoic acid	0.201	0.234	-16.4	100	0.00
33	4-Chloroaniline	0.499	0.476	4.6	79	0.00
34 C	Hexachlorobutadiene	0.266	0.256	3.8	78	0.00
35	Caprolactam	0.128	0.142	-10.9	87	0.00
36 C	4-Chloro-3-methylphenol	0.404	0.398	1.5	79	0.00
37	2-Methylnaphthalene	0.829	0.776	6.4	75	0.00
38	1-Methylnaphthalene	0.779	0.731	6.2	76	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	85	0.00
40	1,2,4,5-Tetrachlorobenzene	0.659	0.597	9.4	82	0.00
41 P	Hexachlorocyclopentadiene	0.360	0.318	11.7	81	0.00
42 S	2,4,6-Tribromophenol	0.262	0.283	-8.0	97	0.00
43 C	2,4,6-Trichlorophenol	0.437	0.421	3.7	86	0.00
44	2,4,5-Trichlorophenol	0.453	0.443	2.2	89	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG031720\
 Data File : BG044836.D
 Acq On : 17 Mar 2020 17:12
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 17 18:14:41 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG031020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 16 17:37:15 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.397	1.250	10.5	79	0.00
46	1,1'-Biphenyl	1.598	1.402	12.3	78	0.00
47	2-Chloronaphthalene	1.221	1.056	13.5	78	0.00
48	2-Nitroaniline	0.427	0.413	3.3	87	0.00
49	Acenaphthylene	1.913	1.702	11.0	80	0.00
50	Dimethylphthalate	1.593	1.481	7.0	83	0.00
51	2,6-Dinitrotoluene	0.321	0.324	-0.9	89	0.00
52 C	Acenaphthene	1.253	1.163	7.2	85	0.00
53	3-Nitroaniline	0.369	0.369	0.0	87	0.00
54 P	2,4-Dinitrophenol	0.143	0.188	-31.5#	124	0.00
55	Dibenzofuran	1.817	1.660	8.6	82	0.00
56 P	4-Nitrophenol	0.282	0.285	-1.1	90	0.00
57	2,4-Dinitrotoluene	0.440	0.470	-6.8	94	0.00
58	Fluorene	1.494	1.380	7.6	83	0.00
59	2,3,4,6-Tetrachlorophenol	0.417	0.430	-3.1	91	0.00
60	Diethylphthalate	1.661	1.569	5.5	84	0.00
61	4-Chlorophenyl-phenylether	0.805	0.758	5.8	86	0.00
62	4-Nitroaniline	0.394	0.388	1.5	88	0.00
63	Azobenzene	1.725	1.487	13.8	77	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	89	0.00
65	4,6-Dinitro-2-methylphenol	0.087	0.110	-26.4#	124	0.00
66 c	n-Nitrosodiphenylamine	0.602	0.539	10.5	84	0.00
67	4-Bromophenyl-phenylether	0.250	0.232	7.2	89	0.00
68	Hexachlorobenzene	0.271	0.253	6.6	89	0.00
69	Atrazine	0.221	0.206	6.8	84	0.00
70 C	Pentachlorophenol	0.149	0.154	-3.4	94	0.00
71	Phenanthrene	1.069	0.976	8.7	86	0.00
72	Anthracene	1.054	0.961	8.8	85	0.00
73	Carbazole	1.012	0.915	9.6	84	0.00
74	Di-n-butylphthalate	1.231	1.140	7.4	85	0.00
75 C	Fluoranthene	1.273	1.173	7.9	85	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	84	0.00
77	Benzidine	0.649	0.636	2.0	77	0.00
78	Pyrene	1.467	1.427	2.7	85	0.00
79 S	Terphenyl-d14	1.069	1.033	3.4	86	0.00
80	Butylbenzylphthalate	0.653	0.605	7.4	80	0.00
81	Benzo(a)anthracene	1.440	1.348	6.4	82	0.00
82	3,3'-Dichlorobenzidine	0.557	0.528	5.2	81	0.00
83	Chrysene	1.376	1.270	7.7	81	0.00
84	Bis(2-ethylhexyl)phthalate	0.901	0.818	9.2	79	0.00
85 c	Di-n-octyl phthalate	1.508	1.402	7.0	80	0.00
86	Indeno(1,2,3-cd)pyrene	1.804	1.621	10.1	80	-0.01
87 I	Perylene-d12	1.000	1.000	0.0	83	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG031720\
Data File : BG044836.D
Acq On : 17 Mar 2020 17:12
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Mar 17 18:14:41 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG031020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 16 17:37:15 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.320	1.208	8.5	80	0.00
89	Benzo(k)fluoranthene	1.250	1.122	10.2	80	0.00
90 C	Benzo(a)pyrene	1.235	1.118	9.5	80	0.00
91	Dibenzo(a,h)anthracene	1.219	1.115	8.5	81	-0.02
92	Benzo(q,h,i)perylene	1.231	1.115	9.4	80	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0