


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG031816\
 Data File : BG021441.D
 Acq On : 18 Mar 2016 8:49
 Operator : UM/SJ
 Sample : DFTPP10
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP10

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG031616.M
 Title : SVOA CALIBRATION
 Last Update : Sat Mar 19 00:12:07 2016



AutoFind: Scans 2554, 2555, 2556; Background Corrected with Scan 2549

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.2	13253	PASS
68	69	0.00	2	0.5	70	PASS
69	198	0.00	100	38.3	14859	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	46.6	18074	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	38789	PASS
199	198	5	9	6.8	2641	PASS
275	198	10	60	28.3	10959	PASS
365	198	1	100	4.0	1549	PASS
441	442	0.01	100	16.6	4161	PASS
442	198	50	100	64.6	25072	PASS
443	442	15	24	18.1	4549	PASS

m/z

Abundance

Instrument :

BNA_G

ClientSampleId :

DFTPP10

m/z	Abundance
38.00	178.0
39.00	950.0
50.00	3722.0
51.00	14030.0
52.00	757.0
55.90	348.0
56.90	892.0
60.90	201.0
62.00	174.0
63.00	638.0
65.00	299.0
68.90	14732.0
73.00	228.0
74.00	1695.0
75.00	2563.0
76.00	969.0
77.00	16992.0
78.00	1366.0
78.90	1155.0
79.90	977.0
80.90	1250.0
81.90	344.0
83.00	306.0
84.90	217.0
85.90	419.0
86.90	179.0
91.00	320.0
92.00	345.0
92.90	2398.0
97.90	1677.0
98.90	1236.0
100.90	692.0
102.90	209.0
103.90	525.0
104.90	428.0
106.90	5832.0
107.90	759.0
108.90	243.0
109.90	9359.0
110.90	1239.0
111.90	159.0
115.90	298.0
116.90	4988.0
117.90	466.0
121.90	435.0
122.90	629.0
123.90	328.0
124.90	226.0
126.90	16280.0
127.90	1326.0
128.90	7859.0
129.90	681.0
130.90	154.0
134.00	269.0
134.90	551.0
135.90	342.0
136.90	402.0
140.90	1199.0
141.90	303.0
142.90	202.0
146.00	174.0
147.00	577.0
147.90	1370.0
148.90	358.0
152.90	321.0
153.90	179.0
154.90	515.0
155.90	678.0
157.00	174.0
157.80	254.0
158.90	206.0
159.90	435.0
160.90	338.0
161.90	189.0
164.90	391.0
166.00	288.0
166.90	1635.0
167.90	909.0
168.90	164.0
173.00	292.0
173.90	431.0
174.90	986.0
175.90	289.0
176.90	308.0
178.90	1752.0
179.90	1034.0
180.90	538.0
184.90	743.0
186.00	5733.0
186.90	1752.0
187.90	156.0

Instrument :
BNA_G
ClientSampleId :
DFTPP10

188.80	355.0
190.90	237.0
191.90	423.0
193.00	615.0
193.90	153.0
195.90	931.0
197.90	33288.0
198.90	2168.0
200.00	226.0
201.60	182.0
202.90	411.0
203.90	1587.0
204.90	2474.0
206.00	9581.0
207.00	1462.0
207.90	522.0
210.90	318.0
215.90	247.0
216.90	2966.0
217.90	348.0
220.90	1330.0
221.80	324.0
222.90	690.0
223.90	5632.0
224.90	1339.0
226.90	2492.0
227.90	335.0
228.90	572.0
230.80	187.0
233.80	252.0
234.90	282.0
235.90	156.0
236.90	300.0
241.90	277.0
242.90	300.0
243.90	4212.0
245.00	626.0
245.90	1031.0
246.90	218.0
248.80	155.0
254.90	21056.0
255.90	3457.0
256.80	282.0
257.90	1331.0
258.90	273.0
264.90	606.0
272.90	713.0
273.90	1676.0
274.90	8629.0
275.90	1335.0
276.90	1033.0
284.90	152.0
293.00	238.0
295.90	3179.0
296.90	364.0
302.90	288.0
314.90	328.0
322.90	816.0
334.00	415.0
340.90	155.0
345.90	203.0
351.90	166.0
352.90	263.0
353.80	169.0
364.90	1054.0
365.90	168.0
371.90	349.0
402.80	211.0
421.90	189.0
422.90	1008.0
440.90	2547.0
441.90	15969.0
442.90	3257.0
443.90	249.0

Instrument :
BNA_G
ClientSampleId :
DFTPP10

m/z	Abundance
38.00	330.0
39.00	1245.0
49.00	171.0
50.00	4577.0
51.00	15367.0
52.00	942.0
55.90	444.0
57.00	1215.0
61.00	235.0
62.00	307.0
63.00	739.0
65.00	281.0
68.90	17472.0
72.90	246.0
74.00	1994.0
75.00	2888.0
76.00	829.0
77.00	19856.0
78.00	1537.0
79.00	1361.0
79.90	1155.0
80.90	1667.0
82.00	456.0
83.00	381.0
84.90	336.0
85.90	500.0
87.00	207.0
90.90	444.0
91.90	354.0
92.90	2999.0
93.90	237.0
97.90	2100.0
98.90	1770.0
101.00	927.0
102.90	364.0
103.90	649.0
104.90	431.0
106.90	7377.0
107.90	1002.0
108.90	195.0
109.90	11635.0
110.90	1543.0
111.90	201.0
115.90	456.0
116.90	6387.0
117.90	518.0
121.90	499.0
123.00	831.0
124.00	342.0
124.90	339.0
126.90	21600.0
127.90	1842.0
128.90	9717.0
129.90	705.0
130.90	187.0
133.90	282.0
135.00	762.0
135.90	340.0
136.90	361.0
138.00	173.0
139.90	171.0
140.90	1426.0
141.90	459.0
142.90	230.0
145.80	238.0
146.90	680.0
147.90	1701.0
148.90	292.0
151.00	188.0
152.90	548.0
153.90	249.0
154.90	727.0
156.00	1026.0
156.80	222.0
157.90	219.0
158.80	258.0
159.90	464.0
160.90	753.0
161.90	207.0
164.90	653.0
166.00	414.0
166.90	2780.0
167.90	1184.0
170.80	176.0
171.90	251.0
172.90	281.0
173.90	646.0
174.90	821.0
175.90	239.0
176.90	703.0
178.00	200.0

Instrument :
BNA_G
ClientSampleId :
DFTPP10

178.90	1947.0
179.90	1463.0
180.90	590.0
183.90	199.0
184.90	1064.0
185.90	6967.0
186.90	2234.0
188.00	300.0
188.90	742.0
190.90	258.0
191.90	624.0
192.90	780.0
193.80	179.0
195.90	1184.0
197.90	46864.0
198.90	3426.0
199.90	283.0
200.90	153.0
202.90	544.0
203.90	2172.0
204.90	3306.0
205.90	13449.0
206.90	2067.0
207.90	643.0
210.00	315.0
210.90	382.0
216.00	350.0
216.90	4076.0
217.90	491.0
220.90	2068.0
221.90	493.0
222.90	841.0
223.90	8389.0
225.00	2099.0
226.00	165.0
226.90	4164.0
227.90	456.0
228.90	630.0
230.90	343.0
233.90	252.0
235.00	257.0
236.90	460.0
240.00	192.0
240.90	223.0
241.80	382.0
242.90	395.0
243.90	6625.0
244.90	724.0
245.90	1385.0
247.00	200.0
248.90	254.0
252.90	153.0
254.90	31856.0
255.90	4735.0
256.90	352.0
257.90	2207.0
258.90	375.0
265.00	874.0
265.90	152.0
271.80	182.0
272.90	920.0
273.90	2360.0
274.90	13127.0
275.90	1832.0
276.90	1409.0
277.90	173.0
284.90	203.0
292.80	308.0
295.90	4144.0
296.90	734.0
303.00	521.0
313.90	259.0
314.90	546.0
315.90	299.0
320.90	174.0
322.90	1159.0
323.90	297.0
326.90	241.0
334.00	583.0
334.90	243.0
345.90	231.0
352.00	308.0
352.90	264.0
353.90	305.0
364.90	2040.0
365.90	186.0
371.90	705.0
401.90	211.0
402.90	242.0
420.90	313.0
421.90	246.0
422.90	1590.0
423.90	385.0
440.90	4927.0
441.90	29200.0

Instrument :
BNA_G
ClientSampleId :
DFTPP10

442.905179.0

443.90563.0

Instrument :
BNA_G
ClientSampleId :
DFTPP10

m/z	Abundance
39.00	717.0
49.00	193.0
50.00	3230.0
51.00	10362.0
52.00	529.0
56.00	317.0
57.00	882.0
62.10	184.0
63.00	462.0
65.00	288.0
68.00	210.0
68.90	12373.0
72.90	237.0
74.00	1438.0
74.90	2304.0
76.00	677.0
77.00	14212.0
78.00	1121.0
79.00	1059.0
79.90	746.0
80.90	1193.0
81.90	248.0
82.80	226.0
85.00	203.0
85.90	344.0
91.00	329.0
92.00	306.0
92.90	1999.0
97.90	1726.0
99.00	1067.0
100.90	589.0
102.90	248.0
103.90	417.0
104.90	310.0
105.80	178.0
106.90	5298.0
107.90	647.0
109.90	7891.0
110.90	1353.0
111.90	173.0
116.00	398.0
116.90	5110.0
117.90	341.0
121.90	447.0
122.90	591.0
123.90	252.0
124.90	212.0
126.90	16342.0
128.00	1439.0
128.90	7415.0
129.90	722.0
133.80	257.0
134.90	559.0
136.00	313.0
136.90	185.0
140.90	941.0
142.00	358.0
142.90	227.0
145.90	239.0
146.90	585.0
147.90	1259.0
148.90	326.0
152.90	382.0
153.90	231.0
154.90	464.0
155.90	881.0
157.90	229.0
159.90	399.0
160.90	450.0
164.90	431.0
165.90	297.0
166.90	1776.0
167.90	716.0
168.90	203.0
171.80	222.0
173.00	385.0
173.90	449.0
175.00	786.0
175.90	181.0
176.80	344.0
178.90	1578.0
179.90	841.0
180.90	464.0
185.00	860.0
185.90	5493.0
186.90	1641.0
187.90	187.0
188.90	511.0
190.90	247.0
191.90	543.0
192.90	567.0

Instrument :
BNA_G
ClientSampleId :
DFTPP10

195.90	1110.0
197.90	36216.0
198.90	2329.0
202.90	269.0
203.90	1672.0
204.90	2553.0
206.00	10411.0
206.90	1579.0
207.90	545.0
208.90	156.0
210.90	377.0
214.90	201.0
215.90	219.0
216.90	2905.0
217.90	443.0
220.90	1373.0
221.90	367.0
222.90	629.0
223.90	6664.0
224.90	1368.0
225.80	206.0
226.90	3362.0
227.90	530.0
228.90	536.0
230.90	221.0
233.80	173.0
234.80	253.0
235.80	153.0
238.80	227.0
241.80	217.0
242.90	343.0
243.90	4476.0
245.00	639.0
245.90	1314.0
246.90	382.0
249.00	155.0
252.80	199.0
254.90	26760.0
255.90	3541.0
257.00	279.0
257.90	1953.0
258.90	365.0
264.90	677.0
272.90	881.0
273.90	2181.0
274.90	11121.0
275.90	1406.0
276.90	1153.0
277.90	244.0
292.90	270.0
295.90	4137.0
296.90	515.0
302.90	462.0
314.00	193.0
314.90	416.0
316.00	214.0
322.90	1000.0
323.90	154.0
326.80	176.0
333.90	756.0
334.90	182.0
345.90	248.0
351.90	397.0
352.90	186.0
353.80	309.0
364.90	1554.0
365.80	201.0
371.90	572.0
382.80	227.0
401.90	259.0
402.80	295.0
420.90	217.0
421.80	241.0
422.90	1618.0
423.90	352.0
440.90	5011.0
441.90	30048.0
442.90	5212.0
443.90	718.0

Instrument :
BNA_G
ClientSampleId :
DFTPP10