

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG031822\
 Data File : BG052710.D
 Acq On : 18 Mar 2022 15:05
 Operator : CG/JU
 Sample : N1882-11
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DBRY1

Quant Time: Mar 18 23:59:53 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG031822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 18 00:55:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.152	152	34212	20.000	ng/u1	0.00
20) Naphthalene-d8	10.966	136	137736	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.766	164	99510	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.510	188	236623	20.000	ng/u1	0.00
79) Chrysene-d12	21.798	240	252091	20.000	ng/u1	0.00
88) Perylene-d12	25.111	264	247529	20.000	ng/u1	-0.01

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.558	96	5156	5.323	ng/uL	0.00
4) Pyridine-d5	3.981	84	62597	25.888	ng/u1	0.00
7) Phenol-d5	7.288	99	85527	27.912	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.465	67	53898	28.800	ng/u1	0.00
11) 2-Chlorophenol-d4	7.676	132	58791	28.256	ng/u1	0.00
15) 4-Methylphenol-d8	8.839	113	61444	27.681	ng/u1	0.00
21) Nitrobenzene-d5	9.309	128	29496	29.749	ng/u1	0.00
24) 2-Nitrophenol-d4	10.032	143	30933	29.579	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.572	165	60112	26.422	ng/u1	0.00
31) 4-Chloroaniline-d4	11.083	131	84022	27.943	ng/u1	0.00
46) Dimethylphthalate-d6	14.167	166	218613	28.596	ng/u1	0.00
49) Acenaphthylene-d8	14.461	160	258736	27.852	ng/u1	0.00
54) 4-Nitrophenol-d4	14.931	143	30947	24.077	ng/u1	0.00
60) Fluorene-d10	15.759	176	188129	28.027	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.859	200	25074	20.823	ng/u1	0.00
73) Anthracene-d10	17.610	188	317867	28.848	ng/u1	0.00
81) Pyrene-d10	19.889	212	396703	28.130	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.888	264	366006	28.617	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG031822\
Data File : BG052710.D
Acq On : 18 Mar 2022 15:05
Operator : CG/JU
Sample : N1882-11
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DBRY1

Quant Time: Mar 18 23:59:53 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG031822.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Mar 18 00:55:42 2022
Response via : Initial Calibration

