

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG031822\
 Data File : BG052720.D
 Acq On : 18 Mar 2022 21:36
 Operator : CG/JU
 Sample : N1882-12
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DBRY2

Quant Time: Mar 19 00:01:46 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG031822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 18 00:55:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.147	152	34877	20.000	ng/u1	0.00
20) Naphthalene-d8	10.966	136	145084	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.767	164	102825	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.510	188	243650	20.000	ng/u1	0.00
79) Chrysene-d12	21.799	240	266730	20.000	ng/u1	0.00
88) Perylene-d12	25.112	264	257026	20.000	ng/u1	-0.01

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.565	96	7093	7.184	ng/uL	0.00
4) Pyridine-d5	3.982	84	83774	33.985	ng/u1	0.00
7) Phenol-d5	7.289	99	115458	36.961	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.465	67	71261	37.351	ng/u1	0.00
11) 2-Chlorophenol-d4	7.677	132	77597	36.584	ng/u1	0.00
15) 4-Methylphenol-d8	8.840	113	83084	36.717	ng/u1	0.00
21) Nitrobenzene-d5	9.304	128	38881	37.229	ng/u1	0.00
24) 2-Nitrophenol-d4	10.038	143	42640	38.709	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.573	165	81132	33.855	ng/u1	0.00
31) 4-Chloroaniline-d4	11.084	131	113529	35.844	ng/u1	0.00
46) Dimethylphthalate-d6	14.168	166	297226	37.626	ng/u1	0.00
49) Acenaphthylene-d8	14.462	160	354317	36.912	ng/u1	0.00
54) 4-Nitrophenol-d4	14.931	143	46266	34.835	ng/u1	0.00
60) Fluorene-d10	15.760	176	257902	37.182	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.860	200	40727	32.846	ng/u1	0.00
73) Anthracene-d10	17.610	188	425033	37.462	ng/u1	0.00
81) Pyrene-d10	19.889	212	538751	36.106	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.889	264	503981	37.949	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG031822\
Data File : BG052720.D
Acq On : 18 Mar 2022 21:36
Operator : CG/JU
Sample : N1882-12
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DBRY2

Quant Time: Mar 19 00:01:46 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG031822.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Mar 18 00:55:42 2022
Response via : Initial Calibration

