

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG031822\
 Data File : BG052730.D
 Acq On : 19 Mar 2022 4:05
 Operator : CG/JU
 Sample : PB143278BS
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SLCS278

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 03/21/2022
 Supervised By :Yogesh Patel 03/24/2022

Quant Time: Mar 19 04:50:59 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG031822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 18 00:55:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.151	152	38130	20.000	ng/u1	0.00
20) Naphthalene-d8	10.965	136	154157	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.772	164	110720	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.515	188	260782	20.000	ng/u1	0.00
79) Chrysene-d12	21.803	240	273345	20.000	ng/u1	0.00
88) Perylene-d12	25.117	264	288653	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.558	96	6467	5.991	ng/uL	0.00
4) Pyridine-d5	3.981	84	85181	31.608	ng/u1	0.00
7) Phenol-d5	7.294	99	113992	33.378	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.470	67	68707	32.940	ng/u1	0.00
11) 2-Chlorophenol-d4	7.676	132	77422	33.387	ng/u1	0.00
15) 4-Methylphenol-d8	8.845	113	83288	33.667	ng/u1	0.00
21) Nitrobenzene-d5	9.309	128	39075	35.213	ng/u1	0.00
24) 2-Nitrophenol-d4	10.037	143	41060	35.081	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.572	165	85384	33.533	ng/u1	0.00
31) 4-Chloroaniline-d4	11.083	131	105216	31.265	ng/u1	0.00
46) Dimethylphthalate-d6	14.167	166	281137	33.052	ng/u1	0.00
49) Acenaphthylene-d8	14.466	160	340325	32.926	ng/u1	0.00
54) 4-Nitrophenol-d4	14.936	143	49654	34.720	ng/u1	0.00
60) Fluorene-d10	15.759	176	246545	33.011	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.864	200	45690	34.428	ng/u1	0.00
73) Anthracene-d10	17.609	188	405045	33.355	ng/u1	0.00
81) Pyrene-d10	19.888	212	527258	34.481	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.893	264	518586	34.770	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.599	88	13396	10.129	ng/uL	95
5) Pyridine	3.998	79	90793	31.180	ng/u1	94
6) Benzaldehyde	7.282	77	75460m	33.237	ng/u1	
8) Phenol	7.323	94	105743	32.127	ng/u1	95
10) Bis(2-Chloroethyl)ether	7.558	93	80381	32.372	ng/u1	98
12) 2-Chlorophenol	7.711	128	78607	33.178	ng/u1	97
13) 2-Methylphenol	8.580	108	84554	33.053	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.674	45	135997	33.696	ng/u1#	96
16) Acetophenone	8.968	105	126070	32.666	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.956	70	72304	32.136	ng/u1	95
18) 4-Methylphenol	8.909	108	89744	33.673	ng/u1	98
19) Hexachloroethane	9.244	117	32294	30.721	ng/u1	92
22) Nitrobenzene	9.350	77	109330	33.718	ng/u1	92
23) Isophorone	9.878	82	199689	32.220	ng/u1	96
25) 2-Nitrophenol	10.066	139	42175	34.748	ng/u1#	92
26) 2,4-Dimethylphenol	10.119	107	94783	31.516	ng/u1	96
27) Bis(2-Chloroethoxy)met...	10.354	93	112026	32.504	ng/u1	97
29) 2,4-Dichlorophenol	10.601	162	79203	31.993	ng/u1	98
30) Naphthalene	11.012	128	254240	31.869	ng/u1#	96
32) 4-Chloroaniline	11.112	127	100694	30.450	ng/u1	97
33) Hexachlorobutadiene	11.306	225	65141	32.218	ng/u1	99
34) Caprolactam	11.870	113	26716m	35.831	ng/u1	
35) 4-Chloro-3-methylphenol	12.216	107	92390	34.403	ng/u1	94
36) 2-Methylnaphthalene	12.610	142	179853	31.946	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG031822\
 Data File : BG052730.D
 Acq On : 19 Mar 2022 4:05
 Operator : CG/JU
 Sample : PB143278BS
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SLCS278

Quant Time: Mar 19 04:50:59 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG031822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 18 00:55:42 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 03/21/2022
 Supervised By :Yogesh Patel 03/24/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.827	142	184100	32.300	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.974	216	113962	31.140	ng/ul	99
40) Hexachlorocyclopentadiene	12.957	237	76192	30.073	ng/ul	97
41) 2,4,6-Trichlorophenol	13.203	196	73703	32.958	ng/ul	98
42) 2,4,5-Trichlorophenol	13.274	196	79471	32.908	ng/ul	98
43) 1,1'-Biphenyl	13.609	154	249528	31.538	ng/ul	99
44) 2-Chloronaphthalene	13.650	162	199545	31.497	ng/ul	97
45) 2-Nitroaniline	13.838	65	65020	35.011	ng/ul	97
47) Dimethylphthalate	14.214	163	256702	31.624	ng/ul	99
48) 2,6-Dinitrotoluene	14.331	165	51069	34.730	ng/ul#	93
50) Acenaphthylene	14.490	152	315800	31.636	ng/ul	99
51) 3-Nitroaniline	14.660	138	49898	33.619	ng/ul	95
52) Acenaphthene	14.831	153	212684	32.613	ng/ul	99
53) 2,4-Dinitrophenol	14.860	184	29583	32.384	ng/ul#	71
55) 4-Nitrophenol	14.954	109	52878	35.133	ng/ul	98
56) Dibenzofuran	15.165	168	306151	31.666	ng/ul	98
57) 2,4-Dinitrotoluene	15.113	165	72076	34.546	ng/ul#	90
58) 2,3,4,6-Tetrachlorophenol	15.389	232	76595	35.187	ng/ul#	97
59) Diethylphthalate	15.577	149	273605	33.281	ng/ul	98
61) Fluorene	15.812	166	252113	31.460	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.800	204	137071	32.144	ng/ul	96
63) 4-Nitroaniline	15.817	138	52502	35.971	ng/ul	93
66) 4,6-Dinitro-2-methylph...	15.876	198	45355	34.929	ng/ul	97
67) N-Nitrosodiphenylamine	16.011	169	228489	32.406	ng/ul	96
68) 4-Bromophenyl-phenylether	16.699	248	85728	33.201	ng/ul	97
69) Hexachlorobenzene	16.822	284	109758	32.906	ng/ul	97
70) Atrazine	16.957	200	100728	34.858	ng/ul	97
71) Pentachlorophenol	17.157	266	67851	33.452	ng/ul	98
72) Phenanthrene	17.556	178	454121	33.180	ng/ul	99
74) Anthracene	17.644	178	445275	32.541	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.573	216	122782	31.612	ng/ul	96
76) Pentachlorobenzene	15.089	250	123556	33.157	ng/ul	99
77) Carbazole	17.909	167	414190	34.250	ng/ul	99
78) Di-n-butylphthalate	18.467	149	493354	34.307	ng/ul	99
80) Fluoranthene	19.559	202	603149	33.471	ng/ul	99
82) Pyrene	19.918	202	591444	33.021	ng/ul	98
83) Butylbenzylphthalate	20.805	149	222143	35.672	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.686	252	197195	32.187	ng/ul	99
85) Benzo(a)anthracene	21.780	228	588074	33.334	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.686	149	305824	35.457	ng/ul	99
87) Chrysene	21.850	228	559569	33.253	ng/ul	99
89) Di-n-octyl phthalate	22.937	149	519264	34.538	ng/ul	100
90) Benzo(b)fluoranthene	24.065	252	623209	33.436	ng/ul	99
91) Benzo(k)fluoranthene	24.130	252	557115	32.386	ng/ul	98
93) Benzo(a)pyrene	24.964	252	581507	33.045	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	28.912	276	670892	33.710	ng/ul	98
95) Dibenzo(a,h)anthracene	28.982	278	568144	33.588	ng/ul	98
96) Benzo(g,h,i)perylene	30.098	276	564751	33.569	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG031822\
 Data File : BG052730.D
 Acq On : 19 Mar 2022 4:05
 Operator : CG/JU
 Sample : PB143278BS
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :

BNA_G

ClientSampleId :

SLCS278

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 03/21/2022

Supervised By :Yogesh Patel 03/24/2022


Quant Time: Mar 19 04:50:59 2022


Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG031822.M

Quant Title : SVOA CALIBRATION

QLast Update : Fri Mar 18 00:55:42 2022


Response via : Initial Calibration





#2
1,4-Dioxane
Concen: 10.129 ng/uL
RT: 3.599 min Scan# 30
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05




Instrument :
BNA_G
ClientSampleId :
SLCS278




Tgt Ion: 88 Resp: 13390
Ion Ratio Lower Upper
88 100
43 37.1 30.6 46.0
58 78.5 58.6 87.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/21/2022
Supervised By :Yogesh Patel 03/24/2022




#5
Pyridine
Concen: 31.180 ng/uL
RT: 3.998 min Scan# 104
Delta R.T. -0.007 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05




Tgt Ion: 79 Resp: 90793
Ion Ratio Lower Upper
79 100
52 76.9 57.2 85.8
51 35.9 31.9 47.9





#6
Benzaldehyde
Concen: 33.237 ng/ul m
RT: 7.282 min Scan# 60
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05




Instrument :
BNA_G
ClientSampleId :
SLCS278




Tgt Ion: 77 Resp: 75460
Ion Ratio Lower Upper
77 100
105 87.3 72.9 109.3
106 86.7 66.5 99.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/21/2022
Supervised By :Yogesh Patel 03/24/2022




#8
Phenol
Concen: 32.127 ng/ul
RT: 7.323 min Scan# 670
Delta R.T. 0.004 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05




Tgt Ion: 94 Resp: 105743
Ion Ratio Lower Upper
94 100
65 30.3 25.9 38.9
66 40.6 35.4 53.2





#10
Bis(2-Chloroethyl)ether
Concen: 32.372 ng/ul
RT: 7.558 min Scan# 710
Delta R.T. -0.007 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05




Instrument :
BNA_G
ClientSampleId :
SLCS278




Tgt Ion: 93 Resp: 8038
Ion Ratio Lower Upper
93 100
63 86.6 68.2 102.4
95 33.4 24.7 37.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/21/2022
Supervised By :Yogesh Patel 03/24/2022




#12
2-Chlorophenol
Concen: 33.178 ng/ul
RT: 7.711 min Scan# 736
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05




Tgt Ion:128 Resp: 78607
Ion Ratio Lower Upper
128 100
64 51.9 39.8 59.6
130 30.6 25.8 38.8





#13
2-Methylphenol
Concen: 33.053 ng/ul
RT: 8.580 min Scan# 84554
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05




Instrument :
BNA_G
ClientSampleId :
SLCS278




Tgt Ion:108 Resp: 84554
Ion Ratio Lower Upper
108 100
107 85.6 68.3 102.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/21/2022
Supervised By :Yogesh Patel 03/24/2022




#14
2,2'-oxybis(1-Chloropropane)
Concen: 33.696 ng/ul
RT: 8.674 min Scan# 900
Delta R.T. -0.007 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05




Tgt Ion: 45 Resp: 135997
Ion Ratio Lower Upper
45 100
77 13.3 11.0 16.4
79 8.7 9.3 13.9#





#16
Acetophenone
Concen: 32.666 ng/ul
RT: 8.968 min Scan# 91
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05




Instrument :
BNA_G
ClientSampleId :
SLCS278



Tgt Ion:105 Resp: 126070
Ion Ratio Lower Upper
105 100
77 91.1 72.9 109.3
51 32.2 25.0 37.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/21/2022
Supervised By :Yogesh Patel 03/24/2022



#17
N-Nitroso-di-n-propylamine
Concen: 32.136 ng/ul
RT: 8.956 min Scan# 948
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05

Tgt Ion: 70 Resp: 72304
Ion Ratio Lower Upper
70 100
42 64.4 47.8 71.6
101 9.3 6.8 10.2
130 17.7 14.7 22.1






#18
4-Methylphenol
Concen: 33.673 ng/ul
RT: 8.909 min Scan# 940
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05

Instrument :

BNA_G

ClientSampleId :

SLCS278





Tgt Ion:108 Resp: 89744
Ion Ratio Lower Upper
108 100
107 105.8 82.7 124.1

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 03/21/2022
Supervised By :Yogesh Patel 03/24/2022



#19
Hexachloroethane
Concen: 30.721 ng/ul
RT: 9.244 min Scan# 997
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05

Tgt Ion:117 Resp: 32294
Ion Ratio Lower Upper
117 100
201 111.5 83.9 125.9
199 72.0 51.4 77.2





#22

Nitrobenzene

Concen: 33.718 ng/ul

RT: 9.350 min Scan# 1015

Delta R.T. -0.001 min

Lab File: BG052730.D

Acq: 19 Mar 2022 4:05

Instrument :

BNA_G

ClientSampleId :

SLCS278





Tgt Ion: 77 Resp: 109330

Ion	Ratio	Lower	Upper
77	100		
123	37.2	25.7	38.5
65	14.1	9.7	14.5

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 03/21/2022
Supervised By :Yogesh Patel 03/24/2022



#23

Isophorone

Concen: 32.220 ng/ul

RT: 9.878 min Scan# 1105




Delta R.T. -0.001 min


Lab File: BG052730.D

Acq: 19 Mar 2022 4:05

Tgt Ion: 82 Resp: 199689

Ion	Ratio	Lower	Upper
82	100		
95	8.1	6.1	9.1
138	15.8	11.0	16.4









#25
2-Nitrophenol
Concen: 34.748 ng/ul
RT: 10.066 min Scan# 1137
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05

Instrument :
BNA_G
ClientSampleId :
SLCS278

Tgt Ion:	139	Resp:	4217
Ion Ratio	Lower	Upper	
139	100		
65	63.3	48.9	73.3
109	45.8	29.0	43.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/21/2022
Supervised By :Yogesh Patel 03/24/2022



#26
2,4-Dimethylphenol
Concen: 31.516 ng/ul
RT: 10.119 min Scan# 1146
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05

Tgt Ion:	107	Resp:	94783
Ion Ratio	Lower	Upper	
107	100		
121	44.7	37.4	56.0
122	75.2	63.0	94.4





#27

Bis(2-Chloroethoxy)methane

Concen: 32.504 ng/ul

RT: 10.354 min Scan# 1186

Delta R.T. -0.001 min

Lab File: BG052730.D

Acq: 19 Mar 2022 4:05

Instrument :

BNA_G

ClientSampleId :

SLCS278

Tgt Ion: 93 Resp: 112020

Ion Ratio Lower Upper

93 100

95 30.0 25.4 38.2





123 9.1 8.2 12.2

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 03/21/2022

Supervised By :Yogesh Patel 03/24/2022



#29

2,4-Dichlorophenol

Concen: 31.993 ng/ul

RT: 10.601 min Scan# 1228

Delta R.T. -0.001 min

Lab File: BG052730.D

Acq: 19 Mar 2022 4:05




Tgt Ion:162 Resp: 79203


Ion Ratio Lower Upper

162 100

164 64.3 53.2 79.8

98 36.9 29.1 43.7









#30
Naphthalene
Concen: 31.869 ng/ul
RT: 11.012 min Scan# 1299
Delta R.T. -0.007 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05

Instrument :
BNA_G
ClientSampleId :
SLCS278

Tgt Ion:128 Resp: 254240
Ion Ratio Lower Upper
128 100
129 11.5 9.0 13.6
127 14.6 9.8 14.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/21/2022
Supervised By :Yogesh Patel 03/24/2022



#32
4-Chloroaniline
Concen: 30.450 ng/ul
RT: 11.112 min Scan# 1315
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05

Tgt Ion:127 Resp: 100694
Ion Ratio Lower Upper
127 100
129 33.5 25.3 37.9





#33
Hexachlorobutadiene
Concen: 32.218 ng/ul
RT: 11.306 min Scan# 11
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05

Instrument :

BNA_G

ClientSampleId :





SLCS278

Tgt Ion:225 Resp: 6514
Ion Ratio Lower Upper
225 100
223 63.5 51.7 77.5
227 61.1 49.6 74.4

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 03/21/2022
Supervised By :Yogesh Patel 03/24/2022



#34
Caprolactam
Concen: 35.831 ng/ul m
RT: 11.870 min Scan# 1444
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05

Tgt Ion:113 Resp: 26716
Ion Ratio Lower Upper
113 100
55 183.9 142.0 213.0
56 160.3 96.1 144.1#









#35
4-Chloro-3-methylphenol
Concen: 34.403 ng/ul
RT: 12.216 min Scan# 1503
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05

Instrument :
BNA_G
ClientSampleId :
SLCS278

Tgt Ion:107 Resp: 92390
Ion Ratio Lower Upper
107 100
144 24.1 21.8 32.6
142 75.7 64.6 97.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/21/2022
Supervised By :Yogesh Patel 03/24/2022



#36
2-Methylnaphthalene
Concen: 31.946 ng/ul
RT: 12.610 min Scan# 1570
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05


Tgt Ion:142 Resp: 179853
Ion Ratio Lower Upper
142 100
141 88.8 71.2 106.8





#37
1-Methylnaphthalene
Concen: 32.300 ng/ul
RT: 12.827 min Scan# 1607
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05




Instrument :
BNA_G
ClientSampleId :
SLCS278




Tgt Ion:142 Resp: 184100
Ion Ratio Lower Upper
142 100
141 91.1 71.4 107.0
116 4.9 3.5 5.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/21/2022
Supervised By :Yogesh Patel 03/24/2022




#39
1,2,4,5-Tetrachlorobenzene
Concen: 31.140 ng/ul
RT: 12.974 min Scan# 1632
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05




Tgt Ion:216 Resp: 113962
Ion Ratio Lower Upper
216 100
214 78.1 62.0 93.0
179 18.4 15.7 23.5
108 16.7 12.9 19.3





#40
Hexachlorocyclopentadiene
Concen: 30.073 ng/ul
RT: 12.957 min Scan# 1629
Delta R.T. -0.007 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05




Instrument :
BNA_G
ClientSampleId :
SLCS278




Tgt Ion:237 Resp: 76192
Ion Ratio Lower Upper
237 100
235 65.8 50.1 75.1
272 13.0 10.2 15.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/21/2022
Supervised By :Yogesh Patel 03/24/2022




#41
2,4,6-Trichlorophenol
Concen: 32.958 ng/ul
RT: 13.203 min Scan# 1671
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05



Tgt Ion:196 Resp: 73703
Ion Ratio Lower Upper
196 100
198 91.8 75.4 113.2
200 31.2 24.9 37.3





#42
2,4,5-Trichlorophenol
Concen: 32.908 ng/ul
RT: 13.274 min Scan# 1683
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05

Instrument :

BNA_G

ClientSampleId :





SLCS278

Tgt Ion:196 Resp: 7947
Ion Ratio Lower Upper
196 100
198 94.3 77.0 115.4
200 30.0 24.2 36.4

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 03/21/2022
Supervised By :Yogesh Patel 03/24/2022



#43
1,1'-Biphenyl
Concen: 31.538 ng/ul
RT: 13.609 min Scan# 1740
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05


Tgt Ion:154 Resp: 249528
Ion Ratio Lower Upper
154 100
153 41.1 33.0 49.4
76 12.0 10.6 15.8





#44
2-Chloronaphthalene
Concen: 31.497 ng/ul
RT: 13.650 min Scan# 1747
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05




Instrument :
BNA_G
ClientSampleId :
SLCS278




Tgt Ion:162 Resp: 19954
Ion Ratio Lower Upper
162 100
164 31.4 26.6 40.0
127 37.3 28.2 42.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/21/2022
Supervised By :Yogesh Patel 03/24/2022




#45
2-Nitroaniline
Concen: 35.011 ng/ul
RT: 13.838 min Scan# 1779
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05



Tgt Ion: 65 Resp: 65020
Ion Ratio Lower Upper
65 100
92 67.4 54.0 81.0
138 90.1 76.0 114.0









#47
Dimethylphthalate
Concen: 31.624 ng/ul
RT: 14.214 min Scan# 1843
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05

Instrument :
BNA_G
ClientSampleId :
SLCS278

Tgt Ion:163 Resp: 25670
Ion Ratio Lower Upper
163 100
194 3.3 2.9 4.3
164 9.9 7.7 11.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/21/2022
Supervised By :Yogesh Patel 03/24/2022



#48
2,6-Dinitrotoluene
Concen: 34.730 ng/ul
RT: 14.331 min Scan# 1863
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05


Tgt Ion:165 Resp: 51069
Ion Ratio Lower Upper
165 100
89 71.5 53.4 80.2
121 20.8 21.2 31.8#





#50
Acenaphthylene
Concen: 31.636 ng/ul
RT: 14.490 min Scan# 1890
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05




Instrument :
BNA_G
ClientSampleId :
SLCS278




Tgt Ion:152 Resp: 315800
Ion Ratio Lower Upper
152 100
151 20.5 16.4 24.6
153 13.0 11.2 16.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/21/2022
Supervised By :Yogesh Patel 03/24/2022




#51
3-Nitroaniline
Concen: 33.619 ng/ul
RT: 14.660 min Scan# 1919
Delta R.T. 0.004 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05



Tgt Ion:138 Resp: 49898
Ion Ratio Lower Upper
138 100
108 11.3 7.8 11.8
92 123.0 103.5 155.3









#52
Acenaphthene
Concen: 32.613 ng/u1
RT: 14.831 min Scan# 1948
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05

Instrument :
BNA_G
ClientSampleId :
SLCS278

Tgt Ion:153 Resp: 212684
Ion Ratio Lower Upper
153 100
152 45.8 38.0 57.0
154 87.0 70.0 105.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/21/2022
Supervised By :Yogesh Patel 03/24/2022



#53
2,4-Dinitrophenol
Concen: 32.384 ng/u1
RT: 14.860 min Scan# 1953
Delta R.T. 0.004 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05


Tgt Ion:184 Resp: 29583
Ion Ratio Lower Upper
184 100
63 59.5 49.3 73.9
154 86.3 108.6 162.8#





#55
4-Nitrophenol
Concen: 35.133 ng/ul
RT: 14.954 min Scan# 1969
Delta R.T. 0.004 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05




Instrument :
BNA_G
ClientSampleId :
SLCS278



Tgt Ion:109 Resp: 52878
Ion Ratio Lower Upper
109 100
139 96.2 75.4 113.2
65 116.2 91.4 137.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/21/2022
Supervised By :Yogesh Patel 03/24/2022



#56
Dibenzofuran
Concen: 31.666 ng/ul
RT: 15.165 min Scan# 2005
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05

Tgt Ion:168 Resp: 306151
Ion Ratio Lower Upper
168 100
139 38.5 31.8 47.6





#57
2,4-Dinitrotoluene
Concen: 34.546 ng/ul
RT: 15.113 min Scan# 1996
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05

Instrument :

BNA_G

ClientSampleId :





SLCS278

Tgt Ion:165 Resp: 72070
Ion Ratio Lower Upper
165 100
63 47.7 33.0 49.4
182 2.9 3.4 5.0

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 03/21/2022
Supervised By :Yogesh Patel 03/24/2022



#58
2,3,4,6-Tetrachlorophenol
Concen: 35.187 ng/ul
RT: 15.389 min Scan# 2043
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05

Tgt Ion:232 Resp: 76595
Ion Ratio Lower Upper
232 100
131 40.3 29.8 44.8
130 2.3 1.0 1.4#
166 27.0 22.0 33.0





#59
Diethylphthalate
Concen: 33.281 ng/ul
RT: 15.577 min Scan# 2075
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05

Instrument :

BNA_G

ClientSampleId :





SLCS278

Tgt Ion:149 Resp: 27360
Ion Ratio Lower Upper
149 100
177 21.0 17.5 26.3
150 13.1 9.5 14.3

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 03/21/2022
Supervised By :Yogesh Patel 03/24/2022



#61
Fluorene
Concen: 31.460 ng/ul
RT: 15.812 min Scan# 2115
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05

Tgt Ion:166 Resp: 252113
Ion Ratio Lower Upper
166 100
165 94.5 75.7 113.5
167 13.2 10.3 15.5









#62
4-Chlorophenyl-phenylether
Concen: 32.144 ng/ul
RT: 15.800 min Scan# 2113
Delta R.T. -0.007 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05

Instrument :
BNA_G
ClientSampleId :
SLCS278

Tgt Ion:204 Resp: 13707
Ion Ratio Lower Upper
204 100
206 31.7 27.4 41.0
141 57.9 44.2 66.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/21/2022
Supervised By :Yogesh Patel 03/24/2022



#63
4-Nitroaniline
Concen: 35.971 ng/ul
RT: 15.817 min Scan# 2116
Delta R.T. 0.004 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05

Tgt Ion:138 Resp: 52502
Ion Ratio Lower Upper
138 100
92 57.1 43.0 64.6
108 114.6 84.6 126.8









#66
4,6-Dinitro-2-methylphenol
Concen: 34.929 ng/ul
RT: 15.876 min Scan# 2126
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05

Instrument :
BNA_G
ClientSampleId :
SLCS278

Tgt Ion:198 Resp: 4535
Ion Ratio Lower Upper
198 100
182 4.1 3.0 4.6
77 23.2 19.7 29.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/21/2022
Supervised By :Yogesh Patel 03/24/2022



#67
N-Nitrosodiphenylamine
Concen: 32.406 ng/ul
RT: 16.011 min Scan# 2149
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05

Tgt Ion:169 Resp: 228489
Ion Ratio Lower Upper
169 100
168 68.2 57.4 86.0
167 37.6 28.1 42.1






#68
4-Bromophenyl-phenylether
Concen: 33.201 ng/ul
RT: 16.699 min Scan# 21
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05

Instrument :

BNA_G

ClientSampleId :

SLCS278






Tgt Ion:248 Resp: 85728
Ion Ratio Lower Upper
248 100
250 95.3 76.3 114.5
141 67.6 58.7 88.1


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 03/21/2022
Supervised By :Yogesh Patel 03/24/2022




#69
Hexachlorobenzene
Concen: 32.906 ng/ul
RT: 16.822 min Scan# 2287
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05




Tgt Ion:284 Resp: 109758
Ion Ratio Lower Upper
284 100
142 28.0 22.3 33.5
249 30.3 21.9 32.9





#70
Atrazine
Concen: 34.858 ng/ul
RT: 16.957 min Scan# 210
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05




Instrument :
BNA_G
ClientSampleId :
SLCS278




Tgt Ion:200 Resp: 100728
Ion Ratio Lower Upper
200 100
173 23.8 19.8 29.6
215 52.0 39.5 59.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/21/2022
Supervised By :Yogesh Patel 03/24/2022




#71
Pentachlorophenol
Concen: 33.452 ng/ul
RT: 17.157 min Scan# 2344
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05




Tgt Ion:266 Resp: 67851
Ion Ratio Lower Upper
266 100
264 63.1 50.3 75.5
268 60.2 45.8 68.8





#72
Phenanthrene
Concen: 33.180 ng/u1
RT: 17.556 min Scan# 2412
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05




Instrument :
BNA_G
ClientSampleId :
SLCS278




Tgt Ion:178 Resp: 45412
Ion Ratio Lower Upper
178 100
179 15.7 11.9 17.9
176 18.7 15.0 22.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/21/2022
Supervised By :Yogesh Patel 03/24/2022




#74
Anthracene
Concen: 32.541 ng/u1
RT: 17.644 min Scan# 2427
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05



Tgt Ion:178 Resp: 445275
Ion Ratio Lower Upper
178 100
179 15.2 12.8 19.2
176 17.9 14.7 22.1









#75
1,2,3,4-Tetrachlorobenzene
Concen: 31.612 ng/uL
RT: 13.573 min Scan# 1734
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05

Instrument :
BNA_G
ClientSampleId :
SLCS278

Tgt Ion:216 Resp: 122782
Ion Ratio Lower Upper
216 100
214 83.1 63.8 95.8
218 51.4 38.3 57.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/21/2022
Supervised By :Yogesh Patel 03/24/2022



#76
Pentachlorobenzene
Concen: 33.157 ng/uL
RT: 15.089 min Scan# 1992
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05

Tgt Ion:250 Resp: 123556
Ion Ratio Lower Upper
250 100
248 61.9 50.4 75.6
252 61.9 49.2 73.8









#77
Carbazole
Concen: 34.250 ng/ul
RT: 17.909 min Scan# 2471
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05

Instrument :
BNA_G
ClientSampleId :
SLCS278

Tgt Ion:167 Resp: 414190
Ion Ratio Lower Upper
167 100
166 21.7 17.3 25.9
139 12.8 9.4 14.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/21/2022
Supervised By :Yogesh Patel 03/24/2022



#78
Di-n-butylphthalate
Concen: 34.307 ng/ul
RT: 18.467 min Scan# 2567
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05


Tgt Ion:149 Resp: 493354
Ion Ratio Lower Upper
149 100
150 8.8 7.5 11.3
104 6.1 4.6 6.8





#80
Fluoranthene
Concen: 33.471 ng/u1
RT: 19.559 min Scan# 2753
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05




Instrument :
BNA_G
ClientSampleId :
SLCS278




Tgt Ion:202 Resp: 603149
Ion Ratio Lower Upper
202 100
101 7.8 6.1 9.1
100 6.4 5.4 8.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/21/2022
Supervised By :Yogesh Patel 03/24/2022




#82
Pyrene
Concen: 33.021 ng/u1
RT: 19.918 min Scan# 2814
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05



Tgt Ion:202 Resp: 591444
Ion Ratio Lower Upper
202 100
101 8.9 8.2 12.4
100 8.3 6.8 10.2





#83
Butylbenzylphthalate
Concen: 35.672 ng/ul
RT: 20.805 min Scan# 2965
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05

Instrument :

BNA_G

ClientSampleId :





SLCS278

Tgt Ion:149 Resp: 22214
Ion Ratio Lower Upper
149 100
91 72.3 57.1 85.7
206 22.4 17.4 26.2

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 03/21/2022
Supervised By :Yogesh Patel 03/24/2022



#84
3,3'-Dichlorobenzidine
Concen: 32.187 ng/ul
RT: 21.686 min Scan# 3115
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05


Tgt Ion:252 Resp: 197195
Ion Ratio Lower Upper
252 100
254 61.2 48.6 73.0
126 7.8 6.2 9.2





#85
Benzo(a)anthracene
Concen: 33.334 ng/ul
RT: 21.780 min Scan# 3131
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05




Instrument :
BNA_G
ClientSampleId :
SLCS278




Tgt Ion:228 Resp: 588074
Ion Ratio Lower Upper
228 100
229 20.3 15.4 23.2
226 27.5 21.4 32.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/21/2022
Supervised By :Yogesh Patel 03/24/2022




#86
Bis(2-ethylhexyl)phthalate
Concen: 35.457 ng/ul
RT: 21.686 min Scan# 3115
Delta R.T. -0.007 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05




Tgt Ion:149 Resp: 305824
Ion Ratio Lower Upper
149 100
167 26.4 21.4 32.2
279 3.7 3.4 5.2





#87
Chrysene
Concen: 33.253 ng/ul
RT: 21.850 min Scan# 3142
Delta R.T. 0.004 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05




Instrument :
BNA_G
ClientSampleId :
SLCS278



Tgt Ion:228 Resp: 559569
Ion Ratio Lower Upper
228 100
226 30.3 23.9 35.9
229 19.6 16.0 24.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/21/2022
Supervised By :Yogesh Patel 03/24/2022



#89
Di-n-octyl phthalate
Concen: 34.538 ng/ul
RT: 22.937 min Scan# 3328
Delta R.T. -0.007 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05

Tgt Ion:149 Resp: 519264





#90
Benzo(b)fluoranthene
Concen: 33.436 ng/ul
RT: 24.065 min Scan# 3519
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05

Instrument :

BNA_G

ClientSampleId :





SLCS278

Tgt Ion:252 Resp: 623209
Ion Ratio Lower Upper
252 100
253 22.1 17.8 26.6
125 6.1 5.5 8.3

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 03/21/2022
Supervised By :Yogesh Patel 03/24/2022



#91
Benzo(k)fluoranthene
Concen: 32.386 ng/ul
RT: 24.130 min Scan# 3531
Delta R.T. -0.007 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05


Tgt Ion:252 Resp: 557115
Ion Ratio Lower Upper
252 100
253 22.5 17.4 26.0
125 6.4 6.0 9.0





#93
Benzo(a)pyrene
Concen: 33.045 ng/ul
RT: 24.964 min Scan# 30
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05




Instrument :
BNA_G
ClientSampleId :
SLCS278




Tgt Ion:252 Resp: 58150
Ion Ratio Lower Upper
252 100
253 21.4 16.5 24.7
125 6.6 6.0 9.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/21/2022
Supervised By :Yogesh Patel 03/24/2022




#94
Indeno(1,2,3-cd)pyrene
Concen: 33.710 ng/ul
RT: 28.912 min Scan# 4345
Delta R.T. -0.001 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05




Tgt Ion:276 Resp: 670892
Ion Ratio Lower Upper
276 100
138 13.1 9.3 13.9
277 24.6 19.4 29.2





#95
Dibenzo(a,h)anthracene
Concen: 33.588 ng/ul
RT: 28.982 min Scan# 41
Delta R.T. -0.007 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05




Instrument :
BNA_G
ClientSampleId :
SLCS278




Tgt Ion:278 Resp: 568144
Ion Ratio Lower Upper
278 100
139 10.2 8.1 12.1
279 24.9 19.0 28.4

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 03/21/2022
Supervised By :Yogesh Patel 03/24/2022



#96
Benzo(g,h,i)perylene
Concen: 33.569 ng/ul
RT: 30.098 min Scan# 4547
Delta R.T. -0.007 min
Lab File: BG052730.D
Acq: 19 Mar 2022 4:05



Tgt Ion:276 Resp: 564751
Ion Ratio Lower Upper
276 100
138 12.0 10.5 15.7
277 25.2 19.0 28.6

