

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG031822\
 Data File : BG052732.D
 Acq On : 19 Mar 2022 6:02
 Operator : CG/JU
 Sample : PB143360BL
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SBLK360

Quant Time: Mar 19 06:48:46 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG031822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 18 00:55:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.145	152	40559	20.000	ng/u1	0.00
20) Naphthalene-d8	10.959	136	160661	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.766	164	111840	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.509	188	272726	20.000	ng/u1	0.00
79) Chrysene-d12	21.797	240	321127	20.000	ng/u1	0.00
88) Perylene-d12	25.116	264	305508	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.557	96	7650	6.662	ng/uL	0.00
4) Pyridine-d5	3.980	84	98292	34.289	ng/u1	0.00
7) Phenol-d5	7.288	99	128293	35.316	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.464	67	81360	36.671	ng/u1	0.00
11) 2-Chlorophenol-d4	7.675	132	89496	36.283	ng/u1	0.00
15) 4-Methylphenol-d8	8.838	113	93249	35.436	ng/u1	0.00
21) Nitrobenzene-d5	9.308	128	44121	38.150	ng/u1	0.00
24) 2-Nitrophenol-d4	10.031	143	49004	40.173	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.571	165	93527	35.244	ng/u1	0.00
31) 4-Chloroaniline-d4	11.083	131	126282	36.005	ng/u1	0.00
46) Dimethylphthalate-d6	14.167	166	321434	37.411	ng/u1	0.00
49) Acenaphthylene-d8	14.460	160	384580	36.835	ng/u1	0.00
54) 4-Nitrophenol-d4	14.930	143	53575	37.087	ng/u1	0.00
60) Fluorene-d10	15.759	176	270238	35.820	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.858	200	47382	34.139	ng/u1	0.00
73) Anthracene-d10	17.609	188	465239	36.634	ng/u1	0.00
81) Pyrene-d10	19.888	212	607857	33.837	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.887	264	575257	36.442	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG031822\
Data File : BG052732.D
Acq On : 19 Mar 2022 6:02
Operator : CG/JU
Sample : PB143360BL
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SBLK360

Quant Time: Mar 19 06:48:46 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG031822.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Mar 18 00:55:42 2022
Response via : Initial Calibration

