

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG031822\
 Data File : BG052739.D
 Acq On : 19 Mar 2022 10:34
 Operator : CG/JU
 Sample : N1883-11MSD
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :

BNA_G

ClientSampleId :

DBSA8MSD

Manual Integrations

APPROVED

Reviewed By :Jagrut
Upadhyay

03/21/2022

Supervised By :Yogesh
Patel

03/24/2022

Quant Time: Mar 21 00:48:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG031822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 18 00:55:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.151	152	45567	20.000	ng/u1	0.00
20) Naphthalene-d8	10.959	136	192794	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.772	164	133539	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.509	188	315220	20.000	ng/u1	0.00
79) Chrysene-d12	21.797	240	304015	20.000	ng/u1	0.00
88) Perylene-d12	25.122	264	317421	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.557	96	5376	4.167	ng/uL	0.00
4) Pyridine-d5	3.980	84	68227	21.185	ng/u1	0.00
7) Phenol-d5	7.294	99	101588	24.891	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.464	67	63271	25.383	ng/u1	0.00
11) 2-Chlorophenol-d4	7.681	132	71903	25.947	ng/u1	0.00
15) 4-Methylphenol-d8	8.844	113	77413	26.185	ng/u1	0.00
21) Nitrobenzene-d5	9.308	128	34653	24.970	ng/u1	0.00
24) 2-Nitrophenol-d4	10.037	143	39717	27.133	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.571	165	81487	25.589	ng/u1	0.00
31) 4-Chloroaniline-d4	11.083	131	93083	22.116	ng/u1	0.00
46) Dimethylphthalate-d6	14.167	166	260983	25.439	ng/u1	0.00
49) Acenaphthylene-d8	14.460	160	312600	25.076	ng/u1	0.00
54) 4-Nitrophenol-d4	14.936	143	42976	24.915	ng/u1	0.00
60) Fluorene-d10	15.759	176	226712	25.168	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.864	200	40148	25.028	ng/u1	0.00
73) Anthracene-d10	17.609	188	357338	24.344	ng/u1	0.00
81) Pyrene-d10	19.888	212	448230	26.355	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.887	264	428379	26.119	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.599	88	19254	12.182	ng/uL	87
5) Pyridine	3.998	79	124352	35.735	ng/u1	94
6) Benzaldehyde	7.282	77	102278m	37.697	ng/u1	
8) Phenol	7.317	94	156906	39.890	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.558	93	112171	37.802	ng/u1	97
12) 2-Chlorophenol	7.711	128	104909	37.053	ng/u1	92
13) 2-Methylphenol	8.580	108	114901	37.586	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.680	45	180419	37.406	ng/u1	97
16) Acetophenone	8.968	105	169334	36.715	ng/u1	96
17) N-Nitroso-di-n-propyla...	8.956	70	100932	37.539	ng/u1	96
18) 4-Methylphenol	8.909	108	121046	38.005	ng/u1	89
19) Hexachloroethane	9.244	117	46641	37.127	ng/u1	94
22) Nitrobenzene	9.350	77	150230	37.047	ng/u1	95
23) Isophorone	9.878	82	281198	36.279	ng/u1	97
25) 2-Nitrophenol	10.066	139	59687	39.321	ng/u1	97
26) 2,4-Dimethylphenol	10.119	107	126322	33.585	ng/u1	96
27) Bis(2-Chloroethoxy)met...	10.354	93	155505	36.077	ng/u1	97
29) 2,4-Dichlorophenol	10.601	162	110097	35.559	ng/u1	99
30) Naphthalene	11.012	128	344091	34.488	ng/u1	97
32) 4-Chloroaniline	11.106	127	135524	32.769	ng/u1	99
33) Hexachlorobutadiene	11.306	225	90373	35.740	ng/u1	95
34) Caprolactam	11.876	113	36186m	38.806	ng/u1	
35) 4-Chloro-3-methylphenol	12.216	107	125661	37.415	ng/u1	90
36) 2-Methylnaphthalene	12.610	142	245578	34.878	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG031822\
 Data File : BG052739.D
 Acq On : 19 Mar 2022 10:34
 Operator : CG/JU
 Sample : N1883-11MSD
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :

BNA_G

ClientSampleId :

DBSA8MSD

Manual Integrations

APPROVED

Reviewed By :Jagrut
Upadhyay

03/21/2022


Supervised By :Yogesh
Patel

03/24/2022

Quant Time: Mar 21 00:48:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG031822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 18 00:55:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.827	142	251445	35.275	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.974	216	156252	35.400	ng/ul	96
40) Hexachlorocyclopentadiene	12.956	237	102780	33.635	ng/ul	98
41) 2,4,6-Trichlorophenol	13.203	196	101451	37.614	ng/ul	96
42) 2,4,5-Trichlorophenol	13.274	196	112956	38.781	ng/ul	95
43) 1,1'-Biphenyl	13.603	154	336772	35.291	ng/ul	99
44) 2-Chloronaphthalene	13.650	162	269909	35.323	ng/ul	98
45) 2-Nitroaniline	13.838	65	91727	40.952	ng/ul	99
47) Dimethylphthalate	14.214	163	355393	36.301	ng/ul	98
48) 2,6-Dinitrotoluene	14.331	165	71405	40.262	ng/ul#	94
50) Acenaphthylene	14.490	152	430469	35.754	ng/ul	99
51) 3-Nitroaniline	14.660	138	67920	37.942	ng/ul#	91
52) Acenaphthene	14.830	153	284901	36.222	ng/ul	98
53) 2,4-Dinitrophenol	14.860	184	44308	40.215	ng/ul#	72
55) 4-Nitrophenol	14.954	109	68296	37.622	ng/ul	95
56) Dibenzofuran	15.165	168	416687	35.735	ng/ul	99
57) 2,4-Dinitrotoluene	15.112	165	101021	40.146	ng/ul#	84
58) 2,3,4,6-Tetrachlorophenol	15.388	232	103013	39.236	ng/ul#	100
59) Diethylphthalate	15.576	149	370166	37.333	ng/ul	98
61) Fluorene	15.811	166	341307	35.312	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.806	204	184989	35.968	ng/ul	96
63) 4-Nitroaniline	15.817	138	70103	39.822	ng/ul	95
66) 4,6-Dinitro-2-methylph...	15.876	198	61532	39.204	ng/ul	93
67) N-Nitrosodiphenylamine	16.011	169	300943	35.310	ng/ul	96
68) 4-Bromophenyl-phenylether	16.698	248	123557	39.587	ng/ul	96
69) Hexachlorobenzene	16.822	284	147177	36.505	ng/ul	95
70) Atrazine	16.957	200	130103	37.248	ng/ul	98
71) Pentachlorophenol	17.157	266	93307	38.058	ng/ul	94
72) Phenanthrene	17.556	178	584068	35.304	ng/ul	100
74) Anthracene	17.644	178	573669	34.684	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.573	216	174441	37.156	ng/uL	97
76) Pentachlorobenzene	15.089	250	172425	38.281	ng/uL	97
77) Carbazole	17.909	167	527648	36.097	ng/ul	98
78) Di-n-butylphthalate	18.467	149	630785	36.289	ng/ul	99
80) Fluoranthene	19.559	202	748570	37.350	ng/ul	99
82) Pyrene	19.918	202	727812	36.536	ng/ul	98
83) Butylbenzylphthalate	20.799	149	273588	39.502	ng/ul	92
84) 3,3'-Dichlorobenzidine	21.686	252	228968	33.603	ng/ul	93
85) Benzo(a)anthracene	21.780	228	706834	36.024	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.686	149	371496	38.726	ng/ul	98
87) Chrysene	21.850	228	665203	35.543	ng/ul	98
89) Di-n-octyl phthalate	22.937	149	628261	38.000	ng/ul	100
90) Benzo(b)fluoranthene	24.065	252	756383	36.903	ng/ul	99
91) Benzo(k)fluoranthene	24.135	252	659421	34.860	ng/ul	98
93) Benzo(a)pyrene	24.964	252	697778	36.059	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	28.917	276	801145	36.607	ng/ul	98
95) Dibenzo(a,h)anthracene	28.988	278	671598	36.105	ng/ul	98
96) Benzo(g,h,i)perylene	30.104	276	673173	36.388	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed




#2
1,4-Dioxane
Concen: 12.182 ng/uL
RT: 3.599 min Scan# 30
Delta R.T. -0.001 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Instrument :

BNA_G

ClientSampleId :

DBSA8MSD



Tgt Ion: 88 Resp: 19254
Ion Ratio Lower Upper
88 100
43 36.1 30.6 46.0
58 87.7 58.6 87.8




Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay

03/21/2022




Supervised By :Yogesh
Patel


03/24/2022



#5
Pyridine
Concen: 35.735 ng/uL
RT: 3.998 min Scan# 104
Delta R.T. -0.007 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34


Tgt Ion: 79 Resp: 124352
Ion Ratio Lower Upper
79 100
52 77.5 57.2 85.8
51 37.6 31.9 47.9





#6
Benzaldehyde
Concen: 37.697 ng/ul m
RT: 7.282 min Scan# 662
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Instrument :
BNA_G
ClientSampleId :
DBSA8MSD





Tgt Ion: 77 Resp: 102278
Ion Ratio Lower Upper
77 100
105 85.8 72.9 109.3
106 82.9 66.5 99.7

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




03/21/2022
Supervised By :Yogesh
Patel


03/24/2022



#8
Phenol
Concen: 39.890 ng/ul
RT: 7.317 min Scan# 669
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Tgt Ion: 94 Resp: 156906
Ion Ratio Lower Upper
94 100
65 33.4 25.9 38.9
66 45.9 35.4 53.2






#10
Bis(2-Chloroethyl)ether
Concen: 37.802 ng/ul
RT: 7.558 min Scan# 710
Delta R.T. -0.007 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Instrument :

BNA_G

ClientSampleId :

DBSA8MSD



Tgt Ion: 93 Resp: 11217
Ion Ratio Lower Upper
93 100
63 83.4 68.2 102.4
95 34.2 24.7 37.1




Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay


03/21/2022

Supervised By :Yogesh
Patel



03/24/2022




#12
2-Chlorophenol
Concen: 37.053 ng/ul
RT: 7.711 min Scan# 736
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34



Tgt Ion:128 Resp: 104909
Ion Ratio Lower Upper
128 100
64 57.6 39.8 59.6
130 30.2 25.8 38.8





#13

2-Methylphenol

Concen: 37.586 ng/ul

RT: 8.580 min Scan# 80

Delta R.T. -0.002 min

Lab File: BG052739.D

Acq: 19 Mar 2022 10:34

Instrument :

BNA_G

ClientSampleId :

DBSA8MSD

Tgt Ion:108 Resp: 11490

Ion Ratio Lower Upper

108 100

107 88.4 68.3 102.5

Manual Integrations





APPROVED

Reviewed By :Jagrut
Upadhyay

03/21/2022

Supervised By :Yogesh
Patel

03/24/2022



#14

2,2'-oxybis(1-Chloropropane)

Concen: 37.406 ng/ul

RT: 8.680 min Scan# 901

Delta R.T. -0.002 min

Lab File: BG052739.D

Acq: 19 Mar 2022 10:34




Tgt Ion: 45 Resp: 180419


Ion Ratio Lower Upper

45 100

77 14.9 11.0 16.4

79 10.8 9.3 13.9






#16
Acetophenone
Concen: 36.715 ng/ul
RT: 8.968 min Scan# 91
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Instrument :

BNA_G

ClientSampleId :

DBSA8MSD



Tgt Ion:105 Resp: 169334
Ion Ratio Lower Upper
105 100
77 86.1 72.9 109.3
51 32.3 25.0 37.6




Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay

03/21/2022




Supervised By :Yogesh
Patel


03/24/2022



#17
N-Nitroso-di-n-propylamine
Concen: 37.539 ng/ul
RT: 8.956 min Scan# 948
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34


Tgt Ion: 70 Resp: 100932
Ion Ratio Lower Upper
70 100
42 63.9 47.8 71.6
101 9.3 6.8 10.2
130 18.0 14.7 22.1





#18
4-Methylphenol
Concen: 38.005 ng/ul
RT: 8.909 min Scan# 940
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Instrument :
BNA_G
ClientSampleId :
DBSA8MSD






Tgt Ion:108 Resp: 121040
Ion Ratio Lower Upper
108 100
107 114.4 82.7 124.1

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




03/21/2022
Supervised By :Yogesh
Patel


03/24/2022



#19
Hexachloroethane
Concen: 37.127 ng/ul
RT: 9.244 min Scan# 997
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Tgt Ion:117 Resp: 46641
Ion Ratio Lower Upper
117 100
201 111.6 83.9 125.9
199 68.0 51.4 77.2






#22
Nitrobenzene
Concen: 37.047 ng/ul
RT: 9.350 min Scan# 1015
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Instrument :

BNA_G

ClientSampleId :

DBSA8MSD



Tgt Ion: 77 Resp: 150230
Ion Ratio Lower Upper
77 100
123 35.2 25.7 38.5
65 13.8 9.7 14.5




Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay

03/21/2022




Supervised By :Yogesh
Patel


03/24/2022



#23
Isophorone
Concen: 36.279 ng/ul
RT: 9.878 min Scan# 1105
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Tgt Ion: 82 Resp: 281198
Ion Ratio Lower Upper
82 100
95 7.8 6.1 9.1
138 15.4 11.0 16.4





#25
2-Nitrophenol
Concen: 39.321 ng/ul
RT: 10.066 min Scan# 1136
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Instrument :

BNA_G

ClientSampleId :

DBSA8MSD

Tgt Ion:139 Resp: 5968
Ion Ratio Lower Upper
139 100
65 61.1 48.9 73.3
109 41.4 29.0 43.6





Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay

03/21/2022




Supervised By :Yogesh
Patel


03/24/2022



#26
2,4-Dimethylphenol
Concen: 33.585 ng/ul
RT: 10.119 min Scan# 1146
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Tgt Ion:107 Resp: 126322
Ion Ratio Lower Upper
107 100
121 43.3 37.4 56.0
122 76.2 63.0 94.4





#27

Bis(2-Chloroethoxy)methane

Concen: 36.077 ng/ul

RT: 10.354 min Scan# 1185

Delta R.T. -0.002 min

Lab File: BG052739.D

Acq: 19 Mar 2022 10:34

Instrument :

BNA_G

ClientSampleId :

DBSA8MSD

Tgt Ion: 93 Resp: 15550

Ion Ratio Lower Upper

93 100

95 30.4 25.4 38.2

123 8.5 8.2 12.2

Manual Integrations

APPROVED

Reviewed By :Jagrut





Upadhyay

03/21/2022

Supervised By :Yogesh

Patel

03/24/2022



#29

2,4-Dichlorophenol

Concen: 35.559 ng/ul

RT: 10.601 min Scan# 1228

Delta R.T. -0.002 min

Lab File: BG052739.D

Acq: 19 Mar 2022 10:34




Tgt Ion:162 Resp: 110097


Ion Ratio Lower Upper

162 100

164 65.5 53.2 79.8


98 35.3 29.1 43.7





#30
Naphthalene
Concen: 34.488 ng/ul
RT: 11.012 min Scan# 1298
Delta R.T. -0.007 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Instrument :
BNA_G
ClientSampleId :
DBSA8MSD






Tgt Ion:128 Resp: 34409
Ion Ratio Lower Upper
128 100
129 10.2 9.0 13.6
127 13.1 9.8 14.6

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




03/21/2022
Supervised By :Yogesh
Patel


03/24/2022



#32
4-Chloroaniline
Concen: 32.769 ng/ul
RT: 11.106 min Scan# 1314
Delta R.T. -0.007 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Tgt Ion:127 Resp: 135524
Ion Ratio Lower Upper
127 100
129 32.4 25.3 37.9






#33
Hexachlorobutadiene
Concen: 35.740 ng/ul
RT: 11.306 min Scan# 11306
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Instrument :

BNA_G

ClientSampleId :

DBSA8MSD



Tgt Ion:225 Resp: 9037
Ion Ratio Lower Upper
225 100
223 61.5 51.7 77.5
227 66.2 49.6 74.4

Manual Integrations




APPROVED

Reviewed By :Jagrut
Upadhyay


03/21/2022

Supervised By :Yogesh
Patel



03/24/2022




#34
Caprolactam
Concen: 38.806 ng/ul m
RT: 11.876 min Scan# 1445
Delta R.T. 0.004 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34



Tgt Ion:113 Resp: 36186
Ion Ratio Lower Upper
113 100
55 188.3 142.0 213.0
56 150.7 96.1 144.1#






#35
4-Chloro-3-methylphenol
Concen: 37.415 ng/ul
RT: 12.216 min Scan# 1502
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Instrument :

BNA_G

ClientSampleId :

DBSA8MSD



Tgt Ion:107 Resp: 12566
Ion Ratio Lower Upper
107 100
144 24.0 21.8 32.6
142 71.1 64.6 97.0




Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay

03/21/2022




Supervised By :Yogesh
Patel


03/24/2022



#36
2-Methylnaphthalene
Concen: 34.878 ng/ul
RT: 12.610 min Scan# 1570
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Tgt Ion:142 Resp: 245578
Ion Ratio Lower Upper
142 100
141 89.2 71.2 106.8





#37
1-Methylnaphthalene
Concen: 35.275 ng/ul
RT: 12.827 min Scan# 1606
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Instrument :

BNA_G

ClientSampleId :

DBSA8MSD

Tgt Ion:142 Resp: 25144
Ion Ratio Lower Upper
142 100
141 89.1 71.4 107.0
116 4.7 3.5 5.3





Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay

03/21/2022




Supervised By :Yogesh
Patel


03/24/2022



#39
1,2,4,5-Tetrachlorobenzene
Concen: 35.400 ng/ul
RT: 12.974 min Scan# 1632
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Tgt Ion:216 Resp: 156252
Ion Ratio Lower Upper
216 100
214 82.1 62.0 93.0
179 20.2 15.7 23.5
108 16.0 12.9 19.3






#40
Hexachlorocyclopentadiene
Concen: 33.635 ng/ul
RT: 12.956 min Scan# 1629
Delta R.T. -0.007 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Instrument :

BNA_G

ClientSampleId :

DBSA8MSD



Tgt Ion:237 Resp: 102780
Ion Ratio Lower Upper
237 100
235 64.3 50.1 75.1
272 12.4 10.2 15.4




Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay

03/21/2022




Supervised By :Yogesh
Patel


03/24/2022



#41
2,4,6-Trichlorophenol
Concen: 37.614 ng/ul
RT: 13.203 min Scan# 1671
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Tgt Ion:196 Resp: 101451
Ion Ratio Lower Upper
196 100
198 98.8 75.4 113.2
200 30.0 24.9 37.3





#42
2,4,5-Trichlorophenol
Concen: 38.781 ng/ul
RT: 13.274 min Scan# 1682
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Instrument :

BNA_G

ClientSampleId :

DBSA8MSD

Tgt Ion:196 Resp: 112950
Ion Ratio Lower Upper
196 100
198 90.7 77.0 115.4
200 31.0 24.2 36.4





Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay

03/21/2022




Supervised By :Yogesh
Patel


03/24/2022



#43
1,1'-Biphenyl
Concen: 35.291 ng/ul
RT: 13.603 min Scan# 1739
Delta R.T. -0.007 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Tgt Ion:154 Resp: 336772
Ion Ratio Lower Upper
154 100
153 41.6 33.0 49.4
76 14.2 10.6 15.8






#44
2-Chloronaphthalene
Concen: 35.323 ng/ul
RT: 13.650 min Scan# 1746
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Instrument :

BNA_G

ClientSampleId :

DBSA8MSD



Tgt Ion:162 Resp: 269909
Ion Ratio Lower Upper
162 100
164 33.0 26.6 40.0
127 37.6 28.2 42.2




Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay


03/21/2022

Supervised By :Yogesh
Patel



03/24/2022




#45
2-Nitroaniline
Concen: 40.952 ng/ul
RT: 13.838 min Scan# 1779
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34



Tgt Ion: 65 Resp: 91727
Ion Ratio Lower Upper
65 100
92 67.7 54.0 81.0
138 92.7 76.0 114.0





#47

Dimethylphthalate

Concen: 36.301 ng/ul

RT: 14.214 min Scan# 1842

Delta R.T. -0.002 min

Lab File: BG052739.D

Acq: 19 Mar 2022 10:34

Instrument :

BNA_G

ClientSampleId :

DBSA8MSD

Tgt Ion:163 Resp: 35539

Ion	Ratio	Lower	Upper
163	100		
194	4.0	2.9	4.3
164	10.5	7.7	11.5





Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay

03/21/2022

Supervised By :Yogesh
Patel

03/24/2022



#48

2,6-Dinitrotoluene

Concen: 40.262 ng/ul

RT: 14.331 min Scan# 1863




Delta R.T. -0.002 min


Lab File: BG052739.D

Acq: 19 Mar 2022 10:34

Tgt Ion:165 Resp: 71405

Ion	Ratio	Lower	Upper
165	100		
89	64.0	53.4	80.2
121	20.4	21.2	31.8






#50
Acenaphthylene
Concen: 35.754 ng/ul
RT: 14.490 min Scan# 1890
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Instrument :

BNA_G

ClientSampleId :

DBSA8MSD



Tgt Ion:152 Resp: 430469
Ion Ratio Lower Upper
152 100
151 21.1 16.4 24.6
153 13.4 11.2 16.8




Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay


03/21/2022

Supervised By :Yogesh
Patel



03/24/2022




#51
3-Nitroaniline
Concen: 37.942 ng/ul
RT: 14.660 min Scan# 1919
Delta R.T. 0.004 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34



Tgt Ion:138 Resp: 67920
Ion Ratio Lower Upper
138 100
108 12.1 7.8 11.8#
92 119.2 103.5 155.3






#52
Acenaphthene
Concen: 36.222 ng/ul
RT: 14.830 min Scan# 1948
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Instrument :

BNA_G

ClientSampleId :

DBSA8MSD



Tgt Ion:153 Resp: 28490
Ion Ratio Lower Upper
153 100
152 46.8 38.0 57.0
154 85.6 70.0 105.0




Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay


03/21/2022

Supervised By :Yogesh
Patel



03/24/2022




#53
2,4-Dinitrophenol
Concen: 40.215 ng/ul
RT: 14.860 min Scan# 1953
Delta R.T. 0.004 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34




Tgt Ion:184 Resp: 44308
Ion Ratio Lower Upper
184 100
63 61.3 49.3 73.9
154 87.9 108.6 162.8#





#55
4-Nitrophenol
Concen: 37.622 ng/ul
RT: 14.954 min Scan# 1967
Delta R.T. 0.004 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Instrument :
BNA_G
ClientSampleId :
DBSA8MSD






Tgt Ion:109 Resp: 68290
Ion Ratio Lower Upper
109 100
139 88.8 75.4 113.2
65 109.4 91.4 137.2

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




03/21/2022
Supervised By :Yogesh
Patel


03/24/2022



#56
Dibenzofuran
Concen: 35.735 ng/ul
RT: 15.165 min Scan# 2005
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Tgt Ion:168 Resp: 416687
Ion Ratio Lower Upper
168 100
139 39.0 31.8 47.6





#57
2,4-Dinitrotoluene
Concen: 40.146 ng/ul
RT: 15.112 min Scan# 1996
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Instrument :

BNA_G

ClientSampleId :

DBSA8MSD

Tgt Ion:165 Resp: 10102
Ion Ratio Lower Upper
165 100
63 52.0 33.0 49.4
182 2.2 3.4 5.0





Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay

03/21/2022




Supervised By :Yogesh
Patel


03/24/2022



#58
2,3,4,6-Tetrachlorophenol
Concen: 39.236 ng/ul
RT: 15.388 min Scan# 2043
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Tgt Ion:232 Resp: 103013
Ion Ratio Lower Upper
232 100
131 37.2 29.8 44.8
130 2.5 1.0 1.4
166 27.5 22.0 33.0





#59
Diethylphthalate
Concen: 37.333 ng/u1
RT: 15.576 min Scan# 2074
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Instrument :

BNA_G

ClientSampleId :

DBSA8MSD

Tgt Ion:149 Resp: 370160
Ion Ratio Lower Upper
149 100
177 21.1 17.5 26.3
150 12.7 9.5 14.3





Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay

03/21/2022




Supervised By :Yogesh
Patel


03/24/2022



#61
Fluorene
Concen: 35.312 ng/u1
RT: 15.811 min Scan# 2115
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34


Tgt Ion:166 Resp: 341307
Ion Ratio Lower Upper
166 100
165 96.2 75.7 113.5
167 14.1 10.3 15.5





#62
4-Chlorophenyl-phenylether
Concen: 35.968 ng/ul
RT: 15.806 min Scan# 2113
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Instrument :
BNA_G
ClientSampleId :
DBSA8MSD






Tgt Ion:204 Resp: 184989
Ion Ratio Lower Upper
204 100
206 32.2 27.4 41.0
141 52.2 44.2 66.2

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




03/21/2022
Supervised By :Yogesh
Patel


03/24/2022



#63
4-Nitroaniline
Concen: 39.822 ng/ul
RT: 15.817 min Scan# 2116
Delta R.T. 0.004 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Tgt Ion:138 Resp: 70103
Ion Ratio Lower Upper
138 100
92 57.1 43.0 64.6
108 110.5 84.6 126.8





#66
4,6-Dinitro-2-methylphenol
Concen: 39.204 ng/ul
RT: 15.876 min Scan# 2125
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Instrument :

BNA_G

ClientSampleId :

DBSA8MSD

Tgt Ion:198 Resp: 6153
Ion Ratio Lower Upper
198 100
182 3.7 3.0 4.6
77 28.8 19.7 29.5





Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay

03/21/2022




Supervised By :Yogesh
Patel


03/24/2022



#67
N-Nitrosodiphenylamine
Concen: 35.310 ng/ul
RT: 16.011 min Scan# 2149
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Tgt Ion:169 Resp: 300943
Ion Ratio Lower Upper
169 100
168 68.5 57.4 86.0
167 38.0 28.1 42.1






#68
4-Bromophenyl-phenylether
Concen: 39.587 ng/ul
RT: 16.698 min Scan# 21
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Instrument :

BNA_G

ClientSampleId :

DBSA8MSD



Tgt Ion:248 Resp: 12355
Ion Ratio Lower Upper
248 100
250 97.3 76.3 114.5
141 67.9 58.7 88.1




Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay

03/21/2022




Supervised By :Yogesh
Patel


03/24/2022



#69
Hexachlorobenzene
Concen: 36.505 ng/ul
RT: 16.822 min Scan# 2287
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Tgt Ion:284 Resp: 147177
Ion Ratio Lower Upper
284 100
142 27.3 22.3 33.5
249 31.6 21.9 32.9





#70
Atrazine
Concen: 37.248 ng/ul
RT: 16.957 min Scan# 2100
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Instrument :

BNA_G

ClientSampleId :

DBSA8MSD

Tgt Ion:200 Resp: 13010
Ion Ratio Lower Upper
200 100
173 24.3 19.8 29.6
215 51.1 39.5 59.3





Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay

03/21/2022




Supervised By :Yogesh
Patel


03/24/2022



#71
Pentachlorophenol
Concen: 38.058 ng/ul
RT: 17.157 min Scan# 2344
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Tgt Ion:266 Resp: 93307
Ion Ratio Lower Upper
266 100
264 67.3 50.3 75.5
268 61.5 45.8 68.8





#72
Phenanthrene
Concen: 35.304 ng/u1
RT: 17.556 min Scan# 2411
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Instrument :

BNA_G

ClientSampleId :

DBSA8MSD

Tgt Ion:178 Resp: 584068
Ion Ratio Lower Upper
178 100
179 15.0 11.9 17.9
176 19.0 15.0 22.6





Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay

03/21/2022




Supervised By :Yogesh
Patel


03/24/2022



#74
Anthracene
Concen: 34.684 ng/u1
RT: 17.644 min Scan# 2427
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34


Tgt Ion:178 Resp: 573669
Ion Ratio Lower Upper
178 100
179 15.4 12.8 19.2
176 17.8 14.7 22.1





#75
1,2,3,4-Tetrachlorobenzene
Concen: 37.156 ng/uL
RT: 13.573 min Scan# 1734
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Instrument :
BNA_G
ClientSampleId :
DBSA8MSD






Tgt Ion:216 Resp: 17444
Ion Ratio Lower Upper
216 100
214 77.0 63.8 95.8
218 45.7 38.3 57.5

Manual Integrations
APPROVED


Reviewed By :Jagrut
Upadhyay

03/21/2022
Supervised By :Yogesh
Patel



03/24/2022




#76
Pentachlorobenzene
Concen: 38.281 ng/uL
RT: 15.089 min Scan# 1992
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34



Tgt Ion:250 Resp: 172425
Ion Ratio Lower Upper
250 100
248 62.5 50.4 75.6
252 65.1 49.2 73.8






#77
 Carbazole
 Concen: 36.097 ng/u1
 RT: 17.909 min Scan# 2471
 Delta R.T. -0.002 min
 Lab File: BG052739.D
 Acq: 19 Mar 2022 10:34

Instrument :

BNA_G

ClientSampleId :

DBSA8MSD



Tgt Ion:167 Resp: 527648
 Ion Ratio Lower Upper
 167 100
 166 22.4 17.3 25.9
 139 12.4 9.4 14.0




Manual Integrations
 APPROVED

Reviewed By :Jagrut
 Upadhyay


03/21/2022

Supervised By :Yogesh
 Patel



03/24/2022




#78
 Di-n-butylphthalate
 Concen: 36.289 ng/u1
 RT: 18.467 min Scan# 2567
 Delta R.T. -0.002 min
 Lab File: BG052739.D
 Acq: 19 Mar 2022 10:34



Tgt Ion:149 Resp: 630785
 Ion Ratio Lower Upper
 149 100
 150 9.4 7.5 11.3
 104 6.2 4.6 6.8






#80
Fluoranthene
Concen: 37.350 ng/u1
RT: 19.559 min Scan# 21
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Instrument :

BNA_G

ClientSampleId :

DBSA8MSD



Tgt Ion:202 Resp: 748570
Ion Ratio Lower Upper
202 100
101 7.4 6.1 9.1
100 6.5 5.4 8.0



Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay

03/21/2022




Supervised By :Yogesh
Patel


03/24/2022



#82
Pyrene
Concen: 36.536 ng/u1
RT: 19.918 min Scan# 2814
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Tgt Ion:202 Resp: 727812
Ion Ratio Lower Upper
202 100
101 9.1 8.2 12.4
100 8.2 6.8 10.2





#83
Butylbenzylphthalate
Concen: 39.502 ng/ul
RT: 20.799 min Scan# 2964
Delta R.T. -0.007 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Instrument :

BNA_G

ClientSampleId :

DBSA8MSD

Tgt Ion:149 Resp: 27358
Ion Ratio Lower Upper
149 100
91 80.0 57.1 85.7
206 22.7 17.4 26.2





Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay

03/21/2022




Supervised By :Yogesh
Patel


03/24/2022



#84
3,3'-Dichlorobenzidine
Concen: 33.603 ng/ul
RT: 21.686 min Scan# 3115
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34


Tgt Ion:252 Resp: 228968
Ion Ratio Lower Upper
252 100
254 66.3 48.6 73.0
126 7.0 6.2 9.2





#85
Benzo(a)anthracene
Concen: 36.024 ng/ul
RT: 21.780 min Scan# 3130
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Instrument :
BNA_G
ClientSampleId :
DBSA8MSD





Tgt Ion:228 Resp: 706834
Ion Ratio Lower Upper
228 100
229 20.2 15.4 23.2
226 26.9 21.4 32.2

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




03/21/2022
Supervised By :Yogesh
Patel


03/24/2022



#86
Bis(2-ethylhexyl)phthalate
Concen: 38.726 ng/ul
RT: 21.686 min Scan# 3115
Delta R.T. -0.007 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34


Tgt Ion:149 Resp: 371496
Ion Ratio Lower Upper
149 100
167 25.7 21.4 32.2
279 4.8 3.4 5.2





#87
Chrysene
Concen: 35.543 ng/u1
RT: 21.850 min Scan# 3142
Delta R.T. 0.004 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Instrument :
BNA_G
ClientSampleId :
DBSA8MSD






Tgt Ion:228 Resp: 66520
Ion Ratio Lower Upper
228 100
226 30.7 23.9 35.9
229 21.0 16.0 24.0

Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




03/21/2022
Supervised By :Yogesh
Patel


03/24/2022



#89
Di-n-octyl phthalate
Concen: 38.000 ng/u1
RT: 22.937 min Scan# 3328
Delta R.T. -0.007 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Tgt Ion:149 Resp: 628261





#90
Benzo(b)fluoranthene
Concen: 36.903 ng/ul
RT: 24.065 min Scan# 3519
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Instrument :

BNA_G

ClientSampleId :

DBSA8MSD

Tgt Ion:252 Resp: 75638
Ion Ratio Lower Upper
252 100
253 21.6 17.8 26.6
125 6.2 5.5 8.3





Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay

03/21/2022




Supervised By :Yogesh
Patel


03/24/2022



#91
Benzo(k)fluoranthene
Concen: 34.860 ng/ul
RT: 24.135 min Scan# 3532
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Tgt Ion:252 Resp: 659421
Ion Ratio Lower Upper
252 100
253 21.3 17.4 26.0
125 6.1 6.0 9.0






#93
Benzo(a)pyrene
Concen: 36.059 ng/u1
RT: 24.964 min Scan# 30
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Instrument :

BNA_G

ClientSampleId :

DBSA8MSD



Tgt Ion:252 Resp: 697778
Ion Ratio Lower Upper
252 100
253 21.9 16.5 24.7
125 7.1 6.0 9.0




Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay


03/21/2022

Supervised By :Yogesh
Patel



03/24/2022




#94
Indeno(1,2,3-cd)pyrene
Concen: 36.607 ng/u1
RT: 28.917 min Scan# 4346
Delta R.T. 0.004 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34



Tgt Ion:276 Resp: 801145
Ion Ratio Lower Upper
276 100
138 12.3 9.3 13.9
277 23.0 19.4 29.2






#95
Dibenzo(a,h)anthracene
Concen: 36.105 ng/ul
RT: 28.988 min Scan# 4358
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Instrument :

BNA_G

ClientSampleId :

DBSA8MSD



Tgt Ion:278 Resp: 671598
Ion Ratio Lower Upper
278 100
139 9.0 8.1 12.1
279 24.6 19.0 28.4




Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay

03/21/2022

Supervised By :Yogesh
Patel

03/24/2022



#96
Benzo(g,h,i)perylene
Concen: 36.388 ng/ul
RT: 30.104 min Scan# 4548
Delta R.T. -0.002 min
Lab File: BG052739.D
Acq: 19 Mar 2022 10:34

Tgt Ion:276 Resp: 673173
Ion Ratio Lower Upper
276 100
138 14.6 10.5 15.7
277 24.3 19.0 28.6

