

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG031822\
 Data File : BG052743.D
 Acq On : 19 Mar 2022 13:09
 Operator : CG/JU
 Sample : N1883-17
 Misc :
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DBSB4

Quant Time: Mar 21 00:49:38 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG031822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 18 00:55:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.149	152	38631	20.000	ng/ul	0.00
20) Naphthalene-d8	10.963	136	154444	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.769	164	109910	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.507	188	257911	20.000	ng/ul	0.00
79) Chrysene-d12	21.795	240	284617	20.000	ng/ul	0.00
88) Perylene-d12	25.114	264	280209	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.561	96	4596	4.202	ng/uL	0.00
4) Pyridine-d5	3.978	84	58461	21.412	ng/ul	0.00
7) Phenol-d5	7.291	99	84597	24.450	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.462	67	53152	25.152	ng/ul	0.00
11) 2-Chlorophenol-d4	7.673	132	56783	24.169	ng/ul	0.00
15) 4-Methylphenol-d8	8.836	113	63923	25.504	ng/ul	0.00
21) Nitrobenzene-d5	9.306	128	28794	25.900	ng/ul	0.00
24) 2-Nitrophenol-d4	10.029	143	31966	27.260	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.569	165	61444	24.086	ng/ul	0.00
31) 4-Chloroaniline-d4	11.086	131	83909	24.887	ng/ul	0.00
46) Dimethylphthalate-d6	14.164	166	223431	26.461	ng/ul	0.00
49) Acenaphthylene-d8	14.458	160	268678	26.186	ng/ul	0.00
54) 4-Nitrophenol-d4	14.928	143	31447	22.151	ng/ul	0.00
60) Fluorene-d10	15.756	176	196823	26.547	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.856	200	27524	20.971	ng/ul	0.00
73) Anthracene-d10	17.607	188	326022	27.146	ng/ul	0.00
81) Pyrene-d10	19.892	212	413772	25.987	ng/ul	0.00
92) Benzo(a)pyrene-d12	24.891	264	399070	27.563	ng/ul	0.00
Target Compounds						
					Qvalue	
8) Phenol	7.315	94	4520	1.355	ng/ul#	83


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG031822\
Data File : BG052743.D
Acq On : 19 Mar 2022 13:09
Operator : CG/JU
Sample : N1883-17
Misc :
ALS Vial : 44 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DBSB4

Quant Time: Mar 21 00:49:38 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG031822.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Mar 18 00:55:42 2022
Response via : Initial Calibration





#8
Phenol
Concen: 1.355 ng/ul
RT: 7.315 min Scan# 60
Delta R.T. -0.004 min
Lab File: BG052743.D
Acq: 19 Mar 2022 13:09

Instrument :
BNA_G
ClientSampleId :
DBSB4

Tgt Ion: 94	Resp:	4520
Ion	Ratio	Lower Upper
94	100	
65	40.4	25.9 38.9#
66	56.3	35.4 53.2#

