

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG031920\
 Data File : BG044903.D
 Acq On : 19 Mar 2020 18:56
 Operator : CG/JU
 Sample : PB127682TB
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PB127682TB

Quant Time: Mar 20 01:56:39 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG031020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 16 17:37:15 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.04	152	76082	20.00	ng	0.00
21) Naphthalene-d8	10.85	136	290237	20.00	ng	0.00
39) Acenaphthene-d10	14.67	164	194824	20.00	ng	-0.01
64) Phenanthrene-d10	17.42	188	447482	20.00	ng	0.00
76) Chrysene-d12	21.69	240	475293	20.00	ng	-0.01
87) Perylene-d12	24.89	264	535016	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.62	112	564255	115.45	ng	0.00
7) Phenol-d6	7.20	99	772240	108.75	ng	0.00
23) Nitrobenzene-d5	9.19	82	499693	75.55	ng	0.00
42) 2,4,6-Tribromophenol	16.16	330	318787	124.97	ng	-0.01
45) 2-Fluorobiphenyl	13.30	172	1073434	78.90	ng	0.00
79) Terphenyl-d14	20.03	244	1940833	76.38	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG031920\
Data File : BG044903.D
Acq On : 19 Mar 2020 18:56
Operator : CG/JU
Sample : PB127682TB
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB127682TB

Quant Time: Mar 20 01:56:39 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG031020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 16 17:37:15 2020
Response via : Initial Calibration

