


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG032015\
Data File : BG016078.D
Acq On : 20 Mar 2015 23:13
Operator : TP/IZ
Sample : DFTPP11
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP11

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM01.2-EPA-BG031715.M
Title : SVOA CALIBRATION
Last Update : Sat Mar 21 04:57:09 2015



AutoFind: Scans 2537, 2538, 2539; Background Corrected with Scan 2532

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.2	15362	PASS
68	69	0.00	2	0.5	90	PASS
69	198	0.00	100	36.8	17030	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	48.3	22341	PASS
197	198	0.00	2	0.4	184	PASS
198	198	100	100	100.0	46216	PASS
199	198	5	9	7.0	3237	PASS
275	198	10	60	25.4	11750	PASS
365	198	1	100	3.3	1528	PASS
441	443	0.01	100	80.9	7106	PASS
442	198	50	100	98.0	45306	PASS
443	442	15	24	19.4	8785	PASS

m/z	Abundance
39.15	998.0
40.05	716.0
44.25	328.0
49.15	156.0
50.15	3464.0
51.15	13926.0
52.15	646.0
56.15	523.0
57.15	1097.0
61.25	169.0
62.05	182.0
63.15	445.0
65.15	419.0
68.15	270.0
69.15	16109.0
74.15	1500.0
75.15	2333.0
76.25	904.0
77.15	18248.0
78.15	1343.0
79.15	850.0
80.05	885.0
81.15	1023.0
82.05	348.0
83.15	396.0
86.05	262.0
91.05	210.0
92.15	381.0
93.15	1681.0
94.05	196.0
98.05	999.0
99.05	1059.0
101.05	561.0
103.05	238.0
104.05	438.0
105.05	517.0
107.15	5295.0
108.05	714.0
110.05	9749.0
111.05	1183.0
112.05	178.0
116.05	226.0
117.05	3391.0
118.05	215.0
122.15	308.0
123.05	463.0
124.05	235.0
125.05	284.0
127.15	19032.0
128.05	1666.0
129.05	7222.0
130.05	743.0
131.05	165.0
134.05	176.0
135.05	690.0
136.05	385.0
137.05	425.0
141.05	899.0
142.05	302.0
143.15	162.0
147.05	653.0
148.05	979.0
148.95	159.0
154.15	300.0
155.05	402.0
156.05	1005.0
157.25	154.0
160.15	348.0
161.05	436.0
165.05	267.0
166.15	282.0
167.05	1627.0
168.05	961.0
169.05	234.0
170.15	171.0
172.05	241.0
173.05	172.0
174.15	262.0
175.05	796.0
176.15	232.0
177.05	309.0
178.05	172.0
179.05	1375.0
180.15	910.0
181.05	396.0
185.05	620.0
186.15	4836.0
187.05	1417.0
189.05	216.0
191.05	166.0
192.05	311.0

Instrument :
BNA_G
ClientSampleId :
DFTPP11

193.05	572.0
196.05	1337.0
196.75	271.0
198.05	35952.0
199.05	2330.0
200.05	193.0
201.45	232.0
203.05	422.0
204.05	1265.0
205.05	1968.0
206.05	7884.0
207.05	1655.0
208.05	440.0
210.15	161.0
211.25	386.0
216.15	212.0
216.95	2451.0
218.05	356.0
221.05	1721.0
223.05	381.0
224.05	4733.0
225.15	1137.0
227.05	1944.0
228.05	323.0
229.05	428.0
235.05	217.0
237.15	164.0
242.05	201.0
243.05	367.0
244.15	3725.0
245.05	595.0
246.05	580.0
247.05	171.0
255.05	18352.0
256.05	2596.0
258.05	867.0
259.05	164.0
265.05	540.0
272.95	516.0
274.05	1660.0
275.05	8653.0
276.05	1083.0
277.05	728.0
296.05	2339.0
297.05	435.0
303.05	175.0
314.05	188.0
315.05	193.0
316.05	181.0
323.05	644.0
324.15	173.0
327.05	204.0
334.05	472.0
346.05	157.0
352.15	249.0
353.05	161.0
354.15	174.0
365.05	885.0
372.15	370.0
402.05	199.0
403.15	186.0
421.05	232.0
422.15	227.0
423.15	1493.0
424.05	246.0
441.15	4214.0
442.15	26416.0
443.15	5076.0
444.15	612.0

Instrument :
BNA_G
ClientSampleId :
DFTPP11

m/z	Abundance
38.15	246.0
39.15	1274.0
40.15	915.0
44.15	304.0
49.15	201.0
50.15	4633.0
51.15	18032.0
52.15	881.0
55.15	212.0
56.15	595.0
57.15	909.0
61.05	236.0
62.05	264.0
63.05	803.0
65.15	447.0
69.05	20392.0
73.15	237.0
74.15	2115.0
75.15	3356.0
76.15	1001.0
77.15	25176.0
78.15	1724.0
79.15	1412.0
80.05	1263.0
81.15	1515.0
82.05	347.0
83.05	397.0
84.05	183.0
85.05	312.0
86.15	650.0
91.15	452.0
92.05	350.0
93.05	2326.0
96.25	199.0
98.05	1674.0
99.05	1247.0
101.05	745.0
103.15	382.0
104.15	508.0
105.15	628.0
106.15	257.0
107.15	6557.0
108.15	1160.0
109.05	209.0
110.05	14733.0
111.15	1944.0
112.05	253.0
116.15	550.0
117.05	4228.0
118.05	364.0
122.05	454.0
123.15	609.0
124.05	378.0
125.05	347.0
127.15	27064.0
128.05	2231.0
129.05	9505.0
130.05	919.0
131.05	189.0
134.15	261.0
135.05	828.0
136.05	288.0
137.05	492.0
141.05	1413.0
142.05	459.0
142.95	345.0
146.05	301.0
147.15	729.0
148.05	1003.0
149.15	290.0
150.85	175.0
152.95	388.0
154.05	373.0
155.15	669.0
156.15	1105.0
157.05	282.0
158.05	240.0
159.15	170.0
159.95	258.0
161.05	586.0
162.15	213.0
165.05	468.0
166.05	331.0
167.05	2658.0
168.05	1455.0
169.05	263.0
171.15	176.0
171.85	219.0
173.05	324.0
174.05	298.0
175.05	849.0

Instrument :
BNA_G
ClientSampleId :
DFTPP11

176.15	309.0
177.05	573.0
178.05	246.0
179.05	1829.0
180.05	1142.0
181.05	587.0
184.05	158.0
185.15	1240.0
186.05	6543.0
187.05	1922.0
188.15	230.0
189.05	481.0
191.05	271.0
192.05	669.0
193.05	678.0
196.05	1555.0
198.05	55208.0
199.05	3681.0
200.05	293.0
201.65	213.0
203.05	398.0
204.05	2115.0
205.05	2945.0
206.15	12508.0
207.15	1775.0
208.15	445.0
210.35	203.0
210.95	513.0
217.05	2841.0
218.05	307.0
221.05	2688.0
223.05	907.0
224.15	7618.0
225.05	1650.0
225.95	261.0
227.05	2799.0
228.05	330.0
228.95	600.0
230.15	153.0
231.05	305.0
233.95	294.0
235.05	153.0
237.05	158.0
239.15	162.0
241.05	303.0
242.05	496.0
243.05	421.0
244.15	5828.0
245.05	932.0
246.05	951.0
246.95	236.0
248.95	228.0
253.15	155.0
255.05	27744.0
256.05	4282.0
257.05	290.0
258.05	1703.0
259.05	276.0
265.05	763.0
273.05	917.0
274.15	2187.0
275.05	14287.0
276.05	1931.0
277.05	916.0
278.05	214.0
293.05	280.0
296.05	3632.0
297.05	650.0
302.95	361.0
313.95	207.0
315.05	314.0
316.05	209.0
323.15	1094.0
324.15	316.0
327.15	376.0
328.05	175.0
334.05	786.0
335.05	225.0
346.05	246.0
352.05	437.0
353.05	314.0
354.15	323.0
365.05	1907.0
365.95	230.0
372.05	857.0
373.05	224.0
382.95	271.0
390.15	190.0
402.05	333.0
403.15	392.0
404.05	215.0
421.15	421.0
422.05	263.0
423.05	2406.0
424.05	779.0

Instrument :
BNA_G
ClientSampleId :
DFTPP11

441.15	8057.0
442.15	50856.0
443.15	10101.0
444.05	989.0

Instrument :
BNA_G
ClientSampleId :
DFTPP11

m/z	Abundance
38.15	177.0
39.15	1071.0
40.15	682.0
44.05	245.0
49.05	321.0
50.15	3263.0
51.15	14130.0
52.15	732.0
56.15	295.0
57.15	911.0
61.15	178.0
62.05	242.0
63.15	534.0
65.15	255.0
69.05	14591.0
73.05	168.0
74.15	1492.0
75.15	2488.0
76.15	918.0
77.15	19040.0
78.15	1415.0
79.15	1126.0
80.05	878.0
81.05	992.0
82.05	244.0
83.15	351.0
84.05	213.0
85.15	223.0
86.05	324.0
87.05	215.0
91.05	381.0
92.05	181.0
93.15	1801.0
96.25	165.0
98.05	1317.0
99.15	936.0
101.05	652.0
103.05	181.0
104.05	411.0
105.05	444.0
107.05	5866.0
108.15	946.0
110.15	11304.0
111.05	1631.0
111.95	246.0
116.15	284.0
117.05	3697.0
118.05	259.0
122.05	366.0
123.05	588.0
124.05	337.0
125.05	293.0
127.15	20928.0
128.15	1683.0
129.05	8036.0
130.05	602.0
131.05	185.0
135.05	554.0
136.05	348.0
137.15	322.0
138.05	187.0
141.05	923.0
142.05	312.0
143.05	230.0
146.05	175.0
147.05	623.0
148.15	1123.0
149.05	236.0
151.05	163.0
153.05	399.0
154.05	217.0
155.05	743.0
156.05	803.0
157.15	210.0
158.05	168.0
159.05	163.0
160.05	379.0
161.05	498.0
162.05	155.0
165.05	506.0
166.05	307.0
167.05	2598.0
168.05	1000.0
169.15	220.0
171.95	167.0
173.15	324.0
174.15	422.0
175.05	938.0
177.05	322.0
179.05	1528.0
180.05	1211.0

Instrument :
BNA_G
ClientSampleId :
DFTPP11

181.05	482.0
185.15	1002.0
186.15	5992.0
187.15	1691.0
189.05	436.0
191.05	366.0
192.05	613.0
193.15	511.0
196.05	1659.0
196.85	281.0
198.05	47488.0
199.05	3700.0
200.05	323.0
201.45	193.0
203.05	344.0
204.05	1867.0
205.05	3176.0
206.05	11346.0
207.05	1599.0
208.05	464.0
210.15	163.0
211.05	533.0
214.85	167.0
216.05	392.0
217.05	2919.0
218.05	400.0
221.05	2684.0
223.05	775.0
224.05	6629.0
225.05	1744.0
226.05	294.0
227.05	2506.0
228.05	500.0
229.05	493.0
231.05	276.0
235.05	186.0
236.15	155.0
237.05	222.0
240.05	175.0
241.05	165.0
242.05	271.0
243.05	432.0
244.05	5072.0
245.05	710.0
246.05	971.0
247.05	238.0
248.95	235.0
252.95	152.0
255.05	27680.0
256.05	3892.0
256.95	257.0
258.05	1339.0
259.05	159.0
265.05	646.0
272.15	219.0
273.05	848.0
274.05	2024.0
275.05	12312.0
276.05	1590.0
277.05	1075.0
277.95	162.0
283.15	203.0
285.05	195.0
293.05	200.0
296.05	3322.0
297.05	449.0
303.05	531.0
314.15	222.0
315.05	291.0
316.05	234.0
323.15	1314.0
324.05	154.0
327.15	153.0
334.05	650.0
335.05	230.0
341.05	169.0
346.05	345.0
352.05	313.0
353.05	244.0
354.15	457.0
365.05	1792.0
366.05	324.0
372.05	705.0
382.95	214.0
401.95	283.0
403.05	517.0
404.05	196.0
421.05	379.0
422.15	509.0
423.15	2727.0
424.15	520.0
441.05	9048.0
442.15	58648.0
443.15	11179.0
444.05	937.0

Instrument :
BNA_G
ClientSampleId :
DFTPP11

Scan 2532 (17.562 min): BG016078.D
DFTPP11

m/z	Abundance
40.15	730.0
44.15	295.0
49.15	150.0
84.05	161.0
207.05	207.0

Instrument :
BNA_G
ClientSampleId :
DFTPP11