

Data Path : Z:\HPCHEM1\BNA G\DATA\BG032017\
 Data File : BG026334.D
 Acq On : 20 Mar 2017 19:45
 Operator : SJ/MA
 Sample : I2194-03DL 5X
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PT-BN-WPDL

Manual Integrations
 APPROVED

mohammad
 3/21/2017 5:22:04 PM

Quant Time: Mar 21 04:55:39 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG032017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 20 17:43:27 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.05	152	124181	20.00	ng	0.00
21) Naphthalene-d8	10.84	136	533664	20.00	ng	0.00
38) Acenaphthene-d10	14.65	164	310033	20.00	ng	0.00
63) Phenanthrene-d10	17.38	188	747704	20.00	ng	0.00
75) Chrysene-d12	21.63	240	888121	20.00	ng	0.00
86) Perylene-d12	24.82	264	847277	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.62	112	213616	30.53	ng	0.00
7) Phenol-d6	7.21	99	291878	31.07	ng	0.00
23) Nitrobenzene-d5	9.21	82	157161	17.71	ng	0.00
41) 2,4,6-Tribromophenol	16.13	330	115621	32.57	ng	0.00
44) 2-Fluorobiphenyl	13.28	172	460020	20.81	ng	0.00
78) Terphenyl-d14	19.98	244	736991	20.15	ng	0.00
Target Compounds						Qvalue
12) 1,3-Dichlorobenzene	7.94	146	299699	31.96	ng	# 85
13) 1,4-Dichlorobenzene	8.08	146	330209m	34.81	ng	
14) 1,2-Dichlorobenzene	8.40	146	96180	10.59	ng	# 86
24) Nitrobenzene	9.25	77	81423	9.29	ng	# 74
25) Isophorone	9.76	82	341507	20.42	ng	# 91
30) 1,2,4-Trichlorobenzene	10.71	180	297297	34.96	ng	98
31) Naphthalene	10.90	128	444030	16.46	ng	99
34) Hexachlorobutadiene	11.19	225	186185	37.11	ng	99
37) 2-Methylnaphthalene	12.50	142	303305	15.35	ng	# 90
48) Acenaphthylene	14.38	152	964259	29.50	ng	98
50) 2,6-Dinitrotoluene	14.22	165	110535	20.24	ng	# 64
51) Acenaphthene	14.72	154	228548	11.35	ng	98
54) Dibenzofuran	15.05	168	812266	28.66	ng	# 88
57) Fluorene	15.70	166	275431	10.92	ng	97
59) Diethylphthalate	15.46	149	868612	36.32	ng	99
60) 4-Chlorophenyl-phenylether	15.69	204	370845	31.03	ng	90
67) Hexachlorobenzene	16.70	284	161426	19.66	ng	92
73) Di-n-butylphthalate	18.34	149	1352183	31.20	ng	# 93
74) Fluoranthene	19.43	202	1835760	38.88	ng	95
77) Pvrene	19.79	202	1936744	37.66	ng	99
79) Butylbenzylphthalate	20.66	149	534563	25.98	ng	86
82) Chrysene	21.68	228	465416	9.75	ng	95
83) Bis(2-ethylhexyl)phthalate	21.51	149	1098191	35.34	ng	99
84) Di-n-octyl phthalate	22.71	149	1862852	35.12	ng	# 82
87) Benzo(b)fluoranthene	23.81	252	1479719	29.63	ng	# 95
88) Benzo(k)fluoranthene	23.87	252	929785	19.20	ng	# 93
89) Benzo(a)pyrene	24.66	252	519890	10.85	ng	# 94
91) Benzo(g,h,i)perylene	29.58	276	1855430	38.03	ng	# 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG032017\
Data File : BG026334.D
Acq On : 20 Mar 2017 19:45
Operator : SJ/MA
Sample : I2194-03DL 5X
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PT-BN-WPDL

Manual Integrations
APPROVED

mohammad
3/21/2017 5:22:04 PM

Quant Time: Mar 21 04:55:39 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG032017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 20 17:43:27 2017
Response via : Initial Calibration

