




Data Path : Z:\HPCHEM1\BNA G\DATA\BG032017\
Data File : BG026336.D
Acq On : 20 Mar 2017 21:04
Operator : SJ/MA
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 21 07:22:11 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG032017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 20 17:43:27 2017
Response via : Initial Calibration



TIC: BG026336.D

(69) Pentachlorophenol (C)




17.036min (-0.006) 625499.82ng

response 272375

Ion	Exp%	Act%
265.70	100	100
268.00	60.70	65.41
264.00	62.60	64.90
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG032017\
Data File : BG026336.D
Acq On : 20 Mar 2017 21:04
Operator : SJ/MA
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 21 07:22:11 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG032017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 20 17:43:27 2017
Response via : Initial Calibration



TIC: BG026336.D

(76) Benzidine

19.604min (-0.000) 116975.47ng

response 1633416

Ion	Exp%	Act%
184.00	100	100
185.00	16.30	17.26
183.00	13.80	13.67
0.00	0.00	0.00