

Data Path : Z:\HPCHEM1\BNA G\DATA\BG032017\
 Data File : BG026330.D
 Acq On : 20 Mar 2017 17:08
 Operator : SJ/MA
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG032017

Quant Time: Mar 20 17:43:42 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG032017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 20 17:43:27 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	94	0.00
2	1,4-Dioxane	40.000	40.726	-1.8	98	0.00
3	Pyridine	40.000	41.725	-4.3	96	0.00
4	n-Nitrosodimethylamine	40.000	40.728	-1.8	95	0.00
5 S	2-Fluorophenol	80.000	81.954	-2.4	96	0.00
6	Aniline	40.000	40.845	-2.1	94	0.00
7 S	Phenol-d6	80.000	82.736	-3.4	96	0.00
8	2-Chlorophenol	40.000	40.874	-2.2	95	0.00
9	Benzaldehyde	40.000	41.511	-3.8	96	0.00
10 C	Phenol	40.000	40.902	-2.3	95	0.00
11	bis(2-Chloroethyl)ether	40.000	41.261	-3.2	98	0.00
12	1,3-Dichlorobenzene	40.000	40.921	-2.3	96	0.00
13 C	1,4-Dichlorobenzene	40.000	41.616	-4.0	97	0.00
14	1,2-Dichlorobenzene	40.000	41.388	-3.5	97	0.00
15	Benzyl Alcohol	40.000	41.590	-4.0	96	0.00
16	2,2'-oxybis(1-Chloropropane)	40.000	39.370	1.6	92	0.00
17	2-Methylphenol	40.000	41.444	-3.6	98	0.00
18	Hexachloroethane	40.000	41.795	-4.5	97	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.558	-1.4	94	0.00
20	3+4-Methylphenols	40.000	41.958	-4.9	95	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	97	0.00
22	Acetophenone	40.000	40.677	-1.7	95	0.00
23 S	Nitrobenzene-d5	80.000	81.738	-2.2	96	0.00
24	Nitrobenzene	40.000	40.417	-1.0	95	0.00
25	Isophorone	40.000	40.632	-1.6	96	0.00
26 C	2-Nitrophenol	40.000	41.253	-3.1	94	0.00
27	2,4-Dimethylphenol	40.000	41.281	-3.2	96	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.486	-1.2	96	0.00
29 C	2,4-Dichlorophenol	40.000	41.492	-3.7	96	0.00
30	1,2,4-Trichlorobenzene	40.000	40.393	-1.0	95	0.00
31	Naphthalene	40.000	40.595	-1.5	97	0.00
32	Benzoic acid	40.000	40.055	-0.1	94	0.00
33	4-Chloroaniline	40.000	40.570	-1.4	95	0.00
34 C	Hexachlorobutadiene	40.000	40.395	-1.0	96	0.00
35	Caprolactam	40.000	41.161	-2.9	97	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.905	-2.3	96	0.00
37	2-Methylnaphthalene	40.000	40.485	-1.2	96	0.00
38 I	Acenaphthene-d10	20.000	20.000	0.0	97	0.00
39	1,2,4,5-Tetrachlorobenzene	40.000	40.609	-1.5	98	0.00
40 P	Hexachlorocyclopentadiene	40.000	40.788	-2.0	97	0.00
41 S	2,4,6-Tribromophenol	80.000	83.313	-4.1	100	0.00
42 C	2,4,6-Trichlorophenol	40.000	40.775	-1.9	97	0.00
43	2,4,5-Trichlorophenol	40.000	40.750	-1.9	97	0.00
44 S	2-Fluorobiphenyl	80.000	80.290	-0.4	97	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG032017\
 Data File : BG026330.D
 Acq On : 20 Mar 2017 17:08
 Operator : SJ/MA
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG032017

Quant Time: Mar 20 17:43:42 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG032017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 20 17:43:27 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	40.000	40.346	-0.9	97	0.00
46	2-Chloronaphthalene	40.000	39.976	0.1	97	0.00
47	2-Nitroaniline	40.000	41.548	-3.9	96	0.00
48	Acenaphthylene	40.000	40.319	-0.8	97	0.00
49	Dimethylphthalate	40.000	41.267	-3.2	99	0.00
50	2,6-Dinitrotoluene	40.000	42.949	-7.4	98	0.00
51 C	Acenaphthene	40.000	40.143	-0.4	97	0.00
52	3-Nitroaniline	40.000	42.167	-5.4	97	0.00
53 P	2,4-Dinitrophenol	40.000	41.063	-2.7	97	0.00
54	Dibenzofuran	40.000	40.689	-1.7	98	0.00
55 P	4-Nitrophenol	40.000	42.948	-7.4	99	0.00
56	2,4-Dinitrotoluene	40.000	42.892	-7.2	99	0.00
57	Fluorene	40.000	40.889	-2.2	98	0.00
58	2,3,4,6-Tetrachlorophenol	40.000	41.149	-2.9	99	0.00
59	Diethylphthalate	40.000	41.047	-2.6	98	0.00
60	4-Chlorophenyl-phenylether	40.000	40.718	-1.8	99	0.00
61	4-Nitroaniline	40.000	42.316	-5.8	98	0.00
62	Azobenzene	40.000	40.684	-1.7	97	0.00
63 I	Phenanthrene-d10	20.000	20.000	0.0	99	0.00
64	4,6-Dinitro-2-methylphenol	40.000	43.298	-8.2	101	0.00
65 c	n-Nitrosodiphenylamine	40.000	40.694	-1.7	99	0.00
66	4-Bromophenyl-phenylether	40.000	41.865	-4.7	100	0.00
67	Hexachlorobenzene	40.000	40.200	-0.5	98	0.00
68	Atrazine	40.000	41.130	-2.8	98	0.00
69 C	Pentachlorophenol	40.000	41.636	-4.1	99	0.00
70	Phenanthrene	40.000	40.314	-0.8	99	0.00
71	Anthracene	40.000	41.056	-2.6	100	0.00
72	Carbazole	40.000	40.595	-1.5	99	0.00
73	Di-n-butylphthalate	40.000	41.132	-2.8	100	0.00
74 C	Fluoranthene	40.000	41.075	-2.7	101	0.00
75 I	Chrysene-d12	20.000	20.000	0.0	100	0.00
76	Benzidine	40.000	41.549	-3.9	96	0.00
77	Pyrene	40.000	41.747	-4.4	101	0.00
78 S	Terphenyl-d14	80.000	82.084	-2.6	101	0.00
79	Butylbenzylphthalate	40.000	41.554	-3.9	100	0.00
80	Benzo(a)anthracene	40.000	40.125	-0.3	99	0.00
81	3,3'-Dichlorobenzidine	40.000	39.671	0.8	96	0.00
82	Chrysene	40.000	40.398	-1.0	100	0.00
83	Bis(2-ethylhexyl)phthalate	40.000	40.678	-1.7	100	0.00
84 c	Di-n-octyl phthalate	40.000	41.090	-2.7	99	0.00
85	Indeno(1,2,3-cd)pyrene	40.000	41.423	-3.6	100	0.00
86 I	Perylene-d12	20.000	20.000	0.0	100	0.00
87	Benzo(b)fluoranthene	40.000	40.629	-1.6	101	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG032017\
Data File : BG026330.D
Acq On : 20 Mar 2017 17:08
Operator : SJ/MA
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG032017

Quant Time: Mar 20 17:43:42 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG032017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 20 17:43:27 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	40.000	40.619	-1.5	98	0.00
89 C	Benzo(a)pyrene	40.000	40.575	-1.4	100	0.00
90	Dibenzo(a,h)anthracene	40.000	40.898	-2.2	100	0.00
91	Benzo(a,h,i)perylene	40.000	40.681	-1.7	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0