

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA G\Data\BG032018\
 Data File : BG033264.D
 Acq On : 20 Mar 2018 9:18
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 20 16:24:09 2018
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG031918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 19 16:48:53 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	110	0.00
2	1,4-Dioxane	40.000	39.682	0.8	110	0.00
3	Pyridine	40.000	40.336	-0.8	100	0.00
4	n-Nitrosodimethylamine	40.000	40.429	-1.1	110	0.00
5 S	2-Fluorophenol	80.000	81.599	-2.0	103	0.00
6	Aniline	40.000	41.755	-4.4	106	0.00
7 S	Phenol-d6	80.000	80.759	-0.9	107	0.00
8	2-Chlorophenol	40.000	40.625	-1.6	103	0.00
9	Benzaldehyde	40.000	36.619	8.5	103	0.00
10 C	Phenol	40.000	40.031	-0.1	104	0.00
11	bis(2-Chloroethyl)ether	40.000	39.581	1.0	100	0.00
12	1,3-Dichlorobenzene	40.000	38.973	2.6	105	0.00
13 C	1,4-Dichlorobenzene	40.000	39.913	0.2	110	0.00
14	1,2-Dichlorobenzene	40.000	40.399	-1.0	106	0.00
15	Benzyl Alcohol	40.000	39.971	0.1	102	0.00
16	2,2'-oxybis(1-Chloropropane)	40.000	40.187	-0.5	108	0.00
17	2-Methylphenol	40.000	37.760	5.6	101	0.00
18	Hexachloroethane	40.000	39.090	2.3	106	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.891	0.3	100	0.00
20	3+4-Methylphenols	40.000	40.361	-0.9	103	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	100	0.00
22	Acetophenone	40.000	41.637	-4.1	96	0.00
23 S	Nitrobenzene-d5	80.000	84.555	-5.7	102	0.00
24	Nitrobenzene	40.000	40.734	-1.8	98	0.00
25	Isophorone	40.000	40.975	-2.4	99	0.00
26 C	2-Nitrophenol	40.000	44.704	-11.8	103	0.00
27	2,4-Dimethylphenol	40.000	41.269	-3.2	105	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.952	-4.9	101	0.00
29 C	2,4-Dichlorophenol	40.000	41.281	-3.2	98	0.00
30	1,2,4-Trichlorobenzene	40.000	41.854	-4.6	103	0.00
31	Naphthalene	40.000	41.815	-4.5	102	0.00
32	Benzoic acid	40.000	39.197	2.0	109	-0.01
33	4-Chloroaniline	40.000	41.008	-2.5	98	0.00
34 C	Hexachlorobutadiene	40.000	41.443	-3.6	105	0.00
35	Caprolactam	40.000	40.950	-2.4	98	-0.01
36 C	4-Chloro-3-methylphenol	40.000	40.449	-1.1	94	0.00
37	2-Methylnaphthalene	40.000	40.315	-0.8	102	0.00
38 I	Acenaphthene-d10	20.000	20.000	0.0	98	0.00
39	1,2,4,5-Tetrachlorobenzene	40.000	40.687	-1.7	99	0.00
40 P	Hexachlorocyclopentadiene	40.000	42.565	-6.4	101	0.00
41 S	2,4,6-Tribromophenol	80.000	81.001	-1.3	96	0.00
42 C	2,4,6-Trichlorophenol	40.000	41.663	-4.2	97	0.00
43	2,4,5-Trichlorophenol	40.000	43.591	-9.0	98	0.00
44 S	2-Fluorobiphenyl	80.000	81.873	-2.3	99	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA G\Data\BG032018\
 Data File : BG033264.D
 Acq On : 20 Mar 2018 9:18
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 20 16:24:09 2018
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG031918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 19 16:48:53 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	40.000	40.489	-1.2	98	0.00
46	2-Chloronaphthalene	40.000	41.233	-3.1	99	0.00
47	2-Nitroaniline	40.000	42.876	-7.2	100	0.00
48	Acenaphthylene	40.000	40.733	-1.8	98	0.00
49	Dimethylphthalate	40.000	40.350	-0.9	98	0.00
50	2,6-Dinitrotoluene	40.000	42.825	-7.1	99	0.00
51 C	Acenaphthene	40.000	40.656	-1.6	97	0.00
52	3-Nitroaniline	40.000	40.226	-0.6	96	0.00
53 P	2,4-Dinitrophenol	40.000	37.064	7.3	90	0.00
54	Dibenzofuran	40.000	39.901	0.2	97	0.00
55 P	4-Nitrophenol	40.000	40.349	-0.9	95	0.00
56	2,4-Dinitrotoluene	40.000	41.718	-4.3	100	0.00
57	Fluorene	40.000	39.805	0.5	98	0.00
58	2,3,4,6-Tetrachlorophenol	40.000	40.104	-0.3	93	0.00
59	Diethylphthalate	40.000	40.916	-2.3	100	0.00
60	4-Chlorophenyl-phenylether	40.000	39.504	1.2	98	0.00
61	4-Nitroaniline	40.000	41.036	-2.6	98	0.00
62	Azobenzene	40.000	41.070	-2.7	99	0.00
63 I	Phenanthrene-d10	20.000	20.000	0.0	98	0.00
64	4,6-Dinitro-2-methylphenol	40.000	39.745	0.6	92	0.00
65 c	n-Nitrosodiphenylamine	40.000	41.393	-3.5	100	0.00
66	4-Bromophenyl-phenylether	40.000	41.114	-2.8	99	0.00
67	Hexachlorobenzene	40.000	41.119	-2.8	100	0.00
68	Atrazine	40.000	41.784	-4.5	100	0.00
69 C	Pentachlorophenol	40.000	42.361	-5.9	102	0.00
70	Phenanthrene	40.000	40.692	-1.7	99	0.00
71	Anthracene	40.000	40.329	-0.8	98	0.00
72	Carbazole	40.000	40.673	-1.7	98	0.00
73	Di-n-butylphthalate	40.000	41.368	-3.4	99	0.00
74 C	Fluoranthene	40.000	40.256	-0.6	98	0.00
75 I	Chrysene-d12	20.000	20.000	0.0	96	0.00
76	Benzidine	40.000	43.478	-8.7	95	0.00
77	Pyrene	40.000	41.261	-3.2	97	0.00
78 S	Terphenyl-d14	80.000	82.382	-3.0	98	0.00
79	Butylbenzylphthalate	40.000	41.835	-4.6	98	0.00
80	Benzo(a)anthracene	40.000	40.326	-0.8	96	0.00
81	3,3'-Dichlorobenzidine	40.000	43.812	-9.5	99	0.00
82	Chrysene	40.000	40.549	-1.4	97	0.00
83	Bis(2-ethylhexyl)phthalate	40.000	42.230	-5.6	99	0.00
84 c	Di-n-octyl phthalate	40.000	42.246	-5.6	97	0.00
85	Indeno(1,2,3-cd)pyrene	40.000	41.763	-4.4	97	0.00
86 I	Perylene-d12	20.000	20.000	0.0	98	0.00
87	Benzo(b)fluoranthene	40.000	39.487	1.3	95	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA G\Data\BG032018\
Data File : BG033264.D
Acq On : 20 Mar 2018 9:18
Operator : SJ/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Mar 20 16:24:09 2018
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG031918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 19 16:48:53 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	40.000	40.040	-0.1	96	0.00
89 C	Benzo(a)pyrene	40.000	40.065	-0.2	96	0.00
90	Dibenzo(a,h)anthracene	40.000	41.633	-4.1	97	0.00
91	Benzo(a,h,i)perylene	40.000	41.392	-3.5	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0