

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA G\Data\BG032018\
 Data File : BG033276.D
 Acq On : 20 Mar 2018 21:42
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 21 03:26:45 2018
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG031918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 19 16:48:53 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	127	0.00
2	1,4-Dioxane	40.000	42.760	-6.9	137	0.00
3	Pyridine	40.000	42.155	-5.4	121	0.00
4	n-Nitrosodimethylamine	40.000	42.304	-5.8	133	0.00
5 S	2-Fluorophenol	80.000	87.950	-9.9	128	0.00
6	Aniline	40.000	42.814	-7.0	125	0.00
7 S	Phenol-d6	80.000	84.664	-5.8	130	0.00
8	2-Chlorophenol	40.000	42.313	-5.8	123	0.00
9	Benzaldehyde	40.000	38.885	2.8	126	0.00
10 C	Phenol	40.000	42.376	-5.9	127	0.00
11	bis(2-Chloroethyl)ether	40.000	42.768	-6.9	124	0.00
12	1,3-Dichlorobenzene	40.000	40.991	-2.5	127	0.00
13 C	1,4-Dichlorobenzene	40.000	40.838	-2.1	129	0.00
14	1,2-Dichlorobenzene	40.000	42.193	-5.5	128	0.00
15	Benzyl Alcohol	40.000	42.387	-6.0	125	0.00
16	2,2'-oxybis(1-Chloropropane)	40.000	41.691	-4.2	129	0.00
17	2-Methylphenol	40.000	40.787	-2.0	126	0.00
18	Hexachloroethane	40.000	41.119	-2.8	129	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.991	-5.0	122	0.00
20	3+4-Methylphenols	40.000	44.490	-11.2	131	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	119	0.00
22	Acetophenone	40.000	43.852	-9.6	121	0.00
23 S	Nitrobenzene-d5	80.000	85.247	-6.6	123	0.00
24	Nitrobenzene	40.000	42.616	-6.5	122	0.00
25	Isophorone	40.000	41.618	-4.0	119	0.00
26 C	2-Nitrophenol	40.000	45.726	-14.3	125	0.00
27	2,4-Dimethylphenol	40.000	42.045	-5.1	127	0.00
28	bis(2-Chloroethoxy)methane	40.000	42.756	-6.9	122	0.00
29 C	2,4-Dichlorophenol	40.000	43.450	-8.6	123	0.00
30	1,2,4-Trichlorobenzene	40.000	41.113	-2.8	120	0.00
31	Naphthalene	40.000	41.179	-2.9	120	0.00
32	Benzoic acid	40.000	35.482	11.3	118	0.00
33	4-Chloroaniline	40.000	39.699	0.8	113	0.00
34 C	Hexachlorobutadiene	40.000	42.289	-5.7	128	0.00
35	Caprolactam	40.000	39.598	1.0	113	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.753	0.6	110	0.00
37	2-Methylnaphthalene	40.000	40.416	-1.0	121	0.00
38 I	Acenaphthene-d10	20.000	20.000	0.0	112	0.00
39	1,2,4,5-Tetrachlorobenzene	40.000	41.726	-4.3	116	0.00
40 P	Hexachlorocyclopentadiene	40.000	44.163	-10.4	120	0.00
41 S	2,4,6-Tribromophenol	80.000	83.774	-4.7	113	0.00
42 C	2,4,6-Trichlorophenol	40.000	44.420	-11.1	118	0.00
43	2,4,5-Trichlorophenol	40.000	45.118	-12.8	116	0.00
44 S	2-Fluorobiphenyl	80.000	84.475	-5.6	116	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA G\Data\BG032018\
 Data File : BG033276.D
 Acq On : 20 Mar 2018 21:42
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 21 03:26:45 2018
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG031918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 19 16:48:53 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	40.000	42.621	-6.6	117	0.00
46	2-Chloronaphthalene	40.000	42.475	-6.2	116	0.00
47	2-Nitroaniline	40.000	42.230	-5.6	112	0.00
48	Acenaphthylene	40.000	40.656	-1.6	112	0.00
49	Dimethylphthalate	40.000	41.867	-4.7	116	0.00
50	2,6-Dinitrotoluene	40.000	42.439	-6.1	112	0.00
51 C	Acenaphthene	40.000	41.720	-4.3	113	0.00
52	3-Nitroaniline	40.000	39.992	0.0	109	0.00
53 P	2,4-Dinitrophenol	40.000	38.711	3.2	108	0.00
54	Dibenzofuran	40.000	40.461	-1.2	112	0.00
55 P	4-Nitrophenol	40.000	40.844	-2.1	109	0.00
56	2,4-Dinitrotoluene	40.000	41.165	-2.9	113	0.00
57	Fluorene	40.000	40.389	-1.0	114	0.00
58	2,3,4,6-Tetrachlorophenol	40.000	43.509	-8.8	116	0.00
59	Diethylphthalate	40.000	41.459	-3.6	115	0.00
60	4-Chlorophenyl-phenylether	40.000	40.434	-1.1	114	0.00
61	4-Nitroaniline	40.000	41.740	-4.4	114	0.00
62	Azobenzene	40.000	41.208	-3.0	113	0.00
63 I	Phenanthrene-d10	20.000	20.000	0.0	113	0.00
64	4,6-Dinitro-2-methylphenol	40.000	40.944	-2.4	109	0.00
65 c	n-Nitrosodiphenylamine	40.000	41.118	-2.8	114	0.00
66	4-Bromophenyl-phenylether	40.000	41.779	-4.4	115	0.00
67	Hexachlorobenzene	40.000	42.113	-5.3	118	0.00
68	Atrazine	40.000	42.545	-6.4	117	0.00
69 C	Pentachlorophenol	40.000	41.943	-4.9	116	0.00
70	Phenanthrene	40.000	40.620	-1.5	113	0.00
71	Anthracene	40.000	40.929	-2.3	114	0.00
72	Carbazole	40.000	40.643	-1.6	112	0.00
73	Di-n-butylphthalate	40.000	42.241	-5.6	116	0.00
74 C	Fluoranthene	40.000	40.575	-1.4	113	0.00
75 I	Chrysene-d12	20.000	20.000	0.0	111	0.00
76	Benzidine	40.000	42.195	-5.5	107	0.00
77	Pyrene	40.000	40.631	-1.6	112	0.00
78 S	Terphenyl-d14	80.000	81.377	-1.7	113	0.00
79	Butylbenzylphthalate	40.000	40.994	-2.5	112	0.00
80	Benzo(a)anthracene	40.000	40.485	-1.2	112	0.00
81	3,3'-Dichlorobenzidine	40.000	42.598	-6.5	113	0.00
82	Chrysene	40.000	39.714	0.7	110	0.00
83	Bis(2-ethylhexyl)phthalate	40.000	41.795	-4.5	114	0.00
84 c	Di-n-octyl phthalate	40.000	42.289	-5.7	114	0.00
85	Indeno(1,2,3-cd)pyrene	40.000	41.358	-3.4	112	0.00
86 I	Perylene-d12	20.000	20.000	0.0	112	0.00
87	Benzo(b)fluoranthene	40.000	40.371	-0.9	111	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA G\Data\BG032018\
Data File : BG033276.D
Acq On : 20 Mar 2018 21:42
Operator : SJ/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Mar 21 03:26:45 2018
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG031918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 19 16:48:53 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	40.000	40.018	-0.0	110	0.00
89 C	Benzo(a)pyrene	40.000	40.486	-1.2	111	0.00
90	Dibenzo(a,h)anthracene	40.000	42.125	-5.3	112	0.01
91	Benzo(a,h,i)perylene	40.000	41.335	-3.3	112	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0