

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG032024\
 Data File : BG060694.D
 Acq On : 20 Mar 2024 16:37
 Operator : MA/JU
 Sample : P1804-01 2X
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 JC-03-03192024-A

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 03/21/2024
 Supervised By :Sohil Jodhani 03/21/2024

Quant Time: Mar 20 17:14:22 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG031324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 18 14:26:31 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.290	152	94856	20.000	ng	0.00
21) Naphthalene-d8	11.134	136	406938	20.000	ng	0.00
39) Acenaphthene-d10	14.924	164	270256	20.000	ng	0.00
64) Phenanthrene-d10	17.673	188	563102	20.000	ng	0.00
76) Chrysene-d12	21.998	240	524674	20.000	ng	0.00
86) Perylene-d12	25.546	264	644153	20.000	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.781	112	252005	41.763	ng	0.00
7) Phenol-d6	7.409	99	373101	42.625	ng	0.00
23) Nitrobenzene-d5	9.489	82	235250	30.086	ng	0.00
42) 2,4,6-Tribromophenol	16.410	330	149942	47.879	ng	0.00
45) 2-Fluorobiphenyl	13.543	172	546118	27.321	ng	0.00
79) Terphenyl-d14	20.253	244	867754	29.895	ng	0.00
Target Compounds						Qvalue
31) Naphthalene	11.187	128	415186	18.266	ng	100
37) 2-Methylnaphthalene	12.767	142	182864	11.648	ng	97
38) 1-Methylnaphthalene	12.991	142	142004m	9.637	ng	
46) 1,1'-Biphenyl	13.760	154	41445	2.018	ng	96
49) Acenaphthylene	14.659	152	100032	3.961	ng	98
52) Acenaphthene	14.988	154	286279	18.969	ng	100
55) Dibenzofuran	15.317	168	318053	13.016	ng	99
58) Fluorene	15.964	166	482706	24.690	ng	97
71) Phenanthrene	17.720	178	3410617	112.439	ng	96
72) Anthracene	17.808	178	1147266	37.407	ng	98
73) Carbazole	18.085	167	332094	12.453	ng	99
75) Fluoranthene	19.718	202	4680511	139.497	ng	88
78) Pyrene	20.082	202	4431169	121.020	ng	# 86
81) Benzo(a)anthracene	21.980	228	2802424	77.575	ng	96
83) Chrysene	22.051	228	2655373m	75.701	ng	
87) Indeno(1,2,3-cd)pyrene	29.665	276	1539256	35.032	ng	# 91
88) Benzo(b)fluoranthene	24.407	252	3457828	97.576	ng	97
89) Benzo(k)fluoranthene	24.471	252	1484740m	39.881	ng	
90) Benzo(a)pyrene	25.388	252	2695572	77.587	ng	96
91) Dibenzo(a,h)anthracene	29.724	278	333161	8.932	ng	# 85
92) Benzo(g,h,i)perylene	30.964	276	1352110	37.440	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG032024\
Data File : BG060694.D
Acq On : 20 Mar 2024 16:37
Operator : MA/JU
Sample : P1804-01 2X
Misc :
ALS Vial : 10 Sample Multiplier: 1


Instrument :
BNA_G
ClientSampleId :
JC-03-03192024-A

Quant Time: Mar 20 17:14:22 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG031324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 18 14:26:31 2024
Response via : Initial Calibration

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/21/2024
Supervised By :Sohil Jodhani 03/21/2024





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.290 min Scan# 81
Delta R.T. 0.002 min
Lab File: BG060694.D
Acq: 20 Mar 2024 16:37




Instrument :
BNA_G
ClientSampleId :
JC-03-03192024-A



Tgt Ion:152 Resp: 94850
Ion Ratio Lower Upper
152 100
150 156.7 126.6 190.0
115 57.5 47.8 71.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/21/2024
Supervised By :Sohil Jodhani 03/21/2024



#5
2-Fluorophenol
Concen: 41.763 ng
RT: 5.781 min Scan# 407
Delta R.T. 0.002 min
Lab File: BG060694.D
Acq: 20 Mar 2024 16:37


Tgt Ion:112 Resp: 252005
Ion Ratio Lower Upper
112 100
64 66.8 54.1 81.1
63 35.1 30.3 45.5





#7
Phenol-d6
Concen: 42.625 ng
RT: 7.409 min Scan# 61
Delta R.T. 0.002 min
Lab File: BG060694.D
Acq: 20 Mar 2024 16:37




Instrument :
BNA_G
ClientSampleId :
JC-03-03192024-A



Tgt Ion: 99 Resp: 37310
Ion Ratio Lower Upper
99 100
42 20.8 17.5 26.3
71 33.3 27.4 41.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/21/2024
Supervised By :Sohil Jodhani 03/21/2024



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 11.134 min Scan# 1318
Delta R.T. -0.004 min
Lab File: BG060694.D
Acq: 20 Mar 2024 16:37

Tgt Ion:136 Resp: 406938
Ion Ratio Lower Upper
136 100
137 11.1 8.4 12.6
54 11.2 8.5 12.7
68 6.6 5.0 7.6





#23

Nitrobenzene-d5

Concen: 30.086 ng

RT: 9.489 min Scan# 1045

Delta R.T. -0.004 min

Lab File: BG060694.D

Acq: 20 Mar 2024 16:37

Instrument :

BNA_G

ClientSampleId :

JC-03-03192024-A





Tgt Ion: 82 Resp: 235250

Ion	Ratio	Lower	Upper
82	100		
128	39.2	31.6	47.4
54	61.2	49.3	73.9

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 03/21/2024
Supervised By :Sohil Jodhani 03/21/2024



#31

Naphthalene

Concen: 18.266 ng

RT: 11.187 min Scan# 1327




Delta R.T. -0.004 min


Lab File: BG060694.D

Acq: 20 Mar 2024 16:37

Tgt Ion:128 Resp: 415186


Ion	Ratio	Lower	Upper
128	100		
129	10.5	8.6	12.8
127	13.6	11.0	16.4





#37
2-Methylnaphthalene
Concen: 11.648 ng
RT: 12.767 min Scan# 11
Delta R.T. -0.004 min
Lab File: BG060694.D
Acq: 20 Mar 2024 16:37




Instrument :
BNA_G
ClientSampleId :
JC-03-03192024-A




Tgt Ion:142 Resp: 182864
Ion Ratio Lower Upper
142 100
141 88.4 72.7 109.1
115 36.1 30.2 45.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/21/2024
Supervised By :Sohil Jodhani 03/21/2024




#38
1-Methylnaphthalene
Concen: 9.637 ng m
RT: 12.991 min Scan# 1634
Delta R.T. 0.002 min
Lab File: BG060694.D
Acq: 20 Mar 2024 16:37




Tgt Ion:142 Resp: 142004
Ion Ratio Lower Upper
142 100
141 95.5 76.3 114.5
116 6.0 3.4 5.2#





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.924 min Scan# 1963
Delta R.T. -0.004 min
Lab File: BG060694.D
Acq: 20 Mar 2024 16:37




Instrument :
BNA_G
ClientSampleId :
JC-03-03192024-A




Tgt Ion:164 Resp: 270250
Ion Ratio Lower Upper
164 100
162 105.9 82.5 123.7
160 47.7 36.4 54.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/21/2024
Supervised By :Sohil Jodhani 03/21/2024




#42
2,4,6-Tribromophenol
Concen: 47.879 ng
RT: 16.410 min Scan# 2216
Delta R.T. 0.002 min
Lab File: BG060694.D
Acq: 20 Mar 2024 16:37




Tgt Ion:330 Resp: 149942
Ion Ratio Lower Upper
330 100
332 97.7 77.8 116.8
141 33.1 25.4 38.2





#45
2-Fluorobiphenyl
Concen: 27.321 ng
RT: 13.543 min Scan# 1728
Delta R.T. -0.004 min
Lab File: BG060694.D
Acq: 20 Mar 2024 16:37




Instrument :
BNA_G
ClientSampleId :
JC-03-03192024-A




Tgt Ion:172 Resp: 546118
Ion Ratio Lower Upper
172 100
171 34.6 29.9 44.9
170 23.2 19.8 29.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/21/2024
Supervised By :Sohil Jodhani 03/21/2024




#46
1,1'-Biphenyl
Concen: 2.018 ng
RT: 13.760 min Scan# 1765
Delta R.T. -0.004 min
Lab File: BG060694.D
Acq: 20 Mar 2024 16:37



Tgt Ion:154 Resp: 41445
Ion Ratio Lower Upper
154 100
153 37.7 20.4 60.4
76 13.0 0.0 34.3





#49

Acenaphthylene

Concen: 3.961 ng

RT: 14.659 min Scan# 1918

Delta R.T. 0.002 min

Lab File: BG060694.D

Acq: 20 Mar 2024 16:37

Instrument :

BNL_G

ClientSampleId :





JC-03-03192024-A

Tgt Ion:	152	Resp:	10003
Ion Ratio	Lower	Upper	
152	100		
151	20.9	17.1	25.7
153	11.9	10.3	15.5

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 03/21/2024
Supervised By :Sohil Jodhani 03/21/2024



#52

Acenaphthene

Concen: 18.969 ng




RT: 14.988 min Scan# 1974


Delta R.T. -0.004 min

Lab File: BG060694.D

Acq: 20 Mar 2024 16:37


Tgt Ion:	154	Resp:	286279
Ion Ratio	Lower	Upper	
154	100		
153	118.4	94.8	142.2
152	55.3	44.8	67.2





#55
Dibenzenofuran
Concen: 13.016 ng
RT: 15.317 min Scan# 2036
Delta R.T. -0.004 min
Lab File: BG060694.D
Acq: 20 Mar 2024 16:37




Instrument :
BNA_G
ClientSampleId :
JC-03-03192024-A



Tgt Ion:168 Resp: 31805
Ion Ratio Lower Upper
168 100
139 38.9 30.9 46.3
169 14.2 10.9 16.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/21/2024
Supervised By :Sohil Jodhani 03/21/2024



#58
Fluorene
Concen: 24.690 ng
RT: 15.964 min Scan# 2140
Delta R.T. -0.004 min
Lab File: BG060694.D
Acq: 20 Mar 2024 16:37

Tgt Ion:166 Resp: 482706
Ion Ratio Lower Upper
166 100
165 97.6 80.5 120.7
167 14.0 9.9 14.9





#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 17.673 min Scan# 2431

Delta R.T. 0.002 min

Lab File: BG060694.D

Acq: 20 Mar 2024 16:37

Instrument :

BNA_G

ClientSampleId :

JC-03-03192024-A

Tgt Ion:188 Resp: 56310

Ion Ratio Lower Upper

188 100

94 9.2 7.9 11.9





80 10.6 8.6 13.0

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 03/21/2024

Supervised By :Sohil Jodhani 03/21/2024



#71

Phenanthrene

Concen: 112.439 ng

RT: 17.720 min Scan# 2439

Delta R.T. 0.002 min

Lab File: BG060694.D

Acq: 20 Mar 2024 16:37




Tgt Ion:178 Resp: 3410617


Ion Ratio Lower Upper

178 100

176 22.2 16.1 24.1


179 17.8 12.9 19.3





#72
 Anthracene
 Concen: 37.407 ng
 RT: 17.808 min Scan# 2454
 Delta R.T. -0.004 min
 Lab File: BG060694.D
 Acq: 20 Mar 2024 16:37




Instrument :
 BNA_G
 ClientSampleId :
 JC-03-03192024-A



Tgt Ion:178 Resp: 1147260
 Ion Ratio Lower Upper
 178 100
 176 19.0 15.8 23.6
 179 16.4 12.6 19.0




Manual Integrations
 APPROVED


Reviewed By :Jagrut Upadhyay 03/21/2024
 Supervised By :Sohil Jodhani 03/21/2024



#73
 Carbazole
 Concen: 12.453 ng
 RT: 18.085 min Scan# 2501
 Delta R.T. 0.002 min
 Lab File: BG060694.D
 Acq: 20 Mar 2024 16:37

Tgt Ion:167 Resp: 332094
 Ion Ratio Lower Upper
 167 100
 166 22.7 18.0 27.0
 139 12.8 10.7 16.1





#75

Fluoranthene

Concen: 139.497 ng

RT: 19.718 min Scan# 21

Delta R.T. 0.008 min

Lab File: BG060694.D

Acq: 20 Mar 2024 16:37

Instrument :

BNA_G

ClientSampleId :

JC-03-03192024-A

Tgt Ion:202 Resp: 468051

Ion Ratio Lower Upper

202 100

101 14.8 0.0 31.2





203 23.8 0.0 38.0

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 03/21/2024

Supervised By :Sohil Jodhani 03/21/2024



#76

Chrysene-d12

Concen: 20.000 ng

RT: 21.998 min Scan# 3167

Delta R.T. 0.008 min

Lab File: BG060694.D

Acq: 20 Mar 2024 16:37




Tgt Ion:240 Resp: 524674


Ion Ratio Lower Upper

240 100

120 11.3 7.8 11.8

236 27.8 20.6 31.0





#78

Pyrene

Concen: 121.020 ng

RT: 20.082 min Scan# 2844

Delta R.T. 0.008 min

Lab File: BG060694.D

Acq: 20 Mar 2024 16:37

Instrument :

BNA_G

ClientSampleId :

JC-03-03192024-A

Tgt Ion:202 Resp: 4431169

Ion Ratio Lower Upper

202 100

200 28.8 17.9 26.9





203 24.2 14.3 21.5

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 03/21/2024

Supervised By :Sohil Jodhani 03/21/2024



#79

Terphenyl-d14

Concen: 29.895 ng

RT: 20.253 min Scan# 2870

Delta R.T. -0.004 min

Lab File: BG060694.D

Acq: 20 Mar 2024 16:37




Tgt Ion:244 Resp: 867754


Ion Ratio Lower Upper

244 100

212 7.0 6.6 9.8


122 9.7 8.5 12.7





#81
Benzo(a)anthracene
Concen: 77.575 ng
RT: 21.980 min Scan# 3164
Delta R.T. 0.008 min
Lab File: BG060694.D
Acq: 20 Mar 2024 16:37




Instrument :
BNA_G
ClientSampleId :
JC-03-03192024-A




Tgt Ion:228 Resp: 2802424
Ion Ratio Lower Upper
228 100
226 29.4 21.8 32.8
229 21.4 15.8 23.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/21/2024
Supervised By :Sohil Jodhani 03/21/2024




#83
Chrysene
Concen: 75.701 ng m
RT: 22.051 min Scan# 3176
Delta R.T. 0.008 min
Lab File: BG060694.D
Acq: 20 Mar 2024 16:37



Tgt Ion:228 Resp: 2655373
Ion Ratio Lower Upper
228 100
226 32.3 24.2 36.2
229 24.4 16.4 24.6





#86

Perylene-d12

Concen: 20.000 ng

RT: 25.546 min Scan# 31

Delta R.T. 0.020 min

Lab File: BG060694.D

Acq: 20 Mar 2024 16:37

Instrument :

BNA_G

ClientSampleId :

JC-03-03192024-A





Tgt Ion:264 Resp: 64415

Ion	Ratio	Lower	Upper
264	100		
260	25.2	19.4	29.2
265	23.7	18.1	27.1

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 03/21/2024
Supervised By :Sohil Jodhani 03/21/2024



#87

Indeno(1,2,3-cd)pyrene

Concen: 35.032 ng

RT: 29.665 min Scan# 4472




Delta R.T. 0.031 min


Lab File: BG060694.D

Acq: 20 Mar 2024 16:37

Tgt Ion:276 Resp: 1539256

Ion	Ratio	Lower	Upper
276	100		
138	18.0	6.6	10.0#
277	21.5	16.7	25.1





#88

Benzo(b)fluoranthene

Concen: 97.576 ng

RT: 24.407 min Scan# 3

Delta R.T. 0.020 min

Lab File: BG060694.D

Acq: 20 Mar 2024 16:37

Instrument :

BNA_G

ClientSampleId :

JC-03-03192024-A

Tgt Ion:252 Resp: 345782

Ion Ratio Lower Upper

252 100

253 23.4 17.6 26.4





125 11.0 7.8 11.6

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 03/21/2024

Supervised By :Sohil Jodhani 03/21/2024



#89

Benzo(k)fluoranthene

Concen: 39.881 ng m

RT: 24.471 min Scan# 3588

Delta R.T. 0.014 min

Lab File: BG060694.D

Acq: 20 Mar 2024 16:37




Tgt Ion:252 Resp: 1484740


Ion Ratio Lower Upper

252 100

253 23.5 17.1 25.7


125 10.3 8.0 12.0





#90
Benzo(a)pyrene
Concen: 77.587 ng
RT: 25.388 min Scan# 31
Delta R.T. 0.031 min
Lab File: BG060694.D
Acq: 20 Mar 2024 16:37




Instrument :
BNA_G
ClientSampleId :
JC-03-03192024-A



Tgt Ion:252 Resp: 269557
Ion Ratio Lower Upper
252 100
253 23.6 17.0 25.6
125 11.4 8.2 12.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/21/2024
Supervised By :Sohil Jodhani 03/21/2024



#91
Dibenzo(a,h)anthracene
Concen: 8.932 ng
RT: 29.724 min Scan# 4482
Delta R.T. 0.008 min
Lab File: BG060694.D
Acq: 20 Mar 2024 16:37


Tgt Ion:278 Resp: 333161
Ion Ratio Lower Upper
278 100
139 16.8 9.8 14.8#
279 31.0 18.4 27.6#





#92
 Benzo(g,h,i)perylene
 Concen: 37.440 ng
 RT: 30.964 min Scan# 4693
 Delta R.T. 0.025 min
 Lab File: BG060694.D
 Acq: 20 Mar 2024 16:37

Instrument :
 BNA_G
 ClientSampleId :
 JC-03-03192024-A



Tgt Ion:276 Resp: 1352110
 Ion Ratio Lower Upper
 276 100
 277 25.0 19.0 28.4
 138 19.9 14.4 21.6

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 03/21/2024
 Supervised By :Sohil Jodhani 03/21/2024

