

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG032221\
 Data File : BG048441.D
 Acq On : 22 Mar 2021 12:19
 Operator : CG/JU
 Sample : PB135247BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PB135247BL

Quant Time: Mar 22 12:56:38 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG031821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 22 11:22:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.104	152	44797	20.00	ng	0.00
21) Naphthalene-d8	10.924	136	181485	20.00	ng	0.00
39) Acenaphthene-d10	14.743	164	146728	20.00	ng	0.00
64) Phenanthrene-d10	17.493	188	365270	20.00	ng	# 0.00
76) Chrysene-d12	21.788	240	383026	20.00	ng	# 0.00
86) Perylene-d12	25.113	264	377382	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.648	112	374636	138.30	ng	0.00
7) Phenol-d6	7.246	99	507429	128.73	ng	0.00
23) Nitrobenzene-d5	9.267	82	386198	85.08	ng	0.00
42) 2,4,6-Tribromophenol	16.235	330	279070	116.95	ng	0.00
45) 2-Fluorobiphenyl	13.368	172	834752	81.84	ng	0.00
79) Terphenyl-d14	20.107	244	1724177	81.18	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG032221\
Data File : BG048441.D
Acq On : 22 Mar 2021 12:19
Operator : CG/JU
Sample : PB135247BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB135247BL

Quant Time: Mar 22 12:56:38 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG031821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 22 11:22:24 2021
Response via : Initial Calibration

