

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG032221\
 Data File : BG048440.D
 Acq On : 22 Mar 2021 10:32
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 22 11:22:58 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG031821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 22 11:22:24 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	105	0.00
2	1,4-Dioxane	40.000	37.456	6.4	105	0.00
3	Pyridine	40.000	40.611	-1.5	112	0.00
4	n-Nitrosodimethylamine	40.000	39.347	1.6	110	0.00
5 S	2-Fluorophenol	80.000	78.983	1.3	111	0.00
6	Aniline	40.000	39.316	1.7	108	0.00
7 S	Phenol-d6	80.000	80.784	-1.0	110	0.00
8	2-Chlorophenol	40.000	40.371	-0.9	111	0.00
9	Benzaldehyde	40.000	41.277	-3.2	107	0.00
10 C	Phenol	40.000	39.013	2.5	109	0.00
11	bis(2-Chloroethyl)ether	40.000	39.853	0.4	110	0.00
12	1,3-Dichlorobenzene	40.000	39.941	0.1	110	0.00
13 C	1,4-Dichlorobenzene	40.000	39.717	0.7	113	0.00
14	1,2-Dichlorobenzene	40.000	39.716	0.7	109	0.00
15	Benzyl Alcohol	40.000	39.531	1.2	109	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.359	4.1	106	0.00
17	2-Methylphenol	40.000	38.740	3.1	110	0.00
18	Hexachloroethane	40.000	39.441	1.4	112	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.196	4.5	104	0.00
20	3+4-Methylphenols	40.000	40.313	-0.8	112	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	109	0.00
22	Acetophenone	40.000	39.195	2.0	110	0.00
23 S	Nitrobenzene-d5	80.000	79.749	0.3	112	0.00
24	Nitrobenzene	40.000	39.342	1.6	109	0.00
25	Isophorone	40.000	38.536	3.7	107	0.00
26 C	2-Nitrophenol	40.000	40.132	-0.3	111	0.00
27	2,4-Dimethylphenol	40.000	40.239	-0.6	113	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.019	2.5	112	0.00
29 C	2,4-Dichlorophenol	40.000	41.157	-2.9	112	0.00
30	1,2,4-Trichlorobenzene	40.000	39.106	2.2	108	0.00
31	Naphthalene	40.000	39.706	0.7	109	0.00
32	Benzoic acid	40.000	39.189	2.0	107	0.00
33	4-Chloroaniline	40.000	38.965	2.6	108	0.00
34 C	Hexachlorobutadiene	40.000	38.289	4.3	105	0.00
35	Caprolactam	40.000	38.308	4.2	108	0.00
36 C	4-Chloro-3-methylphenol	40.000	38.816	3.0	109	0.00
37	2-Methylnaphthalene	40.000	39.201	2.0	110	0.00
38	1-Methylnaphthalene	40.000	38.912	2.7	109	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	110	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.755	0.6	109	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.278	1.8	107	0.00
42 S	2,4,6-Tribromophenol	80.000	76.829	4.0	107	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.878	-2.2	109	0.00
44	2,4,5-Trichlorophenol	40.000	37.932	5.2	103	0.00
45 S	2-Fluorobiphenyl	80.000	79.682	0.4	110	0.00
46	1,1'-Biphenyl	40.000	39.998	0.0	111	0.00
47	2-Chloronaphthalene	40.000	39.998	0.0	110	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG032221\
 Data File : BG048440.D
 Acq On : 22 Mar 2021 10:32
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 22 11:22:58 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG031821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 22 11:22:24 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.862	0.3	107	0.00
49	Acenaphthylene	40.000	39.033	2.4	108	0.00
50	Dimethylphthalate	40.000	38.897	2.8	108	0.00
51	2,6-Dinitrotoluene	40.000	39.964	0.1	108	0.00
52 C	Acenaphthene	40.000	39.137	2.2	107	0.00
53	3-Nitroaniline	40.000	40.216	-0.5	112	0.00
54 P	2,4-Dinitrophenol	40.000	38.368	4.1	102	0.00
55	Dibenzofuran	40.000	39.613	1.0	110	0.00
56 P	4-Nitrophenol	40.000	38.897	2.8	109	0.00
57	2,4-Dinitrotoluene	40.000	40.412	-1.0	108	0.00
58	Fluorene	40.000	39.082	2.3	108	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.282	4.3	104	0.00
60	Diethylphthalate	40.000	37.933	5.2	106	0.00
61	4-Chlorophenyl-phenylether	40.000	39.185	2.0	108	0.00
62	4-Nitroaniline	40.000	39.552	1.1	109	0.00
63	Azobenzene	40.000	38.048	4.9	106	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	106	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.809	-2.0	106	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.542	-1.4	108	0.00
67	4-Bromophenyl-phenylether	40.000	40.146	-0.4	108	0.00
68	Hexachlorobenzene	40.000	39.583	1.0	107	0.00
69	Atrazine	40.000	40.203	-0.5	108	0.00
70 C	Pentachlorophenol	40.000	41.046	-2.6	106	0.00
71	Phenanthrene	40.000	39.342	1.6	105	0.00
72	Anthracene	40.000	40.166	-0.4	107	0.00
73	Carbazole	40.000	39.375	1.6	107	0.00
74	Di-n-butylphthalate	40.000	39.134	2.2	104	0.00
75 C	Fluoranthene	40.000	38.592	3.5	104	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	101	0.00
77	Benzydine	40.000	41.180	-2.9	104	0.00
78	Pyrene	40.000	41.125	-2.8	105	0.00
79 S	Terphenyl-d14	80.000	77.832	2.7	102	0.00
80	Butylbenzylphthalate	40.000	40.302	-0.8	102	0.00
81	Benzo(a)anthracene	40.000	39.729	0.7	102	0.00
82	3,3'-Dichlorobenzidine	40.000	39.721	0.7	103	0.00
83	Chrysene	40.000	40.198	-0.5	103	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	39.992	0.0	101	0.00
85 c	Di-n-octyl phthalate	40.000	39.363	1.6	99	0.00
86 I	Perylene-d12	20.000	20.000	0.0	103	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	38.897	2.8	101	0.00
88	Benzo(b)fluoranthene	40.000	38.031	4.9	99	0.00
89	Benzo(k)fluoranthene	40.000	39.630	0.9	103	0.00
90 C	Benzo(a)pyrene	40.000	38.645	3.4	100	0.00
91	Dibenzo(a,h)anthracene	40.000	38.798	3.0	100	0.00
92	Benzo(g,h,i)perylene	40.000	38.754	3.1	101	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG032221\
Data File : BG048440.D
Acq On : 22 Mar 2021 10:32
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Mar 22 11:22:58 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG031821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 22 11:22:24 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0