

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG032221\
 Data File : BG048451.D
 Acq On : 22 Mar 2021 19:54
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 23 00:47:12 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG031821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 22 11:22:24 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2	1,4-Dioxane	0.510	0.455	10.8	96	0.00
3	Pyridine	1.422	1.315	7.5	97	0.00
4	n-Nitrosodimethylamine	0.827	0.804	2.8	104	0.00
5 S	2-Fluorophenol	1.209	1.139	5.8	101	0.00
6	Aniline	2.081	1.996	4.1	101	0.00
7 S	Phenol-d6	1.760	1.697	3.6	101	0.00
8	2-Chlorophenol	1.261	1.239	1.7	104	0.00
9	Benzaldehyde	1.109	0.965	13.0	100	0.00
10 C	Phenol	1.803	1.705	5.4	101	0.00
11	bis(2-Chloroethyl)ether	1.264	1.225	3.1	102	0.00
12	1,3-Dichlorobenzene	1.500	1.434	4.4	101	0.00
13 C	1,4-Dichlorobenzene	1.489	1.433	3.8	104	0.00
14	1,2-Dichlorobenzene	1.429	1.375	3.8	101	0.00
15	Benzyl Alcohol	1.562	1.524	2.4	102	0.00
16	2,2'-oxybis(1-Chloropropane	1.445	1.308	9.5	95	0.00
17	2-Methylphenol	1.143	1.108	3.1	105	0.00
18	Hexachloroethane	0.573	0.544	5.1	103	0.00
19 P	n-Nitroso-di-n-propylamine	1.306	1.248	4.4	99	0.00
20	3+4-Methylphenols	1.566	1.573	-0.4	106	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
22	Acetophenone	0.562	0.564	-0.4	105	0.00
23 S	Nitrobenzene-d5	0.500	0.506	-1.2	106	0.00
24	Nitrobenzene	0.536	0.534	0.4	104	0.00
25	Isophorone	0.957	0.987	-3.1	107	0.00
26 C	2-Nitrophenol	0.204	0.210	-2.9	107	0.00
27	2,4-Dimethylphenol	0.318	0.328	-3.1	108	0.00
28	bis(2-Chloroethoxy)methane	0.434	0.430	0.9	106	0.00
29 C	2,4-Dichlorophenol	0.369	0.388	-5.1	107	0.00
30	1,2,4-Trichlorobenzene	0.446	0.453	-1.6	105	0.00
31	Naphthalene	1.109	1.125	-1.4	104	0.00
32	Benzoic acid	0.247	0.277	-12.1	114	0.00
33	4-Chloroaniline	0.474	0.480	-1.3	105	0.00
34 C	Hexachlorobutadiene	0.353	0.354	-0.3	102	0.00
35	Caprolactam	0.132	0.135	-2.3	108	0.00
36 C	4-Chloro-3-methylphenol	0.428	0.449	-4.9	110	0.00
37	2-Methylnaphthalene	0.846	0.873	-3.2	108	0.00
38	1-Methylnaphthalene	0.796	0.807	-1.4	106	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	112	0.00
40	1,2,4,5-Tetrachlorobenzene	0.730	0.703	3.7	109	0.00
41 P	Hexachlorocyclopentadiene	0.472	0.442	6.4	104	0.00
42 S	2,4,6-Tribromophenol	0.325	0.326	-0.3	115	0.00
43 C	2,4,6-Trichlorophenol	0.495	0.497	-0.4	110	0.00
44	2,4,5-Trichlorophenol	0.545	0.534	2.0	109	0.00
45 S	2-Fluorobiphenyl	1.390	1.333	4.1	108	0.00
46	1,1'-Biphenyl	1.435	1.356	5.5	108	0.00
47	2-Chloronaphthalene	1.157	1.098	5.1	107	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG032221\
 Data File : BG048451.D
 Acq On : 22 Mar 2021 19:54
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 23 00:47:12 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG031821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 22 11:22:24 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.428	0.432	-0.9	111	0.00
49	Acenaphthylene	1.874	1.776	5.2	108	0.00
50	Dimethylphthalate	1.628	1.552	4.7	109	0.00
51	2,6-Dinitrotoluene	0.350	0.349	0.3	110	0.00
52 C	Acenaphthene	1.259	1.239	1.6	110	0.00
53	3-Nitroaniline	0.335	0.340	-1.5	116	0.00
54 P	2,4-Dinitrophenol	0.214	0.228	-6.5	116	0.00
55	Dibenzofuran	1.876	1.820	3.0	111	0.00
56 P	4-Nitrophenol	0.300	0.297	1.0	114	0.00
57	2,4-Dinitrotoluene	0.497	0.510	-2.6	112	0.00
58	Fluorene	1.574	1.508	4.2	109	0.00
59	2,3,4,6-Tetrachlorophenol	0.538	0.529	1.7	110	0.00
60	Diethylphthalate	1.591	1.535	3.5	110	0.00
61	4-Chlorophenyl-phenylether	0.906	0.882	2.6	110	0.00
62	4-Nitroaniline	0.367	0.355	3.3	109	0.00
63	Azobenzene	1.732	1.613	6.9	106	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	107	0.00
65	4,6-Dinitro-2-methylphenol	0.126	0.141	-11.9	116	0.00
66 c	n-Nitrosodiphenylamine	0.579	0.587	-1.4	109	0.00
67	4-Bromophenyl-phenylether	0.252	0.257	-2.0	110	0.00
68	Hexachlorobenzene	0.270	0.279	-3.3	113	0.00
69	Atrazine	0.233	0.233	0.0	108	0.00
70 C	Pentachlorophenol	0.179	0.192	-7.3	111	0.00
71	Phenanthrene	1.056	1.058	-0.2	108	0.00
72	Anthracene	1.048	1.052	-0.4	108	0.00
73	Carbazole	1.000	0.975	2.5	106	0.00
74	Di-n-butylphthalate	1.106	1.088	1.6	106	0.00
75 C	Fluoranthene	1.407	1.378	2.1	107	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	105	0.00
77	Benzydine	0.599	0.596	0.5	105	0.00
78	Pyrene	1.374	1.374	0.0	106	0.00
79 S	Terphenyl-d14	1.109	1.064	4.1	105	0.00
80	Butylbenzylphthalate	0.488	0.489	-0.2	106	0.00
81	Benzo(a)anthracene	1.387	1.381	0.4	106	0.00
82	3,3'-Dichlorobenzidine	0.501	0.503	-0.4	109	0.00
83	Chrysene	1.330	1.335	-0.4	108	0.00
84	Bis(2-ethylhexyl)phthalate	0.675	0.682	-1.0	107	0.00
85 c	Di-n-octyl phthalate	1.130	1.132	-0.2	105	0.00
86 I	Perylene-d12	1.000	1.000	0.0	108	0.00
87	Indeno(1,2,3-cd)pyrene	1.525	1.511	0.9	108	0.00
88	Benzo(b)fluoranthene	1.409	1.355	3.8	104	0.00
89	Benzo(k)fluoranthene	1.330	1.328	0.2	108	0.00
90 C	Benzo(a)pyrene	1.285	1.272	1.0	107	0.00
91	Dibenzo(a,h)anthracene	1.301	1.283	1.4	106	0.00
92	Benzo(g,h,i)perylene	1.262	1.235	2.1	106	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG032221\
Data File : BG048451.D
Acq On : 22 Mar 2021 19:54
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Mar 23 00:47:12 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG031821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 22 11:22:24 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range SPCC's out = 0 CCC's out = 0