

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG032415\
 Data File : BG016096.D
 Acq On : 24 Mar 2015 10:50
 Operator : TP/IZ
 Sample : PB82193BL
 Misc : SBLK36
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SBLK36

Quant Time: Mar 25 02:28:33 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM01.2-EPA-BG031715.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 25 02:17:41 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.79	152	97872	20.00	ng/ul	0.00
18) Naphthalene-d8	10.58	136	456631	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.42	164	293891	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.16	188	640406	20.00	ng/ul	0.00
75) Chrysene-d12	21.33	240	671642	20.00	ng/ul	0.00
83) Perylene-d12	23.61	264	666434	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	0.00	96	0	0.00	ng/uL	
5) Phenol-d5	6.95	99	234003	26.85	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.12	67	139115	28.64	ng/ul	0.00
9) 2-Chlorophenol-d4	7.32	132	183653	27.14	ng/ul	0.00
13) 4-Methylphenol-d8	8.48	113	188766	28.55	ng/ul	-0.01
19) Nitrobenzene-d5	8.94	128	107927	29.75	ng/ul	0.00
22) 2-Nitrophenol-d4	9.66	143	118235	34.01	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.19	165	168216	26.18	ng/ul	0.00
29) 4-Chloroaniline-d4	10.71	131	297070	33.53	ng/ul	0.00
43) Dimethylphthalate-d6	13.83	166	605822	31.62	ng/ul	0.00
46) Acenaphthylene-d8	14.11	160	733571	27.53	ng/ul	0.00
51) 4-Nitrophenol-d4	14.61	143	137963	30.15	ng/ul	0.00
57) Fluorene-d10	15.41	176	543244	28.35	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.52	200	87522	26.15	ng/ul	0.00
70) Anthracene-d10	17.26	188	833693	28.50	ng/ul	0.00
76) Pyrene-d10	19.54	212	899356	29.94	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.47	264	847657	29.16	ng/ul	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG032415\
Data File : BG016096.D
Acq On : 24 Mar 2015 10:50
Operator : TP/IZ
Sample : PB82193BL
Misc : SBLK36
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SBLK36

Quant Time: Mar 25 02:28:33 2015
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM01.2-EPA-BG031715.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Mar 25 02:17:41 2015
Response via : Initial Calibration

