

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG032416\
 Data File : BG021502.D
 Acq On : 25 Mar 2016 2:21
 Operator : UM/SJ
 Sample : H1909-04
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 A4T40

Quant Time: Mar 25 04:22:38 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG031616.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 25 01:20:42 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	0.00	152	0	0.00	ng/ul	-8.06
18) Naphthalene-d8	10.87	136	1113	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.67	164	784	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.41	188	1722	20.00	ng/ul	0.00
78) Chrysene-d12	21.67	240	2163	20.00	ng/ul	0.00
86) Perylene-d12	24.87	264	2445	20.00	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.46	96	731	0.00	ng/uL	0.00
5) Phenol-d5	7.34	99	20929	0.00	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.39	67	13206	0.00	ng/ul	0.00
9) 2-Chlorophenol-d4	7.64	132	16580	0.00	ng/ul	0.00
13) 4-Methylphenol-d8	8.85	113	16480	0.00	ng/ul	0.00
19) Nitrobenzene-d5	9.23	128	8954	1072.29	ng/ul	0.00
22) 2-Nitrophenol-d4	9.95	143	9779	1024.54	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.57	165	18075	1059.35	ng/ul	0.00
29) 4-Chloroaniline-d4	11.01	131	19504	874.39	ng/ul	0.00
44) Dimethylphthalate-d6	14.08	166	64010	980.54	ng/ul	0.00
47) Acenaphthylene-d8	14.38	160	80077	995.15	ng/ul	0.00
52) 4-Nitrophenol-d4	15.09	143	6056	580.39	ng/ul	0.02
58) Fluorene-d10	15.66	176	57152	971.03	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.79	200	8511	748.13	ng/ul	0.00
71) Anthracene-d10	17.51	188	87110	1032.29	ng/ul	0.00
79) Pyrene-d10	19.79	212	95829	950.07	ng/ul	0.00
90) Benzo(a)pyrene-d12	24.66	264	100290	888.74	ng/ul	0.00
Target Compounds						
21) Isophorone	9.95	82	990	21.77	ng/ul#	69
28) Naphthalene	10.92	128	2216	37.55	ng/ul#	77
34) 2-Methylnaphthalene	12.51	142	321	7.33	ng/ul	91
35) 1-Methylnaphthalene	12.74	142	288	6.78	ng/ul#	59
45) Dimethylphthalate	14.13	163	14432	212.76	ng/ul#	95
48) Acenaphthylene	14.40	152	1654	20.75	ng/ul	96
50) Acenaphthene	14.75	153	833	15.15	ng/ul#	85
54) Dibenzofuran	15.08	168	1045	13.57	ng/ul	89
57) Diethylphthalate	15.47	149	291	4.01	ng/ul#	59
59) Fluorene	15.72	166	1054	15.65	ng/ul#	54
70) Phenanthrene	17.45	178	25681	263.35	ng/ul	98
72) Anthracene	17.45	178	25681	258.81	ng/ul	98
75) Carbazole	17.82	167	1761	20.51	ng/ul	98
76) Di-n-butylphthalate	18.36	149	894	8.07	ng/ul#	77
77) Fluoranthene	19.46	202	55427	470.97	ng/ul#	93
80) Pyrene	19.82	202	46642	361.91	ng/ul#	94
81) Butylbenzylphthalate	20.69	149	225	4.13	ng/ul#	69
82) 3,3'-Dichlorobenzidine	21.71	252	72	1.61	ng/ul#	25
83) Benzo(a)anthracene	21.65	228	26255	201.08	ng/ul	98
84) Bis(2-ethylhexyl)phthalate	21.53	149	2588	32.51	ng/ul#	97
85) Chrysene	21.71	228	29949	242.19	ng/ul	97
87) Di-n-octyl phthalate	22.71	149	329	1.98	ng/ul	100
88) Benzo(b)fluoranthene	23.86	252	43594	277.58	ng/ul	99

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG032416\
Data File : BG021502.D
Acq On : 25 Mar 2016 2:21
Operator : UM/SJ
Sample : H1909-04
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
A4T40

Quant Time: Mar 25 04:22:38 2016
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG031616.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Mar 25 01:20:42 2016
Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
89) Benzo(k)fluoranthene	23.86	252	43594	282.58	ng/ul	99
91) Benzo(a)pyrene	24.73	252	27362	179.47	ng/ul	98
92) Indeno(1,2,3-cd)pyrene	28.53	276	16191	93.44	ng/ul	99
93) Dibenzo(a,h)anthracene	28.57	278	242	1.64	ng/ul#	81
94) Benzo(g,h,i)perylene	29.67	276	9130	64.08	ng/ul	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG032416\
Data File : BG021502.D
Acq On : 25 Mar 2016 2:21
Operator : UM/SJ
Sample : H1909-04
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
A4T40

Quant Time: Mar 25 04:22:38 2016
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG031616.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Mar 25 01:20:42 2016
Response via : Initial Calibration





#21

Isophorone

Concen: 21.77 ng/ul

RT: 9.95 min Scan# 1209

Delta R.T. 0.16 min

Lab File: BG021502.D

Acq: 25 Mar 2016 2:21

Instrument :

BNA_G

ClientSampleId :

A4T40


Tgt Ion: 82 Resp: 990

Ion Ratio Lower Upper




82 100

95 0.0 5.8 8.8#

138 0.0 12.2 18.2#



Abundance Ion 82.00 (81.70 to 82.70): BGC
Ion 95.00 (94.70 to 95.70): BGC
Ion 138.00 (137.70 to 138.70): BGC



#28

Naphthalene

Concen: 37.55 ng/ul

RT: 10.92 min Scan# 1375

Delta R.T. -0.00 min

Lab File: BG021502.D

Acq: 25 Mar 2016 2:21


Tgt Ion: 128 Resp: 2216

Ion Ratio Lower Upper



128 100


129 14.1 8.3 12.5#

127 0.0 10.8 16.2#



Abundance Ion 128.00 (127.70 to 128.70): BGC
Ion 129.00 (128.70 to 129.70): BGC
Ion 127.00 (126.70 to 127.70): BGC









#34
2-Methylnaphthalene
Concen: 7.33 ng/ul
RT: 12.51 min Scan# 1646
Delta R.T. -0.00 min
Lab File: BG021502.D
Acq: 25 Mar 2016 2:21




Instrument :
BNA_G
ClientSampleId :
A4T40


Tgt Ion:142 Resp: 321
Ion Ratio Lower Upper
142 100
141 80.6 71.3 106.9



#35
1-Methylnaphthalene
Concen: 6.78 ng/ul
RT: 12.74 min Scan# 1684
Delta R.T. 0.00 min
Lab File: BG021502.D
Acq: 25 Mar 2016 2:21

Tgt Ion:142 Resp: 288
Ion Ratio Lower Upper
142 100
141 54.6 77.0 115.4#
116 0.0 4.2 6.4#









#45
Dimethylphthalate
Concen: 212.76 ng/ul
RT: 14.13 min Scan# 1921
Delta R.T. -0.00 min
Lab File: BG021502.D
Acq: 25 Mar 2016 2:21

Instrument :
BNA_G
ClientSampleId :
A4T40

Tgt Ion:163 Resp: 14432
Ion Ratio Lower Upper
163 100
194 3.5 3.9 5.9#
164 9.2 8.7 13.1




Abundance Ion 163.00 (162.70 to 163.70): E
Ion 194.00 (193.70 to 194.70): E
Ion 164.00 (163.70 to 164.70): E






#48
Acenaphthylene
Concen: 20.75 ng/ul
RT: 14.40 min Scan# 1968
Delta R.T. 0.00 min
Lab File: BG021502.D
Acq: 25 Mar 2016 2:21

Tgt Ion:152 Resp: 1654
Ion Ratio Lower Upper
152 100
151 19.2 17.3 25.9
153 15.4 11.1 16.7



Abundance Ion 152.00 (151.70 to 152.70): E
Ion 151.00 (150.70 to 151.70): E
Ion 153.00 (152.70 to 153.70): E





#50

Acenaphthene

Concen: 15.15 ng/ul

RT: 14.75 min Scan# 2026

Delta R.T. 0.00 min

Lab File: BG021502.D

Acq: 25 Mar 2016 2:21

Instrument :

BNA_G

ClientSampleId :

A4T40


Tgt Ion:153 Resp: 833

Ion Ratio Lower Upper

153 100

152 60.8 36.5 54.7#

154 74.1 67.3 100.9



Abundance Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E

Ion 154.00 (153.70 to 154.70): E

14.75

400



300

200

100

0

Time--> 14.70 14.75



#54

Dibenzofuran

Concen: 13.57 ng/ul

RT: 15.08 min Scan# 2083

Delta R.T. 0.00 min

Lab File: BG021502.D


Acq: 25 Mar 2016 2:21

Tgt Ion:168 Resp: 1045

Ion Ratio Lower Upper

168 100

139 37.1 35.2 52.8



Abundance Ion 168.00 (167.70 to 168.70): E

Ion 139.00 (138.70 to 139.70): E

15.08


600


400

200

0

Time--> 15.05 15.10







#57
Diethylphthalate
Concen: 4.01 ng/ul
RT: 15.47 min Scan# 2150
Delta R.T. -0.01 min
Lab File: BG021502.D
Acq: 25 Mar 2016 2:21

Instrument :
BNA_G
ClientSampleId :
A4T40



Tgt Ion:149 Resp: 291
Ion Ratio Lower Upper
149 100
177 0.0 17.4 26.2#
150 0.0 10.2 15.2#



Abundance Ion 149.00 (148.70 to 149.70): E
Ion 177.00 (176.70 to 177.70): E
Ion 150.00 (149.70 to 150.70): E




Time--> 15.46 15.48




#59
Fluorene
Concen: 15.65 ng/ul
RT: 15.72 min Scan# 2192
Delta R.T. -0.00 min
Lab File: BG021502.D
Acq: 25 Mar 2016 2:21


Tgt Ion:166 Resp: 1054
Ion Ratio Lower Upper
166 100
165 53.6 81.1 121.7#
167 0.0 10.6 16.0#




Abundance Ion 166.00 (165.70 to 166.70): E
Ion 165.00 (164.70 to 165.70): E
Ion 167.00 (166.70 to 167.70): E



Time--> 15.70 15.72 15.74





#70

Phenanthrene

Concen: 263.35 ng/ul

RT: 17.45 min Scan# 2487

Delta R.T. -0.00 min

Lab File: BG021502.D

Acq: 25 Mar 2016 2:21

Instrument :

BNA_G

ClientSampleId :

A4T40


Tgt Ion:178 Resp: 25681

Ion Ratio Lower Upper

178 100

179 15.9 11.8 17.8




176 18.2 14.8 22.2



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 176.00 (175.70 to 176.70): E



#72

Anthracene

Concen: 258.81 ng/ul

RT: 17.45 min Scan# 2487

Delta R.T. -0.10 min

Lab File: BG021502.D

Acq: 25 Mar 2016 2:21


Tgt Ion:178 Resp: 25681

Ion Ratio Lower Upper

178 100

179 15.9 12.9 19.3



176 18.2 15.6 23.4




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 176.00 (175.70 to 176.70): E





#75

Carbazole

Concen: 20.51 ng/ul

RT: 17.82 min Scan# 2550

Delta R.T. -0.00 min

Lab File: BG021502.D

Acq: 25 Mar 2016 2:21

Instrument :

BNA_G

ClientSampleId :

A4T40


Tgt Ion:167 Resp: 1761

Ion Ratio Lower Upper

167 100

166 24.2 19.0 28.4


139 16.8 12.2 18.4





Abundance Ion 167.00 (166.70 to 167.70): E

Ion 166.00 (165.70 to 166.70): E

Ion 139.00 (138.70 to 139.70): E



Time-->



#76

Di-n-butylphthalate

Concen: 8.07 ng/ul

RT: 18.36 min Scan# 2641

Delta R.T. 0.00 min

Lab File: BG021502.D

Acq: 25 Mar 2016 2:21


Tgt Ion:149 Resp: 894

Ion Ratio Lower Upper

149 100

150 0.0 7.5 11.3#


104 0.0 5.4 8.2#




Abundance Ion 149.00 (148.70 to 149.70): E


Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E



Time-->









#77
Fluoranthene
 Concen: 470.97 ng/ul
 RT: 19.46 min Scan# 2828
 Delta R.T. -0.00 min
 Lab File: BG021502.D
 Acq: 25 Mar 2016 2:21




Instrument :
 BNA_G
 ClientSampleId :
 A4T40


Tgt Ion:202 Resp: 55427
 Ion Ratio Lower Upper
 202 100
 101 8.5 8.9 13.3#
 100 5.8 7.0 10.4#



#80
Pyrene
 Concen: 361.91 ng/ul
 RT: 19.82 min Scan# 2890
 Delta R.T. 0.00 min
 Lab File: BG021502.D
 Acq: 25 Mar 2016 2:21

Tgt Ion:202 Resp: 46642
 Ion Ratio Lower Upper
 202 100
 101 10.1 9.8 14.8
 100 8.9 9.3 13.9#









#81
Butylbenzylphthalate
Concen: 4.13 ng/ul
RT: 20.69 min Scan# 3038
Delta R.T. 0.00 min
Lab File: BG021502.D
Acq: 25 Mar 2016 2:21

Instrument :
BNA_G
ClientSampleId :
A4T40

Tgt Ion	Ratio	Lower	Upper
149	100		
91	56.4	65.8	98.8#
206	0.0	15.5	23.3#




Abundance Ion 149.00 (148.70 to 149.70): E
Ion 91.00 (90.70 to 91.70): BG
Ion 206.00 (205.70 to 206.70): E






#82
3,3'-Dichlorobenzidine
Concen: 1.61 ng/ul
RT: 21.71 min Scan# 3211
Delta R.T. 0.14 min
Lab File: BG021502.D
Acq: 25 Mar 2016 2:21

Tgt Ion	Ratio	Lower	Upper
252	100		
254	0.0	52.8	79.2#
126	0.0	9.8	14.8#



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 254.00 (253.70 to 254.70): E
Ion 126.00 (125.70 to 126.70): E









#83
Benzo(a)anthracene
Concen: 201.08 ng/ul
RT: 21.65 min Scan# 3201
Delta R.T. -0.00 min
Lab File: BG021502.D
Acq: 25 Mar 2016 2:21




Instrument :
BNA_G
ClientSampleId :
A4T40


Tgt Ion: 228 Resp: 26255
Ion Ratio Lower Upper
228 100
229 21.4 16.6 25.0
226 27.4 21.0 31.4



#84
Bis(2-ethylhexyl)phthalate
Concen: 32.51 ng/ul
RT: 21.53 min Scan# 3181
Delta R.T. -0.00 min
Lab File: BG021502.D
Acq: 25 Mar 2016 2:21

Tgt Ion: 149 Resp: 2588
Ion Ratio Lower Upper
149 100
167 26.2 20.5 30.7
279 0.0 3.7 5.5#









#85
 Chrysene
 Concen: 242.19 ng/ul
 RT: 21.71 min Scan# 3212
 Delta R.T. -0.00 min
 Lab File: BG021502.D
 Acq: 25 Mar 2016 2:21




Instrument :
 BNA_G
 ClientSampleId :
 A4T40


Tgt Ion: 228 Resp: 29949
 Ion Ratio Lower Upper
 228 100
 226 27.5 23.4 35.2
 229 21.2 16.2 24.2



#87
 Di-n-octyl phthalate
 Concen: 1.98 ng/ul
 RT: 22.71 min Scan# 3382
 Delta R.T. -0.02 min
 Lab File: BG021502.D
 Acq: 25 Mar 2016 2:21

Tgt Ion: 149 Resp: 329





#88

Benzo(b)fluoranthene

Concen: 277.58 ng/ul

RT: 23.86 min Scan# 3577

Delta R.T. 0.00 min

Lab File: BG021502.D

Acq: 25 Mar 2016 2:21

Instrument :

BNA_G

ClientSampleId :

A4T40


Tgt Ion:252 Resp: 43594

Ion Ratio Lower Upper

252 100

253 22.0 18.1 27.1




125 8.9 7.6 11.4



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#89

Benzo(k)fluoranthene

Concen: 282.58 ng/ul

RT: 23.86 min Scan# 3577

Delta R.T. -0.07 min

Lab File: BG021502.D

Acq: 25 Mar 2016 2:21


Tgt Ion:252 Resp: 43594

Ion Ratio Lower Upper

252 100

253 22.0 17.8 26.8



125 8.9 7.7 11.5




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#91

Benzo(a)pyrene

Concen: 179.47 ng/ul

RT: 24.73 min Scan# 3726

Delta R.T. 0.00 min

Lab File: BG021502.D

Acq: 25 Mar 2016 2:21

Instrument :

BNA_G

ClientSampleId :

A4T40


Tgt Ion:252 Resp: 27362

Ion Ratio Lower Upper

252 100

253 22.6 17.4 26.0




125 9.0 8.2 12.2



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#92

Indeno(1,2,3-cd)pyrene

Concen: 93.44 ng/ul

RT: 28.53 min Scan# 4372

Delta R.T. -0.01 min

Lab File: BG021502.D

Acq: 25 Mar 2016 2:21


Tgt Ion:276 Resp: 16191

Ion Ratio Lower Upper

276 100

138 20.1 15.8 23.6



277 23.2 19.0 28.6




Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E





#93

Dibenzo(a,h)anthracene

Concen: 1.64 ng/ul

RT: 28.57 min Scan# 4379

Delta R.T. -0.02 min

Lab File: BG021502.D

Acq: 25 Mar 2016 2:21

Instrument :

BNA_G

ClientSampleId :

A4T40


Tgt Ion:278 Resp: 242

Ion Ratio Lower Upper

278 100

139 0.0 12.1 18.1#




279 28.8 19.8 29.8



Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E



#94

Benzo(g,h,i)perylene

Concen: 64.08 ng/ul

RT: 29.67 min Scan# 4567

Delta R.T. -0.01 min

Lab File: BG021502.D

Acq: 25 Mar 2016 2:21


Tgt Ion:276 Resp: 9130

Ion Ratio Lower Upper

276 100

138 16.6 13.9 20.9

277 27.3 19.8 29.6



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E

