

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG032422\
 Data File : BG052808.D
 Acq On : 24 Mar 2022 14:26
 Operator : CG/JU
 Sample : PB143535BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PB143535BL

Quant Time: Mar 24 16:50:14 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG031522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 24 16:48:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.142	152	35214	20.000	ng	0.00
21) Naphthalene-d8	10.956	136	140526	20.000	ng	# 0.00
39) Acenaphthene-d10	14.763	164	105859	20.000	ng	0.00
64) Phenanthrene-d10	17.506	188	263988	20.000	ng	0.00
76) Chrysene-d12	21.789	240	295260	20.000	ng	0.00
86) Perylene-d12	25.102	264	284035	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.699	112	285684	141.285	ng	0.00
7) Phenol-d6	7.291	99	415342	136.248	ng	0.00
23) Nitrobenzene-d5	9.300	82	289808	103.508	ng	0.00
42) 2,4,6-Tribromophenol	16.243	330	213753	150.270	ng	0.00
45) 2-Fluorobiphenyl	13.388	172	663003	92.371	ng	0.00
79) Terphenyl-d14	20.109	244	1444931	87.053	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG032422\
Data File : BG052808.D
Acq On : 24 Mar 2022 14:26
Operator : CG/JU
Sample : PB143535BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB143535BL

Quant Time: Mar 24 16:50:14 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG031522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 24 16:48:57 2022
Response via : Initial Calibration

