

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG032422\
 Data File : BG052809.D
 Acq On : 24 Mar 2022 15:04
 Operator : CG/JU
 Sample : PB143570BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PB143570BL

Quant Time: Mar 24 16:50:31 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG031522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 24 16:48:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.144	152	32359	20.000	ng	0.00
21) Naphthalene-d8	10.957	136	129925	20.000	ng	# 0.00
39) Acenaphthene-d10	14.758	164	97827	20.000	ng	0.00
64) Phenanthrene-d10	17.501	188	236713	20.000	ng	0.00
76) Chrysene-d12	21.790	240	271146	20.000	ng	0.00
86) Perylene-d12	25.109	264	265409	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.700	112	266488	143.420	ng	0.00
7) Phenol-d6	7.292	99	387489	138.326	ng	0.00
23) Nitrobenzene-d5	9.301	82	272552	105.287	ng	0.00
42) 2,4,6-Tribromophenol	16.244	330	189782	144.372	ng	0.00
45) 2-Fluorobiphenyl	13.389	172	607296	91.556	ng	0.00
79) Terphenyl-d14	20.110	244	1332309	87.406	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG032422\
Data File : BG052809.D
Acq On : 24 Mar 2022 15:04
Operator : CG/JU
Sample : PB143570BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB143570BL

Quant Time: Mar 24 16:50:31 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG031522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 24 16:48:57 2022
Response via : Initial Calibration

