

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG032422\
 Data File : BG052807.D
 Acq On : 24 Mar 2022 13:42
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 24 16:49:34 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG031522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 24 16:48:57 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	100	0.00
2	1,4-Dioxane	40.000	37.007	7.5	99	0.00
3	Pyridine	40.000	42.041	-5.1	106	0.00
4	n-Nitrosodimethylamine	40.000	42.572	-6.4	110	0.00
5 S	2-Fluorophenol	80.000	78.116	2.4	103	0.00
6	Aniline	40.000	41.429	-3.6	107	0.00
7 S	Phenol-d6	80.000	80.186	-0.2	103	0.00
8	2-Chlorophenol	40.000	40.435	-1.1	103	0.00
9	Benzaldehyde	40.000	39.951	0.1	106	0.00
10 C	Phenol	40.000	40.867	-2.2	107	0.00
11	bis(2-Chloroethyl)ether	40.000	39.585	1.0	105	0.00
12	1,3-Dichlorobenzene	40.000	38.161	4.6	100	0.00
13 C	1,4-Dichlorobenzene	40.000	38.500	3.8	101	0.00
14	1,2-Dichlorobenzene	40.000	38.294	4.3	101	0.00
15	Benzyl Alcohol	40.000	41.837	-4.6	109	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.667	-1.7	106	0.00
17	2-Methylphenol	40.000	39.909	0.2	102	0.00
18	Hexachloroethane	40.000	40.560	-1.4	105	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.275	-3.2	108	0.00
20	3+4-Methylphenols	40.000	41.433	-3.6	107	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	105	0.00
22	Acetophenone	40.000	39.333	1.7	106	0.00
23 S	Nitrobenzene-d5	80.000	90.768	-13.5	119	0.00
24	Nitrobenzene	40.000	43.952	-9.9	116	0.00
25	Isophorone	40.000	40.610	-1.5	109	0.00
26 C	2-Nitrophenol	40.000	50.994	-27.5#	135	0.00
27	2,4-Dimethylphenol	40.000	38.512	3.7	107	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.842	0.4	106	0.00
29 C	2,4-Dichlorophenol	40.000	42.291	-5.7	112	0.00
30	1,2,4-Trichlorobenzene	40.000	40.103	-0.3	105	0.00
31	Naphthalene	40.000	39.217	2.0	105	0.00
32	Benzoic acid	40.000	43.468	-8.7	127	0.00
33	4-Chloroaniline	40.000	40.165	-0.4	106	0.00
34 C	Hexachlorobutadiene	40.000	40.031	-0.1	107	0.00
35	Caprolactam	40.000	54.653	-36.6#	138	0.00
36 C	4-Chloro-3-methylphenol	40.000	42.775	-6.9	111	0.00
37	2-Methylnaphthalene	40.000	39.579	1.1	107	0.00
38	1-Methylnaphthalene	40.000	39.964	0.1	108	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	102	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.994	0.0	107	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.657	0.9	106	0.00
42 S	2,4,6-Tribromophenol	80.000	86.067	-7.6	112	0.00
43 C	2,4,6-Trichlorophenol	40.000	45.318	-13.3	121	0.00
44	2,4,5-Trichlorophenol	40.000	46.634	-16.6	124	0.00
45 S	2-Fluorobiphenyl	80.000	80.606	-0.8	108	0.00
46	1,1'-Biphenyl	40.000	39.844	0.4	107	0.00
47	2-Chloronaphthalene	40.000	39.746	0.6	107	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG032422\
 Data File : BG052807.D
 Acq On : 24 Mar 2022 13:42
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 24 16:49:34 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG031522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 24 16:48:57 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	55.028	-37.6#	138	0.00
49	Acenaphthylene	40.000	39.953	0.1	107	0.00
50	Dimethylphthalate	40.000	40.171	-0.4	106	0.00
51	2,6-Dinitrotoluene	40.000	48.602	-21.5	119	0.00
52 C	Acenaphthene	40.000	42.100	-5.3	113	0.00
53	3-Nitroaniline	40.000	46.533	-16.3	116	0.00
54 P	2,4-Dinitrophenol	40.000	56.955	-42.4#	157	0.00
55	Dibenzofuran	40.000	39.885	0.3	106	0.00
56 P	4-Nitrophenol	40.000	47.180	-17.9	116	0.00
57	2,4-Dinitrotoluene	40.000	50.586	-26.5#	125	0.00
58	Fluorene	40.000	39.145	2.1	106	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	44.629	-11.6	113	0.00
60	Diethylphthalate	40.000	39.704	0.7	104	0.00
61	4-Chlorophenyl-phenylether	40.000	39.433	1.4	106	0.00
62	4-Nitroaniline	40.000	45.567	-13.9	117	0.00
63	Azobenzene	40.000	39.461	1.3	105	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	100	0.00
65	4,6-Dinitro-2-methylphenol	40.000	54.099	-35.2#	137	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.382	1.5	103	0.00
67	4-Bromophenyl-phenylether	40.000	40.316	-0.8	105	0.00
68	Hexachlorobenzene	40.000	40.348	-0.9	104	0.00
69	Atrazine	40.000	41.631	-4.1	104	0.00
70 C	Pentachlorophenol	40.000	43.034	-7.6	106	0.00
71	Phenanthrene	40.000	39.522	1.2	103	0.00
72	Anthracene	40.000	38.934	2.7	102	0.00
73	Carbazole	40.000	38.595	3.5	101	0.00
74	Di-n-butylphthalate	40.000	40.080	-0.2	101	0.00
75 C	Fluoranthene	40.000	39.117	2.2	102	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	101	0.00
77	Benzydine	40.000	27.622	30.9#	73	0.00
78	Pyrene	40.000	38.969	2.6	101	0.00
79 S	Terphenyl-d14	80.000	77.241	3.4	99	0.00
80	Butylbenzylphthalate	40.000	41.297	-3.2	104	0.00
81	Benzo(a)anthracene	40.000	39.178	2.1	103	0.00
82	3,3'-Dichlorobenzidine	40.000	39.469	1.3	102	0.00
83	Chrysene	40.000	39.049	2.4	102	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.619	-4.0	107	0.00
85 c	Di-n-octyl phthalate	40.000	42.440	-6.1	108	0.00
86 I	Perylene-d12	20.000	20.000	0.0	106	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.389	-3.5	115	0.00
88	Benzo(b)fluoranthene	40.000	38.368	4.1	106	0.00
89	Benzo(k)fluoranthene	40.000	39.252	1.9	109	0.00
90 C	Benzo(a)pyrene	40.000	39.552	1.1	108	0.00
91	Dibenzo(a,h)anthracene	40.000	41.102	-2.8	114	0.00
92	Benzo(g,h,i)perylene	40.000	40.898	-2.2	113	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG032422\
Data File : BG052807.D
Acq On : 24 Mar 2022 13:42
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Mar 24 16:49:34 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG031522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 24 16:48:57 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 1