

Data Path : Z:\HPCHEM1\BNA\_G\DATA\BG032515\  
 Data File : BG016115.D  
 Acq On : 25 Mar 2015 22:35  
 Operator : TP/IZ  
 Sample : SSTDCCC040  
 Misc :  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 LabSampleId :  
 SSTDCCC040

Quant Time: Mar 26 15:57:49 2015  
 Quant Method : Z:\HPCHEM1\BNA\_G\METHODS\8270-BG032515.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Mar 26 15:24:59 2015  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound                    | Amount | Calc.  | %Dev | Area% | Dev(min) |
|------|-----------------------------|--------|--------|------|-------|----------|
| 1 I  | 1,4-Dichlorobenzene-d4      | 20.000 | 20.000 | 0.0  | 104   | 0.00     |
| 2    | 1,4-Dioxane                 | 40.000 | 40.025 | -0.1 | 105   | 0.00     |
| 3    | Pyridine                    | 40.000 | 39.500 | 1.3  | 102   | 0.00     |
| 4    | n-Nitrosodimethylamine      | 40.000 | 39.460 | 1.3  | 103   | 0.00     |
| 5 S  | 2-Fluorophenol              | 80.000 | 80.392 | -0.5 | 105   | 0.00     |
| 6    | Aniline                     | 40.000 | 39.812 | 0.5  | 103   | 0.00     |
| 7 S  | Phenol-d6                   | 80.000 | 81.253 | -1.6 | 105   | 0.00     |
| 8    | 2-Chlorophenol              | 40.000 | 40.413 | -1.0 | 106   | 0.00     |
| 9    | Benzaldehyde                | 40.000 | 42.422 | -6.1 | 106   | 0.00     |
| 10 C | Phenol                      | 40.000 | 40.663 | -1.7 | 105   | 0.00     |
| 11   | bis(2-Chloroethyl)ether     | 40.000 | 39.477 | 1.3  | 104   | 0.00     |
| 12   | 1,3-Dichlorobenzene         | 40.000 | 40.212 | -0.5 | 105   | 0.00     |
| 13 C | 1,4-Dichlorobenzene         | 40.000 | 40.448 | -1.1 | 106   | 0.00     |
| 14   | 1,2-Dichlorobenzene         | 40.000 | 39.982 | 0.0  | 105   | 0.00     |
| 15   | Benzyl Alcohol              | 40.000 | 40.368 | -0.9 | 103   | 0.00     |
| 16   | 2,2'-oxybis(1-Chloropropane | 40.000 | 38.069 | 4.8  | 100   | 0.00     |
| 17   | 2-Methylphenol              | 40.000 | 40.597 | -1.5 | 106   | 0.00     |
| 18   | Hexachloroethane            | 40.000 | 40.032 | -0.1 | 106   | 0.00     |
| 19 P | n-Nitroso-di-n-propylamine  | 40.000 | 38.976 | 2.6  | 103   | 0.00     |
| 20   | 3+4-Methylphenols           | 40.000 | 40.891 | -2.2 | 105   | 0.00     |
| 21 I | Naphthalene-d8              | 20.000 | 20.000 | 0.0  | 104   | 0.00     |
| 22   | Acetophenone                | 40.000 | 39.623 | 0.9  | 104   | 0.00     |
| 23 S | Nitrobenzene-d5             | 80.000 | 79.779 | 0.3  | 104   | 0.00     |
| 24   | Nitrobenzene                | 40.000 | 40.170 | -0.4 | 106   | 0.00     |
| 25   | Isophorone                  | 40.000 | 39.027 | 2.4  | 102   | 0.00     |
| 26 C | 2-Nitrophenol               | 40.000 | 40.626 | -1.6 | 103   | 0.00     |
| 27   | 2,4-Dimethylphenol          | 40.000 | 40.322 | -0.8 | 106   | 0.00     |
| 28   | bis(2-Chloroethoxy)methane  | 40.000 | 39.621 | 0.9  | 104   | 0.00     |
| 29 C | 2,4-Dichlorophenol          | 40.000 | 40.779 | -1.9 | 105   | 0.00     |
| 30   | 1,2,4-Trichlorobenzene      | 40.000 | 40.255 | -0.6 | 105   | 0.00     |
| 31   | Naphthalene                 | 40.000 | 39.614 | 1.0  | 104   | 0.00     |
| 32   | Benzoic acid                | 40.000 | 39.316 | 1.7  | 106   | 0.00     |
| 33   | 4-Chloroaniline             | 40.000 | 39.724 | 0.7  | 103   | 0.00     |
| 34 C | Hexachlorobutadiene         | 40.000 | 40.553 | -1.4 | 106   | 0.00     |
| 35   | Caprolactam                 | 40.000 | 39.380 | 1.5  | 102   | 0.00     |
| 36 C | 4-Chloro-3-methylphenol     | 40.000 | 39.952 | 0.1  | 104   | 0.00     |
| 37   | 2-Methylnaphthalene         | 40.000 | 39.692 | 0.8  | 104   | 0.00     |
| 38 I | Acenaphthene-d10            | 20.000 | 20.000 | 0.0  | 104   | 0.00     |
| 39   | 1,2,4,5-Tetrachlorobenzene  | 40.000 | 40.419 | -1.0 | 104   | 0.00     |
| 40 P | Hexachlorocyclopentadiene   | 40.000 | 41.945 | -4.9 | 105   | 0.00     |
| 41 S | 2,4,6-Tribromophenol        | 80.000 | 81.820 | -2.3 | 104   | 0.00     |
| 42 C | 2,4,6-Trichlorophenol       | 40.000 | 41.757 | -4.4 | 107   | 0.00     |
| 43   | 2,4,5-Trichlorophenol       | 40.000 | 41.343 | -3.4 | 106   | 0.00     |
| 44 S | 2-Fluorobiphenyl            | 80.000 | 80.853 | -1.1 | 105   | 0.00     |

Data Path : Z:\HPCHEM1\BNA\_G\DATA\BG032515\  
 Data File : BG016115.D  
 Acq On : 25 Mar 2015 22:35  
 Operator : TP/IZ  
 Sample : SSTDCCC040  
 Misc :  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 LabSampleId :  
 SSTDCCC040

Quant Time: Mar 26 15:57:49 2015  
 Quant Method : Z:\HPCHEM1\BNA\_G\METHODS\8270-BG032515.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Mar 26 15:24:59 2015  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound                   | Amount | Calc.  | %Dev | Area% | Dev(min) |
|------|----------------------------|--------|--------|------|-------|----------|
| 45   | 1,1'-Biphenyl              | 40.000 | 40.227 | -0.6 | 105   | 0.00     |
| 46   | 2-Chloronaphthalene        | 40.000 | 40.214 | -0.5 | 105   | 0.00     |
| 47   | 2-Nitroaniline             | 40.000 | 40.324 | -0.8 | 105   | 0.00     |
| 48   | Acenaphthylene             | 40.000 | 40.308 | -0.8 | 104   | 0.00     |
| 49   | Dimethylphthalate          | 40.000 | 40.115 | -0.3 | 106   | 0.00     |
| 50   | 2,6-Dinitrotoluene         | 40.000 | 41.834 | -4.6 | 107   | 0.00     |
| 51 C | Acenaphthene               | 40.000 | 40.309 | -0.8 | 106   | 0.00     |
| 52   | 3-Nitroaniline             | 40.000 | 39.372 | 1.6  | 101   | 0.00     |
| 53 P | 2,4-Dinitrophenol          | 40.000 | 37.831 | 5.4  | 99    | 0.00     |
| 54   | Dibenzofuran               | 40.000 | 39.964 | 0.1  | 105   | 0.00     |
| 55 P | 4-Nitrophenol              | 40.000 | 41.758 | -4.4 | 104   | 0.00     |
| 56   | 2,4-Dinitrotoluene         | 40.000 | 41.171 | -2.9 | 105   | 0.00     |
| 57   | Fluorene                   | 40.000 | 40.263 | -0.7 | 104   | 0.00     |
| 58   | 2,3,4,6-Tetrachlorophenol  | 40.000 | 40.996 | -2.5 | 107   | 0.00     |
| 59   | Diethylphthalate           | 40.000 | 40.017 | -0.0 | 105   | 0.00     |
| 60   | 4-Chlorophenyl-phenylether | 40.000 | 39.801 | 0.5  | 104   | 0.00     |
| 61   | 4-Nitroaniline             | 40.000 | 40.236 | -0.6 | 102   | 0.00     |
| 62   | Azobenzene                 | 40.000 | 39.283 | 1.8  | 103   | 0.00     |
| 63 I | Phenanthrene-d10           | 20.000 | 20.000 | 0.0  | 103   | 0.00     |
| 64   | 4,6-Dinitro-2-methylphenol | 40.000 | 40.696 | -1.7 | 103   | 0.00     |
| 65 c | n-Nitrosodiphenylamine     | 40.000 | 40.159 | -0.4 | 104   | 0.00     |
| 66   | 4-Bromophenyl-phenylether  | 40.000 | 39.701 | 0.7  | 104   | 0.00     |
| 67   | Hexachlorobenzene          | 40.000 | 39.403 | 1.5  | 103   | 0.00     |
| 68   | Atrazine                   | 40.000 | 39.448 | 1.4  | 103   | 0.00     |
| 69 C | Pentachlorophenol          | 40.000 | 41.387 | -3.5 | 102   | 0.00     |
| 70   | Phenanthrene               | 40.000 | 39.670 | 0.8  | 104   | 0.00     |
| 71   | Anthracene                 | 40.000 | 39.531 | 1.2  | 103   | 0.00     |
| 72   | Carbazole                  | 40.000 | 39.742 | 0.6  | 102   | 0.00     |
| 73   | Di-n-butylphthalate        | 40.000 | 39.614 | 1.0  | 101   | 0.00     |
| 74 C | Fluoranthene               | 40.000 | 40.108 | -0.3 | 103   | 0.00     |
| 75 I | Chrysene-d12               | 20.000 | 20.000 | 0.0  | 101   | 0.00     |
| 76   | Benzidine                  | 40.000 | 40.455 | -1.1 | 98    | 0.00     |
| 77   | Pyrene                     | 40.000 | 40.245 | -0.6 | 102   | 0.00     |
| 78 S | Terphenyl-d14              | 80.000 | 80.840 | -1.1 | 101   | 0.00     |
| 79   | Butylbenzylphthalate       | 40.000 | 41.505 | -3.8 | 103   | 0.00     |
| 80   | Benzo(a)anthracene         | 40.000 | 39.978 | 0.1  | 101   | 0.00     |
| 81   | 3,3'-Dichlorobenzidine     | 40.000 | 40.051 | -0.1 | 100   | 0.00     |
| 82   | Chrysene                   | 40.000 | 39.815 | 0.5  | 101   | 0.00     |
| 83   | Bis(2-ethylhexyl)phthalate | 40.000 | 40.187 | -0.5 | 101   | 0.00     |
| 84 c | Di-n-octyl phthalate       | 40.000 | 40.031 | -0.1 | 100   | 0.00     |
| 85   | Indeno(1,2,3-cd)pyrene     | 40.000 | 39.828 | 0.4  | 101   | 0.00     |
| 86 I | Perylene-d12               | 20.000 | 20.000 | 0.0  | 101   | 0.00     |
| 87   | Benzo(b)fluoranthene       | 40.000 | 39.004 | 2.5  | 100   | 0.00     |

Data Path : Z:\HPCHEM1\BNA\_G\DATA\BG032515\  
Data File : BG016115.D  
Acq On : 25 Mar 2015 22:35  
Operator : TP/IZ  
Sample : SSTDCCC040  
Misc :  
ALS Vial : 5 Sample Multiplier: 1

Instrument :  
BNA\_G  
LabSampleId :  
SSTDCCC040

Quant Time: Mar 26 15:57:49 2015  
Quant Method : Z:\HPCHEM1\BNA\_G\METHODS\8270-BG032515.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Mar 26 15:24:59 2015  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound               | Amount | Calc.  | %Dev | Area% | Dev(min) |
|------|------------------------|--------|--------|------|-------|----------|
| 88   | Benzo(k)fluoranthene   | 40.000 | 40.977 | -2.4 | 103   | 0.00     |
| 89 C | Benzo(a)pyrene         | 40.000 | 39.999 | 0.0  | 101   | 0.00     |
| 90   | Dibenzo(a,h)anthracene | 40.000 | 40.253 | -0.6 | 102   | 0.00     |
| 91   | Benzo(g,h,i)perylene   | 40.000 | 40.037 | -0.1 | 102   | 0.00     |

(#) = Out of Range

SPCC's out = 0 CCC's out = 0