

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG032515\
 Data File : BG016099.D
 Acq On : 25 Mar 2015 13:22
 Operator : TP/IZ
 Sample : SSTDICC2.5
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Manual Integrations
 APPROVED

apatel
 3/27/2015 10:47:18 AM

Quant Time: Mar 26 15:19:26 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG032515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 26 01:24:59 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.79	152	50490	20.00	ng	0.00
21) Naphthalene-d8	10.57	136	228955	20.00	ng	0.00
38) Acenaphthene-d10	14.42	164	148155	20.00	ng	0.00
63) Phenanthrene-d10	17.15	188	323431	20.00	ng	0.00
75) Chrysene-d12	21.32	240	342769	20.00	ng	0.00
86) Perylene-d12	23.61	264	328534	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	5.38	112	16491	5.43	ng	0.00
7) Phenol-d6	6.95	99	21778	4.79	ng	-0.01
23) Nitrobenzene-d5	8.94	82	19729	3.22	ng	0.00
41) 2,4,6-Tribromophenol	15.90	330	8196	2.78	ng	0.00
44) 2-Fluorobiphenyl	13.04	172	48993	5.22	ng	0.00
78) Terphenyl-d14	19.77	244	76840	6.01	ng	0.00

Target Compounds

						Qvalue
6) Aniline	7.12	93	15336	2.39	ng	93
8) 2-Chlorophenol	7.35	128	9620	3.11	ng	98
10) Phenol	6.98	94	11894	2.28	ng	98
11) bis(2-Chloroethyl)ether	7.21	93	10324	2.70	ng	98
12) 1,3-Dichlorobenzene	7.68	146	10582	2.69	ng	97
13) 1,4-Dichlorobenzene	7.82	146	10742	2.64	ng	91
14) 1,2-Dichlorobenzene	8.14	146	10167	2.69	ng	96
16) 2,2'-oxybis(1-Chloropropan	8.32	45	11316	3.73	ng	97
17) 2-Methylphenol	8.22	107	8540	2.44	ng	# 93
18) Hexachloroethane	8.86	117	3936	2.20	ng	93
19) n-Nitroso-di-n-propylamine	8.59	70	8262	2.07	ng	# 96
20) 3+4-Methylphenols	8.55	107	11078	2.34	ng	93
22) Acetophenone	8.61	105	14014	2.13	ng	# 98
26) 2-Nitrophenol	9.70	139	5317	2.50	ng	98
27) 2,4-Dimethylphenol	9.75	122	9457	2.50	ng	93
28) bis(2-Chloroethoxy)methane	9.99	93	13093	2.32	ng	99
31) Naphthalene	10.63	128	34445	2.88	ng	98
33) 4-Chloroaniline	10.74	127	14146	2.74	ng	98
35) Caprolactam	11.47	113	4574	2.36	ng	90
37) 2-Methylnaphthalene	12.24	142	23898	2.56	ng	98
45) 1,1'-Biphenyl	13.25	154	29466	3.16	ng	99
46) 2-Chloronaphthalene	13.29	162	22731	2.85	ng	96
47) 2-Nitroaniline	13.49	65	6278	2.13	ng	95
48) Acenaphthylene	14.13	152	42754	3.17	ng	98
49) Dimethylphthalate	13.87	163	29002	2.60	ng	# 97
50) 2,6-Dinitrotoluene	13.99	165	6019	2.34	ng	92
51) Acenaphthene	14.48	154	25912	3.21	ng	96
52) 3-Nitroaniline	14.32	138	7794	3.31	ng	88
54) Dibenzofuran	14.82	168	36692	2.89	ng	95
55) 4-Nitrophenol	14.62	139	5388	3.01	ng	93
56) 2,4-Dinitrotoluene	14.77	165	8146	2.13	ng	# 94
57) Fluorene	15.46	166	32170	2.75	ng	97
59) Diethylphthalate	15.23	149	29824	2.55	ng	98
61) 4-Nitroaniline	15.47	138	8587	3.02	ng	96

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG032515\
 Data File : BG016099.D
 Acq On : 25 Mar 2015 13:22
 Operator : TP/IZ
 Sample : SSTDICC2.5
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Manual Integrations
 APPROVED


apatel
 3/27/2015 10:47:18 AM

Quant Time: Mar 26 15:19:26 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG032515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 26 01:24:59 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)

62) Azobenzene	15.74	77	30160	2.12	ng	95
65) n-Nitrosodiphenylamine	15.67	169	28001	2.99	ng	97
68) Atrazine	16.61	200	8515	2.19	ng	92
70) Phenanthrene	17.19	178	51994	3.03	ng	96
71) Anthracene	17.28	178	50681	2.98	ng	99
72) Carbazole	17.55	167	49255	3.21	ng	97
73) Di-n-butylphthalate	18.12	149	55294	3.16	ng	98
74) Fluoranthene	19.21	202	55856	2.33	ng	96
76) Benzidine	19.39	184	24318	3.31	ng	99
77) Pyrene	19.57	202	58451	3.56	ng	96
79) Butylbenzylphthalate	20.47	149	22025	3.62	ng	94
80) Benzo(a)anthracene	21.31	228	54706	2.91	ng	97
81) 3,3'-Dichlorobenzidine	21.24	252	17507	2.32	ng	100
82) Chrysene	21.36	228	50462	2.95	ng	97
83) Bis(2-ethylhexyl)phthalate	21.24	149	31677	3.72	ng	97
84) Di-n-octyl phthalate	22.12	149	54080	3.94	ng	100
85) Indeno(1,2,3-cd)pyrene	25.93	276	60260	2.49	ng	99
87) Benzo(b)fluoranthene	22.91	252	51434	2.67	ng	99
88) Benzo(k)fluoranthene	22.96	252	51594	2.82	ng	98
89) Benzo(a)pyrene	23.50	252	47797	2.64	ng	99
90) Dibenzo(a,h)anthracene	25.95	278	49075	2.56	ng	92
91) Benzo(g,h,i)perylene	26.66	276	50005	2.68	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1


1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 7.79 min Scan# 868
Delta R.T. -0.01 min
Lab File: BG016099.D
Acq: 25 Mar 2015 13:22

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5




Tgt Ion	Ratio	Resp	Lower	Upper
152	100	50490		
150	154.9	126.1	189.1	
115	55.0	43.7	65.5	

Manual Integrations
APPROVED

apatel
3/27/2015 10:47:18 AM




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
Concen: 5.43 ng
RT: 5.38 min Scan# 457
Delta R.T. 0.00 min
Lab File: BG016099.D
Acq: 25 Mar 2015 13:22

Tgt Ion	Ratio	Resp	Lower	Upper
112	100	16491		
64	48.3	42.6	64.0	
63	24.5	19.9	29.9	



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BGC
Ion 63.00 (62.70 to 63.70): BGC





#6

Aniline

Concen: 2.39 ng

RT: 7.12 min Scan# 754

Delta R.T. -0.01 min

Lab File: BG016099.D

Acq: 25 Mar 2015 13:22

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tgt Ion:	93	Resp:	15336
Ion	Ratio	Lower	Upper
93	100		
66	38.2	27.8	41.6
65	20.8	13.9	20.9

Manual Integrations
APPROVED

apatel
3/27/2015 10:47:18 AM



#7

Phenol-d6

Concen: 4.79 ng




RT: 6.95 min Scan# 725


Delta R.T. -0.01 min

Lab File: BG016099.D

Acq: 25 Mar 2015 13:22

Tgt Ion:	99	Resp:	21778
Ion	Ratio	Lower	Upper
99	100		
42	12.0	9.4	14.0
71	26.2	19.9	29.9





#8

2-Chlorophenol

Concen: 3.11 ng

RT: 7.35 min Scan# 794

Delta R.T. 0.00 min

Lab File: BG016099.D

Acq: 25 Mar 2015 13:22

Instrument :

BNA_G


ClientSampleId :

SSTDICC2.5




Tgt Ion:	128	Resp:	9620
Ion	Ratio	Lower	Upper
128	100		
130	34.4	12.5	52.5
64	43.7	23.4	63.4

Manual Integrations
APPROVED

apatel
3/27/2015 10:47:18 AM



Abundance Ion 128.00 (127.70 to 128.70): BGC
Ion 130.00 (129.70 to 130.70): BGC
Ion 64.00 (63.70 to 64.70): BGC



#10

Phenol

Concen: 2.28 ng


RT: 6.98 min Scan# 730

Delta R.T. -0.01 min



Lab File: BG016099.D


Acq: 25 Mar 2015 13:22

Tgt Ion:	94	Resp:	11894
Ion	Ratio	Lower	Upper
94	100		
65	23.3	4.0	44.0
66	28.6	7.4	47.4



Abundance Ion 94.00 (93.70 to 94.70): BGC
Ion 65.00 (64.70 to 65.70): BGC
Ion 66.00 (65.70 to 66.70): BGC









#11
 bis(2-Chloroethyl)ether
 Concen: 2.70 ng
 RT: 7.21 min Scan# 770
 Delta R.T. -0.01 min
 Lab File: BG016099.D
 Acq: 25 Mar 2015 13:22

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Tgt Ion:	93	Resp:	10324
Ion	Ratio	Lower	Upper
93	100		
63	63.9	46.5	86.5
95	32.4	12.5	52.5




Manual Integrations
 APPROVED


apatel
 3/27/2015 10:47:18 AM



#12
 1,3-Dichlorobenzene
 Concen: 2.69 ng
 RT: 7.68 min Scan# 849
 Delta R.T. -0.01 min
 Lab File: BG016099.D
 Acq: 25 Mar 2015 13:22

Tgt Ion:	146	Resp:	10582
Ion	Ratio	Lower	Upper
146	100		
148	61.6	51.2	76.8
75	24.8	20.5	30.7









#13
 1,4-Dichlorobenzene
 Concen: 2.64 ng
 RT: 7.82 min Scan# 874
 Delta R.T. -0.01 min
 Lab File: BG016099.D
 Acq: 25 Mar 2015 13:22

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
146	100		
148	71.2	50.3	75.5
111	40.4	29.5	44.3




Manual Integrations
 APPROVED


apatel
 3/27/2015 10:47:18 AM



#14
 1,2-Dichlorobenzene
 Concen: 2.69 ng
 RT: 8.14 min Scan# 928
 Delta R.T. 0.00 min
 Lab File: BG016099.D
 Acq: 25 Mar 2015 13:22

Tgt Ion	Ratio	Lower	Upper
146	100		
148	69.4	51.8	77.8
111	39.1	31.4	47.2









#16
2,2'-oxybis(1-Chloropropane)
Concen: 3.73 ng
RT: 8.32 min Scan# 959
Delta R.T. -0.01 min
Lab File: BG016099.D
Acq: 25 Mar 2015 13:22

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion	Ratio	Resp	Lower	Upper
45	100	11316		
77	15.1	0.0	36.3	
79	12.2	0.0	33.3	




Manual Integrations
APPROVED


apatel
3/27/2015 10:47:18 AM



#17
2-Methylphenol
Concen: 2.44 ng
RT: 8.22 min Scan# 942
Delta R.T. -0.01 min
Lab File: BG016099.D
Acq: 25 Mar 2015 13:22

Tgt Ion	Ratio	Resp	Lower	Upper
107	100	8540		
108	106.6	89.4	134.2	
77	35.2	31.4	47.2	
79	29.4	29.8	44.8	#









#18
Hexachloroethane
Concen: 2.20 ng
RT: 8.86 min Scan# 1051
Delta R.T. 0.00 min
Lab File: BG016099.D
Acq: 25 Mar 2015 13:22

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion	Ratio	Resp	Lower	Upper
117	100	3936		
119	84.0		76.4	114.6
201	88.1		68.6	102.8




Manual Integrations
APPROVED


apatel
3/27/2015 10:47:18 AM



#19
n-Nitroso-di-n-propylamine
Concen: 2.07 ng
RT: 8.59 min Scan# 1004
Delta R.T. -0.01 min
Lab File: BG016099.D
Acq: 25 Mar 2015 13:22

Tgt Ion	Ratio	Resp	Lower	Upper
70	100	8262		
42	40.2		33.4	50.2
101	15.6		9.6	14.4
130	31.9		23.3	34.9






#20
3+4-Methylphenols
Concen: 2.34 ng
RT: 8.55 min Scan# 997
Delta R.T. -0.01 min
Lab File: BG016099.D
Acq: 25 Mar 2015 13:22

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion	Ratio	Resp	Lower	Upper
107	100	11078		
108	80.0	68.4	108.4	
77	27.7	6.9	46.9	
79	24.5	2.3	42.3	

Manual Integrations
APPROVED

apatel
3/27/2015 10:47:18 AM






Abundance

Ion 107.00 (106.70 to 107.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 77.00 (76.70 to 77.70): BGC

Ion 79.00 (78.70 to 79.70): BGC



#21
Naphthalene-d8
Concen: 20.00 ng
RT: 10.57 min Scan# 1342
Delta R.T. -0.01 min
Lab File: BG016099.D
Acq: 25 Mar 2015 13:22

Tgt Ion	Ratio	Resp	Lower	Upper
136	100	228955		
137	10.9	9.0	13.4	
54	7.5	5.5	8.3	
68	7.2	6.0	9.0	




Abundance


Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BGC

Ion 68.00 (67.70 to 68.70): BGC





#22

Acetophenone

Concen: 2.13 ng

RT: 8.61 min Scan# 1007

Delta R.T. -0.01 min

Lab File: BG016099.D


Acq: 25 Mar 2015 13:22

Instrument :

BNA_G

ClientSampleId :

SSTDICC2.5



Tgt Ion: 105 Resp: 14014

Ion Ratio Lower Upper

105 100


71 2.1 2.2 3.2#

51 22.3 17.9 26.9

120 35.8 27.0 40.4

Manual Integrations
APPROVED

apatel
3/27/2015 10:47:18 AM





Abundance Ion 105.00 (104.70 to 105.70): E

Ion 71.00 (70.70 to 71.70): BGC

Ion 51.00 (50.70 to 51.70): BGC

Ion 120.00 (119.70 to 120.70): E



#23

Nitrobenzene-d5

Concen: 3.22 ng

RT: 8.94 min Scan# 1064

Delta R.T. -0.01 min

Lab File: BG016099.D

Acq: 25 Mar 2015 13:22


Tgt Ion: 82 Resp: 19729

Ion Ratio Lower Upper

82 100

128 48.6 40.4 60.6



54 37.7 32.1 48.1




Abundance Ion 82.00 (81.70 to 82.70): BGC

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BGC









#26
 2-Nitrophenol
 Concen: 2.50 ng
 RT: 9.70 min Scan# 1193
 Delta R.T. 0.00 min
 Lab File: BG016099.D
 Acq: 25 Mar 2015 13:22

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
139	100		
109	20.4	15.7	23.5
65	36.5	30.1	45.1




Manual Integrations
 APPROVED


apatel
 3/27/2015 10:47:18 AM



#27
 2,4-Dimethylphenol
 Concen: 2.50 ng
 RT: 9.75 min Scan# 1202
 Delta R.T. 0.00 min
 Lab File: BG016099.D
 Acq: 25 Mar 2015 13:22

Tgt Ion	Ratio	Lower	Upper
122	100		
107	93.8	83.6	125.4
121	55.6	45.6	68.4





#28

bis(2-Chloroethoxy)methane

Concen: 2.32 ng

RT: 9.99 min Scan# 1243

Delta R.T. 0.00 min

Lab File: BG016099.D

Acq: 25 Mar 2015 13:22

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tgt Ion:	93	Resp:	13093
Ion	Ratio	Lower	Upper
93	100		
95	33.4	26.4	39.6
123	14.3	12.3	18.5

Manual Integrations
APPROVED

apatel
3/27/2015 10:47:18 AM



#31

Naphthalene

Concen: 2.88 ng




RT: 10.63 min Scan# 1351


Delta R.T. 0.00 min

Lab File: BG016099.D

Acq: 25 Mar 2015 13:22

Tgt Ion:	128	Resp:	34445
Ion	Ratio	Lower	Upper
128	100		
129	11.3	9.2	13.8
127	12.0	10.9	16.3






#33
4-Chloroaniline
Concen: 2.74 ng
RT: 10.74 min Scan# 1370
Delta R.T. 0.00 min
Lab File: BG016099.D
Acq: 25 Mar 2015 13:22

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5




Tgt Ion	Ratio	Resp	Lower	Upper
127	100	14146		
129	33.9	25.3	37.9	
65	23.7	19.5	29.3	
92	16.4	13.1	19.7	

Manual Integrations
APPROVED

apatel
3/27/2015 10:47:18 AM




Abundance Ion 127.00 (126.70 to 127.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 65.00 (64.70 to 65.70): BGC
Ion 92.00 (91.70 to 92.70): BGC






#35
Caprolactam
Concen: 2.36 ng
RT: 11.47 min Scan# 1495
Delta R.T. -0.03 min
Lab File: BG016099.D
Acq: 25 Mar 2015 13:22

Tgt Ion	Ratio	Resp	Lower	Upper
113	100	4574		
55	106.2	93.9	133.9	
56	75.4	67.1	107.1	



Abundance Ion 113.00 (112.70 to 113.70): E
Ion 55.00 (54.70 to 55.70): BGC
Ion 56.00 (55.70 to 56.70): BGC









#37
 2-Methylnaphthalene
 Concen: 2.56 ng
 RT: 12.24 min Scan# 1625
 Delta R.T. 0.00 min
 Lab File: BG016099.D
 Acq: 25 Mar 2015 13:22

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
142	100		
141	86.3	68.6	103.0
115	27.7	24.0	36.0




Manual Integrations
 APPROVED


apatel
 3/27/2015 10:47:18 AM



#38
 Acenaphthene-d10
 Concen: 20.00 ng
 RT: 14.42 min Scan# 1996
 Delta R.T. 0.00 min
 Lab File: BG016099.D
 Acq: 25 Mar 2015 13:22

Tgt Ion	Ratio	Lower	Upper
164	100		
162	93.2	75.2	112.8
160	42.7	33.9	50.9






#41
 2,4,6-Tribromophenol
 Concen: 2.78 ng
 RT: 15.90 min Scan# 2249
 Delta R.T. 0.00 min
 Lab File: BG016099.D
 Acq: 25 Mar 2015 13:22

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5




Tgt Ion	Ratio	Resp	Lower	Upper
330	100	8196		
332	99.9	76.6	114.8	
141	33.1	26.2	39.4	

Manual Integrations
 APPROVED

apatel
 3/27/2015 10:47:18 AM






Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E




#44
 2-Fluorobiphenyl
 Concen: 5.22 ng
 RT: 13.04 min Scan# 1762
 Delta R.T. 0.00 min
 Lab File: BG016099.D
 Acq: 25 Mar 2015 13:22

Tgt Ion	Ratio	Resp	Lower	Upper
172	100	48993		
171	32.9	30.3	45.5	
170	22.9	19.8	29.8	

Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E









#45
 1,1'-Biphenyl
 Concen: 3.16 ng
 RT: 13.25 min Scan# 1797
 Delta R.T. -0.01 min
 Lab File: BG016099.D
 Acq: 25 Mar 2015 13:22

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Tgt Ion:	154	Resp:	29466
Ion Ratio	Lower	Upper	
154	100		
153	41.0	21.4	61.4
76	14.9	0.0	35.5




Manual Integrations
 APPROVED


apatel
 3/27/2015 10:47:18 AM



#46
 2-Chloronaphthalene
 Concen: 2.85 ng
 RT: 13.29 min Scan# 1805
 Delta R.T. 0.00 min
 Lab File: BG016099.D
 Acq: 25 Mar 2015 13:22

Tgt Ion:	162	Resp:	22731
Ion Ratio	Lower	Upper	
162	100		
127	32.6	29.4	44.2
164	32.6	26.1	39.1





#47

2-Nitroaniline

Concen: 2.13 ng

RT: 13.49 min Scan# 1839

Delta R.T. -0.01 min

Lab File: BG016099.D

Acq: 25 Mar 2015 13:22

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tgt Ion:	65	Resp:	6278
Ion Ratio	Lower	Upper	
65	100		
92	77.7	64.2	96.4
138	130.4	110.4	165.6

Manual Integrations
APPROVED

apatel
3/27/2015 10:47:18 AM



#48

Acenaphthylene

Concen: 3.17 ng




RT: 14.13 min Scan# 1948


Delta R.T. -0.01 min

Lab File: BG016099.D

Acq: 25 Mar 2015 13:22

Tgt Ion:	152	Resp:	42754
Ion Ratio	Lower	Upper	
152	100		
151	21.0	17.0	25.4
153	12.4	11.0	16.4









#49
Dimethylphthalate
Concen: 2.60 ng
RT: 13.87 min Scan# 1903
Delta R.T. -0.01 min
Lab File: BG016099.D
Acq: 25 Mar 2015 13:22

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion	Ratio	Resp	Lower	Upper
163	100	29002		
194	5.3		5.4	8.0#
164	10.0		8.6	13.0




Manual Integrations
APPROVED


apatel
3/27/2015 10:47:18 AM



#50
2,6-Dinitrotoluene
Concen: 2.34 ng
RT: 13.99 min Scan# 1923
Delta R.T. -0.01 min
Lab File: BG016099.D
Acq: 25 Mar 2015 13:22

Tgt Ion	Ratio	Resp	Lower	Upper
165	100	6019		
63	49.9		33.8	50.6
89	47.9		35.8	53.8





#51

Acenaphthene

Concen: 3.21 ng

RT: 14.48 min Scan# 2007

Delta R.T. 0.00 min

Lab File: BG016099.D


Acq: 25 Mar 2015 13:22

Instrument :

BNA_G

ClientSampleId :




SSTDICC2.5



Tgt Ion:	154	Resp:	25912
Ion Ratio	Lower	Upper	
154	100		
153	104.0	87.2	130.8
152	51.5	41.8	62.6

Manual Integrations
APPROVED

apatel
3/27/2015 10:47:18 AM



#52

3-Nitroaniline

Concen: 3.31 ng




RT: 14.32 min Scan# 1979


Delta R.T. -0.01 min

Lab File: BG016099.D

Acq: 25 Mar 2015 13:22

Tgt Ion:	138	Resp:	7794
Ion Ratio	Lower	Upper	
138	100		
108	9.9	7.8	11.6
92	121.4	85.8	128.8





#54

Dibenzofuran

Concen: 2.89 ng

RT: 14.82 min Scan# 2064

Delta R.T. 0.00 min

Lab File: BG016099.D

Acq: 25 Mar 2015 13:22

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tgt Ion:	168	Resp:	36692
Ion Ratio	Lower	Upper	
168	100		
139	33.2	29.0	43.4
169	12.5	11.1	16.7

Manual Integrations
APPROVED

apatel
3/27/2015 10:47:18 AM



#55

4-Nitrophenol

Concen: 3.01 ng




RT: 14.62 min Scan# 2030


Delta R.T. -0.01 min

Lab File: BG016099.D

Acq: 25 Mar 2015 13:22

Tgt Ion:	139	Resp:	5388
Ion Ratio	Lower	Upper	
139	100		
109	45.0	28.7	68.7
65	76.9	50.7	90.7









#56
2,4-Dinitrotoluene
Concen: 2.13 ng
RT: 14.77 min Scan# 2057
Delta R.T. -0.01 min
Lab File: BG016099.D
Acq: 25 Mar 2015 13:22

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
165	100		
63	33.2	27.8	41.6
89	57.7	51.2	76.8
182	6.7	4.4	6.6




Manual Integrations
APPROVED


apatel
3/27/2015 10:47:18 AM



#57
Fluorene
Concen: 2.75 ng
RT: 15.46 min Scan# 2174
Delta R.T. 0.00 min
Lab File: BG016099.D
Acq: 25 Mar 2015 13:22

Tgt Ion	Ratio	Lower	Upper
166	100		
165	95.0	73.9	110.9
167	13.7	11.3	16.9









#59
Diethylphthalate
Concen: 2.55 ng
RT: 15.23 min Scan# 2135
Delta R.T. -0.01 min
Lab File: BG016099.D
Acq: 25 Mar 2015 13:22

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
149	100		
177	22.8	19.5	29.3
150	13.0	10.1	15.1




Manual Integrations
APPROVED


apatel
3/27/2015 10:47:18 AM



#61
4-Nitroaniline
Concen: 3.02 ng
RT: 15.47 min Scan# 2176
Delta R.T. -0.01 min
Lab File: BG016099.D
Acq: 25 Mar 2015 13:22

Tgt Ion	Ratio	Lower	Upper
138	100		
92	46.6	24.2	64.2
108	49.7	32.7	72.7





#62

Azobenzene

Concen: 2.12 ng

RT: 15.74 min Scan# 2222

Delta R.T. -0.01 min

Lab File: BG016099.D

Acq: 25 Mar 2015 13:22

Instrument :

BNA_G


ClientSampleId :

SSTDICC2.5

Tgt Ion	Ion	Ratio	Lower	Upper
77	100			
182	27.5	10.9	50.9	
105	18.2	0.0	40.0	
51	20.1	1.9	41.9	

Manual Integrations
APPROVED

apatel
3/27/2015 10:47:18 AM






Abundance Ion 77.00 (76.70 to 77.70): BGC

Ion 182.00 (181.70 to 182.70): BGC

Ion 105.00 (104.70 to 105.70): BGC

Ion 51.00 (50.70 to 51.70): BGC



#63

Phenanthrene-d10

Concen: 20.00 ng


RT: 17.15 min Scan# 2462

Delta R.T. -0.01 min

Lab File: BG016099.D

Acq: 25 Mar 2015 13:22



Tgt Ion	Ion	Ratio	Lower	Upper
188	100			
94	12.7	9.1	13.7	
80	13.0	9.8	14.8	




Abundance Ion 188.00 (187.70 to 188.70): BGC

Ion 94.00 (93.70 to 94.70): BGC

Ion 80.00 (79.70 to 80.70): BGC









#65
n-Nitrosodiphenylamine
Concen: 2.99 ng
RT: 15.67 min Scan# 2209
Delta R.T. -0.01 min
Lab File: BG016099.D
Acq: 25 Mar 2015 13:22

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion	Ratio	Resp	Lower	Upper
169	100	28001		
168	63.4		53.0	79.4
167	34.3		29.0	43.4




Manual Integrations
APPROVED


apatel
3/27/2015 10:47:18 AM



#68
Atrazine
Concen: 2.19 ng
RT: 16.61 min Scan# 2370
Delta R.T. -0.01 min
Lab File: BG016099.D
Acq: 25 Mar 2015 13:22

Tgt Ion	Ratio	Resp	Lower	Upper
200	100	8515		
173	22.2		7.5	47.5
215	51.1		36.4	76.4





#70

Phenanthrene

Concen: 3.03 ng

RT: 17.19 min Scan# 2469

Delta R.T. -0.01 min

Lab File: BG016099.D

Acq: 25 Mar 2015 13:22

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tgt Ion:	178	Resp:	51994
Ion Ratio	Lower	Upper	
178	100		
176	19.0	16.5	24.7
179	14.7	13.0	19.6

Manual Integrations
APPROVED

apatel
3/27/2015 10:47:18 AM



#71

Anthracene

Concen: 2.98 ng




RT: 17.28 min Scan# 2484


Delta R.T. -0.01 min

Lab File: BG016099.D

Acq: 25 Mar 2015 13:22

Tgt Ion:	178	Resp:	50681
Ion Ratio	Lower	Upper	
178	100		
176	18.6	15.6	23.4
179	16.4	13.3	19.9





#72

Carbazole

Concen: 3.21 ng

RT: 17.55 min Scan# 2530

Delta R.T. -0.01 min

Lab File: BG016099.D

Acq: 25 Mar 2015 13:22

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tgt Ion:	167	Resp:	49255
Ion Ratio	Lower	Upper	
167	100		
166	21.3	18.2	27.2
139	14.2	10.7	16.1

Manual Integrations
APPROVED

apatel
3/27/2015 10:47:18 AM



#73

Di-n-butylphthalate

Concen: 3.16 ng




RT: 18.12 min Scan# 2626


Delta R.T. -0.01 min

Lab File: BG016099.D

Acq: 25 Mar 2015 13:22

Tgt Ion:	149	Resp:	55294
Ion Ratio	Lower	Upper	
149	100		
150	9.1	8.0	12.0
104	4.3	3.8	5.6





#74

Fluoranthene

Concen: 2.33 ng

RT: 19.21 min Scan# 2812

Delta R.T. 0.00 min

Lab File: BG016099.D

Acq: 25 Mar 2015 13:22

Instrument :

BNA_G


ClientSampleId :

SSTDICC2.5




Tgt	Ion	Ratio	Resp	Lower	Upper
202	100		55856		
101	13.9	0.0	35.9		
203	17.4	0.0	38.7		

Manual Integrations
APPROVED

apatel
3/27/2015 10:47:18 AM



Abundance Ion 202.00 (201.70 to 202.70): E
Ion 101.00 (100.70 to 101.70): E
Ion 203.00 (202.70 to 203.70): E



#75

Chrysene-d12

Concen: 20.00 ng


RT: 21.32 min Scan# 3172

Delta R.T. 0.00 min



Lab File: BG016099.D


Acq: 25 Mar 2015 13:22

Tgt	Ion	Ratio	Resp	Lower	Upper
240	100		342769		
120	14.7	13.0	19.4		
236	24.8	20.2	30.2		



Abundance Ion 240.00 (239.70 to 240.70): E
Ion 120.00 (119.70 to 120.70): E
Ion 236.00 (235.70 to 236.70): E





#76

Benzidine

Concen: 3.31 ng

RT: 19.39 min Scan# 2843

Delta R.T. 0.00 min

Lab File: BG016099.D

Acq: 25 Mar 2015 13:22

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tgt Ion:	184	Resp:	24318
Ion Ratio	Lower	Upper	
184	100		
185	16.1	12.6	18.8
183	12.8	10.2	15.4

Manual Integrations
APPROVED

apatel
3/27/2015 10:47:18 AM



#77

Pyrene

Concen: 3.56 ng




RT: 19.57 min Scan# 2873


Delta R.T. -0.01 min

Lab File: BG016099.D

Acq: 25 Mar 2015 13:22

Tgt Ion:	202	Resp:	58451
Ion Ratio	Lower	Upper	
202	100		
200	20.4	18.4	27.6
203	18.3	15.4	23.0





#78

Terphenyl-d14

Concen: 6.01 ng

RT: 19.77 min Scan# 2908

Delta R.T. -0.01 min

Lab File: BG016099.D

Acq: 25 Mar 2015 13:22

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tgt Ion:	244	Resp:	76840
Ion Ratio	Lower	Upper	
244	100		
212	6.2	5.9	8.9
122	13.2	12.2	18.2

Manual Integrations
APPROVED

apatel
3/27/2015 10:47:18 AM



#79

Butylbenzylphthalate

Concen: 3.62 ng




RT: 20.47 min Scan# 3026


Delta R.T. 0.00 min

Lab File: BG016099.D

Acq: 25 Mar 2015 13:22

Tgt Ion:	149	Resp:	22025
Ion Ratio	Lower	Upper	
149	100		
91	64.8	55.8	83.6
206	26.2	19.1	28.7









#80
Benzo(a)anthracene
Concen: 2.91 ng
RT: 21.31 min Scan# 3169
Delta R.T. -0.01 min
Lab File: BG016099.D
Acq: 25 Mar 2015 13:22

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
228	100		
226	26.7	23.0	34.4
229	20.1	16.8	25.2




Manual Integrations
APPROVED


apatel
3/27/2015 10:47:18 AM



#81
3,3'-Dichlorobenzidine
Concen: 2.32 ng
RT: 21.24 min Scan# 3158
Delta R.T. -0.01 min
Lab File: BG016099.D
Acq: 25 Mar 2015 13:22

Tgt Ion	Ratio	Lower	Upper
252	100		
254	64.3	51.7	77.5
126	13.0	10.6	16.0





#82

Chrysene

Concen: 2.95 ng

RT: 21.36 min Scan# 3178

Delta R.T. -0.01 min

Lab File: BG016099.D

Acq: 25 Mar 2015 13:22

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tgt Ion:	228	Resp:	50462
Ion Ratio	Lower	Upper	
228	100		
226	30.6	25.7	38.5
229	20.1	17.0	25.6

Manual Integrations
APPROVED

apatel
3/27/2015 10:47:18 AM



#83

Bis(2-ethylhexyl)phthalate

Concen: 3.72 ng




RT: 21.24 min Scan# 3157


Delta R.T. 0.00 min

Lab File: BG016099.D

Acq: 25 Mar 2015 13:22

Tgt Ion:	149	Resp:	31677
Ion Ratio	Lower	Upper	
149	100		
167	31.5	27.1	40.7
279	7.7	6.4	9.6









#84
Di-n-octyl phthalate
Concen: 3.94 ng
RT: 22.12 min Scan# 3308
Delta R.T. 0.00 min
Lab File: BG016099.D
Acq: 25 Mar 2015 13:22

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion	Ratio	Resp Lower	Resp Upper
149	100		
167	1.7	1.4	2.0
43	7.5	6.1	9.1




Manual Integrations
APPROVED


apatel
3/27/2015 10:47:18 AM



#85
Indeno(1,2,3-cd)pyrene
Concen: 2.49 ng
RT: 25.93 min Scan# 3956
Delta R.T. -0.02 min
Lab File: BG016099.D
Acq: 25 Mar 2015 13:22

Tgt Ion	Ratio	Resp Lower	Resp Upper
276	100		
138	32.7	26.9	40.3
277	25.4	20.7	31.1









#86
Perylene-d12
Concen: 20.00 ng
RT: 23.61 min Scan# 3561
Delta R.T. 0.00 min
Lab File: BG016099.D
Acq: 25 Mar 2015 13:22

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion	Ratio	Resp	Lower	Upper
264	100	328534		
260	22.0		18.4	27.6
265	21.8		17.3	25.9




Manual Integrations
APPROVED


apatel
3/27/2015 10:47:18 AM



#87
Benzo(b)fluoranthene
Concen: 2.67 ng
RT: 22.91 min Scan# 3442
Delta R.T. -0.01 min
Lab File: BG016099.D
Acq: 25 Mar 2015 13:22

Tgt Ion	Ratio	Resp	Lower	Upper
252	100	51434		
253	21.7		17.9	26.9
125	13.6		11.1	16.7









#88
Benzo(k)fluoranthene
Concen: 2.82 ng
RT: 22.96 min Scan# 3450
Delta R.T. -0.01 min
Lab File: BG016099.D
Acq: 25 Mar 2015 13:22

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tgt Ion	Ratio	Lower	Upper
252	100		
253	21.8	18.2	27.4
125	12.9	11.0	16.4




Manual Integrations
APPROVED


apatel
3/27/2015 10:47:18 AM



#89
Benzo(a)pyrene
Concen: 2.64 ng
RT: 23.50 min Scan# 3543
Delta R.T. -0.01 min
Lab File: BG016099.D
Acq: 25 Mar 2015 13:22

Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.3	17.9	26.9
125	15.7	12.5	18.7





#90

Dibenzo(a,h)anthracene

Concen: 2.56 ng

RT: 25.95 min Scan# 3959

Delta R.T. -0.02 min

Lab File: BG016099.D

Acq: 25 Mar 2015 13:22

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tgt Ion:	278	Resp:	49075
Ion Ratio	Lower	Upper	
278	100		
139	18.3	17.8	26.8
279	21.1	19.9	29.9

Manual Integrations
APPROVED

apatel
3/27/2015 10:47:18 AM



#91

Benzo(g,h,i)perylene

Concen: 2.68 ng

RT: 26.66 min Scan# 4080

Delta R.T. -0.02 min

Lab File: BG016099.D

Acq: 25 Mar 2015 13:22

Tgt Ion:	276	Resp:	50005
Ion Ratio	Lower	Upper	
276	100		
277	23.3	18.7	28.1
138	24.4	22.0	33.0

