

Data Path : Z:\SVOASRV\HPCHEM1\BNA\_G\DATA\BG032519\  
 Data File : BG040135.D  
 Acq On : 26 Mar 2019 3:32  
 Operator : JU/SJ  
 Sample : SSTDCCC040  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 LabSampleId :  
 SSTDCCC040

Quant Time: Mar 26 06:46:45 2019  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_G\METHODS\8270-BG030419.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon Mar 04 14:38:15 2019  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound                    | AvgRF | CCRF  | %Dev | Area% | Dev(min) |
|------|-----------------------------|-------|-------|------|-------|----------|
| 1 I  | 1,4-Dichlorobenzene-d4      | 1.000 | 1.000 | 0.0  | 91    | -0.03    |
| 2    | 1,4-Dioxane                 | 0.541 | 0.501 | 7.4  | 88    | -0.03    |
| 3    | Pyridine                    | 1.449 | 1.452 | -0.2 | 86    | -0.02    |
| 4    | n-Nitrosodimethylamine      | 0.687 | 0.676 | 1.6  | 88    | -0.03    |
| 5 S  | 2-Fluorophenol              | 1.172 | 1.167 | 0.4  | 90    | -0.02    |
| 6    | Aniline                     | 2.290 | 2.233 | 2.5  | 89    | -0.03    |
| 7 S  | Phenol-d6                   | 1.748 | 1.765 | -1.0 | 90    | -0.03    |
| 8    | 2-Chlorophenol              | 1.422 | 1.426 | -0.3 | 91    | -0.02    |
| 9    | Benzaldehyde                | 1.128 | 1.026 | 9.0  | 82    | -0.02    |
| 10 C | Phenol                      | 1.894 | 1.898 | -0.2 | 89    | -0.02    |
| 11   | bis(2-Chloroethyl)ether     | 1.476 | 1.429 | 3.2  | 89    | -0.03    |
| 12   | 1,3-Dichlorobenzene         | 1.609 | 1.579 | 1.9  | 89    | -0.02    |
| 13 C | 1,4-Dichlorobenzene         | 1.638 | 1.582 | 3.4  | 89    | -0.02    |
| 14   | 1,2-Dichlorobenzene         | 1.583 | 1.514 | 4.4  | 88    | -0.03    |
| 15   | Benzyl Alcohol              | 1.387 | 1.373 | 1.0  | 87    | -0.03    |
| 16   | 2,2'-oxybis(1-Chloropropane | 2.953 | 2.602 | 11.9 | 81    | -0.03    |
| 17   | 2-Methylphenol              | 1.207 | 1.235 | -2.3 | 91    | -0.02    |
| 18   | Hexachloroethane            | 0.588 | 0.566 | 3.7  | 91    | -0.03    |
| 19 P | n-Nitroso-di-n-propylamine  | 1.258 | 1.182 | 6.0  | 83    | -0.03    |
| 20   | 3+4-Methylphenols           | 1.677 | 1.682 | -0.3 | 89    | -0.02    |
| 21 I | Naphthalene-d8              | 1.000 | 1.000 | 0.0  | 91    | -0.03    |
| 22   | Acetophenone                | 0.537 | 0.527 | 1.9  | 88    | -0.03    |
| 23 S | Nitrobenzene-d5             | 0.370 | 0.372 | -0.5 | 90    | -0.03    |
| 24   | Nitrobenzene                | 0.388 | 0.383 | 1.3  | 88    | -0.02    |
| 25   | Isophorone                  | 0.775 | 0.758 | 2.2  | 87    | -0.02    |
| 26 C | 2-Nitrophenol               | 0.187 | 0.203 | -8.6 | 94    | -0.03    |
| 27   | 2,4-Dimethylphenol          | 0.291 | 0.289 | 0.7  | 87    | -0.03    |
| 28   | bis(2-Chloroethoxy)methane  | 0.471 | 0.451 | 4.2  | 85    | -0.02    |
| 29 C | 2,4-Dichlorophenol          | 0.328 | 0.352 | -7.3 | 93    | -0.02    |
| 30   | 1,2,4-Trichlorobenzene      | 0.369 | 0.365 | 1.1  | 90    | -0.03    |
| 31   | Naphthalene                 | 1.059 | 1.024 | 3.3  | 87    | -0.02    |
| 32   | Benzoic acid                | 0.181 | 0.150 | 17.1 | 74    | -0.02    |
| 33   | 4-Chloroaniline             | 0.449 | 0.456 | -1.6 | 89    | -0.02    |
| 34 C | Hexachlorobutadiene         | 0.252 | 0.255 | -1.2 | 92    | -0.03    |
| 35   | Caprolactam                 | 0.125 | 0.130 | -4.0 | 91    | -0.03    |
| 36 C | 4-Chloro-3-methylphenol     | 0.376 | 0.393 | -4.5 | 91    | -0.02    |
| 37   | 2-Methylnaphthalene         | 0.792 | 0.782 | 1.3  | 89    | -0.02    |
| 38   | 1-Methylnaphthalene         | 0.769 | 0.763 | 0.8  | 88    | -0.02    |
| 39 I | Acenaphthene-d10            | 1.000 | 1.000 | 0.0  | 94    | -0.02    |
| 40   | 1,2,4,5-Tetrachlorobenzene  | 0.678 | 0.665 | 1.9  | 92    | -0.02    |
| 41 P | Hexachlorocyclopentadiene   | 0.363 | 0.323 | 11.0 | 82    | -0.03    |
| 42 S | 2,4,6-Tribromophenol        | 0.233 | 0.248 | -6.4 | 94    | -0.02    |
| 43 C | 2,4,6-Trichlorophenol       | 0.397 | 0.427 | -7.6 | 97    | -0.02    |
| 44   | 2,4,5-Trichlorophenol       | 0.447 | 0.470 | -5.1 | 93    | -0.02    |

Data Path : Z:\SVOASRV\HPCHEM1\BNA\_G\DATA\BG032519\  
 Data File : BG040135.D  
 Acq On : 26 Mar 2019 3:32  
 Operator : JU/SJ  
 Sample : SSTDCCC040  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 LabSampleId :  
 SSTDCCC040

Quant Time: Mar 26 06:46:45 2019  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_G\METHODS\8270-BG030419.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon Mar 04 14:38:15 2019  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound                   | AvgRF | CCRF  | %Dev  | Area% | Dev(min) |
|------|----------------------------|-------|-------|-------|-------|----------|
| 45 S | 2-Fluorobiphenyl           | 1.405 | 1.359 | 3.3   | 91    | -0.03    |
| 46   | 1,1'-Biphenyl              | 1.645 | 1.592 | 3.2   | 91    | -0.02    |
| 47   | 2-Chloronaphthalene        | 1.238 | 1.211 | 2.2   | 92    | -0.02    |
| 48   | 2-Nitroaniline             | 0.398 | 0.406 | -2.0  | 91    | -0.02    |
| 49   | Acenaphthylene             | 2.031 | 2.001 | 1.5   | 91    | -0.02    |
| 50   | Dimethylphthalate          | 1.672 | 1.640 | 1.9   | 90    | -0.02    |
| 51   | 2,6-Dinitrotoluene         | 0.355 | 0.372 | -4.8  | 93    | -0.02    |
| 52 C | Acenaphthene               | 1.223 | 1.183 | 3.3   | 90    | -0.02    |
| 53   | 3-Nitroaniline             | 0.356 | 0.360 | -1.1  | 90    | -0.02    |
| 54 P | 2,4-Dinitrophenol          | 0.197 | 0.064 | 67.5# | 29#   | -0.01    |
| 55   | Dibenzofuran               | 2.006 | 1.976 | 1.5   | 92    | -0.02    |
| 56 P | 4-Nitrophenol              | 0.319 | 0.274 | 14.1  | 77    | -0.01    |
| 57   | 2,4-Dinitrotoluene         | 0.498 | 0.527 | -5.8  | 93    | -0.02    |
| 58   | Fluorene                   | 1.577 | 1.555 | 1.4   | 92    | -0.02    |
| 59   | 2,3,4,6-Tetrachlorophenol  | 0.437 | 0.446 | -2.1  | 91    | -0.02    |
| 60   | Diethylphthalate           | 1.656 | 1.636 | 1.2   | 90    | -0.02    |
| 61   | 4-Chlorophenyl-phenylether | 0.842 | 0.836 | 0.7   | 92    | -0.02    |
| 62   | 4-Nitroaniline             | 0.386 | 0.382 | 1.0   | 87    | -0.02    |
| 63   | Azobenzene                 | 1.501 | 1.478 | 1.5   | 92    | -0.02    |
| 64 I | Phenanthrene-d10           | 1.000 | 1.000 | 0.0   | 94    | -0.02    |
| 65   | 4,6-Dinitro-2-methylphenol | 0.118 | 0.076 | 35.6# | 60    | -0.02    |
| 66 c | n-Nitrosodiphenylamine     | 0.579 | 0.572 | 1.2   | 91    | -0.02    |
| 67   | 4-Bromophenyl-phenylether  | 0.224 | 0.228 | -1.8  | 96    | -0.02    |
| 68   | Hexachlorobenzene          | 0.232 | 0.234 | -0.9  | 95    | -0.02    |
| 69   | Atrazine                   | 0.228 | 0.209 | 8.3   | 85    | -0.02    |
| 70 C | Pentachlorophenol          | 0.147 | 0.120 | 18.4  | 75    | -0.02    |
| 71   | Phenanthrene               | 1.102 | 1.057 | 4.1   | 91    | -0.02    |
| 72   | Anthracene                 | 1.074 | 1.032 | 3.9   | 91    | -0.02    |
| 73   | Carbazole                  | 0.968 | 0.924 | 4.5   | 90    | -0.02    |
| 74   | Di-n-butylphthalate        | 1.139 | 1.104 | 3.1   | 91    | -0.02    |
| 75 C | Fluoranthene               | 1.302 | 1.224 | 6.0   | 90    | -0.02    |
| 76 I | Chrysene-d12               | 1.000 | 1.000 | 0.0   | 94    | -0.03    |
| 77   | Benzidine                  | 0.635 | 0.530 | 16.5  | 70    | -0.02    |
| 78   | Pyrene                     | 1.300 | 1.287 | 1.0   | 91    | -0.02    |
| 79 S | Terphenyl-d14              | 0.908 | 0.895 | 1.4   | 92    | -0.02    |
| 80   | Butylbenzylphthalate       | 0.504 | 0.511 | -1.4  | 91    | -0.02    |
| 81   | Benzo(a)anthracene         | 1.253 | 1.226 | 2.2   | 91    | -0.02    |
| 82   | 3,3'-Dichlorobenzidine     | 0.458 | 0.443 | 3.3   | 86    | -0.02    |
| 83   | Chrysene                   | 1.215 | 1.175 | 3.3   | 91    | -0.02    |
| 84   | Bis(2-ethylhexyl)phthalate | 0.708 | 0.712 | -0.6  | 90    | -0.03    |
| 85 c | Di-n-octyl phthalate       | 1.172 | 1.211 | -3.3  | 91    | -0.04    |
| 86   | Indeno(1,2,3-cd)pyrene     | 1.379 | 1.363 | 1.2   | 90    | -0.06    |
| 87 I | Perylene-d12               | 1.000 | 1.000 | 0.0   | 90    | -0.04    |

Data Path : Z:\SVOASRV\HPCHEM1\BNA\_G\DATA\BG032519\  
Data File : BG040135.D  
Acq On : 26 Mar 2019 3:32  
Operator : JU/SJ  
Sample : SSTDCCC040  
Misc :  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
BNA\_G  
LabSampleId :  
SSTDCCC040

Quant Time: Mar 26 06:46:45 2019  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_G\METHODS\8270-BG030419.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Mon Mar 04 14:38:15 2019  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound               | AvgRF | CCRF  | %Dev | Area% | Dev(min) |
|------|------------------------|-------|-------|------|-------|----------|
| 88   | Benzo(b)fluoranthene   | 1.193 | 1.176 | 1.4  | 89    | -0.04    |
| 89   | Benzo(k)fluoranthene   | 1.110 | 1.100 | 0.9  | 90    | -0.04    |
| 90 C | Benzo(a)pyrene         | 1.102 | 1.112 | -0.9 | 90    | -0.04    |
| 91   | Dibenzo(a,h)anthracene | 1.080 | 1.057 | 2.1  | 89    | -0.07    |
| 92   | Benzo(g,h,i)perylene   | 1.085 | 1.073 | 1.1  | 89    | -0.08    |

(#) = Out of Range

SPCC's out = 0 CCC's out = 0