

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG032615\
 Data File : BG016156.D
 Acq On : 27 Mar 2015 1:01
 Operator : TP/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 27 05:11:19 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG032515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 27 03:23:37 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	0.00
2	1,4-Dioxane	0.483	0.440	8.9	97	0.00
3	Pyridine	1.509	1.404	7.0	98	0.00
4	n-Nitrosodimethylamine	0.444	0.387	12.8	92	0.00
5 S	2-Fluorophenol	1.195	1.129	5.5	100	0.00
6	Aniline	2.373	2.232	5.9	99	0.00
7 S	Phenol-d6	1.663	1.617	2.8	102	0.00
8	2-Chlorophenol	1.418	1.391	1.9	104	0.00
9	Benzaldehyde	0.695	0.672	3.3	98	0.00
10 C	Phenol	1.828	1.772	3.1	102	0.00
11	bis(2-Chloroethyl)ether	1.483	1.380	6.9	99	0.00
12	1,3-Dichlorobenzene	1.532	1.485	3.1	103	0.00
13 C	1,4-Dichlorobenzene	1.590	1.529	3.8	102	0.00
14	1,2-Dichlorobenzene	1.503	1.456	3.1	104	0.00
15	Benzyl Alcohol	1.229	1.128	8.2	95	0.00
16	2,2'-oxybis(1-Chloropropane	1.666	1.436	13.8	92	0.00
17	2-Methylphenol	1.289	1.233	4.3	101	0.00
18	Hexachloroethane	0.570	0.543	4.7	102	0.00
19 P	n-Nitroso-di-n-propylamine	1.219	1.094	10.3	96	0.00
20	3+4-Methylphenols	1.749	1.723	1.5	103	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
22	Acetophenone	0.450	0.434	3.6	101	0.00
23 S	Nitrobenzene-d5	0.323	0.309	4.3	100	0.00
24	Nitrobenzene	0.347	0.329	5.2	99	0.00
25	Isophorone	0.732	0.670	8.5	95	0.00
26 C	2-Nitrophenol	0.188	0.185	1.6	99	0.00
27	2,4-Dimethylphenol	0.315	0.313	0.6	103	0.00
28	bis(2-Chloroethoxy)methane	0.420	0.400	4.8	99	0.00
29 C	2,4-Dichlorophenol	0.277	0.287	-3.6	106	0.00
30	1,2,4-Trichlorobenzene	0.299	0.302	-1.0	105	0.00
31	Naphthalene	1.072	1.055	1.6	103	0.00
32	Benzoic acid	0.193	0.190	1.6	100	0.00
33	4-Chloroaniline	0.469	0.465	0.9	102	0.00
34 C	Hexachlorobutadiene	0.162	0.163	-0.6	104	0.00
35	Caprolactam	0.151	0.140	7.3	96	0.00
36 C	4-Chloro-3-methylphenol	0.327	0.320	2.1	101	0.00
37	2-Methylnaphthalene	0.765	0.764	0.1	104	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	105	0.00
39	1,2,4,5-Tetrachlorobenzene	0.504	0.509	-1.0	105	0.00
40 P	Hexachlorocyclopentadiene	0.295	0.298	-1.0	102	0.00
41 S	2,4,6-Tribromophenol	0.230	0.234	-1.7	104	0.00
42 C	2,4,6-Trichlorophenol	0.360	0.374	-3.9	107	0.00
43	2,4,5-Trichlorophenol	0.392	0.402	-2.6	106	0.00
44 S	2-Fluorobiphenyl	1.195	1.192	0.3	105	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG032615\
 Data File : BG016156.D
 Acq On : 27 Mar 2015 1:01
 Operator : TP/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 27 05:11:19 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG032515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 27 03:23:37 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.467	1.463	0.3	105	0.00
46	2-Chloronaphthalene	1.144	1.149	-0.4	105	0.00
47	2-Nitroaniline	0.329	0.303	7.9	97	0.00
48	Acenaphthylene	2.090	2.080	0.5	104	0.00
49	Dimethylphthalate	1.402	1.373	2.1	104	0.00
50	2,6-Dinitrotoluene	0.333	0.337	-1.2	104	0.00
51 C	Acenaphthene	1.261	1.243	1.4	105	0.00
52	3-Nitroaniline	0.407	0.405	0.5	103	0.00
53 P	2,4-Dinitrophenol	0.179	0.139	22.3	78	0.00
54	Dibenzofuran	1.762	1.751	0.6	105	0.00
55 P	4-Nitrophenol	0.330	0.320	3.0	97	0.00
56	2,4-Dinitrotoluene	0.456	0.457	-0.2	103	0.00
57	Fluorene	1.572	1.541	2.0	102	0.00
58	2,3,4,6-Tetrachlorophenol	0.350	0.355	-1.4	107	0.00
59	Diethylphthalate	1.442	1.387	3.8	101	0.00
60	4-Chlorophenyl-phenylether	0.692	0.681	1.6	103	0.00
61	4-Nitroaniline	0.461	0.447	3.0	99	0.00
62	Azobenzene	1.485	1.407	5.3	100	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	102	0.00
64	4,6-Dinitro-2-methylphenol	0.133	0.123	7.5	93	0.00
65 c	n-Nitrosodiphenylamine	0.638	0.644	-0.9	104	0.00
66	4-Bromophenyl-phenylether	0.205	0.205	0.0	104	0.00
67	Hexachlorobenzene	0.231	0.230	0.4	104	0.00
68	Atrazine	0.190	0.181	4.7	99	0.00
69 C	Pentachlorophenol	0.132	0.133	-0.8	99	0.00
70	Phenanthrene	1.122	1.116	0.5	103	0.00
71	Anthracene	1.116	1.108	0.7	103	0.00
72	Carbazole	1.095	1.089	0.5	102	0.00
73	Di-n-butylphthalate	1.239	1.204	2.8	99	0.00
74 C	Fluoranthene	1.250	1.243	0.6	101	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	103	0.00
76	Benzidine	0.524	0.472	9.9	88	0.00
77	Pyrene	1.233	1.228	0.4	103	0.00
78 S	Terphenyl-d14	0.794	0.792	0.3	101	0.00
79	Butylbenzylphthalate	0.514	0.510	0.8	100	0.00
80	Benzo(a)anthracene	1.134	1.131	0.3	102	0.00
81	3,3'-Dichlorobenzidine	0.394	0.386	2.0	99	0.00
82	Chrysene	1.039	1.032	0.7	103	0.00
83	Bis(2-ethylhexyl)phthalate	0.706	0.678	4.0	98	0.00
84 c	Di-n-octyl phthalate	1.191	1.136	4.6	97	0.00
85	Indeno(1,2,3-cd)pyrene	1.336	1.335	0.1	103	0.00
86 I	Perylene-d12	1.000	1.000	0.0	104	0.00
87	Benzo(b)fluoranthene	1.158	1.160	-0.2	106	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG032615\
Data File : BG016156.D
Acq On : 27 Mar 2015 1:01
Operator : TP/IZ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Mar 27 05:11:19 2015
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG032515.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 27 03:23:37 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.127	1.117	0.9	102	0.00
89 C	Benzo(a)pyrene	1.079	1.073	0.6	103	0.00
90	Dibenzo(a,h)anthracene	1.125	1.125	0.0	104	0.00
91	Benzo(g,h,i)perylene	1.115	1.108	0.6	104	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0