

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA_G\Data\BG032618\
 Data File : BG033324.D
 Acq On : 26 Mar 2018 11:47
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 26 14:17:37 2018
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG031918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 23 15:11:46 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	125	-0.03
2	1,4-Dioxane	40.000	37.788	5.5	119	-0.01
3	Pyridine	40.000	42.580	-6.4	120	-0.02
4	n-Nitrosodimethylamine	40.000	38.654	3.4	119	-0.01
5 S	2-Fluorophenol	80.000	84.686	-5.9	122	-0.02
6	Aniline	40.000	41.012	-2.5	118	-0.02
7 S	Phenol-d6	80.000	75.884	5.1	115	-0.02
8	2-Chlorophenol	40.000	41.445	-3.6	119	-0.01
9	Benzaldehyde	40.000	35.287	11.8	113	-0.02
10 C	Phenol	40.000	39.128	2.2	116	-0.02
11	bis(2-Chloroethyl)ether	40.000	37.388	6.5	107	-0.02
12	1,3-Dichlorobenzene	40.000	38.539	3.7	118	-0.03
13 C	1,4-Dichlorobenzene	40.000	37.794	5.5	118	-0.02
14	1,2-Dichlorobenzene	40.000	38.609	3.5	115	-0.02
15	Benzyl Alcohol	40.000	39.857	0.4	116	-0.02
16	2,2'-oxybis(1-Chloropropane)	40.000	36.178	9.6	110	-0.02
17	2-Methylphenol	40.000	40.134	-0.3	122	-0.02
18	Hexachloroethane	40.000	38.030	4.9	117	-0.02
19 P	n-Nitroso-di-n-propylamine	40.000	39.639	0.9	113	-0.02
20	3+4-Methylphenols	40.000	39.713	0.7	115	-0.02
21 I	Naphthalene-d8	20.000	20.000	0.0	115	-0.03
22	Acetophenone	40.000	42.239	-5.6	113	-0.03
23 S	Nitrobenzene-d5	80.000	77.839	2.7	109	-0.02
24	Nitrobenzene	40.000	39.287	1.8	109	-0.02
25	Isophorone	40.000	40.637	-1.6	113	-0.02
26 C	2-Nitrophenol	40.000	41.304	-3.3	110	-0.03
27	2,4-Dimethylphenol	40.000	42.800	-7.0	125	-0.02
28	bis(2-Chloroethoxy)methane	40.000	41.426	-3.6	115	-0.02
29 C	2,4-Dichlorophenol	40.000	42.859	-7.1	118	-0.02
30	1,2,4-Trichlorobenzene	40.000	39.911	0.2	113	-0.02
31	Naphthalene	40.000	40.132	-0.3	114	-0.02
32	Benzoic acid	40.000	35.751	10.6	115	-0.03
33	4-Chloroaniline	40.000	39.827	0.4	110	-0.02
34 C	Hexachlorobutadiene	40.000	40.591	-1.5	119	-0.02
35	Caprolactam	40.000	42.151	-5.4	116	-0.02
36 C	4-Chloro-3-methylphenol	40.000	40.865	-2.2	109	-0.02
37	2-Methylnaphthalene	40.000	38.348	4.1	112	-0.02
38 I	Acenaphthene-d10	20.000	20.000	0.0	109	-0.02
39	1,2,4,5-Tetrachlorobenzene	40.000	40.580	-1.4	109	-0.02
40 P	Hexachlorocyclopentadiene	40.000	39.896	0.3	105	-0.01
41 S	2,4,6-Tribromophenol	80.000	83.285	-4.1	109	-0.02
42 C	2,4,6-Trichlorophenol	40.000	44.242	-10.6	113	-0.02
43	2,4,5-Trichlorophenol	40.000	44.494	-11.2	110	-0.02
44 S	2-Fluorobiphenyl	80.000	80.646	-0.8	108	-0.02

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA_G\Data\BG032618\
 Data File : BG033324.D
 Acq On : 26 Mar 2018 11:47
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 26 14:17:37 2018
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG031918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 23 15:11:46 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	40.000	40.715	-1.8	108	-0.01
46	2-Chloronaphthalene	40.000	40.607	-1.5	108	-0.02
47	2-Nitroaniline	40.000	40.390	-1.0	104	-0.01
48	Acenaphthylene	40.000	40.221	-0.6	107	-0.01
49	Dimethylphthalate	40.000	40.986	-2.5	110	-0.02
50	2,6-Dinitrotoluene	40.000	40.978	-2.4	105	-0.01
51 C	Acenaphthene	40.000	41.430	-3.6	109	-0.02
52	3-Nitroaniline	40.000	39.524	1.2	104	-0.02
53 P	2,4-Dinitrophenol	40.000	49.304	-23.3	141	-0.02
54	Dibenzofuran	40.000	39.498	1.3	106	-0.02
55 P	4-Nitrophenol	40.000	40.270	-0.7	104	-0.01
56	2,4-Dinitrotoluene	40.000	40.925	-2.3	108	-0.02
57	Fluorene	40.000	39.733	0.7	108	-0.01
58	2,3,4,6-Tetrachlorophenol	40.000	44.001	-10.0	113	-0.01
59	Diethylphthalate	40.000	40.692	-1.7	109	-0.01
60	4-Chlorophenyl-phenylether	40.000	39.909	0.2	109	-0.01
61	4-Nitroaniline	40.000	40.152	-0.4	106	-0.01
62	Azobenzene	40.000	38.319	4.2	102	-0.01
63 I	Phenanthrene-d10	20.000	20.000	0.0	108	-0.01
64	4,6-Dinitro-2-methylphenol	40.000	37.452	6.4	95	-0.01
65 c	n-Nitrosodiphenylamine	40.000	40.430	-1.1	107	-0.01
66	4-Bromophenyl-phenylether	40.000	40.134	-0.3	106	-0.01
67	Hexachlorobenzene	40.000	41.594	-4.0	112	-0.02
68	Atrazine	40.000	41.085	-2.7	108	-0.01
69 C	Pentachlorophenol	40.000	39.686	0.8	105	0.00
70	Phenanthrene	40.000	39.042	2.4	104	-0.01
71	Anthracene	40.000	39.458	1.4	105	-0.01
72	Carbazole	40.000	39.922	0.2	106	-0.01
73	Di-n-butylphthalate	40.000	40.292	-0.7	106	-0.01
74 C	Fluoranthene	40.000	38.811	3.0	104	-0.01
75 I	Chrysene-d12	20.000	20.000	0.0	111	-0.01
76	Benzidine	40.000	39.206	2.0	99	-0.01
77	Pyrene	40.000	37.645	5.9	103	-0.01
78 S	Terphenyl-d14	80.000	77.272	3.4	107	0.00
79	Butylbenzylphthalate	40.000	40.208	-0.5	109	-0.01
80	Benzo(a)anthracene	40.000	38.662	3.3	107	-0.01
81	3,3'-Dichlorobenzidine	40.000	43.433	-8.6	114	0.00
82	Chrysene	40.000	38.847	2.9	107	-0.01
83	Bis(2-ethylhexyl)phthalate	40.000	41.341	-3.4	112	0.00
84 c	Di-n-octyl phthalate	40.000	43.594	-9.0	117	0.00
85	Indeno(1,2,3-cd)pyrene	40.000	43.125	-7.8	117	0.00
86 I	Perylene-d12	20.000	20.000	0.0	116	0.00
87	Benzo(b)fluoranthene	40.000	38.047	4.9	109	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA_G\Data\BG032618\
Data File : BG033324.D
Acq On : 26 Mar 2018 11:47
Operator : SJ/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Mar 26 14:17:37 2018
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG031918.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 23 15:11:46 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	40.000	39.778	0.6	114	0.00
89 C	Benzo(a)pyrene	40.000	39.648	0.9	113	0.00
90	Dibenzo(a,h)anthracene	40.000	42.361	-5.9	117	0.01
91	Benzo(g,h,i)perylene	40.000	41.464	-3.7	117	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0