


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG032816\
Data File : BG021521.D
Acq On : 28 Mar 2016 10:38
Operator : UM/SJ
Sample : DFTPP15
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG031616.M
Title : SVOA CALIBRATION
Last Update : Tue Mar 29 02:01:55 2016



AutoFind: Scans 2550, 2551, 2552; Background Corrected with Scan 2545

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.3	12843	PASS
68	69	0.00	2	0.8	117	PASS
69	198	0.00	100	41.1	14539	PASS
70	69	0.00	2	0.4	52	PASS
127	198	10	80	48.2	17063	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	35416	PASS
199	198	5	9	7.0	2470	PASS
275	198	10	60	28.0	9919	PASS
365	198	1	100	4.8	1698	PASS
441	442	0.01	100	15.1	3447	PASS
442	198	50	100	64.4	22797	PASS
443	442	15	24	19.4	4429	PASS

Scan 2545 (17.790 min): BG021521.D
DFTPP15

m/z	Abundance
41.00	160.0
43.10	188.0
46.90	408.0
47.90	224.0
49.00	1987.0
50.90	472.0
72.90	225.0
83.90	1485.0
85.90	1031.0
207.00	189.0

Instrument :
BNA_G
ClientSampleId :

m/z	Abundance
36.90	191.0
38.00	222.0
39.00	1123.0
41.00	199.0
47.00	435.0
47.90	259.0
49.00	1940.0
50.00	3960.0
51.00	13638.0
52.00	538.0
55.00	249.0
55.90	595.0
57.00	1061.0
60.90	235.0
61.90	217.0
63.00	611.0
65.00	207.0
68.00	352.0
68.90	15522.0
73.00	395.0
74.00	1780.0
75.00	2461.0
76.00	1008.0
77.00	16432.0
78.00	1134.0
78.90	1267.0
80.00	1021.0
80.90	1463.0
81.90	416.0
82.90	372.0
83.90	1885.0
84.90	382.0
85.90	1371.0
86.90	153.0
87.90	213.0
91.00	369.0
92.00	422.0
92.90	2192.0
93.90	201.0
98.00	1964.0
98.90	1229.0
99.90	200.0
100.90	697.0
102.90	274.0
103.90	552.0
104.90	578.0
105.90	172.0
106.90	5484.0
108.00	728.0
108.90	229.0
109.90	9032.0
110.90	1393.0
112.00	164.0
116.00	406.0
116.90	4939.0
117.90	417.0
121.90	369.0
122.90	672.0
123.90	212.0
124.90	203.0
126.90	16496.0
127.90	1634.0
128.90	7386.0
130.00	912.0
133.90	243.0
134.90	543.0
135.90	261.0
136.90	349.0
139.80	174.0
140.90	1111.0
142.00	343.0
143.00	303.0
145.90	205.0
146.90	630.0
147.90	1445.0
149.00	272.0
152.90	397.0
153.90	330.0
154.90	706.0
155.90	854.0
157.90	269.0
158.90	182.0
159.90	227.0
160.90	579.0
164.90	463.0
165.80	389.0
166.90	2007.0
167.90	796.0
169.00	189.0
171.90	173.0
172.90	354.0

Instrument :
BNA_G
ClientSampleId :

174.00	451.0
174.90	800.0
175.90	188.0
176.90	490.0
177.80	178.0
178.90	1680.0
180.00	1014.0
180.90	419.0
185.00	854.0
185.90	5503.0
186.90	1541.0
187.90	198.0
188.90	355.0
190.90	253.0
191.90	485.0
192.90	504.0
195.80	757.0
196.50	231.0
197.90	32224.0
198.90	2545.0
201.20	193.0
202.90	454.0
203.90	1540.0
204.90	2475.0
205.90	9503.0
206.90	1683.0
207.90	445.0
210.80	442.0
214.90	199.0
215.80	210.0
216.90	2882.0
217.90	375.0
220.90	1079.0
221.90	366.0
222.90	679.0
223.90	5533.0
224.90	1522.0
226.90	2676.0
227.90	441.0
228.90	451.0
230.90	216.0
233.80	209.0
234.90	373.0
240.90	164.0
241.90	261.0
242.90	285.0
243.90	3803.0
244.90	566.0
245.90	1101.0
249.00	222.0
254.90	21376.0
255.90	3076.0
256.90	287.0
257.90	1501.0
258.90	271.0
264.90	737.0
271.90	160.0
272.90	706.0
273.90	1411.0
274.90	8646.0
275.90	1265.0
276.90	874.0
277.90	180.0
284.90	222.0
292.90	215.0
295.90	3506.0
296.90	351.0
302.80	342.0
313.90	157.0
314.90	367.0
322.90	770.0
326.90	171.0
334.00	506.0
335.10	155.0
345.90	191.0
352.00	222.0
353.90	284.0
364.90	1486.0
371.90	392.0
402.90	197.0
421.90	155.0
422.90	1014.0
423.90	164.0
440.90	2423.0
441.90	17608.0
442.90	3121.0
443.90	446.0

Instrument :
BNA_G
ClientSampleId :

m/z	Abundance
39.00	1012.0
41.00	194.0
46.90	460.0
49.00	1844.0
50.00	4462.0
51.00	16326.0
51.90	804.0
55.00	287.0
56.00	391.0
57.00	1225.0
60.90	243.0
61.90	292.0
63.00	724.0
65.00	312.0
68.90	16896.0
70.00	156.0
73.00	369.0
74.00	2134.0
75.00	3171.0
76.00	1113.0
77.00	18328.0
78.00	1377.0
78.90	1339.0
79.90	1135.0
80.90	1406.0
81.90	309.0
83.00	406.0
83.90	1656.0
84.90	314.0
85.90	1499.0
86.90	340.0
87.90	311.0
90.90	392.0
91.90	526.0
93.00	2668.0
94.00	171.0
96.00	191.0
97.90	2319.0
98.90	1430.0
101.00	1060.0
102.90	287.0
103.90	535.0
104.90	548.0
106.00	165.0
107.00	6698.0
107.90	1056.0
109.90	10730.0
110.90	1781.0
115.90	501.0
116.90	6203.0
117.90	433.0
121.90	531.0
122.90	790.0
123.90	332.0
124.90	288.0
126.90	21040.0
127.90	1634.0
128.90	9408.0
129.90	867.0
133.90	304.0
134.90	1004.0
136.00	339.0
136.90	415.0
139.90	201.0
140.90	1275.0
141.90	407.0
143.00	282.0
146.00	258.0
146.90	892.0
147.90	1642.0
148.90	331.0
150.90	252.0
153.00	501.0
153.90	253.0
154.90	660.0
155.90	1026.0
157.00	232.0
157.90	178.0
158.80	197.0
159.90	455.0
160.90	641.0
161.80	167.0
164.90	494.0
165.90	419.0
166.90	2401.0
168.00	1133.0
168.90	255.0
171.90	243.0
172.90	347.0
174.00	562.0
175.00	1060.0

Instrument :
BNA_G
ClientSampleId :

176.00	189.0
176.90	410.0
177.90	185.0
178.90	2108.0
179.90	1313.0
180.90	622.0
183.90	178.0
184.90	1047.0
185.90	7097.0
187.00	1936.0
188.00	222.0
188.90	592.0
190.90	196.0
191.90	612.0
192.90	699.0
195.90	1275.0
197.90	42640.0
198.90	2722.0
199.90	204.0
201.20	205.0
202.90	346.0
203.90	1965.0
204.90	3163.0
205.90	13015.0
206.90	1904.0
207.90	461.0
208.80	188.0
209.90	262.0
210.90	500.0
214.90	214.0
215.90	398.0
216.90	3950.0
217.90	445.0
220.90	1908.0
221.60	394.0
221.90	411.0
222.90	904.0
223.90	7770.0
224.90	2095.0
225.90	231.0
226.90	4163.0
227.90	500.0
228.90	800.0
230.00	150.0
230.90	315.0
233.80	254.0
234.90	221.0
236.80	194.0
240.80	212.0
242.00	331.0
242.90	417.0
243.90	5086.0
244.90	723.0
245.90	1319.0
246.90	320.0
252.90	154.0
254.90	31552.0
255.90	4275.0
256.90	378.0
257.90	2052.0
258.90	309.0
264.90	981.0
272.90	1000.0
273.90	2565.0
274.90	12084.0
275.90	1639.0
276.90	1360.0
277.80	229.0
282.90	208.0
283.90	168.0
285.00	330.0
292.90	290.0
295.90	4478.0
296.90	611.0
302.90	303.0
313.80	160.0
314.80	605.0
315.80	244.0
322.90	1154.0
323.90	217.0
326.90	260.0
333.90	1005.0
335.00	164.0
345.90	237.0
351.90	418.0
352.90	286.0
353.90	352.0
364.90	1953.0
365.80	217.0
371.90	556.0
382.80	165.0
401.80	238.0
402.80	258.0
403.90	183.0
420.90	211.0

Instrument :
BNA_G
ClientSampleId :

421.80	231.0
422.90	1518.0
423.90	223.0
440.90	4034.0
441.90	25888.0
442.90	5383.0
443.90	548.0

Instrument :
BNA_G
ClientSampleId :

m/z	Abundance
37.00	253.0
39.00	690.0
41.00	212.0
43.00	170.0
47.00	361.0
47.90	180.0
48.90	1692.0
50.00	2585.0
51.00	9981.0
52.00	514.0
55.90	342.0
57.00	706.0
61.00	174.0
61.90	165.0
63.00	419.0
65.00	292.0
68.90	11199.0
73.00	322.0
74.00	1430.0
74.90	1673.0
76.00	680.0
77.00	11898.0
78.00	700.0
78.90	868.0
79.90	866.0
81.00	1150.0
81.90	353.0
82.90	381.0
83.90	1690.0
84.90	214.0
85.90	1320.0
86.90	151.0
87.90	212.0
90.90	237.0
91.90	241.0
92.90	1783.0
98.00	1426.0
98.90	843.0
100.90	671.0
102.90	238.0
103.90	396.0
104.90	369.0
107.00	4408.0
107.90	587.0
109.90	7078.0
110.90	1052.0
112.00	176.0
115.90	234.0
117.00	4053.0
117.90	349.0
122.00	220.0
122.90	472.0
123.90	253.0
124.90	257.0
126.90	13655.0
127.90	1012.0
128.90	6398.0
129.90	609.0
131.10	155.0
134.90	591.0
135.90	160.0
136.90	420.0
140.90	1133.0
141.90	278.0
143.00	207.0
145.80	194.0
146.90	503.0
147.90	1022.0
148.90	242.0
152.90	351.0
154.90	574.0
155.90	757.0
156.90	173.0
159.90	310.0
160.90	422.0
164.80	446.0
165.90	302.0
166.90	1555.0
167.90	772.0
168.90	157.0
171.90	232.0
172.90	220.0
173.90	324.0
174.90	584.0
175.90	198.0
176.80	346.0
178.00	157.0
178.90	1549.0
179.90	911.0
180.90	379.0
184.90	692.0

Instrument :
BNA_G
ClientSampleId :

185.90	4669.0
186.90	1559.0
188.90	499.0
190.90	220.0
191.90	521.0
193.00	585.0
195.90	882.0
197.90	31384.0
198.80	2144.0
199.90	210.0
202.90	272.0
203.90	1425.0
204.90	2161.0
205.90	9649.0
206.90	1461.0
207.80	387.0
210.90	362.0
215.00	167.0
215.80	246.0
216.90	3112.0
217.90	325.0
220.90	1093.0
221.90	358.0
222.90	565.0
223.90	5479.0
225.00	1388.0
226.00	171.0
226.90	2967.0
228.00	404.0
228.90	599.0
233.80	175.0
234.80	223.0
236.80	214.0
238.90	208.0
241.90	369.0
242.90	282.0
244.00	4319.0
244.90	505.0
245.90	879.0
246.90	319.0
248.90	213.0
254.90	23640.0
255.90	3187.0
256.90	267.0
257.90	1947.0
258.90	297.0
264.90	724.0
272.90	675.0
273.90	1796.0
274.90	9027.0
275.90	1300.0
276.90	1094.0
277.90	208.0
284.90	240.0
292.90	279.0
295.90	3496.0
296.80	499.0
302.90	331.0
313.90	265.0
314.90	422.0
315.90	160.0
320.90	172.0
322.90	835.0
326.90	188.0
333.90	583.0
340.80	154.0
345.90	173.0
351.90	353.0
352.90	262.0
353.90	263.0
364.90	1655.0
365.90	188.0
370.90	176.0
371.90	445.0
402.80	199.0
420.80	157.0
421.90	201.0
422.90	1320.0
440.90	3884.0
441.90	24896.0
442.90	4785.0
443.90	556.0

Instrument :
BNA_G
ClientSampleId :