

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG032822\
 Data File : BG052927.D
 Acq On : 29 Mar 2022 15:09
 Operator : CG/JU
 Sample : N2097-05
 Misc :
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 C0AD4

Quant Time: Mar 30 01:09:32 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG031822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 23 13:16:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.135	152	28262	20.000	ng/u1	-0.01
20) Naphthalene-d8	10.949	136	127723	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.755	164	105043	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.499	188	256123	20.000	ng/u1	0.00
79) Chrysene-d12	21.781	240	274942	20.000	ng/u1	#-0.01
88) Perylene-d12	25.094	264	265454	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.547	96	3190	3.987	ng/uL	0.00
4) Pyridine-d5	3.970	84	19085	9.555	ng/u1	0.00
7) Phenol-d5	7.277	99	21680	8.565	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.448	67	54036	34.952	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.665	132	46668	27.152	ng/u1	-0.01
15) 4-Methylphenol-d8	8.822	113	37532	20.469	ng/u1	-0.01
21) Nitrobenzene-d5	9.292	128	33990	36.970	ng/u1	0.00
24) 2-Nitrophenol-d4	10.021	143	37843	39.024	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.561	165	64972	30.797	ng/u1	0.00
31) 4-Chloroaniline-d4	11.072	131	87190	31.270	ng/u1	0.00
46) Dimethylphthalate-d6	14.150	166	292174	36.206	ng/u1	-0.01
49) Acenaphthylene-d8	14.450	160	320605	32.695	ng/u1	0.00
54) 4-Nitrophenol-d4	14.926	143	8909	6.566	ng/u1	0.00
60) Fluorene-d10	15.742	176	259441	36.615	ng/u1	-0.01
65) 4,6-Dinitro-2-methylph...	15.848	200	43019	33.005	ng/u1	-0.01
73) Anthracene-d10	17.599	188	455904	38.226	ng/u1	0.00
81) Pyrene-d10	19.878	212	554108	36.026	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.859	264	550024	40.101	ng/u1	-0.02

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG032822\
Data File : BG052927.D
Acq On : 29 Mar 2022 15:09
Operator : CG/JU
Sample : N2097-05
Misc :
ALS Vial : 46 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
C0AD4

Quant Time: Mar 30 01:09:32 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG031822.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Mar 23 13:16:32 2022
Response via : Initial Calibration

