

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG032822\
 Data File : BG052921.D
 Acq On : 29 Mar 2022 11:18
 Operator : CG/JU
 Sample : N2097-10
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 C0AD9

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 03/30/2022
 Supervised By :Yogesh Patel 04/02/2022

Quant Time: Mar 30 01:08:29 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG031822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 23 13:16:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.134	152	25123	20.000	ng/u1	-0.01
20) Naphthalene-d8	10.948	136	116775	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.755	164	96125	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.498	188	235156	20.000	ng/u1	0.00
79) Chrysene-d12	21.780	240	250703	20.000	ng/u1	-0.01
88) Perylene-d12	25.088	264	250672	20.000	ng/u1	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.552	96	3159	4.442	ng/uL	0.00
4) Pyridine-d5	3.975	84	11391m	6.415	ng/u1	0.00
7) Phenol-d5	7.282	99	15453	6.868	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.453	67	44843	32.630	ng/u1	0.00
11) 2-Chlorophenol-d4	7.664	132	37024	24.232	ng/u1	-0.01
15) 4-Methylphenol-d8	8.827	113	28220	17.313	ng/u1	0.00
21) Nitrobenzene-d5	9.291	128	28526	33.936	ng/u1	0.00
24) 2-Nitrophenol-d4	10.020	143	29938	33.767	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.560	165	52895	27.423	ng/u1	0.00
31) 4-Chloroaniline-d4	11.071	131	65329	25.627	ng/u1	0.00
46) Dimethylphthalate-d6	14.149	166	236717	32.055	ng/u1	-0.01
49) Acenaphthylene-d8	14.449	160	265015	29.533	ng/u1	0.00
54) 4-Nitrophenol-d4	14.925	143	8489m	6.837	ng/u1	0.00
60) Fluorene-d10	15.741	176	217877	33.601	ng/u1	-0.01
65) 4,6-Dinitro-2-methylph...	15.847	200	41069	34.318	ng/u1	-0.01
73) Anthracene-d10	17.598	188	393708m	35.954	ng/u1	0.00
81) Pyrene-d10	19.877	212	485216	34.597	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.864	264	466406	36.010	ng/u1	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG032822\
Data File : BG052921.D
Acq On : 29 Mar 2022 11:18
Operator : CG/JU
Sample : N2097-10
Misc :
ALS Vial : 40 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
C0AD9

Quant Time: Mar 30 01:08:29 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG031822.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Mar 23 13:16:32 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 03/30/2022
Supervised By :Yogesh Patel 04/02/2022

