

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG032916\
 Data File : BG021572.D
 Acq On : 30 Mar 2016 4:24
 Operator : UM/SJ
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02026

Quant Time: Mar 30 05:09:58 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG031616.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 30 01:34:35 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2	1,4-Dioxane	0.602	0.664	-10.3	99	0.00
3 S	1,4-Dioxane-d8	0.446	0.533	-19.5	94	0.00
4	Benzaldehyde	1.152	1.129	2.0	83	0.00
5 S	Phenol-d5	1.891	1.642	13.2	80	0.00
6	Phenol	1.999	1.743	12.8	83	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.120	1.026	8.4	86	0.00
8	Bis(2-Chloroethyl)ether	1.447	1.295	10.5	81	0.00
9 S	2-Chlorophenol-d4	1.436	1.338	6.8	87	0.00
10	2-Chlorophenol	1.476	1.387	6.0	87	0.00
11	2-Methylphenol	1.538	1.296	15.7	80	0.00
12	2,2'-oxybis(1-Chloropropane	1.745	1.513	13.3	80	0.00
13 S	4-Methylphenol-d8	1.619	1.515	6.4	89	0.00
14	Acetophenone	2.642	2.408	8.9	87	0.00
15 P	N-Nitroso-di-n-propylamine	1.465	1.308	10.7	81	0.00
16	4-Methylphenol	1.745	1.517	13.1	83	0.00
17	Hexachloroethane	0.619	0.631	-1.9	94	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	95	0.00
19 S	Nitrobenzene-d5	0.150	0.141	6.0	83	0.00
20	Nitrobenzene	0.424	0.416	1.9	86	0.00
21	Isophorone	0.817	0.739	9.5	81	0.00
22 S	2-Nitrophenol-d4	0.172	0.157	8.7	82	0.00
23 C	2-Nitrophenol	0.189	0.194	-2.6	91	0.00
24	2,4-Dimethylphenol	0.427	0.393	8.0	81	0.00
25	Bis(2-Chloroethoxy)methane	0.426	0.380	10.8	79	0.00
26 S	2,4-Dichlorophenol-d3	0.307	0.288	6.2	84	0.00
27 C	2,4-Dichlorophenol	0.318	0.309	2.8	87	0.00
28	Naphthalene	1.060	1.005	5.2	85	0.00
29 S	4-Chloroaniline-d4	0.401	0.381	5.0	79	0.00
30	4-Chloroaniline	0.421	0.413	1.9	85	0.00
31 C	Hexachlorobutadiene	0.234	0.224	4.3	85	0.00
32	Caprolactam	0.120	0.106	11.7	84	0.00
33 C	4-Chloro-3-methylphenol	0.389	0.360	7.5	83	0.00
34	2-Methylnaphthalene	0.787	0.729	7.4	82	0.00
35	1-Methylnaphthalene	0.764	0.709	7.2	83	0.00
36 I	Acenaphthene-d10	1.000	1.000	0.0	96	0.00
37	1,2,4,5-Tetrachlorobenzene	0.700	0.671	4.1	86	0.00
38	Hexachlorocyclopentadiene	0.256	0.151	41.0#	75	0.00
39 C	2,4,6-Trichlorophenol	0.412	0.372	9.7	82	0.00
40	2,4,5-Trichlorophenol	0.449	0.430	4.2	87	0.00
41	1,1'-Biphenyl	1.636	1.505	8.0	82	0.00
42	2-Chloronaphthalene	1.280	1.203	6.0	85	0.00
43	2-Nitroaniline	0.446	0.423	5.2	85	0.00
44 S	Dimethylphthalate-d6	1.665	1.559	6.4	85	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG032916\
 Data File : BG021572.D
 Acq On : 30 Mar 2016 4:24
 Operator : UM/SJ
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02026

Quant Time: Mar 30 05:09:58 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG031616.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 30 01:34:35 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	Dimethylphthalate	1.730	1.624	6.1	86	0.00
46	2,6-Dinitrotoluene	0.353	0.346	2.0	87	0.00
47 S	Acenaphthylene-d8	2.053	1.927	6.1	85	0.00
48	Acenaphthylene	2.034	1.903	6.4	84	0.00
49	3-Nitroaniline	0.327	0.338	-3.4	90	0.00
50 C	Acenaphthene	1.403	1.337	4.7	86	0.00
51	2,4-Dinitrophenol	0.178	0.112	37.1#	65	0.00
52 S	4-Nitrophenol-d4	0.266	0.161	39.5#	70	0.00
53	4-Nitrophenol	0.276	0.208	24.6#	70	0.00
54	Dibenzofuran	1.965	1.853	5.7	86	0.00
55	2,4-Dinitrotoluene	0.518	0.512	1.2	88	0.00
56	2,3,4,6-Tetrachlorophenol	0.328	0.297	9.5	81	0.00
57	Diethylphthalate	1.849	1.717	7.1	85	0.00
58 S	Fluorene-d10	1.501	1.403	6.5	85	0.00
59	Fluorene	1.718	1.620	5.7	85	0.00
60	4-Chlorophenyl-phenylether	0.902	0.845	6.3	86	0.00
61	4-Nitroaniline	0.342	0.348	-1.8	91	0.00
62 I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
63 S	4,6-Dinitro-2-methylphenol-	0.132	0.104	21.2#	79	0.00
64	4,6-Dinitro-2-methylphenol	0.140	0.111	20.7#	80	0.00
65	N-Nitrosodiphenylamine	0.620	0.582	6.1	90	0.00
66	4-Bromophenyl-phenylether	0.234	0.218	6.8	89	0.00
67	Hexachlorobenzene	0.246	0.235	4.5	89	0.00
68	Atrazine	0.250	0.241	3.6	92	0.00
69 C	Pentachlorophenol	0.065	0.043	33.8#	73	0.00
70	Phenanthrene	1.133	1.069	5.6	89	0.00
71 S	Anthracene-d10	0.980	0.930	5.1	90	0.00
72	Anthracene	1.152	1.101	4.4	90	0.00
73	1,2,3,4-Tetrachlorobenzene	0.280	0.266	5.0	89	0.00
74	Pentachlorobenzene	0.291	0.255	12.4	82	0.00
75	Carbazole	0.997	0.973	2.4	92	0.00
76	Di-n-butylphthalate	1.286	1.276	0.8	93	0.00
77 C	Fluoranthene	1.367	1.351	1.2	93	0.00
78 I	Chrysene-d12	1.000	1.000	0.0	107	0.00
79 S	Pyrene-d10	0.933	0.871	6.6	91	0.00
80	Pyrene	1.192	1.128	5.4	93	0.00
81	Butylbenzylphthalate	0.503	0.472	6.2	94	0.00
82	3,3'-Dichlorobenzidine	0.415	0.399	3.9	91	0.00
83	Benzo(a)anthracene	1.207	1.148	4.9	92	-0.01
84	Bis(2-ethylhexyl)phthalate	0.736	0.693	5.8	92	0.00
85	Chrysene	1.143	1.089	4.7	94	0.00
86 I	Perylene-d12	1.000	1.000	0.0	108	0.00
87	Di-n-octyl phthalate	1.358	1.241	8.6	93	0.00

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG032916\
Data File : BG021572.D
Acq On : 30 Mar 2016 4:24
Operator : UM/SJ
Sample : SSTDCCC020EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTD02026

Quant Time: Mar 30 05:09:58 2016
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG031616.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Mar 30 01:34:35 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.285	1.206	6.1	93	0.00
89	Benzo(k)fluoranthene	1.262	1.144	9.4	92	0.00
90 S	Benzo(a)pyrene-d12	0.923	0.847	8.2	92	0.00
91 C	Benzo(a)pyrene	1.247	1.147	8.0	92	0.00
92	Indeno(1,2,3-cd)pyrene	1.417	1.300	8.3	92	-0.01
93	Dibenzo(a,h)anthracene	1.206	1.100	8.8	92	-0.01
94	Benzo(g,h,i)perylene	1.166	1.068	8.4	91	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 1