


Data Path : Z:\HPCHEM1\BNA_G\Data\BG032917\
 Data File : BG026477.D
 Acq On : 29 Mar 2017 15:26
 Operator : SJ/MA
 Sample : DFTPP21
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP21

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG032717MA.M
 Title : SVOA CALIBRATION
 Last Update : Thu Mar 30 07:51:13 2017



AutoFind: Scans 2504, 2505, 2506; Background Corrected with Scan 2499

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	22.5	42869	PASS
68	69	0.00	2	0.6	271	PASS
69	198	0.00	100	24.8	47173	PASS
70	69	0.00	2	0.7	328	PASS
127	198	10	80	39.8	75765	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	190378	PASS
199	198	5	9	7.1	13606	PASS
275	198	10	60	25.9	49397	PASS
365	198	1	100	3.0	5666	PASS
441	442	0.01	100	15.6	23684	PASS
442	198	50	100	79.9	152082	PASS
443	442	15	24	19.4	29548	PASS

Scan 2499 (17.770 min): BG026477.D
DFTPP21

m/z	Abundance
43.10	193.0
44.10	1681.0
73.00	172.0
207.10	911.0
209.10	239.0

Instrument :
BNA_G
ClientSampleId :
DFTPP21

m/z	Abundance
38.10	561.0
39.10	1902.0
41.10	155.0
44.10	1465.0
49.00	367.0
50.10	9609.0
51.10	34944.0
52.10	2076.0
55.10	381.0
56.00	1042.0
57.10	2804.0
58.00	326.0
61.10	410.0
62.10	842.0
63.00	2178.0
64.10	191.0
65.10	989.0
68.10	814.0
69.00	36904.0
70.10	295.0
73.00	427.0
74.10	4464.0
75.10	7056.0
76.20	2102.0
77.10	50152.0
78.10	4219.0
79.00	2980.0
80.10	2530.0
81.00	3187.0
82.10	959.0
83.00	661.0
84.10	259.0
85.00	834.0
86.10	993.0
87.00	310.0
91.00	660.0
92.10	1173.0
93.10	5085.0
94.00	390.0
96.10	336.0
98.10	3573.0
99.10	4192.0
100.10	438.0
101.00	2047.0
102.10	156.0
103.10	478.0
104.00	1274.0
105.00	805.0
107.10	14750.0
108.10	2531.0
109.10	525.0
110.10	30336.0
111.10	4335.0
112.10	746.0
113.10	196.0
114.00	151.0
114.90	192.0
116.10	863.0
117.10	9391.0
118.10	607.0
119.00	341.0
120.20	214.0
122.10	1116.0
123.00	1530.0
124.00	1013.0
125.10	497.0
127.10	57672.0
128.10	4379.0
129.00	20944.0
130.00	1926.0
131.00	520.0
132.00	153.0
134.00	663.0
135.00	2087.0
136.10	617.0
137.00	1269.0
138.00	217.0
140.00	417.0
141.00	2041.0
142.00	797.0
143.00	677.0
144.80	235.0
146.00	822.0
147.00	1622.0
148.00	3253.0
149.00	1323.0
150.00	278.0
151.00	312.0
151.80	433.0
153.00	755.0
154.00	913.0

Instrument :
BNA_G
ClientSampleId :
DFTPP21

155.00	2329.0
156.10	2627.0
157.10	383.0
158.10	670.0
159.10	398.0
160.00	796.0
161.00	1173.0
162.20	716.0
164.00	273.0
165.00	1108.0
166.10	1226.0
167.10	5412.0
168.00	4651.0
169.10	1094.0
169.90	342.0
171.10	242.0
172.00	638.0
173.00	758.0
174.00	1362.0
175.10	2058.0
176.10	684.0
177.00	920.0
178.20	193.0
179.00	4785.0
180.00	3490.0
181.10	1450.0
182.80	159.0
184.10	637.0
185.10	1889.0
186.10	15495.0
187.10	5108.0
188.10	587.0
189.00	844.0
190.00	295.0
191.00	585.0
192.10	1218.0
193.10	1170.0
194.10	391.0
196.10	3396.0
198.00	133376.0
199.00	8697.0
200.10	581.0
201.60	543.0
203.20	794.0
204.10	4733.0
205.10	6595.0
206.10	32048.0
207.10	5136.0
208.10	994.0
209.00	362.0
210.00	268.0
211.00	1348.0
215.00	464.0
216.00	798.0
217.00	7963.0
218.00	1179.0
221.00	5873.0
223.10	1988.0
224.10	17472.0
225.10	4997.0
226.00	413.0
227.10	7266.0
228.00	813.0
229.00	1601.0
230.20	207.0
231.10	573.0
231.90	170.0
234.00	369.0
235.10	453.0
235.90	222.0
237.10	713.0
239.00	421.0
239.90	213.0
241.00	450.0
242.10	853.0
243.10	907.0
244.10	13069.0
245.10	2034.0
246.00	2332.0
247.00	601.0
249.00	573.0
252.40	182.0
253.00	298.0
255.10	66264.0
256.00	9863.0
257.10	1116.0
258.00	4048.0
259.10	551.0
261.00	162.0
264.00	159.0
265.10	1548.0
266.00	273.0
271.10	253.0
273.10	2029.0
274.10	4482.0

Instrument :
BNA_G
ClientSampleId :
DFTPP21

275.10	31840.0
276.10	4098.0
277.00	1924.0
278.00	373.0
283.00	220.0
285.10	675.0
289.10	167.0
292.30	231.0
293.00	639.0
294.40	173.0
296.00	6575.0
297.10	1138.0
301.20	158.0
303.10	1012.0
304.00	237.0
310.00	203.0
314.00	594.0
315.10	685.0
316.00	511.0
321.00	387.0
322.00	207.0
323.10	2501.0
324.00	511.0
327.00	395.0
332.00	233.0
332.90	214.0
334.10	1648.0
335.10	305.0
341.10	351.0
346.20	476.0
352.10	888.0
353.00	516.0
354.10	754.0
355.20	204.0
365.00	4036.0
366.00	661.0
371.00	167.0
372.10	1832.0
373.00	327.0
383.10	521.0
391.10	152.0
402.10	622.0
403.10	428.0
404.10	246.0
405.10	279.0
421.10	405.0
422.10	711.0
423.10	4365.0
424.00	1183.0
425.00	262.0
441.10	12005.0
442.10	74680.0
443.10	15109.0
444.10	1329.0

Instrument :
BNA_G
ClientSampleId :
DFTPP21

m/z	Abundance
37.10	266.0
38.10	844.0
39.10	3712.0
40.00	165.0
44.10	1446.0
49.10	530.0
50.10	12831.0
51.10	48576.0
52.10	2415.0
55.10	407.0
56.00	1546.0
57.10	3967.0
58.10	247.0
61.10	827.0
62.10	1091.0
63.10	2615.0
64.10	356.0
65.10	1213.0
69.00	54112.0
70.10	416.0
73.00	485.0
74.10	5399.0
75.10	9783.0
76.10	3504.0
77.10	74104.0
78.10	5672.0
79.00	3936.0
80.00	3026.0
81.00	4372.0
82.10	1346.0
83.10	1326.0
84.00	182.0
85.10	1053.0
86.00	1266.0
87.00	701.0
88.00	180.0
89.20	189.0
91.10	840.0
92.00	1358.0
93.00	6709.0
94.00	574.0
96.10	386.0
98.10	4794.0
99.10	5044.0
100.10	438.0
101.00	3114.0
102.10	301.0
103.00	754.0
104.10	1636.0
105.10	1863.0
106.10	857.0
107.10	21584.0
108.00	4073.0
109.10	859.0
110.00	44752.0
111.10	6647.0
112.00	711.0
113.10	262.0
115.00	155.0
116.10	1308.0
117.00	14607.0
118.10	1343.0
119.10	217.0
120.20	207.0
122.10	1408.0
123.00	2103.0
124.00	1067.0
125.10	1091.0
127.10	87824.0
128.10	7545.0
129.00	31472.0
130.10	2406.0
131.00	502.0
132.30	336.0
133.10	234.0
134.10	908.0
135.00	2799.0
136.00	1170.0
137.00	2075.0
137.90	508.0
138.50	154.0
139.10	157.0
140.00	526.0
141.00	4402.0
142.10	1204.0
143.10	929.0
144.00	436.0
145.00	490.0
146.00	1093.0
147.00	2426.0
148.10	4509.0

Instrument :
BNA_G
ClientSampleId :
DFTPP21

149.00	1020.0
150.10	390.0
151.20	853.0
152.10	429.0
153.10	1700.0
154.00	1076.0
155.00	2636.0
156.10	3922.0
157.10	943.0
158.00	595.0
159.10	596.0
160.00	1619.0
161.00	2013.0
162.00	637.0
163.00	311.0
164.00	497.0
165.10	2154.0
166.10	1492.0
167.10	10241.0
168.10	5732.0
169.10	1096.0
169.90	321.0
171.00	492.0
172.00	1071.0
173.10	1129.0
174.10	2139.0
175.10	3699.0
176.20	1012.0
177.00	1777.0
178.10	734.0
179.00	7302.0
180.00	5445.0
181.10	2223.0
182.10	358.0
183.00	307.0
183.50	437.0
184.10	476.0
185.10	3452.0
186.10	25496.0
187.10	6934.0
188.10	1116.0
189.10	1291.0
190.00	214.0
191.00	757.0
192.00	2093.0
193.10	2053.0
194.10	767.0
195.10	407.0
196.00	5592.0
198.00	219264.0
199.00	17120.0
200.10	1218.0
201.10	455.0
201.80	681.0
203.00	1309.0
204.10	7516.0
205.10	11859.0
206.10	52376.0
207.10	7740.0
208.00	1694.0
209.00	656.0
210.10	997.0
211.00	1823.0
211.80	665.0
212.00	666.0
213.00	217.0
215.00	529.0
216.00	1043.0
217.00	13263.0
218.00	1982.0
219.00	351.0
219.80	340.0
221.10	9042.0
223.00	2995.0
224.10	30080.0
225.10	7898.0
226.10	1353.0
227.00	11553.0
228.10	1901.0
229.00	2502.0
229.90	219.0
231.10	1237.0
232.10	171.0
234.00	693.0
235.00	930.0
236.00	851.0
237.00	1129.0
238.00	237.0
239.10	555.0
240.00	453.0
241.00	690.0
242.10	1214.0
243.10	1536.0
244.10	23904.0
245.10	3140.0

Instrument :
BNA_G
ClientSampleId :
DFTPP21

246.10	4424.0
247.10	626.0
247.90	329.0
249.00	1069.0
250.90	240.0
252.00	283.0
253.10	683.0
255.10	125344.0
256.10	17576.0
257.10	1331.0
258.10	6343.0
259.00	1258.0
260.00	283.0
261.00	351.0
264.20	212.0
265.00	2253.0
266.00	201.0
271.00	341.0
272.00	423.0
273.10	3655.0
274.10	9022.0
275.10	58056.0
276.10	7690.0
277.00	4386.0
278.10	677.0
279.00	158.0
282.00	177.0
283.00	558.0
284.10	476.0
285.10	660.0
286.10	190.0
290.10	187.0
292.10	179.0
293.00	1181.0
294.00	237.0
296.00	13023.0
297.00	2112.0
298.10	234.0
301.10	161.0
302.00	226.0
303.10	1286.0
304.10	402.0
308.20	187.0
309.10	205.0
310.10	160.0
314.00	668.0
315.10	1623.0
316.10	1105.0
316.90	296.0
321.10	623.0
322.10	215.0
323.10	4796.0
324.00	674.0
327.00	1506.0
328.00	353.0
332.10	436.0
333.00	436.0
334.10	3432.0
335.00	793.0
336.20	164.0
339.20	150.0
340.00	154.0
341.20	736.0
346.00	878.0
347.10	220.0
351.00	189.0
352.10	1615.0
353.10	1358.0
354.10	1489.0
355.00	547.0
358.10	185.0
358.80	155.0
365.00	5444.0
366.00	910.0
371.00	368.0
372.10	2286.0
373.00	590.0
383.00	633.0
384.20	206.0
390.00	273.0
392.00	213.0
401.00	254.0
402.00	924.0
403.00	1229.0
404.10	674.0
421.00	1286.0
422.10	1241.0
423.10	9108.0
424.10	2445.0
425.00	562.0
437.90	319.0
439.10	157.0
441.10	26400.0
442.10	166848.0
443.10	33352.0

Instrument :
BNA_G
ClientSampleId :
DFTPP21

444.10

3512.0

Instrument :
BNA_G
ClientSampleId :
DFTPP21

m/z	Abundance
37.10	232.0
38.10	314.0
39.10	2995.0
40.10	272.0
41.10	188.0
43.10	171.0
44.10	1096.0
45.00	206.0
49.10	277.0
50.10	10628.0
51.10	45088.0
52.10	2379.0
55.00	270.0
56.00	1416.0
57.10	3711.0
61.10	701.0
62.10	981.0
63.10	2197.0
64.10	349.0
65.10	1286.0
66.10	231.0
67.30	222.0
69.00	50504.0
70.10	273.0
71.20	206.0
73.10	635.0
74.10	5611.0
75.10	8253.0
76.10	3324.0
77.10	63928.0
78.10	3904.0
79.00	3479.0
80.00	2790.0
81.10	4662.0
82.00	1200.0
83.00	794.0
84.10	274.0
85.10	759.0
86.00	1466.0
87.10	607.0
88.10	299.0
91.10	1107.0
92.10	1263.0
93.10	5998.0
94.10	294.0
95.00	181.0
96.10	432.0
98.10	4545.0
99.10	4378.0
100.00	492.0
101.00	2835.0
103.10	737.0
104.10	1783.0
104.90	1522.0
107.10	19640.0
108.00	2909.0
110.10	39992.0
111.10	5984.0
112.00	744.0
113.00	338.0
116.10	798.0
117.00	12586.0
118.00	1252.0
119.00	344.0
120.10	197.0
122.10	1286.0
123.00	2184.0
124.10	906.0
125.10	1166.0
127.10	81800.0
128.10	6492.0
129.10	28440.0
130.10	2655.0
131.00	714.0
132.10	248.0
134.00	873.0
135.00	2459.0
136.00	988.0
137.10	1672.0
138.00	396.0
139.00	389.0
140.00	339.0
141.00	3689.0
142.10	878.0
143.00	955.0
144.30	283.0
145.00	376.0
146.00	769.0
147.10	1350.0
148.00	4519.0
149.00	1278.0

Instrument :
BNA_G
ClientSampleId :
DFTPP21

150.10	290.0
151.10	532.0
152.00	455.0
153.00	1644.0
154.10	972.0
155.10	2720.0
156.10	3738.0
157.00	641.0
158.10	819.0
159.00	726.0
160.00	1279.0
161.10	1971.0
162.10	362.0
164.20	267.0
165.10	1453.0
166.10	1598.0
167.10	8551.0
168.10	5832.0
169.10	972.0
170.20	280.0
171.00	396.0
172.00	900.0
173.00	1350.0
174.10	2118.0
175.10	3671.0
176.10	1155.0
177.00	2428.0
178.00	609.0
179.00	7464.0
180.00	4336.0
181.10	2225.0
182.10	181.0
182.70	449.0
183.10	436.0
184.10	594.0
185.10	3230.0
186.10	25840.0
187.10	7290.0
188.10	733.0
189.10	1862.0
190.00	291.0
191.10	639.0
192.00	2566.0
193.10	2228.0
194.00	901.0
195.20	562.0
196.00	6201.0
198.00	218496.0
199.00	15002.0
200.10	1057.0
201.50	904.0
203.00	1152.0
204.10	7057.0
205.10	11541.0
206.10	54816.0
207.10	8160.0
208.00	1599.0
209.10	587.0
211.00	1920.0
212.00	563.0
215.10	534.0
216.00	1254.0
217.00	13430.0
217.90	1309.0
219.00	257.0
221.10	9139.0
223.00	3644.0
224.10	28936.0
225.10	8693.0
226.00	942.0
227.10	10809.0
228.00	1894.0
229.00	2376.0
230.00	361.0
231.00	1041.0
232.00	259.0
233.00	198.0
234.00	739.0
235.00	598.0
236.00	279.0
237.00	1361.0
238.10	190.0
239.00	484.0
240.00	474.0
241.00	562.0
242.00	1315.0
243.10	1806.0
244.10	24784.0
245.10	3512.0
246.10	4254.0
247.00	785.0
248.10	300.0
249.00	1323.0
250.00	228.0
251.00	407.0

Instrument :
BNA_G
ClientSampleId :
DFTPP21

252.10	371.0
253.20	690.0
255.10	124592.0
256.00	19624.0
257.10	1670.0
258.00	5985.0
259.00	1299.0
261.00	206.0
264.10	267.0
265.10	2218.0
266.10	330.0
270.00	203.0
271.10	301.0
272.20	488.0
273.10	3809.0
274.00	9748.0
275.10	58296.0
276.10	8467.0
277.00	4069.0
278.00	642.0
279.00	189.0
281.10	230.0
283.00	436.0
284.10	363.0
285.20	739.0
288.90	263.0
291.00	163.0
292.00	162.0
293.10	1127.0
294.10	375.0
296.00	14267.0
297.00	1923.0
301.00	265.0
302.00	371.0
303.10	1959.0
304.00	441.0
308.00	472.0
310.10	237.0
313.10	294.0
314.10	790.0
315.10	1842.0
316.00	1116.0
317.10	219.0
321.10	513.0
322.10	390.0
323.10	5218.0
324.20	1317.0
327.00	691.0
328.00	459.0
332.00	561.0
333.10	549.0
334.10	3397.0
335.00	1018.0
337.20	161.0
341.10	689.0
342.00	268.0
346.00	990.0
352.10	1828.0
353.10	929.0
354.10	2237.0
355.10	689.0
365.00	7520.0
366.00	1208.0
370.10	187.0
371.10	630.0
372.10	3527.0
373.10	1212.0
374.10	174.0
376.90	193.0
383.00	800.0
384.00	346.0
390.00	505.0
391.10	197.0
401.00	397.0
402.10	1404.0
403.10	1737.0
404.00	747.0
410.00	160.0
415.10	248.0
421.10	1983.0
422.10	1546.0
423.10	10697.0
424.10	2167.0
425.10	401.0
425.90	168.0
439.10	183.0
441.10	32648.0
442.10	214720.0
443.10	40184.0
444.10	4060.0

Instrument :
BNA_G
ClientSampleId :
DFTPP21