

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG032924\
 Data File : BG060705.D
 Acq On : 29 Mar 2024 14:55
 Operator : MA/JU
 Sample : SSTDICC060
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC060

Quant Time: Mar 30 01:15:43 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG032924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 30 01:11:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.964	152	59537	20.000	ng	0.00
21) Naphthalene-d8	10.761	136	260207	20.000	ng	0.00
39) Acenaphthene-d10	14.592	164	191108	20.000	ng	0.00
64) Phenanthrene-d10	17.336	188	425089	20.000	ng	0.00
76) Chrysene-d12	21.607	240	382623	20.000	ng	0.00
86) Perylene-d12	24.745	264	448468	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.538	112	444971	124.506	ng	0.00
7) Phenol-d6	7.118	99	649086	122.353	ng	0.00
23) Nitrobenzene-d5	9.116	82	629399	138.023	ng	0.00
42) 2,4,6-Tribromophenol	16.078	330	282083	130.029	ng	0.00
45) 2-Fluorobiphenyl	13.223	172	1553293	119.286	ng	0.00
79) Terphenyl-d14	19.968	244	2452956	112.885	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.464	88	86206	58.833	ng	98
3) Pyridine	3.852	79	265912	60.164	ng	96
4) n-Nitrosodimethylamine	3.764	42	167712	63.481	ng	98
6) Aniline	7.289	93	407983	60.900	ng	99
8) 2-Chlorophenol	7.524	128	252763	62.384	ng	99
9) Benzaldehyde	7.101	77	153188	52.366	ng	100
10) Phenol	7.148	94	334783	61.836	ng	99
11) bis(2-Chloroethyl)ether	7.389	93	261101	59.733	ng	99
12) 1,3-Dichlorobenzene	7.853	146	252426	59.909	ng	99
13) 1,4-Dichlorobenzene	8.000	146	260971	60.710	ng	98
14) 1,2-Dichlorobenzene	8.317	146	254481	60.409	ng	99
15) Benzyl Alcohol	8.194	79	266884	62.106	ng	96
16) 2,2'-oxybis(1-Chloropr...	8.499	45	605796	61.032	ng	99
17) 2-Methylphenol	8.393	107	254040	62.155	ng	97
18) Hexachloroethane	9.046	117	94797	62.092	ng	93
19) n-Nitroso-di-n-propyla...	8.775	70	250998	62.042	ng	# 98
20) 3+4-Methylphenols	8.728	107	352221	62.932	ng	96
22) Acetophenone	8.781	105	447861	62.421	ng	98
24) Nitrobenzene	9.157	77	323577	70.599	ng	97
25) Isophorone	9.686	82	653701	62.239	ng	99
26) 2-Nitrophenol	9.868	139	118314	75.942	ng	93
27) 2,4-Dimethylphenol	9.933	122	263842	62.535	ng	96
28) bis(2-Chloroethoxy)met...	10.174	93	380716	61.397	ng	98
29) 2,4-Dichlorophenol	10.403	162	247126	63.785	ng	98
30) 1,2,4-Trichlorobenzene	10.620	180	278167	63.405	ng	99
31) Naphthalene	10.814	128	855738	60.796	ng	99
32) Benzoic acid	10.074	122	175988	70.805	ng	97
33) 4-Chloroaniline	10.908	127	399171	61.057	ng	99
34) Hexachlorobutadiene	11.108	225	183037	62.602	ng	96
35) Caprolactam	11.690	113	93643	61.026	ng	96
36) 4-Chloro-3-methylphenol	12.030	107	305107	61.646	ng	99
37) 2-Methylnaphthalene	12.418	142	636057	61.465	ng	98
38) 1-Methylnaphthalene	12.635	142	584200	59.760	ng	96
40) 1,2,4,5-Tetrachloroben...	12.782	216	346317	63.034	ng	99
41) Hexachlorocyclopentadiene	12.771	237	188295	67.401	ng	99
43) 2,4,6-Trichlorophenol	13.017	196	231864	65.540	ng	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG032924\
 Data File : BG060705.D
 Acq On : 29 Mar 2024 14:55
 Operator : MA/JU
 Sample : SSTDICC060
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC060


Quant Time: Mar 30 01:15:43 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG032924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 30 01:11:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	13.088	196	272282	64.493	ng	98
46) 1,1'-Biphenyl	13.429	154	812034	60.157	ng	100
47) 2-Chloronaphthalene	13.470	162	625699	61.009	ng	98
48) 2-Nitroaniline	13.664	65	212814	72.464	ng	98
49) Acenaphthylene	14.310	152	1045933	60.696	ng	99
50) Dimethylphthalate	14.057	163	905763	59.963	ng	99
51) 2,6-Dinitrotoluene	14.163	165	169248	71.814	ng	96
52) Acenaphthene	14.657	154	658324	60.845	ng	99
53) 3-Nitroaniline	14.486	138	182910	69.071	ng	98
54) 2,4-Dinitrophenol	14.686	184	76542	72.799	ng	93
55) Dibenzofuran	14.992	168	1015778	59.520	ng	99
56) 4-Nitrophenol	14.786	139	146545	66.363	ng	95
57) 2,4-Dinitrotoluene	14.945	165	222517	72.666	ng	99
58) Fluorene	15.638	166	797836	58.249	ng	100
59) 2,3,4,6-Tetrachlorophenol	15.209	232	231300	62.865	ng	97
60) Diethylphthalate	15.426	149	892051	58.699	ng	99
61) 4-Chlorophenyl-phenyle...	15.638	204	427328	58.972	ng	96
62) 4-Nitroaniline	15.655	138	186503	67.531	ng	99
63) Azobenzene	15.932	77	876527	59.189	ng	98
65) 4,6-Dinitro-2-methylph...	15.708	198	111639	71.568	ng	94
66) n-Nitrosodiphenylamine	15.849	169	750587	61.789	ng	98
67) 4-Bromophenyl-phenylether	16.531	248	275146	63.778	ng	96
68) Hexachlorobenzene	16.643	284	303300	62.278	ng	98
69) Atrazine	16.807	200	254307	59.866	ng	99
70) Pentachlorophenol	16.983	266	217917	68.577	ng	97
71) Phenanthrene	17.383	178	1333071	60.298	ng	100
72) Anthracene	17.471	178	1357365	60.456	ng	99
73) Carbazole	17.735	167	1176302	59.422	ng	99
74) Di-n-butylphthalate	18.323	149	1474609	59.750	ng	98
75) Fluoranthene	19.392	202	1556286	58.012	ng	99
77) Benzidine	19.569	184	604840	60.032	ng	100
78) Pyrene	19.751	202	1593291	59.541	ng	99
80) Butylbenzylphthalate	20.667	149	639315	65.493	ng	99
81) Benzo(a)anthracene	21.584	228	1624581	60.235	ng	100
82) 3,3'-Dichlorobenzidine	21.502	252	560535	61.551	ng	99
83) Chrysene	21.654	228	1564343	60.176	ng	100
84) Bis(2-ethylhexyl)phtha...	21.543	149	969599	62.319	ng	97
85) Di-n-octyl phthalate	22.747	149	1643870	64.855	ng	99
87) Indeno(1,2,3-cd)pyrene	28.329	276	1969352	62.358	ng	99
88) Benzo(b)fluoranthene	23.764	252	1571271	60.986	ng	99
89) Benzo(k)fluoranthene	23.828	252	1626765	61.907	ng	100
90) Benzo(a)pyrene	24.610	252	1554896	62.102	ng	97
91) Dibenzo(a,h)anthracene	28.405	278	1669069	62.695	ng	99
92) Benzo(g,h,i)perylene	29.439	276	1616056	62.403	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Mar 30 01:15:43 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG032924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Mar 30 01:11:12 2024
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 7.964 min Scan# 84
 Delta R.T. 0.003 min
 Lab File: BG060705.D
 Acq: 29 Mar 2024 14:55




Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC060


Tgt Ion:152 Resp: 59537
 Ion Ratio Lower Upper
 152 100
 150 156.2 130.2 195.4
 115 63.9 51.6 77.4



#2
 1,4-Dioxane
 Concen: 58.833 ng
 RT: 3.464 min Scan# 81
 Delta R.T. -0.003 min
 Lab File: BG060705.D
 Acq: 29 Mar 2024 14:55

Tgt Ion: 88 Resp: 86206
 Ion Ratio Lower Upper
 88 100
 58 99.1 80.9 121.3
 43 43.4 35.4 53.0





#3

Pyridine

Concen: 60.164 ng

RT: 3.852 min Scan# 14

Delta R.T. -0.003 min

Lab File: BG060705.D

Acq: 29 Mar 2024 14:55

Instrument :

BNA_G

ClientSampleId :

SSTDICC060





Tgt Ion: 79 Resp: 265912

Ion Ratio Lower Upper

79 100

52 104.6 80.6 120.8

51 51.4 39.1 58.7



#4

n-Nitrosodimethylamine

Concen: 63.481 ng

RT: 3.764 min Scan# 132

Delta R.T. -0.003 min

Lab File: BG060705.D

Acq: 29 Mar 2024 14:55




Tgt Ion: 42 Resp: 167712


Ion Ratio Lower Upper

42 100

74 95.9 77.8 116.8

44 7.0 6.8 10.2









#5
2-Fluorophenol
Concen: 124.506 ng
RT: 5.538 min Scan# 433
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55




Instrument :
BNA_G
ClientSampleId :
SSTDICC060


Tgt Ion:112 Resp: 444971
Ion Ratio Lower Upper
112 100
64 72.5 58.4 87.6
63 38.4 32.7 49.1



#6
Aniline
Concen: 60.900 ng
RT: 7.289 min Scan# 732
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55

Tgt Ion: 93 Resp: 407983
Ion Ratio Lower Upper
93 100
66 44.8 36.1 54.1
65 23.1 18.1 27.1









#7
Phenol-d6
Concen: 122.353 ng
RT: 7.118 min Scan# 702
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55




Instrument :
BNA_G
ClientSampleId :
SSTDICC060


Tgt Ion: 99 Resp: 649086
Ion Ratio Lower Upper
99 100
42 25.5 19.4 29.2
71 35.4 28.2 42.2



#8
2-Chlorophenol
Concen: 62.384 ng
RT: 7.524 min Scan# 772
Delta R.T. -0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55

Tgt Ion: 128 Resp: 252763
Ion Ratio Lower Upper
128 100
130 32.1 12.3 52.3
64 56.0 35.0 75.0









#9
Benzaldehyde
Concen: 52.366 ng
RT: 7.101 min Scan# 700
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55




Instrument :
BNA_G
ClientSampleId :
SSTDICC060


Tgt Ion: 77 Resp: 153188
Ion Ratio Lower Upper
77 100
105 95.2 76.0 116.0
106 90.4 70.3 110.3



#10
Phenol
Concen: 61.836 ng
RT: 7.148 min Scan# 708
Delta R.T. 0.009 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55

Tgt Ion: 94 Resp: 334783
Ion Ratio Lower Upper
94 100
65 32.7 13.3 53.3
66 42.2 22.9 62.9









#11
bis(2-Chloroethyl)ether
Concen: 59.733 ng
RT: 7.389 min Scan# 748
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55




Instrument :
BNA_G
ClientSampleId :
SSTDICC060


Tgt Ion: 93 Resp: 261101
Ion Ratio Lower Upper
93 100
63 93.4 73.7 113.7
95 33.5 11.2 51.2



#12
1,3-Dichlorobenzene
Concen: 59.909 ng
RT: 7.853 min Scan# 828
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55

Tgt Ion:146 Resp: 252426
Ion Ratio Lower Upper
146 100
148 65.5 53.0 79.4
75 34.9 28.1 42.1





#13

1,4-Dichlorobenzene

Concen: 60.710 ng

RT: 8.000 min Scan# 81

Delta R.T. 0.003 min

Lab File: BG060705.D

Acq: 29 Mar 2024 14:55

Instrument :

BNA_G

ClientSampleId :

SSTDICC060





Tgt Ion:146 Resp: 260971

Ion Ratio Lower Upper

146 100

148 62.2 51.1 76.7

111 41.9 35.0 52.6



#14

1,2-Dichlorobenzene

Concen: 60.409 ng

RT: 8.317 min Scan# 907

Delta R.T. 0.003 min

Lab File: BG060705.D

Acq: 29 Mar 2024 14:55




Tgt Ion:146 Resp: 254481


Ion Ratio Lower Upper

146 100

148 62.0 51.0 76.6

111 45.0 36.0 54.0









#15
Benzyl Alcohol
Concen: 62.106 ng
RT: 8.194 min Scan# 885
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55




Instrument :
BNA_G
ClientSampleId :
SSTDICC060


Tgt Ion: 79 Resp: 266884
Ion Ratio Lower Upper
79 100
108 73.9 61.4 92.0
77 63.7 53.6 80.4



#16
2,2'-oxybis(1-Chloropropane)
Concen: 61.032 ng
RT: 8.499 min Scan# 938
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55

Tgt Ion: 45 Resp: 605796
Ion Ratio Lower Upper
45 100
77 9.1 0.0 29.1
79 7.1 0.0 26.5









#17
2-Methylphenol
Concen: 62.155 ng
RT: 8.393 min Scan# 919
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55




Instrument :
BNA_G
ClientSampleId :
SSTDICC060


Tgt Ion:	107	Resp:	254040
Ion Ratio	Lower	Upper	
107	100		
108	111.8	86.4	129.6
77	46.3	38.6	58.0
79	46.6	37.9	56.9



#18
Hexachloroethane
Concen: 62.092 ng
RT: 9.046 min Scan# 1031
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55

Tgt Ion:	117	Resp:	94797
Ion Ratio	Lower	Upper	
117	100		
119	96.5	75.7	113.5
201	115.3	83.4	125.2









#19
n-Nitroso-di-n-propylamine
Concen: 62.042 ng
RT: 8.775 min Scan# 983
Delta R.T. 0.009 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55




Instrument :
BNA_G
ClientSampleId :
SSTDICC060


Tgt Ion: 70 Resp: 250998
Ion Ratio Lower Upper
70 100
42 83.8 68.4 102.6
101 9.3 6.1 9.1#
130 19.1 14.6 21.8



#20
3+4-Methylphenols
Concen: 62.932 ng
RT: 8.728 min Scan# 977
Delta R.T. 0.009 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55

Tgt Ion: 107 Resp: 352221
Ion Ratio Lower Upper
107 100
108 87.6 62.2 102.2
77 34.2 14.1 54.1
79 29.9 7.7 47.7









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.761 min Scan# 1323
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55

Instrument :
BNA_G
ClientSampleId :
SSTDIC060

Tgt Ion:136 Resp: 260207




Ion	Ratio	Lower	Upper
136	100		
137	10.8	8.8	13.2
54	12.4	10.2	15.2
68	6.9	5.2	7.8




#22
Acetophenone
Concen: 62.421 ng
RT: 8.781 min Scan# 986
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55

Tgt Ion:105 Resp: 447861

Ion	Ratio	Lower	Upper
105	100		
71	7.3	4.9	7.3
51	38.3	31.1	46.7
120	20.1	17.0	25.4





#23

Nitrobenzene-d5

Concen: 138.023 ng

RT: 9.116 min Scan# 1043

Delta R.T. 0.003 min

Lab File: BG060705.D

Acq: 29 Mar 2024 14:55

Instrument :

BNA_G

ClientSampleId :

SSTDICC060





Tgt Ion: 82 Resp: 629399

Ion Ratio Lower Upper

82 100

128 38.4 31.2 46.8

54 67.9 53.2 79.8



#24

Nitrobenzene

Concen: 70.599 ng

RT: 9.157 min Scan# 1050

Delta R.T. 0.003 min

Lab File: BG060705.D

Acq: 29 Mar 2024 14:55




Tgt Ion: 77 Resp: 323577


Ion Ratio Lower Upper

77 100

123 42.6 32.7 49.1

65 15.0 11.3 16.9





#25

Isophorone

Concen: 62.239 ng

RT: 9.686 min Scan# 1139

Delta R.T. 0.003 min

Lab File: BG060705.D

Acq: 29 Mar 2024 14:55

Instrument :

BNA_G

ClientSampleId :

SSTDICC060





Tgt Ion: 82 Resp: 653701

Ion Ratio Lower Upper

82 100

95 6.9 5.8 8.8

138 15.6 13.0 19.4



#26

2-Nitrophenol

Concen: 75.942 ng

RT: 9.868 min Scan# 1171

Delta R.T. 0.003 min

Lab File: BG060705.D

Acq: 29 Mar 2024 14:55




Tgt Ion: 139 Resp: 118314


Ion Ratio Lower Upper

139 100

109 31.9 23.9 35.9

65 61.9 44.5 66.7





#27
2,4-Dimethylphenol
Concen: 62.535 ng
RT: 9.933 min Scan# 1181
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55





Instrument :

BNA_G

ClientSampleId :




SSTDICC060


Tgt Ion:122 Resp: 263842
Ion Ratio Lower Upper
122 100
107 117.9 98.6 147.8
121 57.3 48.1 72.1



#28
bis(2-Chloroethoxy)methane
Concen: 61.397 ng
RT: 10.174 min Scan# 1223
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55

Tgt Ion: 93 Resp: 380716
Ion Ratio Lower Upper
93 100
95 32.4 25.1 37.7
123 10.5 8.0 12.0





#29

2,4-Dichlorophenol

Concen: 63.785 ng

RT: 10.403 min Scan# 1261

Delta R.T. 0.003 min

Lab File: BG060705.D

Acq: 29 Mar 2024 14:55

Instrument :

BNA_G

ClientSampleId :

SSTDICC060





Tgt Ion:162 Resp: 247126

Ion Ratio Lower Upper

162 100

164 67.3 45.2 85.2

98 38.1 17.7 57.7



#30

1,2,4-Trichlorobenzene

Concen: 63.405 ng

RT: 10.620 min Scan# 1299

Delta R.T. -0.003 min

Lab File: BG060705.D

Acq: 29 Mar 2024 14:55




Tgt Ion:180 Resp: 278167


Ion Ratio Lower Upper

180 100

182 98.1 79.0 118.4

145 30.1 24.9 37.3









#31
Naphthalene
Concen: 60.796 ng
RT: 10.814 min Scan# 1331
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55




Instrument :
BNA_G
ClientSampleId :
SSTDICC060


Tgt Ion:128 Resp: 855738
Ion Ratio Lower Upper
128 100
129 11.1 8.5 12.7
127 13.3 10.5 15.7



#32
Benzoic acid
Concen: 70.805 ng
RT: 10.074 min Scan# 1206
Delta R.T. 0.033 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55

Tgt Ion:122 Resp: 175988
Ion Ratio Lower Upper
122 100
105 126.3 104.5 144.5
77 99.6 85.3 125.3









#33
4-Chloroaniline
Concen: 61.057 ng
RT: 10.908 min Scan# 1348
Delta R.T. -0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55




Instrument :
BNA_G
ClientSampleId :
SSTDICC060


Tgt Ion:127	Resp: 399171
Ion Ratio	Lower Upper
127 100	
129 32.2	25.3 37.9
65 37.4	30.2 45.2
92 21.0	15.5 23.3



#34
Hexachlorobutadiene
Concen: 62.602 ng
RT: 11.108 min Scan# 1382
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55

Tgt Ion:225	Resp: 183037
Ion Ratio	Lower Upper
225 100	
223 61.5	52.0 78.0
227 63.2	52.5 78.7









#35
Caprolactam
Concen: 61.026 ng
RT: 11.690 min Scan# 1478
Delta R.T. 0.015 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55




Instrument :
BNA_G
ClientSampleId :
SSTDICC060


Tgt Ion:113 Resp: 93643
Ion Ratio Lower Upper
113 100
55 275.5 258.0 298.0
56 197.5 166.3 206.3



#36
4-Chloro-3-methylphenol
Concen: 61.646 ng
RT: 12.030 min Scan# 1539
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55

Tgt Ion:107 Resp: 305107
Ion Ratio Lower Upper
107 100
144 25.0 20.0 30.0
142 74.6 59.2 88.8









#37
2-Methylnaphthalene
Concen: 61.465 ng
RT: 12.418 min Scan# 1605
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55




Instrument :
BNA_G
ClientSampleId :
SSTDICC060


Tgt Ion:142 Resp: 636057
Ion Ratio Lower Upper
142 100
141 89.2 72.7 109.1
115 38.2 29.9 44.9



#38
1-Methylnaphthalene
Concen: 59.760 ng
RT: 12.635 min Scan# 1642
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55

Tgt Ion:142 Resp: 584200
Ion Ratio Lower Upper
142 100
141 93.1 77.7 116.5
116 4.2 3.4 5.2





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.592 min Scan# 1975

Delta R.T. 0.003 min

Lab File: BG060705.D

Acq: 29 Mar 2024 14:55

Instrument :

BNA_G

ClientSampleId :

SSTDICC060





Tgt Ion:164 Resp: 191108

Ion Ratio Lower Upper

164 100

162 104.1 79.5 119.3

160 45.5 35.4 53.0



#40

1,2,4,5-Tetrachlorobenzene

Concen: 63.034 ng

RT: 12.782 min Scan# 1667

Delta R.T. 0.003 min

Lab File: BG060705.D

Acq: 29 Mar 2024 14:55

Tgt Ion:216 Resp: 346317




Ion Ratio Lower Upper


216 100

214 79.9 63.1 94.7

179 20.9 16.7 25.1

108 17.3 13.8 20.6









#41
Hexachlorocyclopentadiene
Concen: 67.401 ng
RT: 12.771 min Scan# 1665
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55




Instrument :
BNA_G
ClientSampleId :
SSTDICC060


Tgt Ion:237 Resp: 188295
Ion Ratio Lower Upper
237 100
235 60.9 42.0 82.0
272 12.8 0.0 32.4



#42
2,4,6-Tribromophenol
Concen: 130.029 ng
RT: 16.078 min Scan# 2228
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55

Tgt Ion:330 Resp: 282083
Ion Ratio Lower Upper
330 100
332 96.3 78.2 117.4
141 32.9 27.0 40.6









#43
2,4,6-Trichlorophenol
Concen: 65.540 ng
RT: 13.017 min Scan# 1719
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55




Instrument :
BNA_G
ClientSampleId :
SSTDICC060


Tgt Ion:196 Resp: 231864
Ion Ratio Lower Upper
196 100
198 97.5 79.1 118.7
200 31.5 24.0 36.0



#44
2,4,5-Trichlorophenol
Concen: 64.493 ng
RT: 13.088 min Scan# 1719
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55

Tgt Ion:196 Resp: 272282
Ion Ratio Lower Upper
196 100
198 94.8 77.8 116.8
97 46.1 38.2 57.4
132 24.8 20.3 30.5









#45
2-Fluorobiphenyl
Concen: 119.286 ng
RT: 13.223 min Scan# 1742
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55




Instrument :
BNA_G
ClientSampleId :
SSTDICC060


Tgt Ion:172 Resp: 1553293
Ion Ratio Lower Upper
172 100
171 36.8 28.7 43.1
170 25.1 19.4 29.0



#46
1,1'-Biphenyl
Concen: 60.157 ng
RT: 13.429 min Scan# 1777
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55

Tgt Ion:154 Resp: 812034
Ion Ratio Lower Upper
154 100
153 40.6 20.8 60.8
76 13.8 0.0 33.9









#47
2-Chloronaphthalene
Concen: 61.009 ng
RT: 13.470 min Scan# 11
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55




Instrument :
BNA_G
ClientSampleId :
SSTDICC060


Tgt Ion:162 Resp: 625699
Ion Ratio Lower Upper
162 100
127 39.1 30.6 46.0
164 34.8 26.9 40.3



#48
2-Nitroaniline
Concen: 72.464 ng
RT: 13.664 min Scan# 1817
Delta R.T. 0.009 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55

Tgt Ion: 65 Resp: 212814
Ion Ratio Lower Upper
65 100
92 59.9 48.6 72.8
138 92.4 72.2 108.4









#49
Acenaphthylene
Concen: 60.696 ng
RT: 14.310 min Scan# 1927
Delta R.T. -0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55




Instrument :
BNA_G
ClientSampleId :
SSTDICC060


Tgt Ion:152 Resp: 1045933
Ion Ratio Lower Upper
152 100
151 20.9 16.4 24.6
153 13.5 10.5 15.7



#50
Dimethylphthalate
Concen: 59.963 ng
RT: 14.057 min Scan# 1884
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55

Tgt Ion:163 Resp: 905763
Ion Ratio Lower Upper
163 100
194 3.5 3.0 4.6
164 10.1 7.9 11.9









#51
2,6-Dinitrotoluene
Concen: 71.814 ng
RT: 14.163 min Scan# 1901
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55




Instrument :
BNA_G
ClientSampleId :
SSTDICC060


Tgt Ion:165 Resp: 169248
Ion Ratio Lower Upper
165 100
63 67.1 58.2 87.4
89 58.7 46.1 69.1



#52
Acenaphthene
Concen: 60.845 ng
RT: 14.657 min Scan# 1986
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55

Tgt Ion:154 Resp: 658324
Ion Ratio Lower Upper
154 100
153 115.7 93.9 140.9
152 55.8 44.1 66.1









#53
3-Nitroaniline
Concen: 69.071 ng
RT: 14.486 min Scan# 1956
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55




Instrument :
BNA_G
ClientSampleId :
SSTDICC060


Tgt Ion:138 Resp: 182910
Ion Ratio Lower Upper
138 100
108 12.2 11.3 16.9
92 118.9 94.1 141.1



#54
2,4-Dinitrophenol
Concen: 72.799 ng
RT: 14.686 min Scan# 1991
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55

Tgt Ion:184 Resp: 76542
Ion Ratio Lower Upper
184 100
63 69.0 51.3 76.9
154 67.9 59.9 89.9









#55
Dibenzofuran
Concen: 59.520 ng
RT: 14.992 min Scan# 2043
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55




Instrument :
BNA_G
ClientSampleId :
SSTDICC060


Tgt Ion:168 Resp: 1015778
Ion Ratio Lower Upper
168 100
139 38.1 30.1 45.1
169 13.9 11.0 16.4



#56
4-Nitrophenol
Concen: 66.363 ng
RT: 14.786 min Scan# 2008
Delta R.T. 0.009 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55

Tgt Ion:139 Resp: 146545
Ion Ratio Lower Upper
139 100
109 82.7 64.0 104.0
65 115.4 104.3 144.3









#57
2,4-Dinitrotoluene
Concen: 72.666 ng
RT: 14.945 min Scan# 2034
Delta R.T. 0.004 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55




Instrument :
BNA_G
ClientSampleId :
SSTDICC060


Tgt Ion:165 Resp: 222517
Ion Ratio Lower Upper
165 100
63 54.2 43.4 65.0
89 78.3 61.0 91.6
182 2.0 1.8 2.6



#58
Fluorene
Concen: 58.249 ng
RT: 15.638 min Scan# 2153
Delta R.T. -0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55


Tgt Ion:166 Resp: 797836
Ion Ratio Lower Upper
166 100
165 102.2 81.8 122.8
167 14.0 10.9 16.3








#59
2,3,4,6-Tetrachlorophenol
Concen: 62.865 ng
RT: 15.209 min Scan# 2080
Delta R.T. -0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55


Instrument :
BNA_G
ClientSampleId :
SSTDICC060





Tgt Ion:232 Resp: 231300
Ion Ratio Lower Upper
232 100
131 42.5 32.2 48.2
130 2.5 2.1 3.1
166 30.2 23.7 35.5




#60
Diethylphthalate
Concen: 58.699 ng
RT: 15.426 min Scan# 2117
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55



Tgt Ion:149 Resp: 892051
Ion Ratio Lower Upper
149 100
177 23.1 18.0 27.0
150 12.9 10.2 15.4









#61
4-Chlorophenyl-phenylether
Concen: 58.972 ng
RT: 15.638 min Scan# 2153
Delta R.T. -0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55




Instrument :
BNA_G
ClientSampleId :
SSTDICC060


Tgt Ion:204 Resp: 427328
Ion Ratio Lower Upper
204 100
206 32.2 28.3 42.5
141 56.9 47.2 70.8



#62
4-Nitroaniline
Concen: 67.531 ng
RT: 15.655 min Scan# 2156
Delta R.T. 0.009 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55

Tgt Ion:138 Resp: 186503
Ion Ratio Lower Upper
138 100
92 56.0 36.8 76.8
108 85.9 66.9 106.9









#63
Azobenzene
Concen: 59.189 ng
RT: 15.932 min Scan# 21
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55




Instrument :
BNA_G
ClientSampleId :
SSTDICC060


Tgt Ion: 77	Resp: 876527
Ion Ratio	Lower Upper
77 100	
182 23.3	3.9 43.9
105 22.0	1.7 41.7
51 40.2	18.4 58.4



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.336 min Scan# 2442
Delta R.T. -0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55

Tgt Ion: 188	Resp: 425089
Ion Ratio	Lower Upper
188 100	
94 8.7	6.6 9.8
80 9.7	7.6 11.4









#65
4,6-Dinitro-2-methylphenol
Concen: 71.568 ng
RT: 15.708 min Scan# 2164
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55




Instrument :
BNA_G
ClientSampleId :
SSTDICC060


Tgt Ion:198 Resp: 111639
Ion Ratio Lower Upper
198 100
51 61.7 36.7 76.7
105 42.5 19.1 59.1



#66
n-Nitrosodiphenylamine
Concen: 61.789 ng
RT: 15.849 min Scan# 2189
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55

Tgt Ion:169 Resp: 750587
Ion Ratio Lower Upper
169 100
168 69.0 54.5 81.7
167 38.1 29.2 43.8









#67
4-Bromophenyl-phenylether
Concen: 63.778 ng
RT: 16.531 min Scan# 2305
Delta R.T. -0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55




Instrument :
BNA_G
ClientSampleId :
SSTDICC060


Tgt Ion:248 Resp: 275146
Ion Ratio Lower Upper
248 100
250 93.6 78.3 117.5
141 69.7 53.7 80.5



#68
Hexachlorobenzene
Concen: 62.278 ng
RT: 16.643 min Scan# 2324
Delta R.T. -0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55

Tgt Ion:284 Resp: 303300
Ion Ratio Lower Upper
284 100
142 30.4 22.9 34.3
249 32.9 27.1 40.7









#69
Atrazine
Concen: 59.866 ng
RT: 16.807 min Scan# 2351
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55




Instrument :
BNA_G
ClientSampleId :
SSTDICC060


Tgt Ion:200 Resp: 254307
Ion Ratio Lower Upper
200 100
173 25.8 5.2 45.2
215 49.4 28.8 68.8



#70
Pentachlorophenol
Concen: 68.577 ng
RT: 16.983 min Scan# 2382
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55

Tgt Ion:266 Resp: 217917
Ion Ratio Lower Upper
266 100
268 63.6 52.2 78.4
264 61.5 51.8 77.6









#71
Phenanthrene
Concen: 60.298 ng
RT: 17.383 min Scan# 2449
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55




Instrument :
BNA_G
ClientSampleId :
SSTDICC060


Tgt Ion:178 Resp: 1333071
Ion Ratio Lower Upper
178 100
176 19.7 15.8 23.6
179 15.6 12.2 18.4



#72
Anthracene
Concen: 60.456 ng
RT: 17.471 min Scan# 2465
Delta R.T. -0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55

Tgt Ion:178 Resp: 1357365
Ion Ratio Lower Upper
178 100
176 19.1 15.5 23.3
179 15.6 12.2 18.2









#73
 Carbazole
 Concen: 59.422 ng
 RT: 17.735 min Scan# 2509
 Delta R.T. 0.003 min
 Lab File: BG060705.D
 Acq: 29 Mar 2024 14:55




Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC060


Tgt Ion:167 Resp: 1176302
 Ion Ratio Lower Upper
 167 100
 166 23.2 18.0 27.0
 139 13.3 10.5 15.7



#74
 Di-n-butylphthalate
 Concen: 59.750 ng
 RT: 18.323 min Scan# 2610
 Delta R.T. -0.003 min
 Lab File: BG060705.D
 Acq: 29 Mar 2024 14:55

Tgt Ion:149 Resp: 1474609
 Ion Ratio Lower Upper
 149 100
 150 10.4 7.7 11.5
 104 5.7 4.2 6.4









#75
Fluoranthene
Concen: 58.012 ng
RT: 19.392 min Scan# 21
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55




Instrument :
BNA_G
ClientSampleId :
SSTDICC060


Tgt Ion:202 Resp: 1556286
Ion Ratio Lower Upper
202 100
101 10.3 0.0 30.8
203 17.7 0.0 37.4



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.607 min Scan# 3169
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55

Tgt Ion:240 Resp: 382623
Ion Ratio Lower Upper
240 100
120 9.6 7.1 10.7
236 25.6 20.7 31.1





#77

Benzidine

Concen: 60.032 ng

RT: 19.569 min Scan# 2822

Delta R.T. -0.003 min

Lab File: BG060705.D

Acq: 29 Mar 2024 14:55

Instrument :

BNA_G

ClientSampleId :

SSTDICC060





Tgt Ion:184 Resp: 604840

Ion Ratio Lower Upper

184 100

185 14.9 11.8 17.6

183 11.9 9.5 14.3



#78

Pyrene

Concen: 59.541 ng

RT: 19.751 min Scan# 2853

Delta R.T. -0.003 min

Lab File: BG060705.D

Acq: 29 Mar 2024 14:55




Tgt Ion:202 Resp: 1593291


Ion Ratio Lower Upper

202 100

200 21.4 17.0 25.4

203 18.3 14.0 21.0





#79

Terphenyl-d14

Concen: 112.885 ng

RT: 19.968 min Scan# 2889

Delta R.T. 0.003 min

Lab File: BG060705.D

Acq: 29 Mar 2024 14:55

Instrument :

BNA_G

ClientSampleId :

SSTDICC060





Tgt Ion:244 Resp: 2452956

Ion Ratio Lower Upper

244 100

212 8.5 6.2 9.2

122 9.7 7.8 11.8



#80

Butylbenzylphthalate

Concen: 65.493 ng

RT: 20.667 min Scan# 3009

Delta R.T. 0.003 min

Lab File: BG060705.D

Acq: 29 Mar 2024 14:55




Tgt Ion:149 Resp: 639315


Ion Ratio Lower Upper

149 100

91 73.9 60.1 90.1

206 21.3 16.8 25.2









#81
Benzo(a)anthracene
Concen: 60.235 ng
RT: 21.584 min Scan# 3165
Delta R.T. -0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55




Instrument :
BNA_G
ClientSampleId :
SSTDICC060


Tgt Ion:228 Resp: 1624581
Ion Ratio Lower Upper
228 100
226 27.1 21.7 32.5
229 19.9 15.8 23.8



#82
3,3'-Dichlorobenzidine
Concen: 61.551 ng
RT: 21.502 min Scan# 3151
Delta R.T. -0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55

Tgt Ion:252 Resp: 560535
Ion Ratio Lower Upper
252 100
254 64.8 52.6 78.8
126 10.5 7.8 11.8





#83

Chrysene

Concen: 60.176 ng

RT: 21.654 min Scan# 3176

Delta R.T. 0.003 min

Lab File: BG060705.D

Acq: 29 Mar 2024 14:55

Instrument :

BNA_G

ClientSampleId :

SSTDICC060





Tgt Ion:228 Resp: 1564343

Ion Ratio Lower Upper

228 100

226 30.6 24.2 36.4

229 19.6 15.5 23.3



#84

Bis(2-ethylhexyl)phthalate

Concen: 62.319 ng

RT: 21.543 min Scan# 3158

Delta R.T. 0.003 min

Lab File: BG060705.D

Acq: 29 Mar 2024 14:55




Tgt Ion:149 Resp: 969599


Ion Ratio Lower Upper

149 100

167 27.4 20.6 30.8

279 4.1 3.0 4.4









#85
Di-n-octyl phthalate
Concen: 64.855 ng
RT: 22.747 min Scan# 3363
Delta R.T. -0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55




Instrument :
BNA_G
ClientSampleId :
SSTDICC060


Tgt Ion:149 Resp: 1643870
Ion Ratio Lower Upper
149 100
167 1.6 1.2 1.8
43 14.6 11.3 16.9



#86
Perylene-d12
Concen: 20.000 ng
RT: 24.745 min Scan# 3703
Delta R.T. -0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55

Tgt Ion:264 Resp: 448468
Ion Ratio Lower Upper
264 100
260 22.4 19.5 29.3
265 21.9 16.9 25.3





#87

Indeno(1,2,3-cd)pyrene

Concen: 62.358 ng

RT: 28.329 min Scan# 4311

Delta R.T. 0.009 min

Lab File: BG060705.D

Acq: 29 Mar 2024 14:55

Instrument :

BNA_G

ClientSampleId :

SSTDICC060





Tgt Ion:276 Resp: 1969352

Ion Ratio Lower Upper

276 100

138 15.1 12.4 18.6

277 25.1 20.2 30.2



#88

Benzo(b)fluoranthene

Concen: 60.986 ng

RT: 23.764 min Scan# 3536

Delta R.T. 0.003 min

Lab File: BG060705.D

Acq: 29 Mar 2024 14:55




Tgt Ion:252 Resp: 1571271


Ion Ratio Lower Upper

252 100

253 22.0 17.8 26.8

125 8.7 7.1 10.7









#89
Benzo(k)fluoranthene
Concen: 61.907 ng
RT: 23.828 min Scan# 31
Delta R.T. 0.003 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55




Instrument :
BNA_G
ClientSampleId :
SSTDICC060


Tgt Ion:252 Resp: 1626765
Ion Ratio Lower Upper
252 100
253 22.0 17.6 26.4
125 8.9 7.5 11.3



#90
Benzo(a)pyrene
Concen: 62.102 ng
RT: 24.610 min Scan# 3680
Delta R.T. 0.009 min
Lab File: BG060705.D
Acq: 29 Mar 2024 14:55

Tgt Ion:252 Resp: 1554896
Ion Ratio Lower Upper
252 100
253 22.9 16.8 25.2
125 10.4 7.6 11.4





#91

Dibenzo(a,h)anthracene

Concen: 62.695 ng

RT: 28.405 min Scan# 4324

Delta R.T. 0.009 min

Lab File: BG060705.D

Acq: 29 Mar 2024 14:55

Instrument :

BNA_G

ClientSampleId :

SSTDICC060





Tgt Ion:278 Resp: 1669069

Ion Ratio Lower Upper

278 100

139 13.2 11.1 16.7

279 24.3 19.8 29.6



#92

Benzo(g,h,i)perylene

Concen: 62.403 ng

RT: 29.439 min Scan# 4502

Delta R.T. 0.015 min

Lab File: BG060705.D

Acq: 29 Mar 2024 14:55

Tgt Ion:276 Resp: 1616056

Ion Ratio Lower Upper

276 100

277 23.1 18.9 28.3

138 18.0 14.0 21.0

