

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG032924\
 Data File : BG060707.D
 Acq On : 29 Mar 2024 17:32
 Operator : MA/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG032924

Quant Time: Mar 30 03:14:17 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG032924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 30 03:12:26 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	114	0.00
2	1,4-Dioxane	40.000	37.757	5.6	111	0.00
3	Pyridine	40.000	40.326	-0.8	107	0.00
4	n-Nitrosodimethylamine	40.000	39.916	0.2	109	0.00
5 S	2-Fluorophenol	80.000	77.422	3.2	110	0.00
6	Aniline	40.000	38.653	3.4	108	0.00
7 S	Phenol-d6	80.000	78.702	1.6	110	0.00
8	2-Chlorophenol	40.000	38.824	2.9	110	0.00
9	Benzaldehyde	40.000	37.692	5.8	112	0.00
10 C	Phenol	40.000	39.048	2.4	108	0.00
11	bis(2-Chloroethyl)ether	40.000	37.903	5.2	108	0.00
12	1,3-Dichlorobenzene	40.000	37.724	5.7	109	0.00
13 C	1,4-Dichlorobenzene	40.000	38.205	4.5	111	0.00
14	1,2-Dichlorobenzene	40.000	38.084	4.8	110	0.00
15	Benzyl Alcohol	40.000	39.384	1.5	110	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.962	5.1	109	0.00
17	2-Methylphenol	40.000	38.695	3.3	109	0.00
18	Hexachloroethane	40.000	38.982	2.5	116	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.731	3.2	109	0.00
20	3+4-Methylphenols	40.000	38.743	3.1	109	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	110	0.00
22	Acetophenone	40.000	39.126	2.2	109	0.00
23 S	Nitrobenzene-d5	80.000	85.723	-7.2	121	0.00
24	Nitrobenzene	40.000	40.303	-0.8	117	0.00
25	Isophorone	40.000	39.435	1.4	108	0.00
26 C	2-Nitrophenol	40.000	43.835	-9.6	130	0.00
27	2,4-Dimethylphenol	40.000	39.269	1.8	110	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.728	3.2	108	0.00
29 C	2,4-Dichlorophenol	40.000	40.950	-2.4	109	0.00
30	1,2,4-Trichlorobenzene	40.000	39.162	2.1	109	0.00
31	Naphthalene	40.000	38.919	2.7	107	0.00
32	Benzoic acid	40.000	41.239	-3.1	119	0.00
33	4-Chloroaniline	40.000	38.603	3.5	105	0.00
34 C	Hexachlorobutadiene	40.000	38.713	3.2	109	0.00
35	Caprolactam	40.000	39.017	2.5	104	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.416	1.5	105	0.00
37	2-Methylnaphthalene	40.000	38.064	4.8	106	0.00
38	1-Methylnaphthalene	40.000	38.396	4.0	109	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	104	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.622	-1.6	109	0.00
41 P	Hexachlorocyclopentadiene	40.000	42.099	-5.2	112	0.00
42 S	2,4,6-Tribromophenol	80.000	88.590	-10.7	116	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.551	-6.4	110	0.00
44	2,4,5-Trichlorophenol	40.000	43.518	-8.8	113	0.00
45 S	2-Fluorobiphenyl	80.000	79.694	0.4	107	0.00
46	1,1'-Biphenyl	40.000	39.169	2.1	106	0.00
47	2-Chloronaphthalene	40.000	39.154	2.1	105	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG032924\
 Data File : BG060707.D
 Acq On : 29 Mar 2024 17:32
 Operator : MA/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG032924

Quant Time: Mar 30 03:14:17 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG032924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 30 03:12:26 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	43.411	-8.5	120	0.00
49	Acenaphthylene	40.000	40.098	-0.2	107	0.00
50	Dimethylphthalate	40.000	40.179	-0.4	106	0.00
51	2,6-Dinitrotoluene	40.000	42.783	-7.0	118	0.00
52 C	Acenaphthene	40.000	39.982	0.0	106	0.00
53	3-Nitroaniline	40.000	43.452	-8.6	116	0.00
54 P	2,4-Dinitrophenol	40.000	46.425	-16.1	124	0.00
55	Dibenzofuran	40.000	40.055	-0.1	107	0.00
56 P	4-Nitrophenol	40.000	46.159	-15.4	117	0.00
57	2,4-Dinitrotoluene	40.000	44.384	-11.0	122	0.00
58	Fluorene	40.000	39.295	1.8	104	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.218	-5.5	106	0.00
60	Diethylphthalate	40.000	40.284	-0.7	107	0.00
61	4-Chlorophenyl-phenylether	40.000	39.457	1.4	106	0.00
62	4-Nitroaniline	40.000	46.488	-16.2	116	0.00
63	Azobenzene	40.000	39.423	1.4	104	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	105	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.553	-3.9	122	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.404	1.5	104	0.00
67	4-Bromophenyl-phenylether	40.000	40.737	-1.8	108	0.00
68	Hexachlorobenzene	40.000	40.236	-0.6	107	0.00
69	Atrazine	40.000	40.553	-1.4	106	0.00
70 C	Pentachlorophenol	40.000	43.911	-9.8	114	0.00
71	Phenanthrene	40.000	39.262	1.8	106	0.00
72	Anthracene	40.000	39.821	0.4	107	0.00
73	Carbazole	40.000	39.106	2.2	106	0.00
74	Di-n-butylphthalate	40.000	39.404	1.5	104	0.00
75 C	Fluoranthene	40.000	38.760	3.1	106	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	104	0.00
77	Benzydine	40.000	44.454	-11.1	102	0.00
78	Pyrene	40.000	39.875	0.3	106	0.00
79 S	Terphenyl-d14	80.000	79.074	1.2	106	0.00
80	Butylbenzylphthalate	40.000	42.295	-5.7	108	0.00
81	Benzo(a)anthracene	40.000	39.795	0.5	104	0.00
82	3,3'-Dichlorobenzidine	40.000	41.581	-4.0	105	0.00
83	Chrysene	40.000	39.623	0.9	104	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.187	-3.0	104	0.00
85 c	Di-n-octyl phthalate	40.000	41.770	-4.4	104	0.00
86 I	Perylene-d12	20.000	20.000	0.0	104	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.559	1.1	104	0.00
88	Benzo(b)fluoranthene	40.000	39.034	2.4	102	0.00
89	Benzo(k)fluoranthene	40.000	39.661	0.8	102	0.00
90 C	Benzo(a)pyrene	40.000	39.448	1.4	102	0.00
91	Dibenzo(a,h)anthracene	40.000	39.573	1.1	104	0.00
92	Benzo(g,h,i)perylene	40.000	39.370	1.6	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG032924\
Data File : BG060707.D
Acq On : 29 Mar 2024 17:32
Operator : MA/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG032924

Quant Time: Mar 30 03:14:17 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG032924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Mar 30 03:12:26 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0