

Data Path : Z:\HPCHEM1\BNA G\DATA\BG033017\  
 Data File : BG026510.D  
 Acq On : 31 Mar 2017 00:00  
 Operator : SJ/MA  
 Sample : I2448-06  
 Misc :  
 ALS Vial : 13 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :

Quant Time: Mar 31 01:26:07 2017  
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM-EPA-BG032717MA.M  
 Quant Title : SVOA CALIBRATION  
 QLast Update : Fri Mar 31 01:08:45 2017  
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.04	152	194450	20.00	ng/ul	0.00
18) Naphthalene-d8	10.84	136	812132	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.64	164	469331	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.38	188	1124198	20.00	ng/ul	0.00
77) Chrysene-d12	21.62	240	1343831	20.00	ng/ul	0.00
85) Perylene-d12	24.80	264	1356774	20.00	ng/ul	0.00

## System Monitoring Compounds

3) 1,4-Dioxane-d8	3.47	96	23543	5.79	ng/uL	0.00
5) Phenol-d5	7.21	99	106522	7.44	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.36	67	258507	32.21	ng/ul	0.00
9) 2-Chlorophenol-d4	7.58	132	342606	26.68	ng/ul	0.00
13) 4-Methylphenol-d8	8.74	113	204357	17.71	ng/ul	0.00
19) Nitrobenzene-d5	9.20	128	197441	34.02	ng/ul	0.00
22) 2-Nitrophenol-d4	9.92	143	237722	34.94	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.46	165	376561	30.16	ng/ul	0.00
29) 4-Chloroaniline-d4	11.00	131	8103m	0.65	ng/ul	0.03
43) Dimethylphthalate-d6	14.04	166	1283998	35.33	ng/ul	0.00
46) Acenaphthylene-d8	14.34	160	1547164	34.17	ng/ul	0.00
51) 4-Nitrophenol-d4	14.85	143	44872m	6.35	ng/ul	0.01
57) Fluorene-d10	15.63	176	1189220	36.02	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.75	200	239324	33.11	ng/ul	0.00
70) Anthracene-d10	17.48	188	1873027	36.69	ng/ul	0.00
78) Pyrene-d10	19.76	212	2123054	35.42	ng/ul	0.00
89) Benzo(a)pyrene-d12	24.59	264	2247521	38.56	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG033017\  
 Data File : BG026510.D  
 Acq On : 31 Mar 2017 00:00  
 Operator : SJ/MA  
 Sample : I2448-06  
 Misc :  
 ALS Vial : 13 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :

Quant Time: Mar 31 01:26:07 2017  
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM-EPA-BG032717MA.M  
 Quant Title : SVOA CALIBRATION  
 QLast Update : Fri Mar 31 01:08:45 2017  
 Response via : Initial Calibration

