

Data Path : Z:\HPCHEM1\BNA G\DATA\BG033017\
 Data File : BG026506.D
 Acq On : 30 Mar 2017 21:22
 Operator : SJ/MA
 Sample : I2448-02
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 G9YB7

Manual Integrations
 APPROVED

mohammad
 3/31/2017 10:31:10 AM

Quant Time: Mar 31 01:18:54 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM-EPA-BG032717MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 31 01:08:45 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.04	152	174793	20.00	ng/ul	0.00
18) Naphthalene-d8	10.84	136	719677	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.65	164	427288	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.38	188	1028818	20.00	ng/ul	0.00
77) Chrysene-d12	21.63	240	1182623	20.00	ng/ul	0.00
85) Perylene-d12	24.80	264	1180207	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.47	96	19548	5.35	ng/uL	0.00
5) Phenol-d5	7.20	99	88778	6.90	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.36	67	232372	32.21	ng/ul	0.00
9) 2-Chlorophenol-d4	7.57	132	297151	25.75	ng/ul	0.00
13) 4-Methylphenol-d8	8.74	113	169839	16.38	ng/ul	0.00
19) Nitrobenzene-d5	9.20	128	177987	34.60	ng/ul	0.00
22) 2-Nitrophenol-d4	9.92	143	212014	35.17	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.46	165	331364	29.95	ng/ul	0.00
29) 4-Chloroaniline-d4	0.00	131	0d	0.00	ng/ul	
43) Dimethylphthalate-d6	14.05	166	1140005	34.45	ng/ul	0.00
46) Acenaphthylene-d8	14.34	160	1318184	31.98	ng/ul	0.00
51) 4-Nitrophenol-d4	14.85	143	33750m	5.24	ng/ul	0.00
57) Fluorene-d10	15.63	176	1053817	35.06	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.75	200	195562	29.56	ng/ul	0.00
70) Anthracene-d10	17.48	188	1626633	34.82	ng/ul	0.00
78) Pyrene-d10	19.75	212	1857256	35.21	ng/ul	0.00
89) Benzo(a)pyrene-d12	24.58	264	1867251	36.82	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG033017\
Data File : BG026506.D
Acq On : 30 Mar 2017 21:22
Operator : SJ/MA
Sample : I2448-02
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
G9YB7

Manual Integrations
APPROVED

mohammad
3/31/2017 10:31:10 AM

Quant Time: Mar 31 01:18:54 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM-EPA-BG032717MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Mar 31 01:08:45 2017
Response via : Initial Calibration

