

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG033022\
 Data File : BG052958.D
 Acq On : 30 Mar 2022 20:39
 Operator : CG/JU
 Sample : PB143704BS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SLCS704

Quant Time: Mar 31 00:54:08 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG031822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 31 00:50:49 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 03/31/2022
 Supervised By :Yogesh Patel 04/06/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.134	152	25294	20.000	ng/u1	0.00
20) Naphthalene-d8	10.948	136	112966	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.754	164	92291	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.498	188	235175	20.000	ng/u1	0.00
79) Chrysene-d12	21.780	240	240094	20.000	ng/u1	# 0.00
88) Perylene-d12	25.087	264	256529	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.546	96	4135	5.774	ng/uL	0.00
4) Pyridine-d5	3.963	84	52953	29.621	ng/u1	0.00
7) Phenol-d5	7.276	99	71402	31.517	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.452	67	47433	34.281	ng/u1	0.00
11) 2-Chlorophenol-d4	7.664	132	49113	31.927	ng/u1	0.00
15) 4-Methylphenol-d8	8.833	113	56982	34.722	ng/u1	0.00
21) Nitrobenzene-d5	9.291	128	27572	33.907	ng/u1	0.00
24) 2-Nitrophenol-d4	10.019	143	31572	36.810	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.560	165	60738	32.551	ng/u1	0.00
31) 4-Chloroaniline-d4	11.071	131	70800	28.709	ng/u1	0.00
46) Dimethylphthalate-d6	14.149	166	234745	33.108	ng/u1	0.00
49) Acenaphthylene-d8	14.449	160	257417	29.878	ng/u1	0.00
54) 4-Nitrophenol-d4	14.925	143	43421m	36.424	ng/u1	0.00
60) Fluorene-d10	15.741	176	200877	32.267	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.847	200	44585	37.253	ng/u1	0.00
73) Anthracene-d10	17.597	188	342248m	31.252	ng/u1	0.00
81) Pyrene-d10	19.877	212	433108	32.246	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.864	264	436524	32.933	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.581	88	8728	9.948	ng/uL#	92
5) Pyridine	3.981	79	60671	31.409	ng/u1	92
6) Benzaldehyde	7.264	77	46867	31.119	ng/u1	90
8) Phenol	7.305	94	78178	35.805	ng/u1	95
10) Bis(2-Chloroethyl)ether	7.540	93	56725	34.438	ng/u1	98
12) 2-Chlorophenol	7.693	128	55403	35.251	ng/u1	93
13) 2-Methylphenol	8.563	108	60620	35.723	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.657	45	95357	35.616	ng/u1	99
16) Acetophenone	8.950	105	94078	36.747	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.933	70	55372	37.100	ng/u1	96
18) 4-Methylphenol	8.892	108	66035	37.350	ng/u1	97
19) Hexachloroethane	9.232	117	23140	33.183	ng/u1	97
22) Nitrobenzene	9.332	77	83487	35.137	ng/u1	94
23) Isophorone	9.861	82	157635	34.709	ng/u1	97
25) 2-Nitrophenol	10.049	139	32379	36.405	ng/u1#	88
26) 2,4-Dimethylphenol	10.102	107	73824	33.498	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.337	93	83862	33.205	ng/u1	98
29) 2,4-Dichlorophenol	10.583	162	61969	34.158	ng/u1	99
30) Naphthalene	10.995	128	189364	32.392	ng/u1	98
32) 4-Chloroaniline	11.094	127	64703	26.701	ng/u1	97
33) Hexachlorobutadiene	11.288	225	48102	32.466	ng/u1	98
34) Caprolactam	11.841	113	23513m	43.034	ng/u1	
35) 4-Chloro-3-methylphenol	12.205	107	75494	38.362	ng/u1	94
36) 2-Methylnaphthalene	12.592	142	140404	34.032	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG033022\
 Data File : BG052958.D
 Acq On : 30 Mar 2022 20:39
 Operator : CG/JU
 Sample : PB143704BS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SLCS704

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 03/31/2022
 Supervised By :Yogesh Patel 04/06/2022

Quant Time: Mar 31 00:54:08 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG031822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 31 00:50:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.810	142	141933	33.982	ng/ul	95
39) 1,2,4,5-Tetrachloroben...	12.957	216	94521	30.985	ng/ul	95
40) Hexachlorocyclopentadiene	12.939	237	56263	26.641	ng/ul	97
41) 2,4,6-Trichlorophenol	13.186	196	59787	32.074	ng/ul	96
42) 2,4,5-Trichlorophenol	13.256	196	67792	33.677	ng/ul	99
43) 1,1'-Biphenyl	13.591	154	201732	30.588	ng/ul	98
44) 2-Chloronaphthalene	13.632	162	161156	30.517	ng/ul	96
45) 2-Nitroaniline	13.820	65	61970	40.032	ng/ul	91
47) Dimethylphthalate	14.196	163	228469	33.766	ng/ul	98
48) 2,6-Dinitrotoluene	14.314	165	47860	39.047	ng/ul	92
50) Acenaphthylene	14.478	152	269699	32.412	ng/ul	98
51) 3-Nitroaniline	14.643	138	44173	35.705	ng/ul	97
52) Acenaphthene	14.819	153	178584	32.853	ng/ul	97
53) 2,4-Dinitrophenol	14.848	184	30392	39.913	ng/ul#	63
55) 4-Nitrophenol	14.942	109	51183	40.797	ng/ul	90
56) Dibenzofuran	15.148	168	273245	33.906	ng/ul	97
57) 2,4-Dinitrotoluene	15.101	165	71311	41.005	ng/ul#	83
58) 2,3,4,6-Tetrachlorophenol	15.371	232	68503	37.753	ng/ul#	98
59) Diethylphthalate	15.559	149	249321	36.383	ng/ul	98
61) Fluorene	15.800	166	221821	33.207	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.788	204	123134	34.642	ng/ul	97
63) 4-Nitroaniline	15.806	138	49229m	40.463	ng/ul	
66) 4,6-Dinitro-2-methylph...	15.864	198	44218	37.761	ng/ul	94
67) N-Nitrosodiphenylamine	16.000	169	204326	32.134	ng/ul	96
68) 4-Bromophenyl-phenylether	16.681	248	89778	38.555	ng/ul	97
69) Hexachlorobenzene	16.810	284	103175	34.301	ng/ul	97
70) Atrazine	16.939	200	70183	26.932	ng/ul	98
71) Pentachlorophenol	17.145	266	61750m	33.759	ng/ul	
72) Phenanthrene	17.539	178	411977	33.378	ng/ul	99
74) Anthracene	17.633	178	403914	32.732	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.556	216	102061	29.139	ng/uL	96
76) Pentachlorobenzene	15.071	250	106731	31.761	ng/uL	99
77) Carbazole	17.891	167	371990	34.110	ng/ul	98
78) Di-n-butylphthalate	18.449	149	443844	34.225	ng/ul	99
80) Fluoranthene	19.542	202	545144	34.442	ng/ul	98
82) Pyrene	19.906	202	520797	33.104	ng/ul	96
83) Butylbenzylphthalate	20.781	149	190723	34.869	ng/ul	91
84) 3,3'-Dichlorobenzidine	21.668	252	168975	31.401	ng/ul	94
85) Benzo(a)anthracene	21.762	228	519097	33.499	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.663	149	268460	35.436	ng/ul	98
87) Chrysene	21.827	228	490830	33.208	ng/ul	99
89) Di-n-octyl phthalate	22.902	149	458407	34.308	ng/ul	100
90) Benzo(b)fluoranthene	24.036	252	557651	33.665	ng/ul	99
91) Benzo(k)fluoranthene	24.106	252	511350	33.449	ng/ul	97
93) Benzo(a)pyrene	24.935	252	523570	33.478	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	28.865	276	622425	35.192	ng/ul	97
95) Dibenzo(a,h)anthracene	28.935	278	515726	34.307	ng/ul	99
96) Benzo(g,h,i)perylene	30.051	276	523011	34.981	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG033022\
 Data File : BG052958.D
 Acq On : 30 Mar 2022 20:39
 Operator : CG/JU
 Sample : PB143704BS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :

BNA_G

ClientSampleId :


SLCS704

Manual Integrations


APPROVED

Reviewed By :Jagrut Upadhyay 03/31/2022
 Supervised By :Yogesh Patel 04/06/2022

Quant Time: Mar 31 00:54:08 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG031822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 31 00:50:49 2022
 Response via : Initial Calibration



Abundance Scan 33 (3.582 min): BG052945.D\data.ms (-28)







#2
1,4-Dioxane
Concen: 9.948 ng/uL
RT: 3.581 min Scan# 31
Delta R.T. -0.001 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39

Instrument :
BNA_G
ClientSampleId :
SLCS704

Tgt Ion: 88 Resp: 8728
Ion Ratio Lower Upper
88 100
43 50.1 30.6 46.0
58 71.1 58.6 87.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/31/2022
Supervised By :Yogesh Patel 04/06/2022



#5
Pyridine
Concen: 31.409 ng/uL
RT: 3.981 min Scan# 101
Delta R.T. -0.001 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39

Tgt Ion: 79 Resp: 60671
Ion Ratio Lower Upper
79 100
52 79.9 57.2 85.8
51 43.1 31.9 47.9









#6
Benzaldehyde
Concen: 31.119 ng/ul
RT: 7.264 min Scan# 660
Delta R.T. -0.001 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39

Instrument :
BNA_G
ClientSampleId :
SLCS704

Tgt Ion: 77 Resp: 4686
Ion Ratio Lower Upper
77 100
105 75.4 72.9 109.3
106 80.5 66.5 99.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/31/2022
Supervised By :Yogesh Patel 04/06/2022



#8
Phenol
Concen: 35.805 ng/ul
RT: 7.305 min Scan# 667
Delta R.T. -0.001 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39

Tgt Ion: 94 Resp: 78178
Ion Ratio Lower Upper
94 100
65 34.4 25.9 38.9
66 48.4 35.4 53.2






#10
Bis(2-Chloroethyl)ether
Concen: 34.438 ng/ul
RT: 7.540 min Scan# 707
Delta R.T. -0.001 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39

Instrument :

BNA_G

ClientSampleId :

SLCS704






Tgt Ion: 93 Resp: 56725
Ion Ratio Lower Upper
93 100
63 82.9 68.2 102.4
95 31.6 24.7 37.1


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 03/31/2022
Supervised By :Yogesh Patel 04/06/2022




#12
2-Chlorophenol
Concen: 35.251 ng/ul
RT: 7.693 min Scan# 733
Delta R.T. -0.001 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39




Tgt Ion:128 Resp: 55403
Ion Ratio Lower Upper
128 100
64 56.7 39.8 59.6
130 31.8 25.8 38.8





#13
2-Methylphenol
Concen: 35.723 ng/ul
RT: 8.563 min Scan# 881
Delta R.T. -0.001 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39




Instrument :
BNA_G
ClientSampleId :
SLCS704




Tgt Ion:108 Resp: 60620
Ion Ratio Lower Upper
108 100
107 84.3 68.3 102.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/31/2022
Supervised By :Yogesh Patel 04/06/2022




#14
2,2'-oxybis(1-Chloropropane)
Concen: 35.616 ng/ul
RT: 8.657 min Scan# 897
Delta R.T. -0.001 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39




Tgt Ion: 45 Resp: 95357
Ion Ratio Lower Upper
45 100
77 14.1 11.0 16.4
79 11.4 9.3 13.9





#16
Acetophenone
Concen: 36.747 ng/ul
RT: 8.950 min Scan# 947
Delta R.T. -0.001 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39




Instrument :
BNA_G
ClientSampleId :
SLCS704



Tgt Ion:105 Resp: 94078
Ion Ratio Lower Upper
105 100
77 91.5 72.9 109.3
51 31.8 25.0 37.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/31/2022
Supervised By :Yogesh Patel 04/06/2022



#17
N-Nitroso-di-n-propylamine
Concen: 37.100 ng/ul
RT: 8.933 min Scan# 944
Delta R.T. 0.005 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39


Tgt Ion: 70 Resp: 55372
Ion Ratio Lower Upper
70 100
42 63.1 47.8 71.6
101 9.3 6.8 10.2
130 16.8 14.7 22.1





#18
4-Methylphenol
Concen: 37.350 ng/ul
RT: 8.892 min Scan# 911
Delta R.T. 0.005 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39



Instrument :
BNA_G
ClientSampleId :
SLCS704



Tgt Ion:108 Resp: 66035
Ion Ratio Lower Upper
108 100
107 106.2 82.7 124.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/31/2022
Supervised By :Yogesh Patel 04/06/2022



#19
Hexachloroethane
Concen: 33.183 ng/ul
RT: 9.232 min Scan# 995
Delta R.T. 0.005 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39

Tgt Ion:117 Resp: 23140
Ion Ratio Lower Upper
117 100
201 101.3 83.9 125.9
199 66.6 51.4 77.2





#22

Nitrobenzene

Concen: 35.137 ng/ul

RT: 9.332 min Scan# 1102

Delta R.T. -0.001 min

Lab File: BG052958.D

Acq: 30 Mar 2022 20:39

Instrument :

BNA_G

ClientSampleId :





SLCS704

Tgt Ion: 77	Resp: 8348
Ion Ratio	Lower Upper
77 100	
123 35.6	25.7 38.5
65 13.8	9.7 14.5

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 03/31/2022
Supervised By :Yogesh Patel 04/06/2022



#23

Isophorone

Concen: 34.709 ng/ul




RT: 9.861 min Scan# 1102


Delta R.T. 0.005 min

Lab File: BG052958.D

Acq: 30 Mar 2022 20:39

Tgt Ion: 82	Resp: 157635
Ion Ratio	Lower Upper
82 100	
95 7.6	6.1 9.1
138 15.6	11.0 16.4









#25
2-Nitrophenol
Concen: 36.405 ng/ul
RT: 10.049 min Scan# 1134
Delta R.T. -0.001 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39

Instrument :
BNA_G
ClientSampleId :
SLCS704

Tgt Ion:139 Resp: 32379
Ion Ratio Lower Upper
139 100
65 68.3 48.9 73.3
109 45.3 29.0 43.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/31/2022
Supervised By :Yogesh Patel 04/06/2022



#26
2,4-Dimethylphenol
Concen: 33.498 ng/ul
RT: 10.102 min Scan# 1143
Delta R.T. -0.001 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39


Tgt Ion:107 Resp: 73824
Ion Ratio Lower Upper
107 100
121 44.5 37.4 56.0
122 77.3 63.0 94.4





#27
Bis(2-Chloroethoxy)methane
Concen: 33.205 ng/ul
RT: 10.337 min Scan# 1183
Delta R.T. -0.001 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39




Instrument :
BNA_G
ClientSampleId :
SLCS704




Tgt Ion: 93 Resp: 83862
Ion Ratio Lower Upper
93 100
95 30.7 25.4 38.2
123 9.1 8.2 12.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/31/2022
Supervised By :Yogesh Patel 04/06/2022




#29
2,4-Dichlorophenol
Concen: 34.158 ng/ul
RT: 10.583 min Scan# 1225
Delta R.T. -0.001 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39




Tgt Ion:162 Resp: 61969
Ion Ratio Lower Upper
162 100
164 65.0 53.2 79.8
98 36.0 29.1 43.7





#30
Naphthalene
Concen: 32.392 ng/ul
RT: 10.995 min Scan# 11095
Delta R.T. -0.001 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39




Instrument :
BNA_G
ClientSampleId :
SLCS704



Tgt Ion:128 Resp: 189364
Ion Ratio Lower Upper
128 100
129 11.0 9.0 13.6
127 13.3 9.8 14.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/31/2022
Supervised By :Yogesh Patel 04/06/2022



#32
4-Chloroaniline
Concen: 26.701 ng/ul
RT: 11.094 min Scan# 1312
Delta R.T. -0.001 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39

Tgt Ion:127 Resp: 64703
Ion Ratio Lower Upper
127 100
129 33.2 25.3 37.9





#33
Hexachlorobutadiene
Concen: 32.466 ng/ul
RT: 11.288 min Scan# 11
Delta R.T. -0.001 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39

Instrument :

BNA_G

ClientSampleId :





SLCS704

Tgt Ion:225 Resp: 4810
Ion Ratio Lower Upper
225 100
223 66.9 51.7 77.5
227 62.1 49.6 74.4

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 03/31/2022
Supervised By :Yogesh Patel 04/06/2022



#34
Caprolactam
Concen: 43.034 ng/ul m
RT: 11.841 min Scan# 1439
Delta R.T. 0.005 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39


Tgt Ion:113 Resp: 23513
Ion Ratio Lower Upper
113 100
55 208.9 142.0 213.0
56 173.5 96.1 144.1#





#35
4-Chloro-3-methylphenol
Concen: 38.362 ng/ul
RT: 12.205 min Scan# 1501
Delta R.T. -0.001 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39




Instrument :
BNA_G
ClientSampleId :
SLCS704



Tgt Ion:107 Resp: 75494
Ion Ratio Lower Upper
107 100
144 24.7 21.8 32.6
142 74.9 64.6 97.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/31/2022
Supervised By :Yogesh Patel 04/06/2022



#36
2-Methylnaphthalene
Concen: 34.032 ng/ul
RT: 12.592 min Scan# 1567
Delta R.T. -0.001 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39


Tgt Ion:142 Resp: 140404
Ion Ratio Lower Upper
142 100
141 88.9 71.2 106.8





#37
1-Methylnaphthalene
Concen: 33.982 ng/ul
RT: 12.810 min Scan# 1604
Delta R.T. 0.005 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39




Instrument :
BNA_G
ClientSampleId :
SLCS704




Tgt Ion:142 Resp: 14193
Ion Ratio Lower Upper
142 100
141 93.8 71.4 107.0
116 4.2 3.5 5.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/31/2022
Supervised By :Yogesh Patel 04/06/2022




#39
1,2,4,5-Tetrachlorobenzene
Concen: 30.985 ng/ul
RT: 12.957 min Scan# 1629
Delta R.T. -0.001 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39




Tgt Ion:216 Resp: 94521
Ion Ratio Lower Upper
216 100
214 71.9 62.0 93.0
179 20.5 15.7 23.5
108 16.3 12.9 19.3





#40
Hexachlorocyclopentadiene
Concen: 26.641 ng/ul
RT: 12.939 min Scan# 1626
Delta R.T. -0.001 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39




Instrument :
BNA_G
ClientSampleId :
SLCS704




Tgt Ion:237 Resp: 5626
Ion Ratio Lower Upper
237 100
235 64.9 50.1 75.1
272 13.3 10.2 15.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/31/2022
Supervised By :Yogesh Patel 04/06/2022




#41
2,4,6-Trichlorophenol
Concen: 32.074 ng/ul
RT: 13.186 min Scan# 1668
Delta R.T. -0.001 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39



Tgt Ion:196 Resp: 59787
Ion Ratio Lower Upper
196 100
198 98.8 75.4 113.2
200 32.4 24.9 37.3





#42
2,4,5-Trichlorophenol
Concen: 33.677 ng/u1
RT: 13.256 min Scan# 1
Delta R.T. -0.001 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39

Instrument :

BNA_G

ClientSampleId :





SLCS704

Tgt Ion:196 Resp: 67792
Ion Ratio Lower Upper
196 100
198 96.6 77.0 115.4
200 29.5 24.2 36.4

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 03/31/2022
Supervised By :Yogesh Patel 04/06/2022



#43
1,1'-Biphenyl
Concen: 30.588 ng/u1
RT: 13.591 min Scan# 1737
Delta R.T. 0.005 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39


Tgt Ion:154 Resp: 201732
Ion Ratio Lower Upper
154 100
153 40.5 33.0 49.4
76 12.0 10.6 15.8





#44
2-Chloronaphthalene
Concen: 30.517 ng/ul
RT: 13.632 min Scan# 1744
Delta R.T. -0.001 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39




Instrument :
BNA_G
ClientSampleId :
SLCS704




Tgt Ion:162 Resp: 161150
Ion Ratio Lower Upper
162 100
164 31.5 26.6 40.0
127 37.4 28.2 42.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/31/2022
Supervised By :Yogesh Patel 04/06/2022




#45
2-Nitroaniline
Concen: 40.032 ng/ul
RT: 13.820 min Scan# 1776
Delta R.T. -0.001 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39




Tgt Ion: 65 Resp: 61970
Ion Ratio Lower Upper
65 100
92 65.0 54.0 81.0
138 82.8 76.0 114.0





#47
Dimethylphthalate
Concen: 33.766 ng/ul
RT: 14.196 min Scan# 1840
Delta R.T. -0.001 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39




Instrument :
BNA_G
ClientSampleId :
SLCS704




Tgt Ion:163 Resp: 228469
Ion Ratio Lower Upper
163 100
194 3.5 2.9 4.3
164 10.7 7.7 11.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/31/2022
Supervised By :Yogesh Patel 04/06/2022




#48
2,6-Dinitrotoluene
Concen: 39.047 ng/ul
RT: 14.314 min Scan# 1860
Delta R.T. -0.001 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39




Tgt Ion:165 Resp: 47860
Ion Ratio Lower Upper
165 100
89 74.6 53.4 80.2
121 24.3 21.2 31.8





#50
Acenaphthylene
Concen: 32.412 ng/ul
RT: 14.478 min Scan# 1887
Delta R.T. 0.005 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39




Instrument :
BNA_G
ClientSampleId :
SLCS704



Tgt Ion:152 Resp: 269699
Ion Ratio Lower Upper
152 100
151 21.4 16.4 24.6
153 13.4 11.2 16.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/31/2022
Supervised By :Yogesh Patel 04/06/2022



#51
3-Nitroaniline
Concen: 35.705 ng/ul
RT: 14.643 min Scan# 1916
Delta R.T. -0.001 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39


Tgt Ion:138 Resp: 44173
Ion Ratio Lower Upper
138 100
108 10.2 7.8 11.8
92 132.9 103.5 155.3





#52
Acenaphthene
Concen: 32.853 ng/ul
RT: 14.819 min Scan# 1945
Delta R.T. 0.005 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39




Instrument :
BNA_G
ClientSampleId :
SLCS704




Tgt Ion:153 Resp: 178584
Ion Ratio Lower Upper
153 100
152 48.6 38.0 57.0
154 91.3 70.0 105.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/31/2022
Supervised By :Yogesh Patel 04/06/2022




#53
2,4-Dinitrophenol
Concen: 39.913 ng/ul
RT: 14.848 min Scan# 1951
Delta R.T. 0.005 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39



Tgt Ion:184 Resp: 30392
Ion Ratio Lower Upper
184 100
63 72.2 49.3 73.9
154 79.0 108.6 162.8#









#55
4-Nitrophenol
Concen: 40.797 ng/ul
RT: 14.942 min Scan# 1966
Delta R.T. 0.005 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39

Instrument :
BNA_G
ClientSampleId :
SLCS704

Tgt Ion:109 Resp: 5118
Ion Ratio Lower Upper
109 100
139 79.2 75.4 113.2
65 108.8 91.4 137.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/31/2022
Supervised By :Yogesh Patel 04/06/2022



#56
Dibenzofuran
Concen: 33.906 ng/ul
RT: 15.148 min Scan# 2002
Delta R.T. -0.001 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39

Tgt Ion:168 Resp: 273245
Ion Ratio Lower Upper
168 100
139 38.2 31.8 47.6









#57
2,4-Dinitrotoluene
Concen: 41.005 ng/ul
RT: 15.101 min Scan# 1994
Delta R.T. -0.001 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39

Instrument :
BNA_G
ClientSampleId :
SLCS704

Tgt Ion:165 Resp: 7131
Ion Ratio Lower Upper
165 100
63 52.8 33.0 49.4
182 3.2 3.4 5.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/31/2022
Supervised By :Yogesh Patel 04/06/2022



#58
2,3,4,6-Tetrachlorophenol
Concen: 37.753 ng/ul
RT: 15.371 min Scan# 2040
Delta R.T. -0.001 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39


Tgt Ion:232 Resp: 68503
Ion Ratio Lower Upper
232 100
131 38.4 29.8 44.8
130 2.2 1.0 1.4
166 26.6 22.0 33.0





#59
Diethylphthalate
Concen: 36.383 ng/ul
RT: 15.559 min Scan# 2071
Delta R.T. 0.005 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39




Instrument :
BNA_G
ClientSampleId :
SLCS704




Tgt Ion:149 Resp: 24932
Ion Ratio Lower Upper
149 100
177 20.9 17.5 26.3
150 12.6 9.5 14.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/31/2022
Supervised By :Yogesh Patel 04/06/2022




#61
Fluorene
Concen: 33.207 ng/ul
RT: 15.800 min Scan# 2113
Delta R.T. 0.005 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39




Tgt Ion:166 Resp: 221821
Ion Ratio Lower Upper
166 100
165 94.0 75.7 113.5
167 12.6 10.3 15.5





#62
4-Chlorophenyl-phenylether
Concen: 34.642 ng/ul
RT: 15.788 min Scan# 2111
Delta R.T. 0.005 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39




Instrument :
BNA_G
ClientSampleId :
SLCS704




Tgt Ion:204 Resp: 123134
Ion Ratio Lower Upper
204 100
206 30.7 27.4 41.0
141 54.2 44.2 66.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/31/2022
Supervised By :Yogesh Patel 04/06/2022




#63
4-Nitroaniline
Concen: 40.463 ng/ul m
RT: 15.806 min Scan# 2114
Delta R.T. 0.005 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39




Tgt Ion:138 Resp: 49229
Ion Ratio Lower Upper
138 100
92 59.3 43.0 64.6
108 122.7 84.6 126.8





#66
4,6-Dinitro-2-methylphenol
Concen: 37.761 ng/ul
RT: 15.864 min Scan# 2124
Delta R.T. -0.001 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39




Instrument :
BNA_G
ClientSampleId :
SLCS704




Tgt Ion:198 Resp: 44218
Ion Ratio Lower Upper
198 100
182 4.4 3.0 4.6
77 28.0 19.7 29.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/31/2022
Supervised By :Yogesh Patel 04/06/2022




#67
N-Nitrosodiphenylamine
Concen: 32.134 ng/ul
RT: 16.000 min Scan# 2147
Delta R.T. 0.005 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39




Tgt Ion:169 Resp: 204326
Ion Ratio Lower Upper
169 100
168 67.8 57.4 86.0
167 36.6 28.1 42.1





#68
4-Bromophenyl-phenylether
Concen: 38.555 ng/ul
RT: 16.681 min Scan# 21
Delta R.T. -0.001 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39




Instrument :
BNA_G
ClientSampleId :
SLCS704




Tgt Ion:248 Resp: 89778
Ion Ratio Lower Upper
248 100
250 96.7 76.3 114.5
141 68.9 58.7 88.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/31/2022
Supervised By :Yogesh Patel 04/06/2022




#69
Hexachlorobenzene
Concen: 34.301 ng/ul
RT: 16.810 min Scan# 2285
Delta R.T. 0.005 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39




Tgt Ion:284 Resp: 103175
Ion Ratio Lower Upper
284 100
142 26.7 22.3 33.5
249 29.4 21.9 32.9





#70
Atrazine
Concen: 26.932 ng/ul
RT: 16.939 min Scan# 21
Delta R.T. 0.005 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39




Instrument :
BNA_G
ClientSampleId :
SLCS704




Tgt Ion:200 Resp: 7018
Ion Ratio Lower Upper
200 100
173 22.3 19.8 29.6
215 48.9 39.5 59.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/31/2022
Supervised By :Yogesh Patel 04/06/2022




#71
Pentachlorophenol
Concen: 33.759 ng/ul m
RT: 17.145 min Scan# 2342
Delta R.T. 0.005 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39



Tgt Ion:266 Resp: 61750
Ion Ratio Lower Upper
266 100
264 64.0 50.3 75.5
268 64.7 45.8 68.8









#72
Phenanthrene
Concen: 33.378 ng/u1
RT: 17.539 min Scan# 2409
Delta R.T. -0.001 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39

Instrument :
BNA_G
ClientSampleId :
SLCS704

Tgt Ion:178 Resp: 41197
Ion Ratio Lower Upper
178 100
179 15.5 11.9 17.9
176 18.7 15.0 22.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/31/2022
Supervised By :Yogesh Patel 04/06/2022



#74
Anthracene
Concen: 32.732 ng/u1
RT: 17.633 min Scan# 2425
Delta R.T. 0.005 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39


Tgt Ion:178 Resp: 403914
Ion Ratio Lower Upper
178 100
179 15.8 12.8 19.2
176 18.1 14.7 22.1





#75
1,2,3,4-Tetrachlorobenzene
Concen: 29.139 ng/uL
RT: 13.556 min Scan# 1731
Delta R.T. -0.001 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39




Instrument :
BNA_G
ClientSampleId :
SLCS704




Tgt Ion:216 Resp: 102061
Ion Ratio Lower Upper
216 100
214 77.4 63.8 95.8
218 44.5 38.3 57.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/31/2022
Supervised By :Yogesh Patel 04/06/2022




#76
Pentachlorobenzene
Concen: 31.761 ng/uL
RT: 15.071 min Scan# 1989
Delta R.T. -0.001 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39




Tgt Ion:250 Resp: 106731
Ion Ratio Lower Upper
250 100
248 64.7 50.4 75.6
252 61.3 49.2 73.8





#77
 Carbazole
 Concen: 34.110 ng/ul
 RT: 17.891 min Scan# 2469
 Delta R.T. -0.001 min
 Lab File: BG052958.D
 Acq: 30 Mar 2022 20:39




Instrument :
 BNA_G
 ClientSampleId :
 SLCS704




Tgt Ion:167 Resp: 371990
 Ion Ratio Lower Upper
 167 100
 166 22.4 17.3 25.9
 139 12.8 9.4 14.0

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 03/31/2022
 Supervised By :Yogesh Patel 04/06/2022




#78
 Di-n-butylphthalate
 Concen: 34.225 ng/ul
 RT: 18.449 min Scan# 2564
 Delta R.T. 0.005 min
 Lab File: BG052958.D
 Acq: 30 Mar 2022 20:39




Tgt Ion:149 Resp: 443844
 Ion Ratio Lower Upper
 149 100
 150 9.5 7.5 11.3
 104 6.4 4.6 6.8





#80
Fluoranthene
Concen: 34.442 ng/u1
RT: 19.542 min Scan# 21
Delta R.T. -0.001 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39




Instrument :
BNA_G
ClientSampleId :
SLCS704




Tgt Ion:202 Resp: 545144
Ion Ratio Lower Upper
202 100
101 7.4 6.1 9.1
100 5.8 5.4 8.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/31/2022
Supervised By :Yogesh Patel 04/06/2022




#82
Pyrene
Concen: 33.104 ng/u1
RT: 19.906 min Scan# 2812
Delta R.T. 0.005 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39



Tgt Ion:202 Resp: 520797
Ion Ratio Lower Upper
202 100
101 8.8 8.2 12.4
100 7.4 6.8 10.2





#83
Butylbenzylphthalate
Concen: 34.869 ng/ul
RT: 20.781 min Scan# 2961
Delta R.T. -0.001 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39

Instrument :

BNA_G

ClientSampleId :





SLCS704

Tgt Ion:149 Resp: 19072
Ion Ratio Lower Upper
149 100
91 80.6 57.1 85.7
206 22.0 17.4 26.2

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 03/31/2022
Supervised By :Yogesh Patel 04/06/2022



#84
3,3'-Dichlorobenzidine
Concen: 31.401 ng/ul
RT: 21.668 min Scan# 3112
Delta R.T. -0.001 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39


Tgt Ion:252 Resp: 168975
Ion Ratio Lower Upper
252 100
254 65.8 48.6 73.0
126 7.5 6.2 9.2





#85
Benzo(a)anthracene
Concen: 33.499 ng/u1
RT: 21.762 min Scan# 3127
Delta R.T. 0.005 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39




Instrument :
BNA_G
ClientSampleId :
SLCS704



Tgt Ion:228 Resp: 51909
Ion Ratio Lower Upper
228 100
229 20.1 15.4 23.2
226 27.5 21.4 32.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/31/2022
Supervised By :Yogesh Patel 04/06/2022



#86
Bis(2-ethylhexyl)phthalate
Concen: 35.436 ng/u1
RT: 21.663 min Scan# 3111
Delta R.T. 0.005 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39


Tgt Ion:149 Resp: 268460
Ion Ratio Lower Upper
149 100
167 25.6 21.4 32.2
279 4.5 3.4 5.2





#87
Chrysene
Concen: 33.208 ng/u1
RT: 21.827 min Scan# 3139
Delta R.T. -0.001 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39




Instrument :
BNA_G
ClientSampleId :
SLCS704



Tgt Ion:228 Resp: 490830
Ion Ratio Lower Upper
228 100
226 30.4 23.9 35.9
229 19.8 16.0 24.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/31/2022
Supervised By :Yogesh Patel 04/06/2022



#89
Di-n-octyl phthalate
Concen: 34.308 ng/u1
RT: 22.902 min Scan# 3322
Delta R.T. 0.005 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39

Tgt Ion:149 Resp: 458407









#90
Benzo(b)fluoranthene
Concen: 33.665 ng/u1
RT: 24.036 min Scan# 3515
Delta R.T. -0.001 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39

Instrument :
BNA_G
ClientSampleId :
SLCS704

Tgt Ion:252 Resp: 55765
Ion Ratio Lower Upper
252 100
253 22.1 17.8 26.6
125 6.2 5.5 8.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 03/31/2022
Supervised By :Yogesh Patel 04/06/2022



#91
Benzo(k)fluoranthene
Concen: 33.449 ng/u1
RT: 24.106 min Scan# 3527
Delta R.T. 0.005 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39


Tgt Ion:252 Resp: 511350
Ion Ratio Lower Upper
252 100
253 22.6 17.4 26.0
125 6.0 6.0 9.0





#93
Benzo(a)pyrene
Concen: 33.478 ng/u1
RT: 24.935 min Scan# 3668
Delta R.T. 0.011 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39




Instrument :
BNA_G
ClientSampleId :
SLCS704




Tgt Ion:252 Resp: 523570
Ion Ratio Lower Upper
252 100
253 22.0 16.5 24.7
125 6.7 6.0 9.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 03/31/2022
Supervised By :Yogesh Patel 04/06/2022




#94
Indeno(1,2,3-cd)pyrene
Concen: 35.192 ng/u1
RT: 28.865 min Scan# 4337
Delta R.T. 0.005 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39




Tgt Ion:276 Resp: 622425
Ion Ratio Lower Upper
276 100
138 11.5 9.3 13.9
277 22.4 19.4 29.2





#95
Dibenzo(a,h)anthracene
Concen: 34.307 ng/ul
RT: 28.935 min Scan# 41
Delta R.T. 0.005 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39




Instrument :
BNA_G
ClientSampleId :
SLCS704




Tgt Ion:278 Resp: 515720
Ion Ratio Lower Upper
278 100
139 9.4 8.1 12.1
279 23.6 19.0 28.4

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 03/31/2022
Supervised By :Yogesh Patel 04/06/2022



#96
Benzo(g,h,i)perylene
Concen: 34.981 ng/ul
RT: 30.051 min Scan# 4539
Delta R.T. 0.017 min
Lab File: BG052958.D
Acq: 30 Mar 2022 20:39



Tgt Ion:276 Resp: 523011
Ion Ratio Lower Upper
276 100
138 12.9 10.5 15.7
277 24.7 19.0 28.6

