


Data Path : Z:\HPCHEM1\BNA_G\Data\BG033115\
 Data File : BG016196.D
 Acq On : 31 Mar 2015 8:48
 Operator : TP/IZ
 Sample : DFTPP15
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP15

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM01.2-EPA-BG033115.M
 Title : SVOA CALIBRATION
 Last Update : Wed Apr 01 04:39:57 2015



AutoFind: Scans 2532, 2533, 2534; Background Corrected with Scan 2526

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	41.7	24574	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	43.0	25328	PASS
70	69	0.00	2	0.9	217	PASS
127	198	10	80	53.5	31514	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	58941	PASS
199	198	5	9	7.0	4121	PASS
275	198	10	60	25.2	14880	PASS
365	198	1	100	4.3	2547	PASS
441	443	0.01	100	82.8	9211	PASS
442	198	50	100	97.3	57365	PASS
443	442	15	24	19.4	11122	PASS

m/z	Abundance
38.00	608.0
39.00	2149.0
39.90	672.0
41.00	313.0
43.00	324.0
43.90	475.0
49.00	770.0
50.00	6672.0
51.00	26360.0
52.00	1157.0
54.90	352.0
56.00	904.0
57.00	1825.0
58.00	182.0
61.00	304.0
61.90	551.0
63.00	982.0
65.00	490.0
68.90	26488.0
70.00	257.0
71.00	172.0
73.00	501.0
74.00	2594.0
75.00	4013.0
77.00	29848.0
78.00	1910.0
78.90	1982.0
79.90	1340.0
80.90	2004.0
82.00	394.0
83.00	436.0
83.90	534.0
85.00	464.0
85.90	923.0
86.90	257.0
90.90	448.0
92.00	535.0
92.90	3282.0
94.00	200.0
95.00	178.0
97.00	166.0
97.90	2265.0
98.90	1917.0
100.00	207.0
100.90	1158.0
102.90	387.0
103.90	781.0
104.90	649.0
107.00	8339.0
107.90	1316.0
109.90	14894.0
111.00	2420.0
111.90	339.0
113.00	240.0
116.00	767.0
116.90	6258.0
117.90	548.0
119.10	231.0
119.90	161.0
121.00	165.0
121.90	504.0
122.90	890.0
123.90	356.0
125.00	604.0
126.90	30976.0
127.90	2429.0
128.90	11512.0
129.90	1212.0
131.00	208.0
133.00	154.0
133.90	363.0
134.90	947.0
135.90	287.0
137.00	497.0
140.00	176.0
140.90	1634.0
141.80	419.0
143.00	431.0
144.00	158.0
145.00	213.0
145.90	430.0
147.00	754.0
147.90	1662.0
149.00	482.0
149.90	179.0
151.00	181.0
152.90	491.0
154.00	393.0
154.90	837.0
155.90	1125.0
156.90	186.0

Instrument :
BNA_G
ClientSampleId :
DFTPP15

157.90	235.0
158.80	210.0
159.90	556.0
160.90	958.0
164.90	675.0
166.00	440.0
166.90	2731.0
167.90	1571.0
168.90	318.0
170.00	168.0
171.80	272.0
172.90	331.0
173.80	709.0
174.90	1032.0
176.00	318.0
176.90	515.0
178.10	202.0
178.90	1785.0
179.90	1538.0
180.90	770.0
182.90	232.0
184.10	188.0
185.00	888.0
185.90	7556.0
186.90	2314.0
187.90	169.0
188.90	419.0
190.80	209.0
191.90	702.0
193.00	968.0
194.00	257.0
194.80	151.0
195.90	1976.0
197.90	54896.0
198.90	4030.0
200.00	423.0
200.90	168.0
201.50	292.0
202.90	628.0
204.00	2201.0
205.00	3528.0
206.00	13314.0
206.90	2193.0
207.90	711.0
208.90	187.0
210.00	240.0
210.30	253.0
211.00	587.0
211.70	177.0
214.80	219.0
215.90	372.0
216.90	4079.0
217.90	491.0
220.90	2405.0
221.80	576.0
223.00	971.0
223.90	7560.0
225.00	1670.0
225.90	246.0
226.90	4020.0
227.90	442.0
228.90	635.0
230.00	209.0
230.90	337.0
233.90	259.0
234.90	238.0
236.00	228.0
237.00	202.0
241.00	288.0
242.00	495.0
243.20	690.0
244.00	6144.0
245.00	837.0
245.90	1107.0
246.90	267.0
248.90	291.0
250.80	161.0
253.00	409.0
254.90	30248.0
255.90	4384.0
257.00	476.0
257.90	1521.0
258.90	331.0
264.90	901.0
273.00	1008.0
273.90	2846.0
274.90	12701.0
275.90	2019.0
276.90	1043.0
277.90	272.0
283.00	180.0
292.80	333.0
295.90	4402.0
296.90	515.0
302.90	519.0

Instrument :
BNA_G
ClientSampleId :
DFTPP15

314.00	195.0
314.80	494.0
315.90	204.0
323.00	1525.0
323.90	250.0
326.80	299.0
327.90	158.0
334.00	778.0
335.00	296.0
341.00	221.0
345.90	516.0
352.00	612.0
352.90	371.0
354.00	389.0
355.00	165.0
364.90	1919.0
366.00	311.0
372.00	1003.0
372.80	228.0
382.90	290.0
401.90	308.0
402.90	506.0
403.90	198.0
420.80	396.0
421.90	269.0
423.00	1976.0
423.90	553.0
436.70	196.0
437.80	151.0
438.90	274.0
440.90	6145.0
442.00	40000.0
443.00	8597.0
444.00	866.0

Instrument :
BNA_G
ClientSampleId :
DFTPP15

m/z	Abundance
38.00	395.0
39.00	2345.0
39.90	620.0
41.00	256.0
43.00	274.0
43.90	498.0
44.90	160.0
49.00	917.0
50.00	7744.0
51.00	29624.0
52.00	1600.0
55.00	378.0
55.90	807.0
57.00	2205.0
60.10	162.0
61.00	448.0
61.90	424.0
63.00	1127.0
65.00	793.0
67.00	272.0
68.90	30376.0
70.00	183.0
73.00	538.0
74.00	3232.0
75.00	4605.0
76.00	2170.0
77.00	37256.0
78.00	2345.0
79.00	1934.0
79.90	1774.0
80.90	2523.0
81.90	652.0
83.00	708.0
83.90	589.0
84.90	339.0
85.90	843.0
86.90	244.0
90.90	623.0
91.90	590.0
92.90	3485.0
94.00	267.0
95.90	291.0
97.10	207.0
97.90	2634.0
98.90	2358.0
99.90	213.0
100.90	1434.0
102.90	372.0
103.90	988.0
104.90	801.0
106.90	10102.0
108.00	1845.0
108.90	378.0
109.90	20136.0
110.90	2628.0
111.90	290.0
112.90	232.0
116.00	798.0
116.90	8117.0
117.90	542.0
119.90	198.0
121.90	901.0
122.90	1041.0
123.80	598.0
124.90	654.0
127.00	37400.0
128.00	2967.0
128.90	14066.0
130.00	1131.0
130.90	270.0
133.00	153.0
134.00	382.0
134.90	983.0
135.90	413.0
137.00	517.0
138.10	225.0
139.80	162.0
141.00	1615.0
141.90	573.0
142.80	460.0
145.90	565.0
147.00	1226.0
147.90	1920.0
148.90	584.0
151.10	217.0
151.40	208.0
151.80	213.0
152.90	577.0
153.90	505.0
154.90	1173.0
155.90	1426.0

Instrument :
BNA_G
ClientSampleId :
DFTPP15

156.90	469.0
157.90	418.0
159.00	231.0
159.90	559.0
160.90	836.0
162.00	339.0
163.90	179.0
164.90	725.0
166.00	700.0
167.00	4034.0
167.90	1758.0
168.90	205.0
170.00	171.0
171.90	369.0
172.90	437.0
174.00	733.0
174.90	1559.0
175.90	564.0
176.90	747.0
178.00	289.0
178.90	2825.0
180.00	1625.0
180.90	904.0
181.90	247.0
183.90	311.0
185.00	1581.0
185.90	9401.0
187.00	2668.0
188.00	275.0
188.90	1010.0
190.80	331.0
191.90	700.0
193.00	1076.0
193.90	290.0
195.90	2317.0
197.90	69928.0
198.90	5148.0
199.90	244.0
201.40	369.0
202.80	585.0
204.00	2808.0
204.90	4506.0
206.00	18512.0
207.00	2719.0
207.90	687.0
209.00	287.0
209.90	336.0
210.90	764.0
216.00	543.0
216.90	5124.0
218.00	879.0
220.90	3667.0
221.90	770.0
222.90	1361.0
224.00	10345.0
224.90	2814.0
226.00	378.0
226.90	4678.0
227.90	573.0
228.90	1072.0
229.90	155.0
231.00	366.0
233.90	351.0
234.90	231.0
235.80	262.0
237.00	390.0
238.80	300.0
239.80	199.0
240.80	264.0
241.90	799.0
242.90	689.0
244.00	7831.0
245.00	1043.0
245.90	1686.0
246.90	270.0
248.80	364.0
251.80	243.0
253.00	369.0
254.90	42760.0
256.00	6140.0
256.90	476.0
257.90	2204.0
258.90	463.0
264.90	1082.0
265.90	175.0
270.00	153.0
270.90	225.0
272.00	247.0
272.90	1641.0
273.90	3483.0
274.90	16944.0
275.90	2234.0
276.90	1530.0
277.90	248.0
283.00	224.0

Instrument :
BNA_G
ClientSampleId :
DFTPP15

284.80	312.0
292.90	515.0
295.90	5926.0
296.90	962.0
301.90	152.0
302.90	754.0
304.00	286.0
314.00	291.0
314.90	970.0
316.00	505.0
320.90	228.0
321.90	216.0
323.00	2035.0
324.00	304.0
326.90	385.0
327.90	204.0
333.90	1253.0
335.00	332.0
340.90	261.0
341.90	159.0
345.90	622.0
352.00	431.0
353.00	319.0
353.90	697.0
364.90	3137.0
365.90	476.0
370.70	263.0
371.90	1085.0
372.90	346.0
382.90	269.0
390.00	159.0
402.00	361.0
402.90	727.0
403.90	208.0
420.90	494.0
421.90	445.0
423.00	3605.0
423.90	859.0
435.70	169.0
437.10	272.0
438.10	268.0
438.90	656.0
439.40	542.0
441.00	11143.0
442.00	66408.0
443.00	13130.0
444.00	1226.0

Instrument :
BNA_G
ClientSampleId :
DFTPP15

m/z	Abundance
37.90	289.0
39.00	1357.0
39.90	738.0
41.00	345.0
42.90	156.0
44.00	467.0
49.00	641.0
50.00	5156.0
51.00	18280.0
52.00	849.0
55.00	468.0
55.90	597.0
57.00	1270.0
58.00	209.0
61.00	207.0
61.90	230.0
62.90	854.0
65.00	314.0
68.90	19800.0
70.00	212.0
71.00	243.0
73.00	327.0
74.00	2007.0
75.00	2949.0
76.00	1317.0
77.00	23064.0
78.00	1663.0
78.90	1562.0
79.90	977.0
81.00	1354.0
81.90	388.0
83.00	447.0
84.00	607.0
84.90	386.0
86.00	823.0
86.90	253.0
91.00	496.0
91.90	486.0
92.90	2441.0
94.00	182.0
95.00	196.0
96.00	207.0
98.00	1810.0
98.90	1202.0
99.90	204.0
100.90	1087.0
103.00	234.0
103.90	666.0
105.00	558.0
106.90	7224.0
107.90	1137.0
109.90	12669.0
110.90	1931.0
113.10	185.0
114.90	159.0
116.90	5580.0
117.90	364.0
122.00	345.0
122.90	672.0
123.90	374.0
124.90	343.0
126.90	26168.0
128.00	2093.0
128.90	9746.0
129.90	864.0
130.80	193.0
133.90	186.0
135.00	589.0
135.90	356.0
136.90	311.0
137.70	217.0
139.90	211.0
140.90	1094.0
142.00	424.0
143.00	371.0
146.90	652.0
147.90	1484.0
148.90	407.0
149.90	151.0
151.30	158.0
152.90	396.0
153.90	242.0
155.00	910.0
156.00	842.0
157.40	198.0
158.00	259.0
159.00	193.0
160.00	455.0
161.00	683.0
162.00	231.0
165.00	502.0

Instrument :
BNA_G
ClientSampleId :
DFTPP15

165.90	386.0
166.90	2790.0
167.90	1090.0
169.00	245.0
172.00	209.0
172.90	276.0
173.90	543.0
175.00	866.0
175.90	346.0
176.90	547.0
177.70	294.0
178.90	2343.0
179.90	1109.0
180.90	584.0
184.10	187.0
185.00	929.0
186.00	7059.0
186.90	2097.0
187.90	231.0
188.90	684.0
190.90	269.0
191.90	705.0
192.90	877.0
194.00	193.0
195.90	1645.0
197.90	52000.0
198.90	3185.0
199.90	289.0
201.30	286.0
201.90	231.0
202.80	333.0
203.90	2073.0
204.90	3831.0
206.00	13926.0
207.00	2317.0
207.80	556.0
210.30	404.0
210.80	529.0
211.50	178.0
214.90	231.0
215.90	397.0
216.90	3863.0
217.90	611.0
220.90	3348.0
223.00	924.0
224.00	8313.0
225.00	2216.0
226.00	263.0
226.90	3285.0
227.90	511.0
228.90	525.0
230.90	467.0
233.90	228.0
234.90	282.0
236.90	268.0
239.80	166.0
240.80	256.0
241.80	525.0
243.00	557.0
244.00	6458.0
245.00	847.0
245.90	1326.0
246.90	412.0
248.90	171.0
252.40	199.0
253.00	272.0
254.90	33216.0
255.90	4960.0
256.80	362.0
257.90	1918.0
258.90	379.0
264.90	841.0
265.80	167.0
271.90	241.0
272.90	1099.0
273.90	3173.0
274.90	14997.0
275.90	1855.0
276.90	1611.0
277.80	231.0
282.90	196.0
284.90	301.0
290.00	158.0
292.90	329.0
295.90	4812.0
296.90	821.0
302.90	587.0
313.90	279.0
315.00	696.0
316.00	227.0
321.00	205.0
323.00	1741.0
323.90	309.0
326.90	303.0
327.90	178.0

Instrument :
BNA_G
ClientSampleId :
DFTPP15

331.90	173.0
332.90	162.0
334.00	1095.0
334.90	171.0
340.80	197.0
345.90	442.0
351.90	611.0
353.00	515.0
354.00	448.0
364.90	2587.0
366.00	287.0
370.90	213.0
372.00	915.0
373.00	296.0
374.10	153.0
382.80	355.0
401.90	394.0
402.90	415.0
403.90	246.0
421.00	703.0
421.90	582.0
423.00	3670.0
424.00	712.0
436.50	197.0
437.10	304.0
437.60	217.0
438.60	355.0
439.40	538.0
441.00	10345.0
442.00	65688.0
443.00	11641.0
444.00	1314.0

Instrument :
BNA_G
ClientSampleId :
DFTPP15

m/z	Abundance
39.90	709.0
40.90	420.0
43.10	378.0
44.00	436.0
49.00	545.0
50.90	180.0
55.10	230.0
57.00	207.0
69.00	226.0
71.00	215.0
73.00	173.0
83.10	226.0
83.90	466.0
85.90	269.0
97.10	157.0
167.00	402.0
206.90	513.0

Instrument :
BNA_G
ClientSampleId :
DFTPP15