


Data Path : Z:\HPCHEM1\BNA_G\Data\BG033115\
 Data File : BG016202.D
 Acq On : 31 Mar 2015 15:26
 Operator : TP/IZ
 Sample : DFTPP16
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP16

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM01.2-EPA-BG033115.M
 Title : SVOA CALIBRATION
 Last Update : Wed Apr 01 04:39:57 2015



AutoFind: Scans 2532, 2533, 2534; Background Corrected with Scan 2527

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	41.6	24671	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	44.2	26160	PASS
70	69	0.00	2	0.2	52	PASS
127	198	10	80	53.5	31712	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	59237	PASS
199	198	5	9	6.6	3916	PASS
275	198	10	60	25.7	15239	PASS
365	198	1	100	4.0	2386	PASS
441	443	0.01	100	83.3	8716	PASS
442	198	50	100	92.0	54522	PASS
443	442	15	24	19.2	10464	PASS

m/z	Abundance
38.00	283.0
39.00	1828.0
39.90	753.0
41.00	258.0
44.00	326.0
48.90	440.0
50.00	5371.0
51.00	22048.0
52.00	1067.0
55.00	257.0
55.90	488.0
57.00	1708.0
60.90	339.0
61.90	343.0
63.00	642.0
65.00	374.0
68.90	22048.0
73.00	304.0
74.00	2392.0
75.00	3824.0
76.00	1341.0
77.00	26544.0
78.00	2006.0
78.90	1399.0
79.90	1189.0
81.00	1716.0
82.00	503.0
82.90	420.0
83.90	338.0
84.90	327.0
85.90	674.0
87.00	207.0
90.90	364.0
92.00	507.0
92.90	2473.0
93.90	158.0
96.00	250.0
97.90	1817.0
98.90	1704.0
99.90	166.0
100.90	1058.0
102.90	327.0
103.90	568.0
105.00	668.0
106.90	6963.0
107.90	1070.0
109.90	12748.0
110.90	2245.0
111.90	168.0
114.90	194.0
116.00	428.0
116.90	5076.0
118.00	444.0
121.90	401.0
122.90	680.0
123.90	341.0
124.80	381.0
126.90	25992.0
128.00	1851.0
128.90	9780.0
129.90	1018.0
131.00	172.0
132.00	173.0
134.00	338.0
134.90	845.0
135.90	393.0
136.90	449.0
140.90	1171.0
141.90	411.0
142.90	225.0
145.90	275.0
146.90	543.0
147.90	1723.0
148.90	344.0
152.90	310.0
153.90	309.0
155.00	687.0
155.90	1169.0
157.00	342.0
157.80	168.0
158.80	182.0
160.00	353.0
160.90	468.0
161.90	363.0
164.90	342.0
166.00	305.0
166.90	2229.0
167.90	1200.0
168.90	348.0
171.80	298.0
172.90	199.0

Instrument :
BNA_G
ClientSampleId :
DFTPP16

173.90	490.0
174.90	1095.0
175.90	217.0
176.80	603.0
178.90	2022.0
179.90	1245.0
180.90	583.0
184.90	1034.0
185.90	6114.0
186.90	1655.0
187.90	237.0
188.90	563.0
190.90	208.0
191.30	175.0
191.90	515.0
193.00	653.0
196.00	1138.0
197.90	43512.0
198.90	3033.0
199.90	337.0
201.60	215.0
202.90	404.0
204.00	1887.0
204.90	2877.0
206.00	10395.0
206.90	1690.0
207.90	674.0
210.30	195.0
211.00	522.0
214.90	186.0
216.90	2775.0
217.90	395.0
220.90	2680.0
223.00	580.0
224.00	5819.0
224.90	1566.0
225.90	197.0
226.90	2536.0
227.80	380.0
228.90	580.0
230.90	286.0
234.90	223.0
236.90	289.0
238.90	168.0
242.00	396.0
242.90	431.0
244.00	4682.0
244.90	510.0
245.90	956.0
246.90	203.0
249.00	151.0
252.90	169.0
253.90	453.0
254.90	23088.0
255.90	3649.0
257.00	310.0
257.90	1445.0
258.90	236.0
264.80	638.0
272.20	182.0
272.90	829.0
273.90	1940.0
274.90	9701.0
275.90	1392.0
276.90	1016.0
292.90	361.0
295.90	3141.0
296.90	498.0
302.90	197.0
313.90	202.0
315.00	301.0
315.90	169.0
323.00	917.0
324.00	270.0
326.80	161.0
333.90	771.0
334.90	170.0
345.80	150.0
352.00	249.0
354.10	247.0
364.90	1420.0
365.90	258.0
370.80	153.0
371.90	432.0
373.00	214.0
382.90	155.0
401.90	272.0
402.80	310.0
420.90	306.0
421.90	291.0
422.90	1574.0
423.90	313.0
439.00	177.0
439.20	186.0
439.60	253.0

Instrument :
BNA_G
ClientSampleId :
DFTPP16

441.00	4494.0
442.00	29360.0
443.00	5903.0
444.00	556.0

Instrument :
BNA_G
ClientSampleId :
DFTPP16

m/z	Abundance
37.90	297.0
39.00	2718.0
39.90	1007.0
41.00	329.0
44.00	372.0
49.00	604.0
50.00	7586.0
51.00	29296.0
52.00	1735.0
55.00	211.0
56.00	833.0
57.00	1956.0
60.90	447.0
62.10	444.0
62.90	1219.0
64.00	221.0
65.00	617.0
68.90	32176.0
70.00	156.0
72.90	294.0
74.00	3427.0
75.00	4648.0
77.00	38096.0
78.00	2599.0
78.90	1979.0
79.90	1627.0
81.00	2527.0
82.00	626.0
83.00	731.0
83.80	377.0
84.90	523.0
85.90	593.0
87.00	281.0
91.00	501.0
91.90	700.0
93.00	3504.0
94.00	267.0
96.00	197.0
98.00	2602.0
98.90	2165.0
99.90	261.0
100.90	1179.0
102.90	466.0
103.90	626.0
105.00	799.0
106.90	10971.0
107.90	1503.0
109.90	19168.0
110.90	3230.0
111.90	469.0
113.00	202.0
115.90	671.0
116.90	7888.0
117.90	729.0
119.90	197.0
122.00	695.0
122.90	1098.0
123.90	483.0
124.90	649.0
126.90	36832.0
128.00	3269.0
128.90	15166.0
129.90	1122.0
130.90	275.0
133.90	429.0
134.90	1017.0
136.00	551.0
136.90	700.0
137.90	228.0
140.00	279.0
140.90	2035.0
141.90	674.0
142.90	423.0
146.00	274.0
147.00	1228.0
147.90	2209.0
148.90	507.0
150.80	317.0
151.90	179.0
152.90	520.0
153.90	394.0
155.00	1135.0
156.00	1602.0
156.90	274.0
157.80	303.0
158.90	222.0
159.90	489.0
160.90	827.0
162.00	343.0
164.90	700.0
165.90	494.0

Instrument :
BNA_G
ClientSampleId :
DFTPP16

166.90	3664.0
167.90	1767.0
168.90	237.0
172.00	320.0
172.90	582.0
174.00	900.0
175.00	1232.0
175.90	373.0
176.90	662.0
177.90	309.0
178.90	2820.0
179.90	1951.0
180.90	736.0
184.10	281.0
185.00	1477.0
185.90	9968.0
186.90	2846.0
187.90	260.0
189.00	694.0
191.00	268.0
191.90	937.0
192.90	951.0
193.90	217.0
195.90	2420.0
197.90	70920.0
198.90	4556.0
199.90	424.0
200.80	174.0
201.60	203.0
202.90	501.0
203.90	3025.0
205.00	4778.0
206.00	18632.0
206.90	3174.0
207.90	717.0
209.10	204.0
209.90	462.0
211.00	1043.0
214.90	366.0
216.00	569.0
216.90	4677.0
217.90	529.0
220.90	3310.0
221.90	698.0
222.90	1300.0
223.90	10371.0
224.90	2492.0
226.00	398.0
226.90	4290.0
228.00	625.0
228.90	995.0
231.00	277.0
233.90	240.0
234.70	201.0
235.80	216.0
236.90	312.0
238.60	184.0
240.70	249.0
241.90	596.0
242.90	908.0
244.00	8445.0
245.00	1387.0
245.90	1795.0
246.90	296.0
248.90	448.0
252.00	326.0
252.90	348.0
254.90	41096.0
255.90	6625.0
257.00	453.0
257.90	2566.0
258.90	535.0
264.90	1340.0
266.00	210.0
269.80	180.0
271.90	309.0
272.90	1579.0
273.90	3951.0
274.90	18336.0
275.90	2554.0
276.90	1654.0
278.00	413.0
282.80	264.0
285.00	351.0
291.90	164.0
292.90	359.0
293.90	197.0
294.90	197.0
295.90	5733.0
296.90	823.0
301.90	185.0
302.90	538.0
303.90	215.0
314.00	226.0
314.80	797.0

Instrument :
BNA_G
ClientSampleId :
DFTPP16

316.00	324.0
320.90	247.0
322.90	1913.0
323.90	345.0
326.90	457.0
328.00	230.0
331.90	169.0
333.90	1359.0
334.90	425.0
340.90	197.0
345.90	479.0
352.00	619.0
352.90	514.0
354.00	501.0
364.90	2344.0
365.90	450.0
371.00	176.0
371.90	1230.0
372.90	457.0
382.80	308.0
390.00	174.0
401.90	539.0
402.90	546.0
404.00	196.0
420.90	522.0
421.90	629.0
423.00	3404.0
423.90	774.0
436.30	186.0
437.00	205.0
438.00	230.0
439.00	303.0
441.00	10241.0
442.00	62448.0
443.00	11850.0
443.90	1307.0

Instrument :
BNA_G
ClientSampleId :
DFTPP16

m/z	Abundance
37.00	151.0
37.90	239.0
39.00	1911.0
39.90	641.0
41.00	259.0
44.00	311.0
49.00	582.0
50.00	5793.0
51.00	23504.0
52.00	1115.0
55.90	798.0
57.00	1607.0
60.90	303.0
62.00	279.0
63.00	849.0
64.30	157.0
65.00	357.0
68.90	24256.0
72.90	316.0
74.00	2308.0
75.00	3722.0
76.00	1365.0
77.00	29888.0
78.00	1839.0
79.00	1523.0
79.90	1476.0
81.00	2259.0
82.00	653.0
83.00	436.0
83.90	318.0
85.00	336.0
85.90	877.0
87.00	211.0
90.90	380.0
92.00	347.0
92.90	2819.0
93.70	248.0
95.90	195.0
97.90	2029.0
98.90	1791.0
101.00	874.0
102.80	187.0
103.10	195.0
103.90	680.0
104.90	765.0
106.90	8265.0
107.90	1317.0
109.90	15346.0
110.90	2105.0
111.90	315.0
115.90	555.0
116.90	6769.0
117.80	481.0
121.90	739.0
122.90	926.0
123.90	459.0
125.10	339.0
126.90	32312.0
128.00	2209.0
128.90	12358.0
129.90	928.0
131.00	274.0
133.90	269.0
134.90	973.0
135.90	402.0
136.90	518.0
140.90	1605.0
141.90	698.0
142.90	343.0
146.10	292.0
146.90	979.0
147.90	1610.0
148.90	360.0
151.00	239.0
152.90	587.0
153.90	272.0
155.00	885.0
156.00	1227.0
156.90	209.0
158.00	169.0
158.90	152.0
159.90	502.0
160.90	561.0
162.00	195.0
164.90	707.0
166.00	520.0
166.90	2672.0
167.90	1438.0
168.90	204.0
171.80	330.0
172.90	401.0

Instrument :
BNA_G
ClientSampleId :
DFTPP16

174.00	507.0
174.90	1002.0
176.00	271.0
177.00	758.0
177.90	220.0
178.90	2134.0
179.90	1699.0
181.00	698.0
181.90	236.0
183.80	276.0
185.00	1495.0
186.00	9664.0
187.00	2412.0
187.90	201.0
188.90	648.0
190.80	238.0
191.90	1104.0
192.90	811.0
193.90	233.0
194.90	182.0
195.90	1984.0
197.90	63280.0
198.90	4160.0
200.00	314.0
201.50	306.0
203.00	449.0
204.00	2523.0
204.90	4665.0
206.00	15912.0
206.90	2183.0
208.00	417.0
209.90	259.0
211.00	610.0
214.80	255.0
216.00	530.0
216.90	4650.0
218.00	614.0
220.90	3182.0
221.90	479.0
223.00	1319.0
224.00	9481.0
224.90	2089.0
226.00	378.0
226.90	3828.0
227.90	584.0
228.90	794.0
230.90	454.0
233.90	192.0
234.90	335.0
235.80	370.0
236.80	264.0
238.90	168.0
240.80	177.0
242.00	341.0
243.00	633.0
243.90	8142.0
245.00	1107.0
245.90	1575.0
246.90	216.0
248.90	269.0
252.00	255.0
252.80	318.0
254.90	39392.0
255.90	5854.0
256.90	374.0
257.90	2380.0
258.80	391.0
264.90	1172.0
270.80	263.0
271.90	195.0
272.90	1484.0
273.90	4136.0
274.90	17680.0
275.90	2524.0
276.90	1637.0
278.00	407.0
283.00	199.0
283.80	173.0
285.00	397.0
292.90	362.0
293.80	198.0
295.90	5381.0
296.90	677.0
302.90	633.0
307.60	162.0
313.90	381.0
314.90	770.0
315.90	259.0
317.00	178.0
320.90	205.0
322.00	164.0
322.90	1970.0
323.90	346.0
326.80	268.0
328.00	201.0

Instrument :
BNA_G
ClientSampleId :
DFTPP16

331.90	229.0
333.00	181.0
333.90	1134.0
334.70	249.0
340.90	266.0
345.90	605.0
352.00	665.0
352.90	436.0
354.00	726.0
364.90	3394.0
365.90	527.0
370.90	205.0
371.90	1223.0
373.00	163.0
382.90	484.0
401.90	492.0
402.90	614.0
403.90	327.0
420.90	651.0
422.00	574.0
422.90	3994.0
424.00	892.0
436.90	177.0
438.60	263.0
441.00	11414.0
442.00	71760.0
442.90	13640.0
443.90	1210.0

Instrument :
BNA_G
ClientSampleId :
DFTPP16

m/z	Abundance
39.90	494.0
44.00	378.0
48.90	448.0
51.00	278.0
55.00	151.0
73.00	195.0
83.80	359.0
85.90	270.0
95.80	174.0
167.00	663.0
207.00	577.0
207.90	192.0

Instrument :
BNA_G
ClientSampleId :
DFTPP16