


Data Path : Z:\HPCHEM1\BNA_G\Data\BG033116\
 Data File : BG021591.D
 Acq On : 31 Mar 2016 9:13
 Operator : UM/SJ
 Sample : DFTPP18
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP18

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG031616.M
 Title : SVOA CALIBRATION
 Last Update : Fri Apr 01 01:28:23 2016



AutoFind: Scans 2549, 2550, 2551; Background Corrected with Scan 2543

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.5	12534	PASS
68	69	0.00	2	1.9	273	PASS
69	198	0.00	100	41.0	14067	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	50.0	17171	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	34309	PASS
199	198	5	9	7.3	2502	PASS
275	198	10	60	28.6	9824	PASS
365	198	1	100	4.7	1622	PASS
441	442	0.01	100	15.8	3652	PASS
442	198	50	100	67.3	23093	PASS
443	442	15	24	19.9	4595	PASS

m/z	Abundance
38.00	223.0
39.00	1069.0
50.00	4056.0
51.00	14643.0
52.00	612.0
55.90	499.0
56.90	902.0
60.90	195.0
62.00	221.0
63.00	785.0
64.00	215.0
65.00	237.0
68.00	356.0
68.90	15414.0
73.00	308.0
74.00	1779.0
74.90	2380.0
76.00	1065.0
77.00	17728.0
78.00	1211.0
79.00	1382.0
79.90	947.0
80.90	1429.0
81.90	488.0
82.90	324.0
85.00	228.0
85.90	345.0
91.00	332.0
91.90	378.0
92.90	2458.0
94.00	190.0
97.90	1855.0
98.90	1432.0
100.90	811.0
102.80	300.0
103.90	635.0
104.90	502.0
106.90	6178.0
107.90	748.0
109.90	9336.0
110.90	1567.0
112.00	246.0
115.90	393.0
116.90	5496.0
117.90	365.0
121.90	416.0
122.90	746.0
123.90	283.0
124.90	270.0
126.90	18288.0
127.90	1408.0
128.90	8370.0
129.90	850.0
130.90	194.0
133.90	257.0
135.00	606.0
135.90	203.0
136.90	216.0
140.90	1100.0
142.00	278.0
142.90	286.0
145.90	200.0
146.90	545.0
147.90	1366.0
148.90	268.0
151.20	190.0
152.90	328.0
153.90	374.0
154.90	700.0
155.90	1008.0
157.90	279.0
158.90	176.0
159.90	532.0
160.90	575.0
164.90	489.0
166.00	254.0
167.00	1937.0
168.00	923.0
168.90	254.0
171.90	257.0
172.90	206.0
174.00	424.0
174.90	778.0
175.80	166.0
176.90	450.0
178.90	1531.0
179.90	973.0
180.90	490.0
183.90	216.0
184.90	684.0
185.90	5619.0

Instrument :
BNA_G
ClientSampleId :
DFTPP18

186.90	1564.0
187.90	176.0
188.90	465.0
190.80	164.0
191.90	489.0
192.90	569.0
195.90	1128.0
197.90	33928.0
198.90	2610.0
199.90	178.0
202.90	349.0
204.00	1569.0
205.00	2535.0
205.90	9930.0
206.90	1587.0
207.90	447.0
209.00	216.0
210.00	173.0
210.90	418.0
214.90	170.0
216.00	289.0
216.90	3148.0
217.90	389.0
220.90	1482.0
221.90	361.0
222.80	533.0
223.90	5678.0
224.90	1263.0
225.90	232.0
226.90	3350.0
227.90	453.0
228.90	562.0
230.90	254.0
233.90	198.0
234.90	210.0
236.80	190.0
241.00	180.0
241.90	355.0
242.90	308.0
243.90	4453.0
244.90	504.0
245.90	1015.0
246.90	256.0
248.80	184.0
254.90	23528.0
255.90	3448.0
256.80	297.0
257.90	1884.0
258.90	324.0
264.90	605.0
272.80	581.0
273.90	1597.0
274.90	8606.0
275.90	1387.0
276.90	1000.0
277.90	160.0
292.90	157.0
295.90	3464.0
296.90	515.0
302.90	365.0
314.80	344.0
315.90	241.0
322.90	851.0
326.90	166.0
333.90	555.0
334.90	248.0
345.90	248.0
351.80	234.0
352.80	189.0
354.00	218.0
364.90	1204.0
365.90	184.0
371.90	383.0
402.80	216.0
420.90	182.0
422.90	863.0
440.90	2719.0
441.90	17672.0
442.90	3720.0
443.90	285.0

Instrument :
BNA_G
ClientSampleId :
DFTPP18

m/z	Abundance
38.00	157.0
39.00	1031.0
50.00	3939.0
51.00	14469.0
52.00	755.0
55.90	549.0
57.00	1245.0
60.90	287.0
62.00	230.0
62.90	749.0
64.90	234.0
68.00	464.0
68.90	16592.0
73.00	237.0
74.00	2176.0
74.90	3021.0
76.00	872.0
77.00	18648.0
78.00	1257.0
79.00	1260.0
79.90	973.0
80.90	1399.0
81.90	412.0
82.90	265.0
84.90	230.0
85.90	393.0
87.00	162.0
90.90	237.0
91.90	375.0
92.90	2533.0
93.90	192.0
97.90	2112.0
98.90	1582.0
100.90	817.0
102.90	353.0
103.90	553.0
105.00	569.0
107.00	6664.0
108.00	936.0
108.90	265.0
109.90	10950.0
110.90	1632.0
111.90	216.0
116.00	499.0
116.90	6420.0
117.90	463.0
121.90	482.0
122.90	708.0
123.90	345.0
125.00	343.0
126.90	19680.0
127.90	1482.0
128.90	8397.0
129.90	786.0
133.90	349.0
134.90	911.0
135.80	284.0
137.00	393.0
140.90	1347.0
141.90	460.0
142.90	325.0
146.00	239.0
146.90	849.0
147.90	1415.0
148.90	327.0
150.90	227.0
152.90	294.0
153.90	429.0
154.90	685.0
156.00	933.0
156.90	159.0
157.80	292.0
159.80	337.0
160.90	572.0
161.90	150.0
164.90	504.0
165.90	381.0
166.90	2416.0
168.00	1177.0
168.90	265.0
172.00	226.0
173.10	235.0
173.90	552.0
174.90	954.0
175.90	231.0
176.90	530.0
177.80	186.0
178.90	1712.0
179.90	1247.0
180.90	578.0
183.90	188.0

Instrument :
BNA_G
ClientSampleId :
DFTPP18

184.90	1087.0
185.90	6671.0
186.90	1932.0
187.90	333.0
188.80	509.0
190.90	259.0
191.90	610.0
193.00	883.0
193.90	172.0
195.90	1347.0
197.90	40856.0
198.90	3097.0
199.90	161.0
201.50	220.0
202.90	329.0
203.90	1602.0
204.90	3292.0
205.90	12576.0
206.90	1776.0
207.90	516.0
208.80	224.0
209.90	215.0
210.90	537.0
214.90	162.0
216.00	258.0
216.90	4007.0
217.90	389.0
220.90	1834.0
221.90	594.0
222.90	780.0
223.90	7630.0
224.90	2142.0
226.00	303.0
226.90	4208.0
227.90	561.0
228.90	724.0
230.90	308.0
233.80	296.0
234.90	281.0
235.90	179.0
236.90	233.0
240.90	193.0
241.90	526.0
243.00	436.0
243.90	5378.0
244.90	883.0
245.90	1378.0
246.80	367.0
248.90	206.0
254.90	30016.0
255.90	4233.0
256.90	291.0
257.90	2222.0
258.90	363.0
264.90	939.0
272.90	1011.0
273.90	1955.0
274.90	12064.0
275.90	1590.0
276.90	1124.0
277.90	204.0
284.90	230.0
292.90	369.0
295.90	4492.0
296.90	585.0
302.90	596.0
314.90	424.0
315.90	174.0
322.90	1232.0
324.00	188.0
326.80	221.0
333.90	738.0
334.90	310.0
341.00	175.0
345.80	265.0
352.00	410.0
354.00	299.0
364.90	2103.0
365.90	263.0
371.90	522.0
382.80	216.0
401.90	280.0
402.90	249.0
403.80	159.0
420.90	176.0
421.90	261.0
423.00	1430.0
423.90	341.0
440.90	4409.0
441.90	27992.0
442.90	5078.0
443.90	553.0

Instrument :
BNA_G
ClientSampleId :
DFTPP18

m/z	Abundance
39.00	426.0
50.00	2647.0
51.00	8490.0
52.00	430.0
55.90	286.0
56.90	503.0
63.00	414.0
65.00	260.0
68.90	10196.0
73.00	220.0
74.00	1274.0
74.90	1699.0
76.00	557.0
77.00	10688.0
78.00	988.0
78.90	944.0
79.90	770.0
80.90	1184.0
81.90	252.0
82.90	177.0
84.90	218.0
85.90	280.0
86.90	162.0
90.90	250.0
91.90	238.0
92.90	1860.0
97.90	1368.0
98.90	920.0
100.90	587.0
102.90	220.0
103.80	298.0
104.90	359.0
106.90	4230.0
108.00	584.0
109.90	6696.0
110.90	1094.0
111.90	181.0
115.90	320.0
116.90	4074.0
117.90	259.0
121.90	337.0
123.00	487.0
126.90	13546.0
127.90	1092.0
128.90	5680.0
129.90	576.0
133.90	266.0
134.90	564.0
136.00	224.0
136.90	300.0
140.90	896.0
142.00	338.0
142.90	189.0
145.90	191.0
146.90	508.0
147.90	1066.0
148.90	255.0
152.90	458.0
153.90	153.0
154.90	486.0
155.90	668.0
157.80	164.0
159.00	157.0
159.90	426.0
160.90	373.0
164.90	245.0
166.00	257.0
166.90	1734.0
167.90	760.0
173.00	194.0
173.90	295.0
174.90	793.0
176.00	206.0
176.90	338.0
177.90	185.0
178.90	1281.0
179.90	985.0
180.90	464.0
184.90	723.0
185.90	4990.0
186.90	1212.0
188.90	476.0
191.90	532.0
193.00	485.0
195.90	749.0
197.90	28144.0
198.90	1801.0
199.90	151.0
202.80	401.0
203.90	1628.0
205.00	2163.0

Instrument :
BNA_G
ClientSampleId :
DFTPP18

206.00	8524.0
206.90	1004.0
208.00	191.0
209.90	164.0
210.90	328.0
216.00	224.0
216.90	2563.0
218.00	240.0
220.90	1186.0
222.80	615.0
223.90	5311.0
224.90	1461.0
226.00	165.0
226.90	2466.0
227.90	353.0
228.90	527.0
233.90	151.0
236.90	273.0
238.90	156.0
240.90	207.0
242.00	290.0
242.90	344.0
243.90	3914.0
244.90	577.0
245.90	934.0
246.90	269.0
248.90	157.0
254.90	20912.0
255.90	2873.0
256.90	299.0
257.90	1312.0
258.90	359.0
264.90	746.0
272.90	644.0
273.90	1540.0
274.90	8803.0
275.90	1268.0
276.90	1059.0
277.90	179.0
283.00	163.0
292.80	207.0
295.90	3811.0
296.80	452.0
303.00	364.0
314.00	195.0
314.90	334.0
315.80	189.0
323.00	776.0
326.90	220.0
333.90	659.0
345.90	277.0
351.90	248.0
353.00	237.0
353.90	269.0
364.90	1559.0
365.90	244.0
371.90	507.0
402.90	230.0
420.80	152.0
421.90	243.0
422.90	1351.0
423.90	283.0
440.90	3828.0
441.90	23616.0
442.90	4989.0
443.90	346.0

Instrument :
BNA_G
ClientSampleId :
DFTPP18